Magnetic ionic liquids as non-conventional extraction solvents for the determination of polycyclic aromatic hydrocarbons

María J. Trujillo-Rodríguez¹, Omprakash Nacham², Kevin D. Clark²,
Verónica Pino^{*1}, Jared L. Anderson^{*2}, Juan H. Ayala¹, Ana M. Afonso¹

¹*Departamento de Química (Área de Química Analítica), Universidad de La Laguna (ULL), La Laguna (Tenerife), 38206 Spain*

²*Department of Chemistry, Iowa State University, Ames, Iowa 50011 USA*

Abstract

This work describes the applicability of magnetic ionic liquids (MILs) in the analytical determination of a group of heavy polycyclic aromatic hydrocarbons. Three different MILs, namely, benzyltrioctylammonium bromotrichloroferrate (III) (MIL A), methoxybenzyltrioctylammonium bromotrichloroferrate (III) (MIL B), and 1,12-di(3-benzylbenzimidazolium)dodecanebis[(trifluoromethyl)sulfonyl]imide bromotrichloroferrate (III) (MIL C), were designed to exhibit hydrophobic properties, and their performance examined in a microextraction method for hydrophobic analytes. The magnet-assisted approach with these MILs was performed in combination with high performance liquid chromatography and fluorescence detection. The study of the extraction performance showed that MIL A was the most suitable solvent for the extraction of polycyclic aromatic hydrocarbons and under optimum conditions the fast extraction step required ~20 µL of MIL A for 10 mL of aqueous sample, 24 mmol·L⁻¹ NaOH, high ionic strength content of NaCl (25% (w/v)), 500 µL of acetone as dispersive solvent, and 5 min of vortex. The desorption step required the aid of an external magnetic field with a strong NdFeB magnet (the separation requires few seconds), two back-extraction steps for polycyclic aromatic hydrocarbons retained in the MIL droplet with n-hexane, evaporation and reconstitution with acetonitrile. The overall method presented limits of detection down to 5 ng·L⁻¹, relative recoveries ranging from 91.5 to 119%, and inter-day reproducibility values (expressed as relative standard deviation) lower than 16.4% for a spiked level of 0.4 µg·L⁻¹ (n = 9). The method was also applied for the analysis of real samples, including tap water, wastewater, and tea infusion.

Keywords: Ionic liquids; Magnetic ionic liquids; Magnetic-assisted microextraction; Polycyclic aromatic hydrocarbons; High-performance liquid chromatography

---

*Corresponding author:

Tel. +34 922318990. E-mail: veropino@ull.edu.es

**Co-corresponding author:

Tel. +1 5152948356. E-mail: andersoj@iastate.edu

M.J. Trujillo-Rodríguez   mtrujill@ull.edu.es

Tel. +34 922845200

41	O.Nacham	onacham@iastate.edu	Tel. +1 5152948356
42	K.D. Clark	kdclark@iastate.edu	Tel. +1 5152948356
43	J.H. Ayala	jayala@ull.edu.es	Tel. +34 922318044
44	A.M. Afonso	aafonso@ull.es	Tel. +34 922318039
45			

1. Introduction

Trends in sample preparation are focused on the development of non-conventional solvents as well as novel solid materials in extraction and preconcentration schemes to improve limitations and toxicity issues of current organic solvents and common sorbents [1-4]. Within non-conventional solvents, ionic liquids (ILs) have been widely studied in recent years due to their interesting structures and unusual properties [5]. These non-molecular solvents exhibit melting points below 100 °C, low to negligible vapor pressure at room temperature, high chemical stability as well as wide electrochemical windows. They have been also pointed out as green solvents because select classes of ILs do not generate volatile organic compounds and are hydrolytically stable [6,7]. Their structures can be easily modified by changing the nature of the cation/s and/or anion/s of the IL, or by incorporating different functional groups to the cationic/anionic moieties [8,9]. This simple structural versatility is accompanied by important modification of their properties, from low to high viscosity, moderate to high conductivity, and miscibility in water or in organic solvents, among others [10].

Several derivatives of ILs have been also designed with the purpose of combining some of their unique properties with those derived from other types of materials. Thereby, polymeric ionic liquids (PILs) combine the inherent properties of polymers as well as IL character [11,12], while IL-based surfactants offer the possibility to form a micellar media when dissolved in an aqueous media at a certain concentration [13,14]. Following this idea, magnetic ionic liquids (MILs) exhibit paramagnetic behavior under the application of an external magnetic field [15], and can be designed to incorporate paramagnetic metal anions or metal complexes. Tetrachloroferrate(III) ($[\text{FeCl}_4]^-$) and tetrabromoferrate(III) ($[\text{FeBr}_4]^-$)-based MILs are among those that are more commonly studied [15]. In recent years, other MILs containing anions based on

Fe(III) [16,17], Co(II) [17], Mn(II) [17,18] and lanthanide [17,19] complexes have also been reported.

Magnet-based separations constitute a quite interesting and evolving advancement in sample preparation [20]. These methods utilize an extractant material with magnetic properties. Analytes are enriched in the material, which can be further separated from the remaining components of the sample with the aid of a strong magnet [21,22]. Analytes are finally eluted from the material and subjected to quantification. Thus, the overall method is quite simple because it does not require centrifugation or filtration to separate the magnetic material from the sample once extraction has been accomplished. ILs [23,24] and IL-based surfactants [25-27] have been described as extractants presenting a magnetic core (normally magnetite: α -Fe₃O₄) in a number of magnet-based extraction approaches. Bare Fe₃O₄ MNPs tend to aggregate, they easily oxidize in air, or they experience biodegradation, and thus coatings are required in an attempt to increase their stability [20]. Applications using Fe₃O₄@ILs (or @IL-based surfactants) require more than one synthetic step (that of the IL and that of the Fe₃O₄ magnetic nanoparticles, MNPs); and in some cases the whole extraction efficiency of the hybrid material is lower than that of the neat IL (or IL-based surfactant). In some approaches, dispersive solvents or agitation may be needed to disperse the extraction phase and improve the mass transfer of the analytes to the micro-amounts of material [20].

MILs have been utilized as extraction solvents for the preconcentration of DNA from a cell lysate [28]. They have also been shown to be highly compatible with the polymerase chain reaction thereby permitting high throughput DNA amplification from template DNA enriched in the MIL microdroplet [29]. MILs have been also used for the removal of certain compounds [17,30-34]. In these cases, there was no elution of the

trapped analytes from the MIL prior to downstream analysis. In fact, there are few studies in the literature that exploit the use of neat MILs (paramagnetic MILs that do not combine ILs with $\alpha\text{-Fe}_3\text{O}_4$) in analytical microextraction approaches involving further determination of analytes [35-37]. Two of these studies required the addition of carbonyl iron powder to the MIL because the paramagnetism of the tested MILs (1-hexyl-3-methylimidazolium[FeCl₄][35] and 1-butyl-3-methylimidazolium-[FeCl₄] [36]) was not sufficient to enable the magnetic separation under the application of an external magnetic field. The remaining report was applied to the determination of polar analytes (phenols) using methyltriocetylammmonium[FeCl₄] [37].

A main challenge encountered when working with MILs in microextraction approaches is the need to minimize water solubility of the MIL. This can largely be achieved by incorporating nonpolar moieties into the cation and/or using non-coordinating anions. The main aim of this work is the development of a MIL-based analytical microextraction approach devoted to the determination of hydrophobic analytes because so far the unique analytical determination study for small molecules with neat MILs has been devoted to the determination of polar compounds (phenols) [37]; and the majority of studies with MILs do not imply the further analytical determination of target compounds [17,30-34]. A group of five heavy polycyclic aromatic hydrocarbons (PAHs) was selected as test analytes, as model hydrophobic compounds widely determined using ILs or any other novel material [38-41]. The method was combined with high performance liquid chromatography and fluorescence detection (HPLC-FD) for the determination of these compounds within various aqueous samples.

2. Experimental

121

122 **2.1. Chemicals, reagents, materials, and samples**

123 Benzo(a)anthracene (BaA, $\geq 99.0\%$), chrysene (Chy, $\geq 99.0\%$), and
124 benzo(a)pyrene (BaPy, $\geq 99.0\%$) were supplied by Sigma-Aldrich (St. Louis, MO,
125 USA). Benzo(b)fluoranthene (BbF, $\geq 99.0\%$) was purchased from Supelco (Bellefonte,
126 PA, USA) whereas benzo(k)fluoranthene (BkF, $\geq 99.0\%$) was purchased from Fluka (St
127 Louis, MO, USA). Ultrapure water ($18.2 \text{ M}\Omega\cdot\text{cm}$) was obtained from a Milli-Q water
128 purification system (Millipore, Watford, UK).

129 All PAHs were individually dissolved in acetonitrile ($\geq 99.9\%$), supplied by
130 VWR International Eurolab S.L. (Barcelona, Spain) at a concentration of $1200 \text{ mg}\cdot\text{L}^{-1}$
131 for BaA, $760 \text{ mg}\cdot\text{L}^{-1}$ for Chy, and $1000 \text{ mg}\cdot\text{L}^{-1}$ for the remaining PAHs. Three
132 intermediate solutions of all PAHs at $1 \text{ mg}\cdot\text{L}^{-1}$, $0.5 \text{ mg}\cdot\text{L}^{-1}$, and $0.25 \text{ mg}\cdot\text{L}^{-1}$, in
133 acetonitrile, were prepared by dilution of the individual standards. All experiments were
134 carried out by dilution of the intermediate solutions, or by spiking water with these
135 intermediate solutions.

136 Calibration curves of the overall MIL-based microextraction-HPLC-FD method
137 were carried out preparing standards of PAHs in ultrapure water, at concentrations
138 between 0.05 and $5.0 \mu\text{g}\cdot\text{L}^{-1}$.

139 The extraction method requires the use of NaOH ($\geq 98\%$), NaCl ($\geq 99.5\%$),
140 acetone ($\geq 99.9\%$), n-hexane ($> 97\%$), and acetonitrile ($\geq 99\%$), which were supplied by
141 Sigma-Aldrich. Other reagents utilized during optimization of the method include
142 chloroform ($\geq 99.5\%$), and dichloromethane (99.8%), obtained from Sigma-Aldrich, and
143 the IL-based surfactants 1-hexadecyl-3-methylimidazolium bromide ($[\text{C}_{16}\text{MIm}][\text{Br}]$,
144 with a critical micelle concentration (CMC) value of $0.7 \text{ mmol}\cdot\text{L}^{-1}$ in pure water) and

1,3-didodecylimidazolium bromide ($[\text{C}_{12}\text{C}_{12}\text{Im}][\text{Br}]$, $\text{CMC} = 0.1 \text{ mmol}\cdot\text{L}^{-1}$ in pure water). Both ILs were synthesized and fully characterized as previously reported [42].

Three different MILs were utilized in the analytical microextraction approach. These include two monocationic MILs, benzyltrioctylammoniumbromotrichloroferrate(III) ($[\text{N}_{8,8,8,\text{B}}][\text{FeCl}_3\text{Br}]$ denoted as MIL A) and methoxybenzyltrioctylammoniumbromotrichloroferrate(III) ($[\text{N}_{8,8,8,\text{MOB}}][\text{FeCl}_3\text{Br}]$ denoted as MIL B), and a dicationic MIL, 1,12-di(3-benzylbenzimidazolium)dodecanebis[(trifluoromethyl)sulfonyl]imide bromotrichloroferrate(III) ($[(\text{BBnIm})_2\text{C}_{12}][\text{NTf}_2][\text{FeCl}_3\text{Br}]$ denoted as MIL C). These MILs were synthesized and characterized according to a recent report by Nacham *et al.* [16]. Table 1 shows their chemical structures together with some of their properties. Figure ESM-1 of the Electronic Supplementary Material (ESM) includes their ^1H NMR and ^{13}C NMR spectra.

0.22 μm polyvinylidene difluoride (PVDF) membrane filters (Whatman, Freiburg, Germany) were used before HPLC injection, while all aqueous samples were filtered through 0.45 μm PVDF Durapore membrane filters (Millipore, Darmstadt, Germany) before analysis.

Aqueous samples include tap water taken in the laboratory, wastewater collected from a wastewater treatment plant (WWTP) located in Tenerife (Spain), and a tea infusion of wild fruits tea bag bought in a local supermarket. Wastewater was subjected to 5 min of sonication and stored in an amber glass bottle at 4 $^{\circ}\text{C}$ until analysis. Tea infusion was prepared to mimic common preparation of tea [43]: a tea bag was placed in a glass vessel containing 200 mL of previously boiled mineral water (acquired in a local supermarket) and kept at room temperature for 10 min.

2.2. Instrumentation

All MIL-based microextraction experiments were carried out in Pyrex® glass tubes of 25 mL (Staffordshire, United Kingdom). A cylindrical NdFeB magnet (5.08 cm diameter × 5.08 cm thickness; B = 0.9 T) was purchased from K&J Magnetics (Pipersville, PA, USA). Vortex was applied with a Reax-Control Heidolph GMBH (Schwabach, Germany). An ultrasonic bath KM (Shenzhen Codyson Electrical Co., Ltd., Shenzhen, China) was also employed. The evaporation step was performed in a Visiprep TM vacuum system of Supelco.

Separation and detection of the PAHs were carried out in a HPLC system equipped with a solvent delivery Varian ProStar 230 SDM module (Palo Alto, CA, USA) and a Rheodyne 7725i injection valve obtained from Supelco (Bellefonte, PA, USA) with a 5 µL loop. A fluorescence detection system (FD) Waters 474 (Waters, Milford, MA, USA) was used and connected through a Varian Star 800 module interfaced to the HPLC. The separation was achieved using a SUPELCOSIL™ LC-PAH column (25 cm × 2.1 mm I.D. × 5 µm particle size) supplied by Supelco and protected by a Pelliguard LC-18 guard column purchased from Supelco. The temperature of the column was controlled in the thermostat of a Varian ProStar 410 autosampler. LC workstation software (version 6.41) from Varian was used for data acquisition.

2.3. Procedures

2.3.1. HPLC-FD

The HPLC separation was achieved using a binary mobile phase of acetonitrile:water at a constant flow rate of $0.5 \text{ mL} \cdot \text{min}^{-1}$, and the following linear elution gradient: 3 min with 75% of acetonitrile (v/v), then up to 100% of acetonitrile (v/v) in 17 min, and finally kept for 5 min. The analytical column was maintained at a constant temperature of $30 \text{ }^{\circ}\text{C}$.

Fluorescence detection required the use of excitation and emission splits of 18 nm and the following wavelength program: initially, 248/370 (nm) (excitation wavelength/emission wavelength: $\lambda_{\text{ex}}/\lambda_{\text{em}}$) with a gain of 10, a change to 273/384 (nm) and to a gain of 100 at a chromatographic time of 6.5 min, following by a change to 254/451 (nm) at 9.9 min, then change to 288/406 (nm) at 12.3 min, and finally changed to 248/370 (nm) and to a gain of 10 at 16.3 min. Figure ESM-2(A) of the ESM shows a representative chromatogram of the PAHs (obtained without any preconcentration step, only HPLC-FD). PAHs eluted at the retention times reported in Table ESM-1 of the ESM, with RSD values lower than 1.2%.

2.3.2. MIL-based microextraction method

The MIL-based microextraction method under optimum conditions required 10 mL of the aqueous sample involving ultrapure water, tap water, wastewater, or tea infusion (with or without spiking PAHs, depending on the experiment). The NaCl content was adjusted to 25% (w/v), and 0.5 mL of NaOH $0.5 \text{ mol} \cdot \text{L}^{-1}$ was added (final content in the extraction tube: $24 \text{ mmol} \cdot \text{L}^{-1}$). The optimum extraction step required $\sim 20 \mu\text{L}$ of MIL A (as extraction solvent) and $500 \mu\text{L}$ of acetone (as dispersive solvent), which were sequentially added to the tube containing the aqueous sample, followed by 5 min of vortex to ensure proper dispersion of the MIL. Then, the cylindrical magnet was placed at one side of the extraction tube. The MIL was immediately attracted to the

magnet and retained on the wall of the tube, allowing easy removal of the supernatant using a pipette. Once PAHs were enriched in the MIL droplet already separated from the remaining aqueous sample, the optimum back-extraction required 500 μ L of n-hexane, 1 min of vortex, and separation of the eluent with a pipette and the aid of the magnet. This procedure was repeated twice. The total volume of eluent ($\sim$ 1 mL) was evaporated to dryness using a stream of air ($\sim$ 10 min). The residue was diluted up to 300 μ L with acetonitrile, and then filtered through 0.2 μ m PVDF membrane filters. Finally, a volume of 5 μ L was injected in the HPLC-FD system. The overall procedure under optimum condition is summarized in Figure 1.

All variables were properly studied in order to achieve the optimum conditions. Each optimization experiment was repeated in triplicate ($n = 3$).

3. Results and discussion

3.1. Use of MILs in the microextraction method for determining PAHs

3.1.1. Previous considerations of the MIL-based microextraction method

A group of three different MILs, described in Section 2.1, were examined as candidates for the magnet-based microextraction of PAHs. As previously discussed, the intrinsic hydrophobicity and high viscosity of the MILs studied exert an important influence in the method development.

Given the hydrophobicity of the MILs, it was challenging to homogeneously disperse the neat MIL into the aqueous solution by using stirring media (e.g., sonication,

vortex, or manual stirring) or by incorporating any dispersive solvent. Furthermore, once dispersed, MILs easily stick on the walls of the tube making the recovery and back-extraction complicated. It was observed that the addition of sodium hydroxide before introduction of the MIL was beneficial for MIL dispersion (helping in the formation of multiple microdroplets) and subsequent recovery. A concentrated NaOH solution (spiked volume: 0.5 mL) was added to the aqueous sample. Two concentration levels of NaOH, $4.8 \text{ mmol}\cdot\text{L}^{-1}$ and $24 \text{ mmol}\cdot\text{L}^{-1}$, were tested in the final aqueous solution. Figure ESM-3 of the ESM shows that similar extraction efficiencies (expressed as peak area of the PAHs) were obtained with both NaOH concentration values, even slightly higher for the lower NaOH content for BaA and BkF. However, the reproducibility was worse at lower values, and therefore the highest concentration of NaOH ($24 \text{ mmol}\cdot\text{L}^{-1}$) was selected as the optimum condition and was used in the rest of the study.

Finally, it is important to evaluate the influence of the ionic strength in the MIL-based microextraction approach. The use of high ionic strength is normally beneficial for exploiting the salting-out effect in the majority of liquid-phase microextraction methods [44]. In magnet-based extraction approaches using bare Fe_3O_4 as core of the magnetic material, the use of high ionic strength is normally not convenient when using ILs [45,46] or IL-based surfactants [25,27]. The use of high ionic strength can be accomplished only when Fe_3O_4 coated with silica ($\text{Fe}_3\text{O}_4@\text{SiO}_2$) is used, but requires an additional step in the preparation of the material [24,47]. Figure ESM-4 of the ESM shows the influence of the ionic strength (expressed as NaCl content), evaluated in the range of 0 to 25% (w/v) of NaCl. Clearly, an increase of the NaCl content increases the extraction efficiency for all PAHs. It is important to note that even when no NaCl is added, the net ionic strength of the sample is not zero due to the contribution of the

NaOH (24 mmol·L⁻¹). From these results, 25% (w/v) of NaCl was selected as optimum value for the further experiments. The ability to utilize high ionic strength in MIL-based extraction is advantageous compared to other magnet-based extraction methods in which the addition of salts has a negative effect in the extraction efficiency[25,27,45,46]). The MIL-based microextraction method is therefore clearly valid for aqueous samples with a high salt content.

3.1.2. Screening of MILs in the microextraction method for determining PAHs

The three hydrophobic MILs were tested for the determination of PAHs, using the fixed conditions described above in Section 3.2.1. The normalized peak area for each PAH was introduced as a tool to compensate for possible differences in weighing the MIL (as described in Section 3.2.1) when different MILs are compared. Thus, the normalized peak area is defined as the ratio between the chromatographic peak area and the amount of MIL weighed. The obtained results are shown in Figure 2. It can be observed that better results are achieved with MIL A for all PAHs tested. MIL B showed adequate extraction efficiency (but lower than that obtained with MIL A), whereas MIL C poorly extracted the PAHs. It appears that MILs belonging to the ammonium family may be more suitable than those of the benzimidazolium family for the determination of PAHs. These results can be linked to the different viscosities of the MILs in which ammonium-based MILs (A and B) are less viscous than MIL C and, consequently, they are more easily dispersed in the bulk of the sample, thus favoring the mass transfer of PAHs to the MIL microdroplets. Interestingly, MIL C was designed to possess higher hydrophobicity (which *a priori* could be beneficial for PAHs) but its higher viscosity hindered its performance within the magnet-based microextraction approach.

291

292 **3.2. Optimization of the MIL-based microextraction-HPLC-FD method**

293   The remaining parameters that have an influence in the MIL-based
294 microextraction-HPLC-FD procedure were optimized using a simple factor by factor
295 approach, but other mathematical models can be used in more complex optimizations
296 [48]. The procedure was optimized with MIL A, which provided the best extraction
297 efficiencies. The optimization study was carried out using a spiked level of $5 \mu\text{g}\cdot\text{L}^{-1}$ for
298 the PAHs in ultrapure water, with the remaining conditions presented in Section 3.2
299 (NaOH $24 \text{ mmol}\cdot\text{L}^{-1}$, 25% (w/v) of NaCl, and $20 \mu\text{L}$ of MIL A).

300   The hydrophobic MILs used in this work possess high viscosity so their direct
301 injection into HPLC is not recommended to avoid issues with high back pressure.
302 Therefore, the MIL-based microextraction method was accomplished by designing a
303 back-extraction step followed with reconstitution with acetonitrile. Regarding the back-
304 extraction step, several parameters were fixed, including the volume of back-extraction
305 solvent ($500 \mu\text{L}$, to ensure an adequate desorption volume to make contact with $20 \mu\text{L}$
306 of the MIL), together with the aid of vortex (1 min) to aid in the process. Once back-
307 extracted, the time required to evaporate the solvent to dryness with the stream of
308 air was dictated by the nature of the elution solvent. Finally, acetonitrile was selected to
309 reconstitute the dry residue due to its HPLC compatibility.

310   The optimized parameters include the nature and volume of the dispersive
311 solvent, the type and time of the stirring method during the extraction step, and the nature
312 of the back-extraction solvent as well as the number of back-extraction steps.

313

314 **3.2.1. Optimization of the extraction step**

The dispersion of MIL A was studied using various commonly used organic dispersive solvents (e.g., acetonitrile, methanol, and acetone) as well as two different IL-based surfactants ([C₁₆MIm][Br] and [C₁₂C₁₂Im][Br], working in both cases at concentrations in the extraction tube that correspond with their CMC values). Acetonitrile and methanol caused the MIL to be soluble in the aqueous phase, precluding phase separation upon application of the external magnetic field. For the remaining solvents examined, the best results were achieved with acetone which produced peak areas around 27 times higher than [C₁₆MIm][Br] and around 3 times higher than [C₁₂C₁₂Im][Br], as can be observed in Figure 3. Thus, acetone was selected for the subsequent experiments.

The volume of acetone was also studied in the range 0.25–1.5 mL (see Figure ESM-5 of the ESM). The use of acetone volumes equal to or higher than 1.0 mL caused the MIL to be soluble in the aqueous phase. The use of 500 µL of acetone provided best extraction efficiencies for PAHs (with the exception of BbF). A volume of 500 µL of acetone was selected as optimum value to avoid any partial solubility of the MIL in the aqueous phase.

Vortex and sonication were tested as dispersion modes to increase the mass transfer of PAHs to the MIL Amicrodroplet while also considering the stirring time. Results are included in Figure ESM-6 of the ESM. Vortex and sonication both gave higher extraction efficiency compared to when no stirring of any kind was applied. However, the application of sonication provoked partial solubility of MIL A in the aqueous solution causing lower extraction efficiency compared with vortex stirring. Therefore, the application of vortex was chosen to improve the efficiency of the method. The effect of the vortex time was then studied in the time range from 2.5 to 10 min. It was observed that a vortex of 10 min was adequate for the majority of PAHs

(except for BbF). However, a vortexing time of 5 min was chosen with the goal of developing a fast extraction method but also considering that higher times are usually not recommended due to the health of laboratory personnel.

3.2.2. Optimization of the MIL-DLLME back-extraction step

The solvent used to accomplish the back-extraction step of PAHs from MIL A should be immiscible with the MIL and should be a good solubilizing medium for PAHs. Several solvents were tested, including dichloromethane, chloroform, and n-hexane. Of these solvents, only n-hexane satisfied the above conditions as the chlorinated solvents were partially miscible with MIL A. Thus, n-hexane was selected as the back-extraction solvent.

Regarding the number of back-extraction steps, successive back-extraction steps ranging from 1-4 steps using 500 μ L of n-hexane and 1 min of vortex, followed by magnetic separation, were undertaken. The obtained results are shown in Figure ESM-7 of the ESM. The majority of PAHs were back-extracted with only 2 steps, with the exception of BbF (which required only one single step). Therefore, two successive back-extraction steps were chosen as optimum.

In order to increase the enrichment factor of the method, the final volume of reconstitution with acetonitrile was decreased from 500 μ L to 300 μ L without any loss of efficiency. A summary of the overall optimized MIL-based microextraction-HPLC-FD method is shown in Figure 1.

3.3. Analytical performance of the MIL-based microextraction-HPLC-FD method

Calibration curves of the overall MIL-based microextraction-HPLC-FD method were obtained by applying the optimized method to standards of PAHs in ultrapure water. Figure ESM-2(B) of the ESM shows a representative chromatogram obtained after the application of the entire method. Table 2 includes several quality analytical parameters of the method such as the linearity ranges, calibration levels, correlation coefficients, slopes and intercepts, and limits of detection and quantification.

The obtained calibration curves show a wide linearity range, from 0.05–5.0 $\mu\text{g}\cdot\text{L}^{-1}$ for Chy and from 0.05–3.5 $\mu\text{g}\cdot\text{L}^{-1}$ for the remaining PAHs, with correlation coefficients (R) for the entire method ranging from 0.994 to 0.996.

Limits of detection (LODs) and limits of quantification (LOQs) of the entire method were estimated at three and ten times the signal-to-noise ratio, respectively. They were verified by preparing standards at those levels, which were then subjected to the whole microextraction-HPLC-FD method. Thus, LODs ranged from 0.005 $\mu\text{g}\cdot\text{L}^{-1}$ for BbF to 0.02 $\mu\text{g}\cdot\text{L}^{-1}$ for BaPy, while LOQs were always lower than 0.06 $\mu\text{g}\cdot\text{L}^{-1}$. The LODs of the overall method were ~13 to 54 times lower than those only associated with the HPLC-FD method (Table ESM-2 of the ESM). It is also interesting to mention that the PAHs are dissolved in acetonitrile for injection in the HPLC-FD after the MIL-based microextraction approach, and so this method is also suitable for LC-MS.

Table 2 also summarizes results regarding the reproducibility and extraction performance of the method. The reproducibility was evaluated as the relative standard deviation (RSD) using intra-day ($n = 3$) and inter-day ($n = 9$) experiments. Two spiked levels (0.4 and 1.5 $\mu\text{g}\cdot\text{L}^{-1}$) were selected for intra-day experiments. The obtained intra-day precision results ranged from 1.0% for Chy to 13% for BkF at the low spiked level, and from 7.5% for BbF to 13% for BkF for the intermediate spiked level (1.5 $\mu\text{g}\cdot\text{L}^{-1}$). The inter-day precision was evaluated at the low spiked level (0.4 $\mu\text{g}\cdot\text{L}^{-1}$) and during

three non-consecutive days. RSD values ranged between 7.1% for BaA and 16% for BbF.

The extraction performance of the method was evaluated by means of the enrichment factor (E_F) and the extraction efficiency (E_R , in %) of each PAH (see Table 2). They were determined by performing experiments in ultrapure water at two spiked levels (0.4 and 1.5 $\mu\text{g}\cdot\text{L}^{-1}$, with $n = 3$).

The E_F of the overall MIL-based microextraction-HPLC-FD method was calculated as the ratio between the predicted concentration obtained with chromatographic calibrations, without performing the MIL-based microextraction approach (see Table ESM-2 of the ESM) and the spiked concentration of each PAH. The obtained E_F values ranged from 8.84 for BkF to 16.6 for BaPy at the low spiked level, and between 8.70 for BkF and 12.3 for Chy for the intermediate spiked level.

E_R values were estimated as a percentage value by the ratio of the E_F and the theoretical maximum preconcentration factor ($E_{F,\text{max}}$) of the overall method. The $E_{F,\text{max}}$ was calculated as the ratio of the initial volume in the extraction tube (10.5 mL) and the final volume after reconstitution with acetonitrile (0.3 mL). Following this consideration, the $E_{F,\text{max}}$ of the method is ~35.0. This value indicates the maximum preconcentration that can be achieved if all PAH are effectively extracted in the MIL and back-extracted by n-hexane in the method. Regarding E_R values, they ranged from 25.3% to 47.3% at the low spiked level, and from 24.9% to 35.3% at the intermediate spiked level. This means that, on average, around 35.6% and 31.2% of the PAHs could be extracted at each spiked level, respectively. It is important to mention that E_R values approaching ~100% are difficult to achieve in any microextraction procedure [49]. Thus, the obtained values are valid as long as the other analytical parameters of the method (mainly sensitivity and reproducibility) are sufficient for the intended analytical

application. In any case, higher extraction efficiencies were obtained in this method using a hydrophobic MIL than those reported for the determination of phenols using a non-hydrophobic MIL, with calculated E_R values ranging from 17.9 to 18.8% [37].

The accuracy of the method was evaluated by the determination of the relative recovery (RR, in %). RR was calculated as the ratio of the predicted concentration obtained in the calibration curves of the overall method (Table 2) and the initial spiked concentration. The obtained RR values were expressed as a percentage for the studied two spiked levels. RR values ranged from 91.5% to 119% at the low spiked level. The obtained RR values in this study are comparable to those obtained in other studies with non-hydrophobic MILs [35-37].

3.4. Analysis of real aqueous samples

Once the overall MIL-based microextraction-HPLC-FD method was optimized and validated, it was applied for the analysis of real aqueous samples, including tap water, wastewater, and a tea infusion. The sampling and pretreatment of both samples prior to the analysis was performed as explained in Section 2.1. The studied PAHs were not detected in any of the samples analyzed. A representative chromatogram obtained with a non-spiked wastewater subjected to the whole method is included in Figure ESM-8 of the ESM.

The matrix effect of the method was evaluated with wastewater and tea infusion samples by the estimation of the reproducibility (as RSD) and the relative recovery. With this purpose, PAHs were spiked in these samples at $0.4 \mu\text{g}\cdot\text{L}^{-1}$ and experiments were repeated in triplicate ($n = 3$). A representative chromatogram obtained with a

spiked wastewater subjected to the whole method is included in Figure ESM-8 of the ESM.

Regarding the reproducibility studies, low RSD values were obtained with both samples, with values ranging from 4.3% to 9.5%, and 3.5% to 17% when wastewater and tea infusion were analyzed, respectively. RR values ranged from 85.5% to 135% when analyzing wastewater; and from 46.2% to 116% when tea infusion was analyzed. Clearly, matrix effects were present but the method could still be successfully performed in these complex samples. For these complex samples, it is advisable the utilization of matrix-matched calibrations and, if possible, LC-MS.

This MIL-based microextraction approach provides similar extraction performance to other reported microextraction methods in the literature for determination of PAHs by HPLC-FD, in terms of extraction time, extraction solvent consumption, LODs, RR and reproducibility (see Table ESM-3 of the ESM) [39-41]. Similar conclusions can be derived from Table ESM-4 of the ESM, in which the proposed method is compared with other MIL-based approaches [35-37].

Conclusions

Magnetic ionic liquids have been successfully studied for first time in a magnet-based microextraction approach for the determination of a group of heavy hydrophobic PAHs. The extraction efficiency of three different hydrophobic and viscous MILs was compared and the results indicate that $[N_{8,8,8,B}][FeCl_3Br]$, denoted as MIL A, provided the best extraction performance for the group of PAHs.

The novel microextraction approach in combination with HPLC-FD was properly optimized and the method validated using MIL A. The proposed method is

capable of performing adequate quantitation of heavy PAHs, in terms of high sensitivity (LODs down to 5 ng L^{-1}), adequate reproducibility and efficiency, while requiring low sample volumes, low amounts of MIL, low consumption of organic solvent, while being simple and fast.

Ongoing work is currently being conducted in our laboratories to prepare less viscous MILs while maintaining their hydrophobic nature. By doing this, a better dispersion of the MIL in the sample matrix can be obtained, thus decreasing sources of error and increasing the extraction efficiency and precision.

Acknowledgements

MJT-R thanks the Agencia Canaria de Investigación, Innovación y Sociedad de la Información (ACIISI), co-funded by the European Social Fund, for her FPI PhD fellowship. VP acknowledges funding from the Spanish Ministry of Economy and Competitiveness (MINECO) projects ref. MAT2013-43101-R and MAT2014-57465-R. JHA and VP also thank funding from the Fundación Caja Canarias, project ref. SPDs-Agua05. J.L.A. acknowledges funding from the Chemical Measurement and Imaging Program at the National Science Foundation (grant number CHE-1413199).

References

- [1] M. Espino, M.A. Fernández, F.J.V. Gomez, M.F. Silva, Natural designer solvents for greening analytical chemistry, *Trends Anal. Chem.* 76 (2016) 126–136.
- [2] H. Qiang, W. Zonghua, Z. Xiaoqiong, D. Mingyu, Graphene and its composites in sample preparation, *Prog. Chem.* 26 (2014) 820–833.
- [3] B.H. Fumes, M. Ribeiro Silva, F.N. Andrade, C. E. DominguesNazario, F.M. Lanças, Recent advances and future trends in new materials in sample preparation, *Trends Anal. Chem.* 71 (2015) 9–25.
- [4] L. Xu, X. Qi, X. Li, Y. Bai, H. Liu, Recent advances in applications of nanomaterials for sample preparation, *Talanta* 146 (2016) 714–726.
- [5] J.P. Hallet, T. Welton, Room-temperature ionic liquids: solvents for synthesis and catalysis. 2, *Chem. Rev.* 111 (2011) 3508–3576.
- [6] D. Coleman, N. Gathergood, Biodegradation studies of ionic liquids, *Chem. Soc. Rev.* 39 (2010) 600–637.
- [7] Y. Deng, I. Beadham, M. Ghavre, M.F. Costa Gomes, N. Gathergood, P. Husson, B. Légeret, B. Quilty, M. Sancelme, P. Besse-Hoggan, When can ionic liquids be considered readily biodegradable? Biodegradation pathways of pyridinium, pyrrolidinium and ammonium-based ionic liquids, *Green Chem.* 17 (2015) 1479–1491.
- [8] T.D. Ho, C. Zhang, L.W. Hantao, J.L. Anderson, Ionic liquids in analytical chemistry: fundamentals, advances, and perspectives, *Anal. Chem.* 86 (2014) 262–285.

- 500 [9] Q. Liang, Y. Shi, W. Ma, Z. Li, X. Yang, Uniform square-like BaFBr:Eu²⁺
501 microplates: controlled synthesis and photoluminescence properties, RSC Adv. 2
502 (2012) 5403–5410.
- 503 [10] D. Rooney, J. Jacquemin, R. Gardas, Thermophysical properties of ionic liquids,
504 in: B. Kirchner (Ed.), Topics in Current Chemistry, vol. 290: Ionic Liquids,
505 Springer-Verlag, Heidelberg, Berlin, 2009, pp. 185–212.
- 506 [11] J. Yuan, D. Mecerreyes, M. Antonietti, Poly(ionic liquid)s: An update, Prog.
507 Polym. Sci. 38 (2013) 1009–1036.
- 508 [12] H. Yu, T.D. Ho, J.L. Anderson, Ionic liquid and polymeric ionic liquid coatings in
509 solid-phase microextraction, Trends Anal. Chem. 45 (2013) 219–232.
- 510 [13] M.J. Trujillo-Rodríguez, P. González-Hernández, V. Pino, Analytical applications
511 of ionic liquid-based surfactants in separation science, in: B.K. Paul, S.P. Moulik
512 (Eds.), Ionic liquid-based surfactant self-assemblies: formulation,
513 characterization, and applications, Wiley series on surface and interfacial
514 chemistry, John Wiley & Sons, Hoboken, New Jersey, USA, 2015, pp. 475–502.
- 515 [14] V. Pino, M. Germán-Hernández, A. Martín-Pérez, J.L. Anderson, Ionic liquid-
516 based surfactants in separation science, Sep. Sci. Technol. 47 (2012) 264–276.
- 517 [15] E. Santos, J. Albo, A. Irabien, Magnetic ionic liquids: synthesis, properties and
518 applications, RSC Adv. 4 (2014) 40008–40018.
- 519 [16] O. Nacham, K.D. Clark, H. Yu, J.L. Anderson, Synthetic strategies for tailoring
520 the physicochemical and magnetic properties of hydrophobic magnetic ionic
521 liquids, Chem. Mater. 27 (2015) 923–931.

- 522 [17] E. Santos, J. Albo, A. Rosatella, C.A.M. Afonso, A. Irabien, Synthesis and
523 characterization of magnetic ionic liquids (MILs) for CO₂ separation, J. Chem.
524 Technol. Biotechnol. 89 (2014) 866–871.
- 525 [18] S. Pitula, A.-V. Mudring, Synthesis, structure, and physico-optical properties of
526 manganate(II)-based ionic liquids, Chem. Eur. J. 16 (2010) 3355–3365.
- 527 [19] B. Mallick, B. Balke, C. Felser, A.-V. Mudring, Dysprosium room-temperature
528 ionic liquids with strong luminescence and response to magnetic fields, Angew.
529 Chem. Int. Ed. 47 (2008) 7635–7638.
- 530 [20] P. Rocío-Bautista, V. Pino, Extraction methods facilitated by the use of magnetic
531 nanoparticles, in: J.L. Anderson, A. Berthod, V. Pino, A. Stalcup (Eds.),
532 Analytical Separation Science, vol. 5, John Wiley & Sons, Hoboken, New Jersey,
533 USA, 2015, pp. 1681–1724.
- 534 [21] M.K. Bojdi, M. Behbahani, G. Hesam, M.H. Mashhadizadeh, Application of
535 magnetic lamotrigine-imprinted polymer nanoparticles as an electrochemical
536 sensor for trace determination of lamotrigine in biological samples, RSC Adv. 6
537 (2016) 32374–32380.
- 538 [22] M. Behbahani, S. Bagheri, M.M. Amini, H.S. Abandansari, H.R. Moazami, A.
539 Bagheri, Application of a magnetic molecularly imprinted polymer for the
540 selective extraction and trace detection of lamotrigine in urine and plasma samples,
541 J. Sep. Sci. 37 (2014) 1610–1616.
- 542 [23] M.D. Farahani, F. Shemirani, Supported hydrophobic ionic liquid on magnetic
543 nanoparticles as a new sorbent for separation and preconcentration of lead and
544 cadmium in milk and water samples, Microchim. Acta 179 (2012) 219–226.

- 545 [24] J. Chen, X. Zhu, Magnetic solid phase extraction using ionic liquid-coated core–
546 shell magnetic nanoparticles followed by high-performance liquid
547 chromatography for determination of rhodamine B in food samples, *Food Chem.*
548 200 (2016) 10–15.
- 549 [25] M.J. Trujillo-Rodríguez, V. Pino, J.L. Anderson, J.H. Ayala, A.M. Afonso,
550 Double salts of ionic-liquid-based surfactants in microextraction: application of
551 their mixed hemimicelles as novel sorbents in magnetic-assisted micro-dispersive
552 solid-phase extraction for the determination of phenols, *Anal. Bioanal. Chem.* 407
553 (2015) 8753–8764.
- 554 [26] S. Dong, G. Huang, X. Wang, Q. Hu, T. Huang, Magnetically mixed
555 hemimicelles solid-phase extraction based on ionic liquid-coated Fe_3O_4
556 nanoparticles for the analysis of trace organic contaminants in water, *Anal.*
557 *Methods* 6 (2014) 6783–6788.
- 558 [27] Q. Cheng, F. Qu, N.B. Li, H.Q. Luo, Mixed hemimicelles solid-phase extraction
559 of chlorophenols in environmental water samples with 1-hexadecyl-3-
560 methylimidazolium bromide-coated Fe_3O_4 magnetic nanoparticles with high-
561 performance liquid chromatographic analysis, *Anal. Chim. Acta* 715 (2012) 113–
562 119.
- 563 [28] K.D. Clark, O. Nacham, H. Yu, T. Li, M.M. Yamsek, D.R. Ronning, J.L.
564 Anderson, Extraction of DNA by magnetic ionic liquids: tunable solvents for
565 rapid and selective DNA analysis, *Anal. Chem.* 87 (2015) 1552–1559.
- 566 [29] K.D. Clark, M.M. Yamsek, O. Nacham, J.L. Anderson, Magnetic ionic liquids as
567 PCR-compatible solvents for DNA extraction from biological samples, *Chem.*
568 *Commun.* 51 (2015) 16771–16773.

- 569 [30] G. Cheng, Q. Zhang, B. Bai, Removal of Hg^0 from flue gas using Fe-based ionic
570 liquid, *Chem. Eng. J.* 252 (2014) 159–165.
- 571 [31] P. Brown, A. Bushmelev, C.P. Butts, J.Cheng, J. Eastoe, I. Grillo, R.K. Heenan,
572 A.M. Schmidt, Magnetic control over liquid surface properties with responsive
573 surfactants, *Angew. Chem. In. Ed.* 51 (2012) 2414–2416.
- 574 [32] J. Wang, H. Yao, Y. Nie, L. Bai, X. Zhang, J. Li, Application of iron-containing
575 magnetic ionic liquids in extraction process of coal direct liquefaction residues,
576 *Ind. Eng. Chem. Res.* 51 (2012) 3776–3782.
- 577 [33] W. Jiang, W. Zhu, H. Li, J. Xue, J. Xiong, Y. Chang, H. Liu, Z. Zhao, Fast
578 oxidative removal of refractory aromatic sulfur compounds by a magnetic ionic
579 liquid, *Chem. Eng. Technol.* 37 (2014) 36–42.
- 580 [34] N. Deng, M. Li, L. Zhao, C. Lu, S.L. de Rooy, I.M. Warner, Highly efficient
581 extraction of phenolic compounds by use of magnetic room temperature ionic
582 liquids for environmental remediation, *J. Hazard. Mater.* 192 (2011) 1350–1357.
- 583 [35] Y. Wang, Y. Sun, B. Xu, X. Li, R. Jin, H. Zhang, D. Song, Magnetic ionic liquid-
584 based dispersive liquid–liquid microextraction for the determination of triazine
585 herbicides in vegetable oils by liquid chromatography, *J. Chromatogr. A* 1373
586 (2014) 9–16.
- 587 [36] Y. Wang, Y. Sun, B. Xu, X. Li, X. Wang, H. Zhang, D. Song, Matrix solid-phase
588 dispersion coupled with magnetic ionic liquid dispersive liquid-liquid
589 microextraction for the determination of triazine herbicides in oilseeds, *Anal.*
590 *Chim.Acta* 888 (2015) 67–74.
- 591 [37] T. Chatzimitakos, C. Binellas, K. Maidatsi, C. Stalikas, Magnetic ionic liquid in
592 stirring-assisted drop-breakup microextraction: Proof-of-concept extraction of

593 phenolic endocrine disrupters and acidic pharmaceuticals, *Anal. Chim. Acta* 910
 594 (2016) 53–59.

595 [38] J.-T. Liu, G.-B. Jiang, Y.-G. Chi, Y.-Q. Cai, Q.-X. Zhou, J.-T. Hu, Use of ionic
 596 liquids for liquid-phase microextraction of polycyclic aromatic hydrocarbons,
 597 *Anal. Chem.* 75 (2003) 5870–5876.

598 [39] Y. Shi, H. Wu, C. Wang, X. Guo, J. Du, L. Du, Determination of polycyclic
 599 aromatic hydrocarbons in coffee and tea samples by magnetic solid-phase
 600 extraction coupled with HPLC–FLD, *Food Chem.* 199 (2016) 75–80.

601 [40] C.-E. Nika, E. Yiantzi, E. Pisillakis, Plastic pellets sorptive extraction: Low-cost,
 602 rapid and efficient extraction of polycyclic aromatic hydrocarbons from
 603 environmental waters. *Anal. Chim. Acta* 922 (2016) 30–36.

604 [41] M. Fernández-Amado, M.C. Prieto-Blanco, P. López-Mahía, S. Muniategui-
 605 Lorenzo, D. Prada-Rodríguez, A novel and cost-
 606 effective method for the determination of fifteen polycyclic
 607 aromatic hydrocarbons in low volume rainwater samples, *Talanta* 155 (2016) 175–
 608 184.

609 [42] Q.Q. Baltazar, J. Chandawalla, K. Sawyer, J.L. Anderson, Interfacial and micellar
 610 properties of imidazolium-based monocationic and dicationic ionic liquids,
 611 *Colloid Surf. A-Physicochem. Eng. Asp.* 302 (2007) 150–156.

612 [43] M. Germán-Hernández, P. Crespo-Llabrés, V. Pino, J.H. Ayala, A.M. Afonso,
 613 Utilization of an ionic liquid *in situ* preconcentration method for the determination
 614 of the 15 + 1 European Union polycyclic aromatic hydrocarbons in drinking water
 615 and fruit-tea infusions, *J. Sep. Sci.* 36 (2013) 2496–2506.

- 616 [44] H. Yan, H. Wang, Recent development and applications of dispersive liquid-
617 liquid microextraction, *J. Chromatogr. A* 1295 (2013) 1–15.
- 618 [45] E. Yilmaz, M. Soylak, Ionic liquid-linked dual magnetic microextraction of
619 lead(II) from environmental samples prior to its micro-sampling flame atomic
620 absorption spectrometric determination, *Talanta* 116 (2013) 882–886.
- 621 [46] S. Kamran, M. Asadi, G. Absalan, Adsorption of folic acid, riboflavin, and
622 ascorbic acid from aqueous samples by Fe_3O_4 magnetic nanoparticles using ionic
623 liquid as modifier, *Anal. Methods* 6 (2014) 798–806.
- 624 [47] F. Galán-Cano, M.C. Alcudia-León, R. Lucena, S. Cárdenas, M. Valcárcel, Ionic
625 liquid coated magnetic nanoparticles for the gaschromatography/mass
626 spectrometric determination of polycyclicaromatic hydrocarbons in waters, *J.*
627 *Chromatogr. A* 1300 (2013) 134–140.
- 628 [48] M. Valipour, Optimization of neutral networks for precipitation analysis in a
629 humid region to detect drought and wet year alarms, *Meteorol. Appl.* 23 (2016)
630 91–100.
- 631 [49] AOAC Peer-Verified Methods Program, Manual on policies and procedures,
632 AOAC International, Gaithersburg, 1998.

Figure Captions

Fig.1 Scheme of the overall MIL-based microextraction method-HPLC-FD performed under optimum conditions.

Fig.2 Comparison of the extraction efficiency of the tested MILs in the microextraction-HPLC-FD method for PAHs determination. The normalized peak area is defined as the chromatographic peak area of each PAH divided by the amount of MIL weighed in each experiment. Fixed experimental conditions ($n = 3$): 10 mL of ultrapure water containing $5 \mu\text{g}\cdot\text{L}^{-1}$ of PAHs, spiked with 0.5 mL of NaOH solution of $0.5 \text{ mol}\cdot\text{L}^{-1}$; 25% (w/v) NaCl; 20 μL of MIL A, MIL B or MIL C (depending on the experiment); 500 μL of acetone as dispersive solvent; 5 min of vortex; magnetic separation (few seconds); two back-extraction steps using 500 μL of n-hexane and 1 min of vortex; evaporation; reconstitution in 500 μL of acetonitrile; filtration; and HPLC injection.

Fig.3 Study of the effect of the nature of the dispersive solvent in the MIL-based microextraction-HPLC-FD method. Fixed experimental conditions ($n = 3$): 10 mL of ultrapure water containing $5 \mu\text{g}\cdot\text{L}^{-1}$ of PAHs, spiked with 0.5 mL of NaOH solution of $0.5 \text{ mol}\cdot\text{L}^{-1}$; 25% (w/v) NaCl; 20 μL of MIL A; 500 μL of $15 \text{ mmol}\cdot\text{L}^{-1}$ $[\text{C}_{16}\text{MIm}][\text{Br}]$, $2.1 \text{ mmol}\cdot\text{L}^{-1}$ of $[\text{C}_{12}\text{C}_{12}\text{Im}][\text{Br}]$, or acetone; 5 min of vortex; magnetic separation (few seconds); two back-extraction steps with 500 μL of n-hexane and 1 min of vortex; evaporation; reconstitution in 500 μL of acetonitrile; filtration; and HPLC injection.

Figure 1

Extraction

Back-extraction

Figure 2

Figure 3

656 **Table 1.** Chemical structures and properties of the three MILs studied [16].

MIL	Chemical formula	MW (g·mol ⁻¹)	Thermal stability (° C)	μ_{eff}^a (μ_B)	Structure
MIL A	[N _{8,8,8,B}][FeCl ₃ Br]	689.9	258	5.26	
MIL B	[N _{8,8,8,MOB}][FeCl ₃ Br]	716.9	203	5.60	
MIL C	[(BBnIm) ₂ C ₁₂][NTf ₂][FeCl ₃ Br]	1107.1	314	5.45	

^aeffective magnetic moment, measured at 295 K

Table 2 Analytical performance of the overall MIL-based microextraction-HPLC-FD method. The linearity range was 0.05–5.0 $\mu\text{g}\cdot\text{L}^{-1}$ for Chy and 0.05–3.5 $\mu\text{g}\cdot\text{L}^{-1}$ for the remaining PAHs, using 7 calibration levels.

PAH	(Slope $\pm$ SD ^a) $\cdot 10^{-6}$	R ^b	LOD ^c ($\mu\text{g}\cdot\text{L}^{-1}$)	LOQ ^d ($\mu\text{g}\cdot\text{L}^{-1}$)	Low spiked level (0.4 $\mu\text{g}\cdot\text{L}^{-1}$)				Intermediate spiked level (1.5 $\mu\text{g}\cdot\text{L}^{-1}$)			
					E _F ^e	E _R ^f (%)	RR ^g (%)	RSD intra-day ^h / inter-day ⁱ (%)	E _F ^e	E _R ^f (%)	RR ^g (%)	RSD intra-day ^h (%)
BaA	1.5 $\pm$ 0.1	0.996	0.01	0.03	10.3	29.3	105	5.3 / 7.1	11.2	32.0	111	9.1
Chy	0.20 $\pm$ 0.01	0.995	0.01	0.05	13.9	39.6	98.9	1.0 / 10	12.3	35.3	107	9.6
BbF	0.43 $\pm$ 0.02	0.994	0.005	0.02	12.7	36.3	119	3.0 / 16	10.7	30.6	120	7.5
BkF	1.9 $\pm$ 0.1	0.994	0.01	0.04	8.84	25.3	91.5	13 / 16	8.70	24.9	99.0	13
BaPy	1.2 $\pm$ 0.1	0.994	0.02	0.05	16.6	47.3	107	6.2 / 12	11.6	33.2	89.8	9.2

^a Standard deviation of slope.

^b Correlation coefficient.

^c Limits of detection, calculated as 3 times the signal to noise ratio.

^d Limits of quantification, calculated as 10 times the signal to noise ratio.

^e Enrichment factor.

^f Extraction efficiency (E_{F,max} = 35.0).

^g Relative recovery.

^h Relative standard deviation (n = 3).

ⁱ Relative standard deviation (n = 9), calculated in 3 non-consecutive days.