Title: Impact of different eddy covariance sensors, site set-up, and maintenance on the annual balance of CO₂ and CH₄ in the harsh Arctic environment

Running head: Arctic eddy covariance sensor comparison

¹Goodrich JP, ^{1,2}Oechel WC, ³Gioli B, ^{1,4}Moreaux V, ¹Murphy PC, ⁵Burba G, ^{1,6}Zona D

¹Global Change Research Group, Dept. of Biology, San Diego State University, San Diego, CA, 92182, USA

²Department of Environment, Earth and Ecosystems, The Open University, Milton Keynes, England, UK MK7 6AA

³IBIMET-CNR, Istituto di Biometeorologia, Consiglio Nazionale delle Ricerche, Via G. Caproni 8, 50145 Firenze, Italy

⁴Now at: INRA. UMR 1137. Forest Ecology and Ecophysiology, Centre de Nancy, F-54280 Champenoux, France

⁵Science & Technology, LI-COR Biosciences, 4421 Superior St., Lincoln, NE, 68504, USA

⁶Dept. of Animal and Plant Sciences, University of Sheffield, Western Bank, Sheffield, S10 2TN, United Kingdom

Corresponding author:

Jordan P. Goodrich

jpgoodrich@mail.sdsu.edu

(603) 686-4971

Key words: Arctic, eddy covariance, open-path, closed-path, sonic anemometer

Abstract

Improving year-round data coverage for CO₂ and CH₄ fluxes in the Arctic is critical for refining the global C budget but continuous measurements are very sparse due to the remote location limiting instrument maintenance, to low power availability, and to extreme weather conditions. The need for tailoring instrumentation, site set up, and maintenance at different sites can add uncertainty to estimates of annual C budgets from different ecosystems. In this study, we investigated the influence of different sensor combinations on fluxes of sensible heat, CO₂, latent heat (LE), and CH₄, and assessed the differences in annual CO₂ and CH₄ fluxes estimated with different instrumentation at the same sites. Using data from four sites across the North Slope of Alaska, we found that annual CO₂ fluxes estimated with heated ($7.5 \pm 1.4 \text{ gC m}^{-2} \text{ yr}^{-1}$) and non-heated ($7.9 \pm 1.3 \text{ gC m}^{-2} \text{ yr}^{-1}$) anemometers were within uncertainty bounds. Similarly, despite elevated noise in 30-min flux data, we found that summer CO₂ fluxes from open ($-17.0 \pm 1.1 \text{ gC m}^{-2} \text{ yr}^{-1}$) and close-path ($-14.2 \pm 1.7 \text{ gC m}^{-2} \text{ yr}^{-1}$) gas analyzers were not significantly different. Annual CH₄ fluxes were also within uncertainty bounds when comparing both open ($4.5 \pm 0.31 \text{ gC m}^{-2} \text{ yr}^{-1}$) and closed-path ($4.9 \pm 0.27 \text{ gC m}^{-2} \text{ yr}^{-1}$) gas analyzers as well as heated ($3.7 \pm 0.26 \text{ gC m}^{-2} \text{ yr}^{-1}$) and non-heated ($3.7 \pm 0.28 \text{ gC m}^{-2} \text{ yr}^{-1}$) anemometers. A continuously heated anemometer increased data coverage (64%) relative to non-heated anemometers (47-52%). However, sensible heat fluxes were over-estimated by 12%, on average, with the heated anemometer, contributing to the overestimation of CO₂, CH₄, and LE fluxes (mean biases of $-0.03 \text{ } \mu\text{mol m}^{-2} \text{ s}^{-1}$, $-0.05 \text{ mgC m}^{-2} \text{ hr}^{-1}$, and -3.77 W m^{-2} , respectively). To circumvent this potential bias and reduce power consumption, we implemented an intermittent heating strategy whereby activation only occurred when ice or snow blockage of the transducers was detected. This resulted in comparable coverage (50%) during winter to the continuously heated anemometer (46%), while avoiding flux over-estimation. Closed and open-path analyzers showed good

54 agreement, but data coverage was generally greater when using closed-path, especially during
55 winter. Winter data coverage of 26-32% was obtained with closed-path devices, vs 10-14%
56 for the open-path devices with unheated anemometers or up to 46% and 35% using closed
57 and open-path analyzers, respectively with heated anemometers. Accurate estimation of LE
58 remains difficult in the Arctic due to strong attenuation in closed-path systems, even when
59 intake tubes are heated, and due to poor data coverage from open-path sensors in such a harsh
60 environment.

1. Introduction

Assessment of Arctic ecosystem-atmosphere carbon (C) exchange is critical for refining the global C budget (IPCC 2007, Fisher et al. 2014). Despite the importance of both CO₂ and CH₄ emissions from the Arctic and their sensitivity to climate change (Mastepanov et al., 2013; Ueyama et al., 2013), their annual balances are still largely uncertain (Melton et al., 2013; Fisher et al. 2014). Although some researchers have had success measuring ecosystem-scale Arctic CO₂ and CH₄ fluxes (Oechel et al., 2014; Kutzbach et al., 2007; Parmentier et al. 2011, Emmerton et al., 2015 Zona et al. 2014, 2016), spatial and temporal data coverage is still sparse and year-round coverage is especially lacking (Wille et al., 2008; Oechel et al., 2014; Euskirchen et al., 2012; Luers et al., 2014). The scarcity of continuous, year-round measurements in the Arctic is due to the extremely harsh environmental conditions, especially in winter, and relative lack of infrastructure in these remote sites, preventing regular maintenance of the instruments or making it prohibitively expensive. These challenges have limited our ability to obtain accurate annual C budgets and assess interannual variability in greenhouse gas fluxes from Arctic regions.

Currently, important scientific questions on the terrestrial C cycle are based on multi-site syntheses of flux data from eddy covariance tower networks (e.g. FLUXNET, AmeriFlux, ICOS, AsiaFlux). Although there are ongoing efforts to standardize the sensors used in these networks (i.e. ICOS, NEON, etc.), a variety of different instruments, and site-dependent processing methods are still employed (Fratini et al. 2014). In addition, as technology improves and new sensor models become available, instrumentation is often upgraded after years of field deployment at long-term sites (e.g., Burns et al., 2014). Such methodological and instrumental differences and updates may contribute to observed differences in seasonal and annual C budgets, and add uncertainty when comparing C fluxes from different sites or ecosystems, as well as among years at long-term sites. Bias in

measurements due to instrumentation changes becomes particularly important in the Arctic where C fluxes can be very small (especially in winter), and can be considerably influenced by post-processing corrections (Oechel et al., 2014).

Additionally, the configuration of gas analyzers and sonic anemometers at a given site often needs to be tailored to the specific requirements of that environment and available infrastructure, and therefore may vary in different sites even across standardized networks (e.g. the boreal sites in ICOS will need to rely on heated sonic anemometers, which will not be needed in the more southern sites). With respect to gas analyzers, both open- and closed-path sensors have been deployed worldwide for CO₂ and H₂O flux measurements, and both have shown good performance in the Arctic (Nakai et al. 2011, 2013, Oechel et al. 2014; Zona et al., 2016). Multiple studies have compared open-path and closed-path analyzers in different conditions for CO₂ fluxes (Leuning and King 1992, Lee et al. 1994, Jarvi et al. 2009, Ueyama et al. 2012, Burns et al. 2014). Open-path analyzers typically require less overall maintenance, have considerably smaller power demand, better time response and smaller frequency corrections than closed-path systems due to the absence of pumps, filters, and intake tubes (Massman 1991, 2000). However, they require larger corrections for density fluctuations (Webb-Pearman-Leuning correction, WPL, Webb et al. 1980), and older models may need surface heating corrections, particularly during winter (Grelle and Burba, 2007; Burba et al., 2008; Oechel et al., 2014). Open-path sensors also lose more data during precipitation and under high humidity or fog (Jarvi et al. 2009). The annual data coverage of closed-path systems in harsh environments has been up to 70% (Goulden et al. 2006), while open-path designs resulted in overall annual data coverage of 44-68% (Oechel et al. 2014; Euskirchen et al., 2012) but as low as 15% during winter (Oechel et al. 2014). Increased maintenance frequency or winterization of the instrument can increase open-path data coverage, but these are often costly or impractical.

For CH₄ flux measurements, successful inter-comparisons of open-path and closed-path analyzers have also been performed (Detto et al. 2011, Peltola et al. 2013, Iwata et al. 2014). While the open-path design (LI-7700) can lead to substantial data losses due to precipitation, with data coverage as low as 25% in the harsh Arctic environment (Sturtevant et al., 2012), it is usually the only option for CH₄ flux measurements at remote sites due to low power consumption and autonomous operation (McDermitt et al., 2011; Burba, 2013). Furthermore, this sensor was successfully deployed in an alpine wetland (mean annual temperature = -1.1 °C) attaining data coverage up to 66% (Song et al., 2015). Generally, closed-path systems have better data coverage in the Arctic with 66 - 85% (Zona et al., 2016; Zona et al., 2009, Sachs et al., 2008), although depending on the set-up and maintenance schedule, associated technical and power supply issues can reduce closed-path data capture to 12 - 26% in Arctic and sub-Arctic sites (Hanis et al., 2013; Wille et al., 2008).

Various sonic anemometers have been used extensively in cold environments (Gazovic et al., 2013; Jackowicz-Korczynsky et al., 2010; Sturtevant et al., 2012, Zona et al., 2009, 2010, Rinne et al., 2007), but a major challenge for measuring fluxes in these regions is anemometer performance in extreme weather conditions when water, snow, and ice can block or divert the sonic signals from the transducers. In order to measure fluxes outside the summer period, the transducers of the sonic anemometer need to be maintained ice-free. Heating systems for these sensors have generally utilized heating tape wrapped around the anemometer and, less commonly, the hot film technology (Lekakis et al., 1989, Skelly et al., 2002). Since it has been shown that continuous heating of the anemometer may increase the apparent sensible heat fluxes (Tammelin et al., 1998; Skelly et al., 2002), an intermittent heating strategy needs to be explored at cold sites. Although multiple cross-comparisons of sonic anemometers have been performed in the past (e.g. Kochendorfer et al. 2012, Frank et al. 2013, El-Madany et al. 2013, Nakai et al. 2014), none have yet tested the impact of

heating on the sensible heat and gas fluxes in Arctic sites. One of the few commercially available self-heating anemometers is provided by Metek GmbH (uSonic-3 Class A), which initiates heating based on air temperature thresholds, whereby the heater is maintained on for temperatures below 4.5 °C. In the Arctic, this temperature-activated heating would result in continuous heating for the entire autumn, winter and spring, potentially affecting the fluxes during the most critical and uncertain periods, emphasizing the need for a better heating strategy.

To investigate how the choice of gas analyzers and sonic anemometers influences fluxes of CO₂ (F_{CO_2}), latent heat (LE), sensible heat (H), and CH₄ (F_{CH_4}) and to understand the potential influence of different configurations on the long term C budget, we compared several recently installed instrument sets to the historically operating sets at four flux sites in Arctic Alaska. Instrument configuration at each site was selected depending on availability of line power, climate conditions, and site accessibility. This study reports on the comparability over half-hourly scales for F_{CO_2} , LE, H, and F_{CH_4} , and on annual totals of F_{CO_2} and F_{CH_4} . The comparisons are organized as follows:

- i) Comparisons of fluxes derived from heated and non-heated anemometers from half-hourly (all fluxes) to annual time scales (for F_{CO_2} and F_{CH_4}).
- ii) Comparisons of F_{CO_2} and LE obtained from closed-path and (en)closed-path analyzers.
- iii) Comparisons of F_{CO_2} , and F_{CH_4} obtained from open-path and closed-path analyzers.

In addition to direct comparisons of final fluxes, we investigated the influence of site and sensor specific spectral corrections, which are known to vary greatly depending on the site (e.g., Ibrom et al., 2007; Leuning and King, 1992). The results of this study should provide useful and practical considerations for future studies as well as for ongoing instrumental

updates at long-term measurement sites. These may be particularly helpful for experimental design, station setup, configuration and maintenance planning, as well as data interpretation at other remote, high-latitude, cold sites experiencing long periods of small C fluxes.

2. Materials and methods

2.1. Study sites

The eddy covariance (EC) flux towers were located at four sites on the North Slope of Alaska: two in Barrow (CMDL and BES), one in Atkasuk (ATQ) and one in Ivotuk (IVO). The CMDL site (71°19'21.10'' N: 156°36'33.04'' W, 1 m elevation above sea level) is about 2 km south of the Arctic Ocean. This wet sedge tundra site is characterized by low species diversity, dominance of grasses and sedges, rare occurrences of tussock, and a lack of shrubs (Brown et al., 1980). Further site details can be found in Kwon et al. (2006). The BES site is 6.5 km south of the Arctic Ocean (71°16'51.17'' N, 156°35'47.28'' W, 3 m elevation) and dominant vegetation includes grasses, sedges, and mosses along with a few prostrate dwarf shrubs (Mullier et al. 1999, Raynolds et al. 2005, Zona et al. 2011). More site details are available in Zona et al. (2012). Both CMDL and BES are located near Barrow, have grid power and relatively easy access for instrumentation maintenance including during the winter. The ATQ site (70°28'10.64'' N: 157°24'32.21'' W, 24m elevation) is located approximately 100 km south of Barrow, also has access to line power, and is fairly accessible during the winter. This site is characterized by moist coastal sedge tundra with moist-tussock vegetation (Kwon et al. 2006). More information can be found in Oechel et al. (2014). The IVO site (68°29'11.36'' N: 155°45'0.79'' W, 543m elevation) is located approximately 300 km south of Barrow, at the foothills of the Brooks Mountain Range. Dominant vegetation includes tussock species, dwarf and creeping shrubs, mosses and lichen. More site information is provided in Zona et al., (2016). This site does not have line power access and

thus is powered by combination of two diesel generators, twelve solar panels, and a wind turbine. Access to IVO requires chartered flights to a remote air strip, limiting instrument maintenance to the summer period only.

2.2. Long-term instrumentation at study sites

The initial selection of instrumentation at each site was made largely based on the need to limit maintenance and servicing of the sensors as well as on power restrictions. We also endeavored to deploy the most appropriate instrumentation commercially available at the time of each set-up (spanning approximately 15 years). This necessitated different instrument configurations over time as well as among the sites (Table 1).

The CMDL site, established in 1997, was upgraded with an LI-7500 CO₂/H₂O analyzer in 2001 and with a Gill WindMaster Pro in 2012. The LI-7500 was then replaced by the (en)closed-path LI-7200 (Burba et al., 2010, 2012) gas analyzer in 2011, and an open-path LI-7700 CH₄ analyzer was installed in 2013. BES was established in summer 2005, and was initially equipped with a Gill WindMaster Pro anemometer and open-path LI-7500 CO₂/H₂O analyzer (Zona et al. 2012). ATQ was established in 1999, initially equipped with a Gill R3 anemometer and updated with an open-path LI-7500 CO₂/H₂O analyzer in 2001. The gas analyzer was then replaced by an (en)closed-path LI-7200 in 2011. The IVO site was first installed in 2003 and included an LI-7500 gas analyzer and Gill R3 anemometer. Data acquisition was nearly continuous at all of these sites, except IVO where data collection was stopped in summer 2008 and re-started in summer 2013. In all sites, new instrument models were added in summer/fall 2013 as described below. Data from October 2013 to July 2015 were used for the comparisons of sonic anemometers and gas analyzers in this study.

2.3. Instrumentation for sensor comparison

In 2013, a Metek uSonic-3 Class A non-orthogonal ultrasonic anemometer with self-heating feature (hereafter referred to as Metek) and a CSAT3 anemometer (non-orthogonal)

were installed at ATQ for comparison along with the non-orthogonal Gill R3. At ATQ, the Metek was heated continuously until March 2015, when the heating was switched off, whereas both the Gill R3 and CSAT3 were not heated (Table 1).

At the remote IVO site, a Metek anemometer was also deployed in 2013. To limit power consumption, we developed an intermittent heating strategy such that heating was activated only when the transducers were blocked as reported by analog data quality indicators, rather than the default activation scheme based on the sonic temperature. The impact of this power-efficient heating strategy was investigated with respect to its effectiveness for de-icing the anemometer, the resulting data coverage during cold-periods, and the quality of sensible heat flux measurements.

As part of a larger effort toward upgrading instrumentation, we installed closed-path LGR-FGGA-24EP CO₂/H₂O/CH₄ analyzers at the CMDL, BES, and ATQ sites. To assess data continuity and comparability, we compared CO₂ and LE fluxes from the LGR-FGGA-24EP at ATQ to those from the (en)closed-path LI-7200 system (Table 1). At BES, the CO₂ fluxes from the LGR-FGGA-24EP were compared to those from the open-path LI-7500 (Table 1). Due to condensation issues within the long inlet tube at BES (the wettest, inundated site), the signal lag between vertical wind and H₂O vapor concentrations became very large and insurmountable for spectral corrections. LE fluxes therefore, could not be used from the LGR-FGGA-24EP for direct comparison with the LI-7500 at that site. Finally, at CMDL, CH₄ fluxes from the LGR-FGGA-24EP were compared to those measured by the open-path LI-7700 analyzer (Table 1).

2.4. Instrumentation set-up

At ATQ, the LI-7200 analyzer utilized insulated unheated tube and rain cup of the larger pre-2013 design, which were subsequently replaced by a smaller improved design on 2 July 2014. We found that the larger intake tube and rain cup initially used, resulted in

substantially under-estimated turbulent exchange and fluxes of H₂O relative to the LGR-FGGA-24EP with heated tubing (Figures S1 and S2e). Therefore, only data collected after this change were used for the comparison (LI-7200/LGR-FGGA-24EP). The LI-7200-101 flow module was used to automatically regulate and maintain the flow rate at 15 l min⁻¹. The LGR-FGGA-24EP analyzer and associated dry-scroll vacuum pump, sampling air at a rate of 20 l min⁻¹ (N 940.5 APE-W, KNF Neuberger AG, Balterswil, Switzerland) were housed inside water-proof, insulated boxes (Grizzly Coolers, Decorah, Iowa). To minimize condensation inside the inlet tube of the LGR-FGGA-24EP analyzer, the tubing (PFA Tubing, 3/8 in. OD x 0.062 in., Swagelok, Solon, Ohio) was wrapped in heating tape, and both the inlet tube and the heating tape were insulated. The inlet tubing was terminated with an inverted funnel and protected with flexible mosquito netting to prevent water intake, ice formation at the inlet of the tubing, and mosquitos from entering the sample line. A 2µm stainless filter (SS-4FW-2 1/4T x 1/4T, Swagelok) was installed before the analyzer inlet to prevent sample cell contamination. For the cross-comparison of CO₂ and LE fluxes from the LI-7200 and LGR-FGGA-24EP analyzers at ATQ, we used three-dimensional wind speed from the CSAT3. Raw signals were collected at 10 Hz using a CR3000 (Campbell Scientific, Logan, UT, USA).

At BES, the set-up for the LGR-FGGA-24EP was similar to that in ATQ. The open-path LI-7500 at BES was mounted at a 20° angle to minimize water build-up on the windows of the analyzer. Three-dimensional wind speed was measured with the CSAT3 anemometer and raw signals of each instrument were collected at 10Hz using a CR3000.

At CMDL, the wind components from the Gill WindMaster Pro were used for flux calculations. Raw signals of the LI-7700 and WindMaster Pro were collected at 10Hz using the LI-7550 Analyzer Interface Unit, while a CR3000 also recorded data from the LGR-

FGGA-24EP and WindMaster Pro at 10 Hz. The automated LI7700 sensor mirror washer was activated based on signal strength of the analyzer to maximize data quality.

At IVO wind components from the intermittently heated Metek uSonic-3 Class A anemometer were used for flux calculations with CO₂/H₂O data from an (en)closed-path LI-7200 and CH₄ from an open-path LI-7700. Raw 10 Hz signals were collected with the LI-7550 Analyzer Interface Unit. A 5-gallon jug was installed to supplement the washer fluid basin used with the automated LI-7700 sensor mirror washer, which was activated based on the signal strength of the analyzer.

2.5. Data quality control, processing, and analyses

The LGR-FGGA-24EP sensors were calibrated by the manufacturer just before being shipped to Alaska. The LI-7500 and LI-7200 analyzers were all calibrated at least twice per year in the laboratory in Barrow during 2013 – 2015. Half-hourly eddy covariance fluxes were calculated using EddyPro® (www.licor.com/eddypro). De-spiking and absolute limit determination were included in the preliminary processing of raw signals (Vickers and Mahrt, 1997) and outliers were discarded. Angle of attack errors were corrected according to Nakai et al. (2006) and Nakai and Shimoyama (2012), respectively for the Gill R3 and WindMaster Pro anemometers. A double axis rotation of the wind vector was performed (Wilczac et al. 2001) and the block-averaging method was used to extract turbulent fluctuations from time series (Gash and Culf, 1996). Time lags between vertical wind speed and the variable of interest were determined for each averaging period by covariance maximization. Low frequency spectral corrections were applied according to the analytic method described in Moncrieff et al. (2004). High frequency spectral corrections were applied depending on the setup: the fully analytic method of Moncrieff et al. (1997) was adopted for the open-path systems (LI-7500, LI-7700), which includes a correction for sensor separation effects; an *in-situ* spectral correction method (Ibrom et al. 2007) was used for the closed-path analyzers

(LI-7200, LGR-FGGA-24EP) as it is a more suitable method to describe attenuation along the intake tube walls. For the closed-path analyzers, a correction was also applied to account for sonic anemometer and analyzer separation according to Horst and Lenschow (2009).

For the open-path LI-7500 analyzer at BES, the WPL correction (Webb et al. 1980) and the self-heating correction adapted to Arctic conditions were also applied (Burba et al. 2008, Oechel et al. 2014). Mixing ratio data were used from closed-path analyzers (LI7200) or data were converted to mixing ratios (LGR-FGGA), thus avoiding the need for WPL corrections. For the open-path LI-7700, a spectroscopic correction was computed with the WPL correction to account for the modification in the shape and width of the absorption line due to changes in temperature-pressure-water vapor (McDermitt et al. 2011). As a QA/QC test of the final fluxes, we used the standard flags (0-1-2) defined by Mauder and Foken (2004) and data with a flag = 2 were discarded. Remaining flux data were filtered for insufficient atmospheric turbulence (Reichstein et al., 2005) with a friction velocity threshold of 0.1 m s^{-1} . Remaining spikes were removed using a 30-day moving window that advanced one day at a time and any half-hours that exceeded ± 2 standard deviations from the mean for that half-hour were discarded.

In ATQ, the anemometer flux comparisons were limited to the wind sector between the roughly perpendicularly oriented CSAT3 and Metek (Table 1) to minimize flow distortion effects. We performed orthogonal regression analyses, as there was no true dependent variable or a control, and we therefore needed to account for errors in both flux estimates. Since R^2 values cannot be obtained from orthogonal regression, we have reported Pearson's correlation coefficient (r) associated with each comparison. In addition to the 1:1 regression comparisons, we analyzed the differences between flux measurement pairs by plotting the distribution of differences (Δ flux values) and fitting Laplace (double exponential) probability

density functions to obtain the mean difference (bias) and spread (variance) of differences for each comparison.

To compare annual estimates of CO₂ derived from various sensor combinations, we filled gaps in the half-hourly fluxes using the online eddy covariance gap-filling tool (<http://www.bgc-jena.mpg.de/~MDIwork/eddyproc>) which employs standard methods of Reichstein et al. (2005). Methane fluxes were gap-filled using an artificial neural network (ANN) (Dengel et al., 2009; Zona et al., 2016). Meteorological inputs to the ANN included air temperature, soil temperature at 10cm depth, photosynthetic photon flux density (PPFD), vapor pressure deficit (VPD), and two sonic cross wind components, (u and v). The ANN was run 25 times and the median value was used to fill gaps in half-hourly flux time series.

To assess the uncertainty associated with each annual sum, we applied the ‘paired days’ approach to estimate random uncertainty of measured fluxes (Hollinger and Richardson, 2005). We used a Monte Carlo simulation to sample randomly from double-exponential distributions defined by the sigma values that resulted from the paired-days analysis (Dragoni et al., 2007). The median half-hourly uncertainty from 250 simulations was used in the annual assessment. The uncertainty associated with the gap-filling approach was assessed by simulating gaps and comparing synthetic data to observations (Reichstein et al., 2005), and half-hourly uncertainties were propagated in quadrature (Taylor, 1997) to obtain annual values. All data analyses were performed with Matlab (R2014a, MathWorks, Natick, MA, USA).

3. Results

3.1. Sonic anemometer comparisons

The half-hourly sensible heat fluxes (H) calculated with the unheated anemometers (CSAT3 and Gill R3) revealed a good comparison, with a slope of 1.01 and Pearson’s r of

0.96 (Figure 1a). The sensible heat fluxes derived from the continuously heated Metek were higher on average than the (unheated) CSAT3, with a slope of 1.12 and intercept of 7.27 W m⁻² (Figure 1b). Further comparison between these two sonic anemometers, revealed that differences in the variance in sonic temperature (T_s) were negligible (Figure 2a), whereas there was higher variance in the vertical wind component (w) measured by the heated Metek than the CSAT3 (red line; slope = 1.17 and intercept = -0.01 m² s⁻²) (Figure 2b).

We explored the effect of heating the Metek at ATQ and potential influence of over-estimated fluctuations in w on the gas fluxes by comparing F_{CO_2} , LE, and F_{CH_4} derived from both the CSAT3 and the heated Metek, paired with the LGR-FGGA-24EP closed-path gas analyzer. We found that the heated Metek resulted in higher LE than the unheated CSAT3-derived fluxes, with a slope of 1.19 (Figure 3a) and a mean bias (ΔLE) of -3.77 W m⁻² (Figure 4a). Discrepancies in CO₂ and CH₄ fluxes from the two sensor pairs were smaller, with slopes of 1.09 and 1.07 in the regressions, respectively (Figure 3c and e), and delta flux values were also small for F_{CO_2} (-0.03 $\mu\text{mol m}^{-2} \text{s}^{-1}$) and F_{CH_4} (-0.05 mgC m⁻² hr⁻¹) (Figure 4c and e). Given this smaller offset in the direct comparison of C fluxes, the annual sums of F_{CO_2} from these two sensor pairs were also very similar (within 5%). Specifically, from 1 October 2013 to 30 September 2014 at ATQ, the heated Metek – LGR sensor pair resulted in an estimated loss of $7.5 \pm 1.4 \text{ gC-CO}_2 \text{ m}^{-2} \text{ yr}^{-1}$ and the CSAT3 – LGR sensor pair resulted in an estimated loss of $7.9 \pm 1.3 \text{ gC-CO}_2 \text{ m}^{-2} \text{ yr}^{-1}$ (Table 2). This small difference in annual F_{CO_2} resulted from the compensating effect of slightly higher uptake during the day and slightly higher losses at night estimated from the Metek-derived fluxes, a consequence of the higher variance in the vertical wind component (w). For example, winter CO₂ losses were 9.2% higher and summer uptake was 6.7% higher when estimated from the Metek – LGR sensor pair (Figure 5a). Additionally, when isolating data from June, July and August 2014, the

mean daytime CO₂ uptake and mean night time CO₂ losses were also approximately 0.1 μmol m⁻² s⁻¹ larger from the Metek – LGR pair.

The annual F_{CH_4} at ATQ from the heated Metek and unheated CSAT3 were identical (3.7 gC m⁻² yr⁻¹) (Table 2). Furthermore, there was very little temporal shift in the F_{CH_4} cumulative time series from these two sensor pairs (Figure 5b). Therefore, timing and seasonal comparisons are also constrained for F_{CH_4} between the heated and unheated setup.

To separate the impact of heating from the potential differences in the performance of the Metek and CSAT3, we de-activated heating of the Metek from 17 March 2015 to the present day. Using the same QA/QC protocol, we found better agreement in H, whereby the regression slope was reduced 1.06 (Figure 1b), suggesting the continuous heating was responsible for at least half of the over-estimation of $\sigma^2 w$ (Figure 2b), in addition to reducing heating-related errors. However, over-estimation of LE persisted even after the heating of the Metek was deactivated (slope = 1.20, Figure 3b), with only a slight improvement in the mean bias to -3.21 W m⁻² (Figure 4b). Finally, the comparison of F_{CO_2} and F_{CH_4} between the heated and unheated setup remained very good (slopes of 1.06 and 0.99, respectively) (Figure 3d and f). The mean bias for F_{CO_2} dropped to 0, while that for F_{CH_4} was reduced to -0.01 mgC m⁻² hr⁻¹ (Figure 4d and f).

The intermittent heating system for the Metek at IVO successfully de-iced the transducers during cold periods (Figure 6), bypassing the observed biases to the fluxes. The very noisy flux data associated with ice build-up was minimized during the short periods when heating was activated, and the post-heating sensible heat fluxes were comparable to pre-heating values (Figure 6), suggesting only a temporary influence, and no long-term biases on the fluxes. Despite losing 3.2% of data from the Metek at IVO during heating activation (which we removed during post-processing), the annual data coverage of H from that site was the same as that of the continuously heated Metek at ATQ (64%) and the winter data

coverage was similar (50% and 46% for IVO and ATQ, respectively) (Table 3). However, it should be noted that due to unrelated technical issues at ATQ that led to data losses during winter, the available data before QA/QC and filtering was lower at that site than IVO (Table 3).

3.2. Gas analyzer comparisons

3.2.1. Closed-path and (en)closed-path CO₂ and H₂O analyser comparison

We compared fluxes from the closed-path LGR-FGGA-24EP and the (en)closed-path LI-7200 at ATQ using the CSAT3 anemometer. F_{CO_2} was very similar between the two sensor pairs with a slope of 0.99 and -0.01 intercept ($r = 0.87$, $P < 0.001$) (Figure 7a). The ΔF_{CO_2} distribution also suggested very low bias ($\mu = 0.01 \text{ } \mu\text{mol m}^{-2} \text{ s}^{-1}$) with generally tight spread ($\sigma = 0.53 \text{ } \mu\text{mol m}^{-2} \text{ s}^{-1}$) (Figure 8a). The resulting annual total CO₂ fluxes derived from the two sensor pairs were within $1.4 \text{ gC m}^{-2} \text{ yr}^{-1}$, with the LI-7200 resulting in slightly larger CO₂ loss ($9.3 \pm 1.1 \text{ gC m}^{-2} \text{ yr}^{-1}$) than the LGR-FGGA-24EP ($7.9 \pm 1.3 \text{ gC m}^{-2} \text{ yr}^{-1}$) (Table 2). Primary differences in the cumulative CO₂ flux trajectories were slightly higher respiratory losses of late autumn CO₂ from the LGR-FGGA-24EP with subsequent greater spring CO₂ uptake (Figure 5a). Latent heat fluxes at ATQ from the LI-7200 were generally somewhat higher than from the LGR-FGGA-24EP with a slope of 0.86 from the regression ($r = 0.98$, $P < 0.001$) (Figure 7c). This was also reflected in the mean bias in ΔLE of 3.22 W m^{-2} , with tight spread ($\sigma = 7.60 \text{ W m}^{-2}$) (Figure 8c).

3.2.2. Open-path and closed-path CO₂ and CH₄ analyser comparison

We used data from BES to compare F_{CO_2} derived from the closed-path LGR-FGGA-24-EP and the open-path LI-7500, both paired with the CSAT3 anemometer. There were less available data from the LI-7500 and more noise in fluxes from both sensor pairs at BES than was observed at ATQ, and thus the amount of data used in these comparisons was smaller

(Table 3). As a result, the regression comparison showed considerable scatter with a slope of 1.28 and 0.29 $\mu\text{mol m}^{-2} \text{s}^{-1}$ intercept ($r = 0.61$, $P < 0.001$) (Figure 7b). Despite the high degree of spread in the distribution of ΔF_{CO_2} ($\sigma = 1.13 \mu\text{mol m}^{-2} \text{s}^{-1}$), the mean bias was low ($\mu = -0.04 \mu\text{mol m}^{-2} \text{s}^{-1}$) (Figure 8b).

Ice and snow build-up on the windows of the open-path LI-7500, for which no winterization was attempted during 2013-2015, led to particularly poor data coverage during cold periods at BES and as little as 10% of CO_2 flux data remained after quality control filtering for the period from 1 October 2013 to 23 Apr 2015 (i.e. winter). We therefore estimated integral CO_2 fluxes only for late spring and summer (1 May to 31 October 2015) at BES. Total flux calculated for this period from the LGR-FGGA-24-EP was $-14.2 \pm 1.7 \text{ gC m}^{-2} \text{ summer}^{-1}$ and from the LI-7500 was $-17.0 \pm 1.1 \text{ gC m}^{-2} \text{ summer}^{-1}$, with most of this difference derived from slightly larger spring thaw-out CO_2 losses and summer uptake measured by the LI-7500 (Figure 5c). Generally, despite potentially large differences in instantaneous measurements between these two set-ups, and larger spectral corrections required for LGR-FGGA-24EP flux calculations (Table 4), mean flux values from longer periods (seasonal) converge to values within uncertainty ranges.

Methane fluxes were compared at CMDL using the LGR-FGGA-24EP closed-path analyser and the LI-7700 open-path sensor. The regression comparison resulted in a slope of 0.86 with an intercept of 0.05 $\text{mgC m}^{-2} \text{ hr}^{-1}$ ($r = 0.68$, $P < 0.001$) (Figure 7d). The 0.02 $\text{mgC m}^{-2} \text{ hr}^{-1}$ mean ΔF_{CH_4} suggested a very small bias toward higher fluxes derived from the open-path LI-7700 (Figure 8d). However, the data coverage in CMDL was markedly better from the LGR-FGGA-24EP with 54% coverage over the entire measurement year, whereas coverage was 26% for the open-path LI-7700 (Table 3). As expected, the seasonality in flux data coverage from these two sensors showed much higher cold-season data capture rates with the closed-path LGR-FGGA-24EP. Using the seasonal breakdown based on PPFD and

temperature described in Oechel et al. (2014), we found the LGR-FGGA-24EP data coverage was 40, 58, 81, and 65% for winter, spring, summer, and fall, respectively. Whereas the data capture rates from the LI-7700 at CMDL during those seasons were 14, 46, 49, and 27%, respectively. On the other hand, at IVO the annual data coverage obtained by the open-path LI-7700 paired with the intermittently heated Metek anemometer was 47% and winter coverage was 35% (Table 3), indicating that good coverage is possible with open-path CH_4 analysers farther inland from the coast, with lower humidity (annual average RH of 75% and 87% at IVO and CMDL, respectively) and less influence of salt spray. Furthermore, despite the small bias in the fluxes at CMDL toward higher LI-7700 F_{CH_4} , and lower data coverage during cold periods, the annual sums from the open and closed-path set-ups were not significantly different (Table 2), with only slight deviations in the cumulative trajectory in winter and autumn (Figure 5d).

Table 1. Sensor configurations and distances between analyzers at each site. Heated intake tubes are indicated by subscript ‘H’. The LGR-FGGA-24EP was manufactured by Los Gatos Research Inc., CA, USA; the LI-7500, LI-7200, and LI-7700 by LI-COR Biosciences Inc., NE, USA; the CSAT3 by Campbell Scientific Inc., UT, USA; the Gill WMP and R3 by Gill Instruments Ltd., Hampshire, UK; and the Metek uSonic-3 Class A by Metek GmbH, Elmshorn, DE.

Sonic anemometers

Site	Model	Height (m)	Orientation (°)	Heating (dist. to Metek (m))
CMDL	Gill WMP ^a	4.17	35	non-heated
BES	CSAT3 ^a	2.18	60	non-heated
	CSAT3 ^a	2.28	175	non-heated (0.35)
ATQ	Gill R3	2.28	0	non-heated (0.56)
	Metek	2.28	94	continuous
IVO	Metek ^a	3.42	205	intermittent

Gas analyzers

Site	Model	Height (m)	Gas species	Tube length (m) (dist. to CP (m)) ^b	Dist. to primary anemometer (m)
CMDL	LGR-FGGA-24EP	4.18	CO ₂ /H ₂ O/CH ₄	5.71 _H	0.18
	LI-7700	4.12	CH ₄	(0.23)	0.23
BES	LI-7500	1.60	CO ₂ /H ₂ O	(0.37)	0.20
	LGR-FGGA-24EP	2.00	CO ₂ /H ₂ O/CH ₄	4.50 _H	0.17
ATQ	LI-7200	2.25	CO ₂ /H ₂ O	1.20 (0.07)	0.44
	LGR-FGGA-24EP	2.30	CO ₂ /H ₂ O/CH ₄	3.12 _H	0.43
IVO	LI-7200	3.22	CO ₂ /H ₂ O	1.20	0.14
	LI-7700	3.12	CH ₄	-	0.45

^aPrimary anemometer used for flux calculation in gas analyzer comparisons.

^bDistance to the closed or (en)closed-path (CP) analyzer used in comparisons.

444
445
446
447

Table 2. Annual total F_{CO_2} and F_{CH_4} estimated with different sensor combinations at ATQ, BES, and CMDL. See section 2.5 for a description of the uncertainty estimation.

Site	Gas analyzer	Anemometer	Annual NEE [gC m ⁻² yr ⁻¹]	Annual F_{CH_4} [gC m ⁻² yr ⁻¹]
ATQ	LGR	CSAT	7.9 (±1.3)	3.7 (±0.28)
	LGR	Metek (heated)	7.5 (±1.4)	3.7 (±0.26)
	LI-7200	CSAT	9.3 (±1.1)	-
BES*	LGR	CSAT	-14.2 (±1.7)	-
	LI-7500	CSAT	-17.0 (±1.1)	-
CMDL	LGR	Gill WMP	-	4.9 (±0.27)
	LI-7700	Gill WMP	-	4.5 (±0.31)

* Flux integrals for BES were calculated for 1 May to 31 October.

Table 3. Data coverage for fluxes calculated with each sensor configuration. The continuously heated Metek is indicated by (con), the intermittently heated Metek is indicated by (int), closed- and (en)closed-path gas analyzers are indicated by (CP), and open-path analyzers are indicated by (OP). The available data column excludes only those for which a flux could not be calculated due to loss of power (< 5% for all sites), instrument malfunction, and active rain or heavy fog, which causes high frequency data to be out of range.

Site	Flux	Sensor (pair)	Available data (%)	Annual coverage following QA/QC, spike removal (%)	Winter coverage (%)
ATQ	H	CSAT	66	52	31
	H	Metek (con)	76	64	46
	H	Gill R3	64	47	25
IVO	H	Metek (int)	91	64	50
BES	F_{CO_2}	LI-7500 (OP) - CSAT	51	30	10
	F_{CO_2}	LGR (CP) - CSAT	88	70	41
ATQ	F_{CO_2}	LI-7200 (CP) - CSAT	66	48	27
	F_{CO_2}	LGR (CP) - CSAT	53	53	32
	F_{CO_2}	LGR (CP) - MTK (con)	76	61	46
ATQ	LE	LI-7200 (CP) - CSAT	66	46	26
	LE	LGR (CP) - CSAT	66	52	35
	LE	LGR (CP) - MTK (con)	76	67	46
ATQ	F_{CH_4}	LGR (CP) - CSAT	66	52	31
	F_{CH_4}	LGR (CP) - MTK (con)	76	61	44
BRW	F_{CH_4}	LI-7700 (OP) - Gill WP	27	26	14
	F_{CH_4}	LGR (CP) - Gill WP	55	54	40
IVO	F_{CH_4}	Metek (int)	65	47	35

Table 4. Comparison of summary statistics associated with spectral correction factors applied to fluxes at each site with different gas analyzers.

Site	Flux & sensor	Spectral correction factors		
		Median	1st quartile	3rd quartile
BES	F_{CO_2} LI-7500	1.14	1.13	1.16
	F_{CO_2} LGR	1.29	1.23	1.35
ATQ	F_{CO_2} LI-7200	1.15	1.11	1.20
	F_{CO_2} LGR	1.34	1.25	1.47
BES	LE LI-7500	1.14	1.13	1.15
	LE LGR	3.33	2.63	3.99
ATQ	LE LI-7200	2.34	1.99	2.89
	LE LGR	2.07	1.78	2.48
CMDL	F_{CH_4} LI-7700	1.15	1.13	1.17
	F_{CH_4} LGR	1.49	1.38	1.65

460
461
462

Figure captions

Figure 1. Comparisons of half-hourly sensible heat fluxes (H) at ATQ derived from the (unheated) CSAT3 anemometer and the (unheated) Gill R3 ($n = 634$) (a) and the heated Metek ($n = 634$) (b) from 1 October 2013 to 30 September 2014. The red line in panel b corresponds to the heated Metek data while the grey line and symbols represent data after heating of the Metek at ATQ was fully de-activated from 17 March to 11 June 2015 ($n = 581$). Regression coefficients with the ‘con’ subscript indicates results from the continuously heated Metek data. In this and all of the following regression figures, ‘I’ denotes the fitted intercept in units given on the x and y-axes, and ‘S’ denotes the slope of the regression.

Figure 2. Comparisons of the CSAT3 and the heated Metek at ATQ of half-hourly variance in T_s ($n = 634$) (a) and w ($n = 554$) (b) from 1 October 2013 to 30 September 2014. The red line in panel b corresponds to the heated Metek data while the grey line and symbols represent data collected after heating of the Metek at ATQ was fully de-activated from 17 March to 11 June 2015 ($n = 543$). Regression coefficients with the ‘con’ subscript represent results from the continuously heated Metek data.

Figure 3. Comparison of 30-min average LE ($n = 2316$) (a), F_{CO_2} ($n = 2167$) (b), and F_{CH_4} ($n = 3563$) (c) fluxes derived from the heated Metek sonic anemometer (y-axis) and the unheated CSAT3 anemometer, and LE ($n = 2969$) (b), F_{CO_2} ($n = 2689$) (d), and F_{CH_4} ($n = 2058$) (f) from the unheated Metek and unheated CSAT3, in all cases paired with the LGR closed-path analyzer at ATQ from 1 October 2013 to 22 July 2015.

Figure 4. Distributions of flux differences (Δflux) from the analyzer pairs being compared in Fig. 3. Delta values were calculated as CSAT3 LE – Metek LE (a, b), CSAT3 F_{CO_2} – Metek F_{CO_2} (c, d), and CSAT3 F_{CH_4} – Metek F_{CH_4} (e, f). Distributions were fitted with Laplace (double exponential) maximum likelihood $ML = \frac{1}{2}\beta e^{-|x-\mu|/\beta}$, where $\sigma = \sqrt{2}\beta$.

Figure 5. Cumulative CO_2 from the open and closed-path analyzers in BES paired with the CSAT3 (Table 2) (a), F_{CO_2} from the closed and (en)closed-path analyzers and heated and unheated anemometers in ATQ (c). Cumulative CH_4 fluxes from the open and closed-path analyzers in CMDL (b), and from the heated and non-heated anemometers in ATQ (d).

Figure 6. Sensible heat fluxes from IVO during a heating activation of the intermittently heated Metek anemometer. The solid red line shows the duration of heating as 30-min average values of the heating flag, which was set to 0 (off) or 1 (on) depending on the activation status. Grey points show noisy data resulting from the build-up of ice and snow and the resulting activation of anemometer heating.

Figure 7. Comparisons of 30-min average F_{CO_2} ($n = 5036$) (a) and LE ($n = 2837$) (c) between closed and (en)closed-path sensors ATQ, and comparisons of F_{CO_2} ($n = 790$) (b) and F_{CH_4} ($n = 1681$) (d) between open and closed-paths sensors at BES and CMDL, respectively.

Figure 8. Distributions of flux differences (Δflux) from the analyzer pairs being compared in Fig. 6. Delta values were calculated as (en)closed-path – closed-path fluxes (ba and c) and open-path – closed path fluxes (b and d).

4. Discussion

4.1. Sonic anemometer comparisons

Continuous heating of the Metek sonic anemometer at ATQ considerably increased the data coverage, especially in winter, relative to the unheated CSAT3 and Gill R3 anemometers (Table 3), however sensible heat and gas fluxes were over-estimated with the heated Metek. Recently, Frank et al. (2013) suggested that non-orthogonal sonic anemometers may under-estimate the vertical wind component and H, which was also found by Mauder et al. (2007) for two non-orthogonal designs. However, Loescher et al. (2005) found higher H from the CSAT3 compared to orthogonal anemometers. All of the anemometers compared in this study were non-orthogonal, however our results suggest that the anemometer geometry (horizontal head vs. vertical head) had a relatively small impact, on wind components and sensible heat and C fluxes at ATQ. The anemometer heating had the most important influence on comparisons. A similar result was reported from the BOREAS IV campaign, where several heated anemometers, including an earlier Metek model (USA-1 55W), were compared to unheated anemometers (including a Gill R3) and a heightened sensitivity to the vertical wind component was noted in the heated models (Tammelin et al., 1998). Nonetheless, the observed over-estimation of vertical wind variations by the heated Metek in this study led only to a relatively small increase in F_{CO_2} (both uptake and losses) and F_{CH_4} (Figure 3c and e). As a result, there was a small difference in the estimated annual CO_2 fluxes compared to the CSAT-derived fluxes, and the two estimates had overlapping uncertainties (Table 2). This result suggests that despite a small over-estimation of C gas fluxes derived from a continuously heated anemometer, the increase in data coverage with continuous heating and the comparable annual sums that result can still lead to defensible annual estimates of C gas fluxes using heated sonic anemometers.

Despite improved data coverage and relatively small effects on long-term fluxes observed at ATQ, continuous heating may not be the optimal choice, particularly for remote flux towers where power consumption is critical for site operation. An intermittent heating approach, such as that devised in this study, provides a better alternative. This minimizes power consumption, reduces heating-related errors in fluxes, and retains very similar data coverage compared to continuously heated systems (Figures 2 and S3; Table 3). We recommend that intermittent heating be implemented in all sites operating in cold climates, to increase data coverage during the cold season and avoid overestimating fluxes in all periods.

4.2. Gas analyzer comparisons

We found that despite the potential for relatively large instantaneous differences in fluxes as well as the larger spectral corrections required for longer-tube LGR-FGGA-24EP fluxes due to sensor time response, tubing and filter attenuation, and intake assembly, the resulting annual CO₂ fluxes did not differ significantly from the shorter-tube LI-7200. Burns et al. (2014) also found good comparisons between the LI-7200 and an older model LI-6262 close-path analyzer with a longer intake tube after data treatment and post-processing were properly accounted for, despite differing spectral response. These results provide a measure of confidence in carbon flux estimates obtained by closed-path gas analyzers of various designs when upgrading long-term sites or planning new experiments, provided that careful spectral corrections are applied to each design using data-driven, *in-situ* methods. On the other hand, latent heat fluxes from the LGR-FGGA-24EP with heated tube in this study were slightly under-estimated compared to the LI-7200 with a shorter, unheated tube (ATQ) (Figures 7c and 8c). The longer tube length (3.12 m compared to 1.2 m) of the LGR-FGGA-24EP (Table 1) at ATQ likely contributed to more attenuation of the H₂O signal compared to the LI-7200, leading to spectral correction factors that did not fully compensate for this signal loss (Table 3, Figure S2). This was noted in other comparisons (Burns et al., 2014), and the

attenuation of H₂O signals due to intake tubing remains a trade-off in eddy covariance systems where increased data coverage is desired (Lenschow and Raupach, 1991; Leuning and Judd, 1996). At the wettest site, BES, H₂O signal attenuation was so severe that spectral corrections factors of 300-400% applied to the closed-path LGR-FGGA-24EP LE fluxes (Table 3) were insufficient to calculate comparable direct comparisons with the LI-7500 (not shown). Other studies have found similar under-estimation of LE from closed-path systems relative to open-path (Halswanter et al., 2009; Ueyama et al., 2012). The BES site exhibits more extensive inundated areas relative to the other study sites (Zona et al., 2009; Zona et al., 2016), and the attenuation of H₂O signals can be strongly influenced by relative humidity (Runkle et al., 2012), especially if the heating of the intake tube is not uniform (Mammarella et al., 2009). Reductions in intake tube lengths and funnel sizes as well as improvements to the tube heating system should be investigated in order to improve H₂O flux measurements in future work at these cold, inundated sites.

Fluxes of CO₂ from the open and closed-path systems at BES compared well after applying differing spectral correction factors (Table 4), while more noise in the fluxes was observed at this site leading to a large spread in ΔF_{CO_2} . A primary concern between sensor configurations for CO₂ fluxes at BES was data coverage during cold periods. The lack of reliable CO₂ flux data from the non-winterized open-path analyzer for a large part of the winter precluded the estimation of a full annual CO₂ budget at this site. Despite this, comparable flux integrals were obtained during warmer seasons and estimates from the different sensor combinations had overlapping uncertainties. Oechel et al. (2014) noted the difficulty in maintaining open-path analyzers during the Arctic winter and found that data capture rates fell below 15% during this period. This was, in part, due to the inability to safely access and maintain instrumentation at the tower, which is especially necessary for open-path sensors to keep the windows clean and unobstructed by snow and ice.

The open and closed-path F_{CH_4} compared well at CMDL (Figure 5d and 7d). A number of eddy covariance CH_4 sensors have also successfully measured comparable fluxes (Detto et al., 2011; Peltola et al., 2011; Iwata et al. 2014). Therefore the choice of sensor may be adapted to the specific conditions and requirements (e.g. power availability) at a site if the research focus is on C- CH_4 gas fluxes and budgets. However, with increasing evidence for the importance of cold season emissions in the Arctic (Mastepanov et al., 2008; Mastepanov et al., 2013; Sturtevant et al., 2012; Zona et al., 2016), the most reliable data coverage during such periods currently requires closed-path analyzers, although heavily winterized open-path analyzers were not examined in this study. Coastal areas are particularly problematic in this regard given the higher humidity and the salt spray that affect the mirrors of open-path analyzers, leading to substantially higher data losses than areas farther inland, even when automated mirror cleaning is employed. For the inland sites, where low power availability prevents the use of close path CH_4 analyzers, we have found in this, and other studies (Zona et al., 2016), that the open path LI-7700 was able to successfully measure fluxes year-round with reasonable data coverage.

5. Conclusions

Using different eddy covariance sensor combinations on the same towers, we have shown that seasonal and annual CO_2 and CH_4 fluxes estimated using different gas analyzers and from heated and non-heated anemometers were within uncertainties. The remote locations and harsh winter conditions in the Arctic often necessitate the use of different site-specific instrumentation at each site. However, this does not preclude the ability to obtain comparable C fluxes if instruments are properly setup and fluxes are carefully processed. Heating sonic anemometers intermittently for de-icing, and excluding minimal data during the heating (~3% of the total data set), is the optimal solution for minimizing biases in wind components while maximizing data coverage (e.g. obtaining similar data coverage to a

continuously heated anemometer). If sufficient power is available at the site, closed-path and (en)closed-path gas analyzers should be used to provide better data coverage than non-winterized open-path analyzers, especially in cold and wet environments, where 90% of data can be lost.

Acknowledgements

This work was funded by the Division of Polar Programs of the National Science Foundation (NSF) (Award 1204263); Carbon in Arctic Reservoirs Vulnerability Experiment (CARVE), an Earth Ventures (EV-1) investigation, under contract with the National Aeronautics and Space Administration (NASA); and Department of Energy (DOE) Grant DE-SC005160. Logistical support was funded by the NSF Division of Polar Programs. The co-authors are also grateful to the United States Permafrost Association Early Career Grant for their financial support on field trips. This research was conducted on land owned by the Ukpeagvik Inupiat Corporation (UIC). We would like to thank the Global Change Research Group at San Diego State University, UMIAQ, UIC, CPS for logistical support, and Salvatore Losacco, Owen Hayman, and Herbert Njuabe for the help with the field data collection.

References

- Auble DL, Meyers TP (1992) An open path, fast response infrared absorption gas analyzer for H₂O and CO₂. *Boundary-Layer Meteorology*, 59, 243-256
- Brown J, Everett KR, Webber PJ, MacLean Jr SF, Murray DF (1980) An Arctic Ecosystem: The Coastal Tundra at Barrow, Alaska, (eds. J. Brown et al.) pp. 1 –29, Dowden, Hutchinson and Ross, Stroudsburg, PA.

631 Burba GG, McDermitt DK, Grelle A, Anderson DJ, Xu L (2008) Addressing the influence of
632 instrument surface heat exchange on the measurements of CO₂ flux from open-path
633 gas analyzers. *Global Change Biology*, 14, 1–23.

634 Burba GG, Schmidt A, Scott RL et al. (2012) Calculating CO₂ and H₂O eddy covariance
635 fluxes from an enclosed gas analyzer using an instantaneous mixing ratio. *Global*
636 *Change Biology*, 18, 385–399.

637 Burba GG, McDermitt DK, Anderson DJ, Furtaw MD, Eckles RD (2010) Novel design of an
638 enclosed CO₂/H₂O gas analyser for eddy covariance flux measurements. *Tellus*
639 *B*, 62(5), 743-748.

640 Burba GG (2013) Eddy Covariance Method for Scientific, Industrial, Agricultural and
641 Regulatory Applications: a Field Book on Measuring Ecosystem Gas Exchange and
642 Areal Emission Rates. LI-COR Biosciences, Lincoln, USA.

643 Burns S, Metzger S, Blanken P et al. (2014) A comparison of infrared gas analyzers above a
644 subalpine forest in complex terrain. American Meteorological Society Committee on
645 Atmospheric Measurements: 21st Conference on Applied Climatology/17th
646 Symposium on Meteorological Observation and Instrumentation. Westminster,
647 Colorado, 9-13 June: 21.

648 Detto M, Verfaillie J, Anderson F, Xu L, Baldocchi D (2011) Comparing laser-based open-
649 and closed-path gas analyzers to measure methane fluxes using the eddy covariance
650 method, *Agricultural and Forest Meteorology*, 151, 1312–1324.

651 Dragoni D, Schmid HP, Grimmond CSB, Loescher HW (2007) Uncertainty of annual net
652 ecosystem productivity estimated using eddy covariance flux measurements. *Journal*
653 *of Geophysical Research-Atmospheres* 112(D17), doi:10.1029/2006JD008149.

654 El-Madany T S, Griessbaum F, Fratini G, Juang JY, Chang SC, Klemm O (2013)
655 Comparison of sonic anemometer performance under foggy conditions. *Agricultural*
656 *and Forest Meteorology*, 173, 63-73.

657 Emmerton CA, St. Louis VL, Humphreys ER, Gamon JA, Barker JD, Pastorello GZ (2015)
658 Net ecosystem exchange of Co₂ with rapidly changing high Arctic landscapes. *Global*
659 *Change Biology*, doi: 10.1111/gcb.13064.

660 Euskirchen ES, Bret-Harte MS, Scott GJ, Edgar C, Shaver GR (2012) Seasonal patterns of
661 carbon dioxide and water fluxes in three representative tundra ecosystems in northern
662 Alaska. *Ecosphere* 3(1), <http://dx.doi.org/10.1890/ES11-00202>.

663 Fisher JB, Sikka M, Oechel WC et al. (2014) Carbon cycle uncertainty in the Alaskan Arctic.
664 *Biogeosciences Discussions*, 11, 2887-2932, doi:10.5194/bgd-11-2887-2014.

665 Frank JM, Massman WJ, Ewers BE (2013) Underestimates of sensible heat flux due to
666 vertical velocity measurement errors in non-orthogonal sonic anemometers.
667 *Agricultural and Forest Meteorology*, 171, 72-81.

668 Fratini G, McDermitt DK, Papale D (2014) Eddy-covariance flux errors due to biases in gas
669 concentration measurements: origins, quantification and correction. *Biogeosciences*,
670 11, 1037-1051. Doi: 10.5194/bg-11-1037-2014.

671 Gash JHC, Culf AD (1996) Applying linear de-trend to eddy correlation data in real time.
672 *Boundary-Layer Meteorology*, 79: 301-306.

673 Gažovič M, Forbrich I, Jager DF, Kutzbach L, Wille C, Wilmking M (2013) Hydrology-
674 driven ecosystem respiration determines the carbon balance of a boreal
675 peatland. *Science of the Total Environment*, 463, 675-682.

676 Goulden ML, Winston GC, McMillan AMS, Litvak ME, Read EL, Rocha AV, Elliot JR
677 (2006) An eddy covariance mesonet to measure the effect of forest age on land-
678 atmosphere exchange. *Global Change Biology*, 12, 2146–2162.

679 Grelle A, Burba G (2007). Fine-wire thermometer to correct CO₂ fluxes by open-path
680 analyzers for artificial density fluctuations. *Agricultural and Forest*
681 *Meteorology*, 147(1), 48-57.

682 Hanis KL, Tenuta M, Amiro BD, Papakyriakou TN (2013). Seasonal dynamics of methane
683 emissions from a subarctic fen in the Hudson Bay Lowlands. *Biogeosciences*, 10(7),
684 4465-4479.

685 Haslwanter A, Hammerle A, Wohlfahrt G (2009) Open-path vs. closed-path eddy covariance
686 measurements of the net ecosystem carbon dioxide and water vapor exchange: a long-
687 term perspective. *Agricultural and Forest Meteorology*, 149(2), 291-302.

688 Hollinger DY, Richardson AD (2005) Uncertainty in eddy covariance measurements and its
689 application to physiological models. *Tree physiology*, 25(7), 873-885.

690 Horst TW, Lenschow DH (2009) Attenuation of scalar fluxes measured with spatially-
691 displaced sensors. *Boundary-layer meteorology*, 130(2), 275-300.

692 Ibrom A, Dellwik E, Flyvbjerg H, Jensen NO, Pilegaard K (2007) Strong low-pass filtering
693 effects on water vapor flux measurements with closed-path eddy correlation systems,
694 *Agricultural and Forest Meteorology*, 147, 140-156.

695 IPCC (2007) Climate Change 2007: The Physical Science Basis. Contribution of Working
696 Group I to the Fourth Assessment Report of the Intergovernmental Panel on Climate
697 Change [Solomon S, Qin D, Manning M, Chen Z, Marquis M, Averyt KB, Tignor M,
698 Miller HL (eds.)]. Cambridge University Press, Cambridge, United Kingdom and
699 New York, NY, USA.

700 Iwata H, Kosugi Y, Ono K, Mano M, Sakabe A, Miyata A, Takahashi K (2014) Cross-
701 validation of open-path and closed-path eddy-covariance techniques for observing
702 methane fluxes. *Boundary-Layer Meteorology*, 151: 95–118. Doi: 10.1007/s10546-
703 013-9890-2.

704 Jackowicz-Korczynski M, Christensen TR, Backstrand K, Crill P, Friborg T, Mastepanov M,
 705 Strom L (2010) Annual cycle of methane emission from a subarctic peatland, *Journal*
 706 *of Geophysical Research*, 115(G02009), doi:10.1029/2008JG000913.

707 Järvi L, Mammarella I, Eugster W, Ibrom A, Siivola E, Dellwik E, Keronen P, Burba G,
 708 Vesala T (2009) Comparison of net CO₂ fluxes measured with open- and closed-path
 709 infrared gas analyzers in an urban complex environment. *Boreal Environmental*
 710 *Research*, 14: 499–514.

711 Kochendorfer JP, Meyers TP, Frank JM, Massman WJ, Heuer MW (2012) How well can we
 712 measure the vertical wind speed? Implications for fluxes of energy and mass.
 713 *Boundary-Layer Meteorology*, 145 (2), 383–398.

714 Kutzbach L, Wille C, Pfeiffer E-M (2007) The exchange of carbon dioxide between wet
 715 arctic tundra and the atmosphere at the Lena River Delta, Northern Siberia.
 716 *Biogeosciences*, 4: 869-890.

717 Kwon HJ, Oechel WC, Zulueta RC, Hastings SJ (2006) Effects of climate variability on
 718 carbon sequestration among adjacent wet sedge tundra and moist tussock tundra
 719 ecosystems. *Journal of Geophysical Research-Biogeosciences*, 111: 2005-2012.

720 Lee X, Black TA, Novak MD (1994) Comparison of flux measurements with open- and
 721 closed-path gas analyzers above an agricultural field and a forest floor. *Boundary-*
 722 *Layer Meteorology*, 67(1-2), 195-202.

723 Lekakis IC, Adrian RJ, Jones BG (1989) Measurement of velocity vectors with orthogonal
 724 and non-orthogonal triple-sensor probes. *Experiments in Fluids*, 7, 228–40.

725 Lenschow DH, Raupach MR (1991) The attenuation of fluctuations in scalar concentrations
 726 through sampling tubes. *Journal of Geophysical Research-Atmospheres*, 96(D8),
 727 15259-15268.

728 Leuning R, Judd MJ (1996) The relative merits of open-and closed-path analysers for
 729 measurement of eddy fluxes. *Global Change Biology*, 2(3), 241-253.

730 Leuning R, King KM (1992) Comparison of eddy-covariance measurements of CO₂ fluxes
 731 by open- and closed-path CO₂ analyzers. *Boundary-Layer Meteorology*, 59: 297–311.

732 Loescher HW, Ocheltree T, Tanner B et al. (2005) Comparison of temperature and wind
 733 statistics in contrasting environments among different sonic anemometer–
 734 thermometers. *Agricultural and forest meteorology*, 133(1), 119-139.

735 Lüers J, Westermann S, Piel K, Boike J (2014). Annual CO₂ budget and seasonal CO₂
 736 exchange signals at a High Arctic permafrost site on Spitsbergen, Svalbard
 737 archipelago. *Biogeosciences*, 11(22), 6307-6322.

738 Mammarella I, Launiainen S, Gronholm T, Keronen P, Pumpanen J, Rannik Ü, Vesala T
 739 (2009) Relative humidity effect on the high-frequency attenuation of water vapor flux
 740 measured by a closed-path eddy covariance system. *Journal of Atmospheric and*
 741 *Oceanic Technology*, 26: 1856-1866.

742 Massman WJ (1991) The attenuation of concentration fluctuations in turbulent flow through a
 743 tube. *Journal of Geophysical Research*, 96, 15269– 15273.

744 Massman WJ (2000) A simple method for estimating frequency response corrections for
 745 eddy-covariance systems. *Agricultural and Forest Meteorology*, 104: 185–198.

746 Mastepanov M, Sigsgaard C, Dlugokencky EJ, Houweling S, Ström L, Tamstorf MP,
 747 Christensen TR (2008) Large tundra methane burst during onset of
 748 freezing. *Nature*, 456(7222), 628-630.

749 Mastepanov M, Sigsgaard C, Tagesson HT, Ström L, Tamstorf MP, Lund M, Christensen TR
 750 (2013) Revisiting factors controlling methane emissions from high-Arctic
 751 tundra. *Biogeosciences*, 10(7), 5139-5158.

752 Mauder M, Oncley SP, Vogt R, Weidinger T, Ribeiro L, Bernhofer C, Foken T, Kohsiek W,
753 De Bruin HAR, Liu H (2007) The energy balance experiment EBEX-2000. Part II:
754 Intercomparison of eddy-covariance sensors and post-field data processing
755 methods. *Boundary-Layer Meteorology*, 123(1), 29-54.

756 Mauder M, Foken T (2006) Impact of post-field data processing on eddy covariance flux
757 estimates and energy balance closure. *Meteorologische Zeitschrift*, 15: 597-609.

758 McDermitt D, Burba G, Xu L, et al. (2011) A new low-power, open-path instrument for
759 measuring methane flux by eddy covariance. *Applied Physics B*, doi: 10.1007/s00340-
760 010-4307-0.

761 Melton JR, Wania R, Hodson EL et al. (2013) Present state of global wetland extent and
762 wetland methane modelling: conclusions from a model intercomparison project
763 (WETCHIMP). *Biogeosciences*, 10, 753-788.

764 Moncrieff JB, Massheder JM, de Bruin H et al. (1997) A system to measure surface fluxes of
765 momentum, sensible heat, water vapor and carbon dioxide. *Journal of Hydrology*,
766 188–189: 589–611.

767 Moncrieff JB, Clement R, Finnigan J, Meyers T (2004) Averaging, detrending and filtering
768 of eddy covariance time series, in Handbook of micrometeorology: a guide for surface
769 flux measurements, eds. Lee X, Massman WJ, Law BE Dordrecht: Kluwer Academic,
770 7-31.

771 Mullier SV, Racoviteanu AE, Walker DA (1999) Landsat MSS-derived land-cover map of
772 northern Alaska: Extrapolation methods and a comparison with photo-interpreted and
773 AVHRR-derived maps, *International Journal of Remote Sensing*, 20, 2921–2946.

774 Nakai T, van der Molen MK, Gash JHC, Kodama Y (2006) Correction of sonic anemometer
775 angle of attack errors. *Agricultural and Forest Meteorology*, 136, 19–30. doi:
776 10.1016/j.agrformet.2006.01.006.

777 Nakai T, Iwata H, Harazono Y (2011) Importance of mixing ratio for a long-term CO₂ flux
 778 measurement with a closed-path system. *Tellus B*, 63(3), 302-308.

779 Nakai T, Kim Y, Busey RC et al. (2013) Characteristics of evapotranspiration from a
 780 permafrost black spruce forest in interior Alaska. *Polar Science*, 7, 136–148.

781 Nakai T, Iwata H, Harazono Y, Ueyama M (2014). An inter-comparison between Gill and
 782 Campbell sonic anemometers. *Agricultural and Forest Meteorology*, 195, 123-131.

783 Nakai T, Shimoyama K (2012) Ultrasonic anemometer angle of attack errors under turbulent
 784 conditions. *Agricultural and Forest Meteorology*, 162, 14-26.

785 Oechel WC, Laskowski CA, Burba G, Gioli B, Kalhori AAM (2014) Annual patterns and
 786 budget of CO₂ flux in an Arctic tussock tundra ecosystem. *Journal of Geophysical*
 787 *Research-Biogeosciences*, 119, doi: 10.1002/2013JG002431.

788 Parmentier FJW, van Huissteden J, van der Molen MK, Dolman AJ, Schaepman-Strub G,
 789 Karsanaev SA, Maximov TC (2011) Spatial and temporal dynamics in eddy
 790 covariance observations of methane fluxes at a tundra site in northeastern Siberia.
 791 *Journal of Geophysical Research*, 116, G03016. doi:10.1029/2010JG001637.

792 Peltola O, Mammarella I, Haapanala S, Burba G, Vesala T (2013) Field intercomparison of
 793 four methane gas analyzers suitable for eddy covariance flux measurements.
 794 *Biogeosciences* 10:3749-3765.

795 Reynolds MK, Walker DA, Maier HA (2005) Plant community-level mapping of arctic
 796 Alaska based on the Circumpolar Arctic Vegetation Map. *Phytocoenologia*, 35(4):
 797 821-848.

798 Reichstein M, Falge E, Baldocchi D et al. (2005) On the separation of net ecosystem
 799 exchange into assimilation and ecosystem respiration: review and improved
 800 algorithm. *Global Change Biology*, 11: 1-11, doi: 10.1111/j.1365-
 801 2486.2005.001002.x.

802 Rinne J, Riutta T, Pihlatie M et al. (2007) Annual cycle of methane emission from a boreal
803 fen measured by the eddy covariance technique, *Tellus*, 59B, 449-457.

804 Runkle BRK, Wille C, Gazovic M, Kutzbach L (2012) Attenuation correction procedures for
805 water vapour fluxes from closed-path eddy-covariance systems. *Boundary-Layer*
806 *Meteorol*, 142: 401-423.

807 Sachs T, Wille C, Boike J, Kutzbach L (2008) Environmental controls on ecosystem-scale
808 CH₄ emission from polygonal tundra in the Lena River Delta, Siberia. *Journal of*
809 *Geophysical Research-Biogeosciences*, 113: 1395-1408.

810 Skelly BT, Miller DR, Meyer TH (2002) Triple-hot-film anemometer performance in
811 CASES-99 and a comparison to sonic anemometer measurements, *Boundary-Layer*
812 *Meteorology*, 105: 275–304.

813 Song W, Wang H, Wang G, Chen L, Jin Z, Zhuang Q, He J-S (2015) Methane emissions
814 from an alpine wetland on the Tibetan Plateau: neglected but vital contribution of
815 non-growing season, *Journal of Geophysical Research-Biogeosciences*, doi:
816 10.1002/2015JG003043.

817 Sturtevant CS, Oechel WC, Zona D, Kim Y, Emerson CE (2012) Soil moisture control over
818 autumn season methane flux, Arctic Coastal Plain of Alaska. *Biogeosciences*, 9, 1423-
819 1440.

820 Tammelin B, Cavaliere M, Kimura S, Morgan C (1998) Ice free anemometers. BOREAS IV
821 31 March – 2 April 1998, Hetta, Finland, 239-252.

822 Taylor, J.R. (1997) *Error Analysis; The study of uncertainties in physical measurements* 2nd
823 Ed. University Science Books, Sausalito, CA, US.

824 Ueyama M, Hirata R, Mano M et al. (2012) Influences of various calculation options on heat,
825 water and carbon fluxes determined by open- and closed-path eddy covariance
826 methods. *Tellus B*, doi: 10.3402/tellusb.v64i0.19048.

827 Ueyama M, Iwata H, Harazono Y, Euskirchen ES, Oechel WC, Zona D (2013) Growing
 828 season and spatial variations of carbon fluxes of Arctic and boreal ecosystems in
 829 Alaska. *Ecological Applications*, 23: 1798-1816.

830 Vickers D, Mahrt L (1997) Quality control and flux sampling problems for tower and aircraft
 831 data. *Journal of Atmospheric and Oceanic Technology*, 14: 512-526.

832 Webb EK, Pearman GI, Leuning R (1980) Correction of flux measurements for density
 833 effects due to heat and water vapor transfer. *Quarterly Journal of the Royal*
 834 *Meteorological Society*, 106: 85–100.

835 Wilczak JM, Oncley SP, Stage SA (2001) Sonic anemometer tilt correction algorithms.
 836 *Boundary-Layer Meteorology*, 99: 127-150.

837 Wille C, Kutzbach L, Sachs T, Wagner D, Pfeiffer E (2008) Methane emission from Siberian
 838 arctic polygonal tundra: eddy covariance measurements and modeling. *Global*
 839 *Change Biology*, 14(6), 1395-1408.

840 Zona, D., Gioli, B., Commane, R., Lindaas, J., Wofsy, S.C., Miller, C.E., Dinardo, S.J.,
 841 Dengel, S., Sweeney, C., Karion, A., Chang, R.Y.-W., Henderson, J.M., Murphy,
 842 P.C., Goodrich, J.P., Moreaux, V., Liljedahl, A., Watts, J.D., Kimball, J.S., Lipson,
 843 D.A., Oechel, W.C. (2016) Cold season emission dominate the Arctic tundra methane
 844 budget. *Proceedings of the National Academy of Sciences*, 113(1), 40-45.

845 Zona D, Lipson DA, Richards JH et al. (2014) Delayed responses of an Arctic ecosystem to
 846 an extreme summer: impacts on net ecosystem exchange and vegetation
 847 functioning. *Biogeosciences*, 11(20), 5877-5888.

848 Zona D, Oechel WC, Kochendorfer J et al. (2009) Methane fluxes during the initiation of a
 849 large-scale water table manipulation experiment in the Alaskan Arctic tundra. *Global*
 850 *Biogeochemical Cycles*, 23(2), doi: 10.1029/2009GB003487.

851 Zona D, Oechel WC, Peterson KM, Clements RJ, Paw U KT and Ustin SL (2010)
852 Characterization of the carbon fluxes of a vegetated drained lake basin
853 chronosequence on the Alaskan Arctic Coastal Plain. *Global Change Biology*, 16,
854 1870-1882. doi: 10.1111/j.1365-2486.2009.02107.

855 Zona D, Oechel WC, Richards JH, Hastings S, Kopetz I, Ikawa H, Oberbauer S (2011) Light-
856 stress avoidance mechanisms in a Sphagnum-dominated wet coastal Arctic tundra
857 ecosystem in Alaska, *Ecology*, 92, 633–644, doi: 10.1890/10-0822.1.

858 Zona D, Lipson DA, Paw U KT, Oberbauer SF, Olivas P, Gioli B, Oechel WC (2012)
859 Increased CO₂ loss from vegetated drained lake tundra ecosystems due to flooding.
860 *Global Biogeochemical Cycles* 26(GB2013), doi:10.1029/2011GB004037.

861

862

Figure 1

Figure 2

Figure 3

Unheated CSAT3 anemometer

Heated Metek anemometer

Unheated Metek anemometer

Figure 4

Heated Metek anemometer

Unheated Metek anemometer

Figure 5

Figure 6

Figure 7

Closed-path LGR-FGGA-24EP

(en)Closed-path LI7200

S = 0.99; I = -0.01
r = 0.87; P < 0.001

Open-path LI7500

S = 1.28; I = 0.29
r = 0.61; P < 0.001

S = 0.86; I = 0.35
r = 0.98; P < 0.001

Open-path LI7700

S = 0.86; I = 0.05
r = 0.68; P < 0.001

Figure 8

(en) Closed-path comparisons

Open-path comparisons