The Detector System of the Daya Bay Reactor Neutrino Experiment

F. P. An^a, J. Z. Bai^b, A. B. Balantekin^c, H. R. Band^d, D. Beavis^e,
W. Beriguete^e, M. Bishai^e, S. Blyth^{f,g}, R. L. Brown^{e,1}, I. Butorov^h, D. Caoⁱ,
G. F. Cao^b, J. Cao^b, R. Carr^j, W. R. Cen^b, W. T. Chan^e, Y. L. Chan^k,
J. F. Chang^b, L. C. Chang^l, Y. Chang^g, C. Chasman^e, H. Y. Chen^l,
H. S. Chen^b, M. J. Chen^b, Q. Y. Chen^m, S. J. Chenⁱ, S. M. Chenⁿ,
X. C. Chen^k, X. H. Chen^b, X. S. Chen^b, Y. X. Chen^o, Y. Chen^p,
J. H. Cheng^l, J. Cheng^m, Y. P. Cheng^b, J. J. Cherwinka^c, S. Chidzik^q,
K. Chow^r, M. C. Chu^k, J. P. Cummings^s, J. de Arcos^t, Z. Y. Deng^b,
X. F. Ding^b, Y. Y. Ding^b, M. V. Diwan^e, L. Dong^b, J. Dove^u, E. Draeger^t,
X. F. Du^b, D. A. Dwyer^r, W. R. Edwards^r, S. R. Ely^u, S. D. Fangⁱ,
J. Y. Fu^b, Z. W. Fuⁱ, L. Q. Ge^v, V. Ghazikhanian^w, R. Gill^e, J. Goett^x,
M. Gonchar^h, G. H. Gongⁿ, H. Gongⁿ, Y. A. Gornushkin^h, M. Grassi^b,
L. S. Greenler^c, W. Q. Gu^y, M. Y. Guan^b, R. P. Guo^b, X. H. Guo^z,
R. W. Hackenburg^e, R. L. Hahn^e, R. Han^o, S. Hans^e, M. He^b, Q. He^q,
W. S. He^f, K. M. Heeger^d, Y. K. Heng^b, A. Higuera^{aa}, P. Hinrichs^c,
T. H. Ho^f, M. Hoff^r, Y. K. Hor^{ab}, Y. B. Hsiung^f, B. Z. Hu^f, L. M. Hu^e,
L. J. Hu^z, T. Hu^b, W. Hu^b, E. C. Huang^u, H. Z. Huang^w, H. X. Huang^{ac},
P. W. Huangⁱ, X. Huang^{aa}, X. T. Huang^m, P. Huber^{ab}, G. Hussainⁿ,
Z. Isvan^e, D. E. Jaffe^e, P. Jaffke^{ab}, K. L. Jen^l, S. Jetter^b, X. P. Ji^{ad,n},
X. L. Ji^b, H. J. Jiang^v, W. Q. Jiang^b, J. B. Jiao^m, R. A. Johnson^{ae},
J. Joseph^r, L. Kang^{af}, S. H. Kettell^e, S. Kohn^{ag}, M. Kramer^{r,ag},
K. K. Kwan^k, M. W. Kwok^k, T. Kwok^{ah}, C. Y. Lai^f, W. C. Lai^v,
W. H. Lai^l, T. J. Langford^d, K. Lau^{aa}, L. Lebanowski^{n,aa}, J. Lee^r,
M. K. P. Lee^{ah}, R. T. Lei^{af}, R. Leitner^{ai}, J. K. C. Leung^{ah}, C. A. Lewis^c,
B. Li^b, C. Li^m, D. J. Li^{aj}, F. Li^b, G. S. Li^y, J. Li^b, N. Y. Li^r, Q. J. Li^b,
S. F. Li^{af}, S. C. Li^{ah}, W. D. Li^b, X. B. Li^b, X. N. Li^b, X. Q. Li^{ad}, Y. Li^{af},
Y. F. Li^b, Z. B. Li^{ak}, H. Liang^{aj}, J. Liang^b, C. J. Lin^r, G. L. Lin^l, P. Y. Lin^l,
S. X. Lin^{af}, S. K. Lin^{aa}, Y. C. Lin^v, J. J. Ling^{ak,e,u}, J. M. Link^{ab},
L. Littenberg^e, B. R. Littlejohn^{ae,t,c}, B. J. Liu^k, C. Liu^b, D. W. Liu^{aa,r,u},
H. Liu^{aa}, J. L. Liu^y, J. C. Liu^b, S. Liu^r, S. S. Liu^{ah}, X. Liu^{b,1}, Y. B. Liu^b,
C. Lu^q, H. Q. Lu^b, J. S. Lu^b, A. Luk^k, K. B. Luk^{ag,r}, T. Luo^b, X. L. Luo^b,

¹Deceased.

L. H. Ma^b, Q. M. Ma^b, X. Y. Ma^b, X. B. Ma^o, Y. Q. Ma^b, B. Mayes^{aa},
 K. T. McDonald^a, M. C. McFarlane^c, R. D. McKeown^{j,al}, Y. Meng^{ab},
 I. Mitchell^{aa}, D. Mohapatra^{ab}, J. Monari Kebwaro^{am}, J. E. Morgan^{ab},
 Y. Nakajima^r, J. Napolitano^{an}, D. Naumov^h, E. Naumova^h, C. Newsom^{aa},
 H. Y. Ngai^{ah}, W. K. Ngai^u, Y. B. Nie^{ac}, Z. Ning^b, J. P. Ochoa-Ricoux^{ao,r},
 A. Olshevskiy^h, A. Pagac^c, H.-R. Pan^f, S. Patton^r, C. Pearson^e, V. Pec^{ai},
 J. C. Peng^u, L. E. Pilonen^{ab}, L. Pinsky^{aa}, C. S. J. Pun^{ah}, F. Z. Qi^b, M. Qiⁱ,
 X. Qian^e, N. Raper^x, B. Ren^{af}, J. Ren^{ac}, R. Rosero^e, B. Roskovec^{ai},
 X. C. Ruan^{ac}, W. R. Sands III^a, B. Seilhan^t, B. B. Shaoⁿ, K. Shih^k,
 W. Y. Song^b, H. Steiner^{ag,r}, P. Stoler^x, M. Stuart^r, G. X. Sun^b, J. L. Sun^{ap},
 N. Tagg^e, Y. H. Tam^k, H. K. Tanaka^e, W. Tang^e, X. Tang^b,
 D. Taychenachev^h, H. Themann^e, Y. Torun^t, S. Trentalange^w, O. Tsai^w,
 K. V. Tsang^r, R. H. M. Tsang^j, C. E. Tull^r, Y. C. Tung^f, N. Viaux^{ao},
 B. Viren^e, S. Virotek^r, V. Vorobel^{ai}, C. H. Wang^g, L. S. Wang^b,
 L. Y. Wang^b, L. Z. Wang^o, M. Wang^m, N. Y. Wang^z, R. G. Wang^b,
 T. Wang^b, W. Wang^{ak,al}, W. W. Wangⁱ, X. T. Wang^b, X. Wang^{aq},
 Y. F. Wang^b, Z. Wangⁿ, Z. Wang^b, Z. M. Wang^b, D. M. Webber^c,
 H. Y. Weiⁿ, Y. D. Wei^{af}, L. J. Wen^b, D. L. Wenman^c, K. Whisnant^{ar},
 C. G. White^t, L. Whitehead^{aa}, C. A. Whitten Jr.^{w,1}, J. Wilhelm^{ian},
 T. Wise^{c,d}, H. C. Wong^{ah}, H. L. H. Wong^{ag,r}, J. Wong^k, S. C. F. Wong^k,
 E. Worcester^e, F. F. Wu^j, Q. Wu^{m,t}, D. M. Xia^{b,as}, J. K. Xia^b,
 S. T. Xiang^{aj}, Q. Xiao^c, Z. Z. Xing^b, G. Xu^{aa}, J. Y. Xu^k, J. L. Xu^b, J. Xu^z,
 W. Xu^w, Y. Xu^{ad}, T. Xueⁿ, J. Yan^{am}, C. G. Yang^b, L. Yang^{af}, M. S. Yang^b,
 M. T. Yang^m, M. Ye^b, M. Yeh^e, Y. S. Yeh^l, K. Yip^e, B. L. Young^{ar},
 G. Y. Yuⁱ, Z. Y. Yu^b, S. Zeng^b, L. Zhan^b, C. Zhang^e, F. H. Zhang^b,
 H. H. Zhang^{ak}, J. W. Zhang^b, K. Zhang^e, Q. X. Zhang^v, Q. M. Zhang^{am},
 S. H. Zhang^b, X. T. Zhang^b, Y. C. Zhang^{aj}, Y. H. Zhang^b, Y. M. Zhangⁿ,
 Y. X. Zhang^{ap}, Y. M. Zhang^{ak}, Z. J. Zhang^{af}, Z. Y. Zhang^b, Z. P. Zhang^{aj},
 J. Zhao^b, Q. W. Zhao^b, Y. F. Zhao^o, Y. B. Zhao^b, L. Zheng^{aj},
 W. L. Zhong^{b,r}, L. Zhou^b, N. Zhou^{aj}, Z. Y. Zhou^{ac}, H. L. Zhuang^b,
 S. Zimmerman^r, J. H. Zou^b

^a*Institute of Modern Physics, East China University of Science and Technology, Shanghai*

^b*Institute of High Energy Physics, Beijing*

^c*University of Wisconsin, Madison, Wisconsin, USA*

^d*Department of Physics, Yale University, New Haven, Connecticut, USA*

^e*Brookhaven National Laboratory, Upton, New York, USA*

^f*Department of Physics, National Taiwan University, Taipei*

^gNational United University, Miao-Li
^hJoint Institute for Nuclear Research, Dubna, Moscow Region
ⁱNanjing University, Nanjing
^jCalifornia Institute of Technology, Pasadena, California, USA
^kChinese University of Hong Kong, Hong Kong
^lInstitute of Physics, National Chiao-Tung University, Hsinchu
^mShandong University, Jinan
ⁿDepartment of Engineering Physics, Tsinghua University, Beijing
^oNorth China Electric Power University, Beijing
^pShenzhen University, Shenzhen
^qJoseph Henry Laboratories, Princeton University, Princeton, New Jersey, USA
^rLawrence Berkeley National Laboratory, Berkeley, California, USA
^sSiena College, Loudonville, New York, USA
^tDepartment of Physics, Illinois Institute of Technology, Chicago, Illinois, USA
^uDepartment of Physics, University of Illinois at Urbana-Champaign, Urbana, Illinois, USA
^vChengdu University of Technology, Chengdu
^wUniversity of California, Los Angeles, California, USA
^xDepartment of Physics, Applied Physics, and Astronomy, Rensselaer Polytechnic Institute, Troy, New York, USA
^yDepartment of Physics and Astronomy, Shanghai Jiao Tong University, Shanghai
Laboratory for Particle Physics and Cosmology, Shanghai
^zBeijing Normal University, Beijing
^{aa}Department of Physics, University of Houston, Houston, Texas, USA
^{ab}Center for Neutrino Physics, Virginia Tech, Blacksburg, Virginia, USA
^{ac}China Institute of Atomic Energy, Beijing
^{ad}School of Physics, Nankai University, Tianjin
^{ae}Department of Physics, University of Cincinnati, Cincinnati, Ohio, USA
^{af}Dongguan University of Technology, Dongguan
^{ag}Department of Physics, University of California, Berkeley, California, USA
^{ah}Department of Physics, The University of Hong Kong, Pokfulam, Hong Kong
^{ai}Charles University, Faculty of Mathematics and Physics, Prague, Czech Republic
^{aj}University of Science and Technology of China, Hefei
^{ak}Sun Yat-Sen (Zhongshan) University, Guangzhou
^{al}College of William and Mary, Williamsburg, Virginia, USA
^{am}Xi'an Jiaotong University, Xi'an
^{an}Department of Physics, College of Science and Technology, Temple University, Philadelphia, Pennsylvania, USA
^{ao}Instituto de Física, Pontificia Universidad Católica de Chile, Santiago, Chile
^{ap}China General Nuclear Power Group
^{aq}College of Electronic Science and Engineering, National University of Defense Technology, Changsha
^{ar}Iowa State University, Ames, Iowa, USA
^{as}Chongqing University, Chongqing

Abstract

The Daya Bay experiment was the first to report simultaneous measurements of reactor antineutrinos at multiple baselines leading to the discovery of $\bar{\nu}_e$ oscillations over km-baselines. Subsequent data has provided the world's most precise measurement of $\sin^2 2\theta_{13}$ and the effective mass splitting Δm_{ee}^2 . The experiment is located in Daya Bay, China where the cluster of six nuclear reactors is among the world's most prolific sources of electron antineutrinos. Multiple antineutrino detectors are deployed in three underground water pools at different distances from the reactor cores to search for deviations in the antineutrino rate and energy spectrum due to neutrino mixing. Instrumented with photomultiplier tubes, the water pools serve as shielding against natural radioactivity from the surrounding rock and provide efficient muon tagging. Arrays of resistive plate chambers over the top of each pool provide additional muon detection. The antineutrino detectors were specifically designed for measurements of the antineutrino flux with minimal systematic uncertainty. Relative detector efficiencies between the near and far detectors are known to better than 0.2%. With the unblinding of the final two detectors' baselines and target masses, a complete description and comparison of the eight antineutrino detectors can now be presented. This paper describes the Daya Bay detector systems, consisting of eight antineutrino detectors in three instrumented water pools in three underground halls, and their operation through the first year of eight detector data-taking.

Keywords: neutrino oscillation, neutrino mixing, reactor, Daya Bay

PACS: 14.60Pq, 29.40Mc, 28.50Hw, 13.15+g

1. Introduction

The Daya Bay experiment was the first to measure a nonzero θ_{13} with a significance of ≥ 5 standard deviations [1] and continues to provide the most precise measurement of θ_{13} [2, 3]. In the generally accepted three-neutrino framework, neutrino oscillations are described by three mixing angles (θ_{12} , θ_{23} , and θ_{13}), two mass-squared differences (Δm_{31}^2 and Δm_{21}^2), and a phase in the Pontecorvo-Maki-Nakagawa-Sakata matrix [4, 5]. Prior to Daya Bay, θ_{13} was poorly known, constrained by a $\sin^2 2\theta_{13} \leq 0.17$ limit from the Chooz experiment [6] and several low statistics measurements by T2K [7], MINOS [8] and Double Chooz [9] that indicated that θ_{13} could be non-zero.

For reactor-based experiments, θ_{13} can be extracted from the survival probability of the electron antineutrino $\bar{\nu}_e$ at distances a few km from the reactor:

$$P_{\text{sur}} = 1 - \sin^2 2\theta_{13} \sin^2 \Delta_{ee} - \cos^4 \theta_{13} \sin^2 2\theta_{12} \sin^2 \Delta_{21} \quad (1)$$

where $\Delta_{ee} = \Delta_{32} \pm \phi/2$, with $\phi = \arctan \left(\frac{\sin 2\Delta_{21}}{\cos 2\Delta_{21} + \tan^2 \theta_{12}} \right)$ for normal mass hierarchy (+) and inverted mass hierarchy (-), $\Delta_{ji} \equiv 1.267 \Delta m_{ji}^2 (\text{eV}^2) L(\text{m}) / E(\text{MeV})$, with Δm_{ji}^2 the difference between the mass-squares of the mass eigenstates ν_j and ν_i , E is the $\bar{\nu}_e$ energy and L is the distance between the $\bar{\nu}_e$ source and the detector (baseline). The effective mass-squared difference Δm_{ee}^2 has been measured in Daya Bay [3].

Antineutrinos with energy ≥ 1.8 MeV are detected via the inverse beta-decay (IBD) reaction on “free” protons in the liquid scintillating target:

$$\bar{\nu}_e + p \rightarrow e^+ + n. \quad (2)$$

22 The positron, carrying most of the antineutrino energy, deposits its en-
 23 ergy and annihilates with an electron producing a prompt signal with energy
 24 ranging from 1 MeV to about 8 MeV. The neutron, after thermalizing, cap-
 25 tures on a gadolinium or hydrogen nucleus, producing a delayed signal with
 26 energy of ~ 8 MeV or 2.2 MeV, typically within $100 \mu s$. The correlation in
 27 time and space between the prompt and delayed signals provides a distinctive
 28 $\bar{\nu}_e$ signature, greatly suppressing backgrounds.

29 The number of detected antineutrinos N_{det} is given by

$$N_{\text{det}} = \frac{N_p \epsilon}{4\pi L^2} \int \sigma(E) P_{\text{sur}}(E, L, \theta_{13}) S(E) dE \quad (3)$$

30 where N_p is the number of free protons in the target, ϵ is the efficiency of
 31 detecting an IBD, σ is the total cross section of the IBD process, P_{sur} is
 32 the $\bar{\nu}_e \rightarrow \bar{\nu}_e$ survival probability (Eq. 1), and S is the differential energy
 33 distribution of the antineutrinos.

34 Using Eq. 3 to precisely measure θ_{13} from either a deficit in the observed
 35 antineutrino rate or a distortion in the energy spectrum requires that all
 36 systematic uncertainties be well controlled. Background corrections to the
 37 observed antineutrino spectrum can be reduced with overburden and detector
 38 shielding. Detector efficiency uncertainties in the 2-3% range were achieved
 39 in previous experiments. The reactor antineutrino flux and spectrum can
 40 also be calculated to 2-3% precision [10, 11].

41 To eliminate the dependence on the neutrino flux, Mikaelyan and Sinev
 42 proposed the use of two antineutrino detectors [12]. The first detector, lo-
 43 cated close to the reactor, measures the antineutrino flux as a function of

energy while a second detector further from the reactor is located close to the oscillation minimum of P_{sur} . For this situation, one near and one far detector observing antineutrino events from a single reactor, the ratio of antineutrinos observed in a specific energy range at the two detectors is given by

$$\frac{N_f}{N_n} = \left(\frac{N_{p,f}}{N_{p,n}} \right) \left(\frac{L_n}{L_f} \right)^2 \left(\frac{\epsilon_f}{\epsilon_n} \right) \left[\frac{P_{\text{sur}}(E, L_f, \theta_{13})}{P_{\text{sur}}(E, L_n, \theta_{13})} \right] \quad (4)$$

where $N_{p,f}$ and $N_{p,n}$ refer to the numbers of target protons at the far and near sites, and L_f and L_n are the distances of the far and near detectors from the reactor core. The ratio of the detector efficiencies (ϵ_f/ϵ_n) can be determined more precisely than the individual efficiency. The near-far relative measurement cancels nearly all reactor-related and detector-related systematic errors.

The Daya Bay experiment uses eight antineutrino detectors (ADs) to monitor 6 nuclear reactor cores. To minimize efficiency differences between the near and far detectors, interchangeable ADs of identical design were built and assembled above ground using standardized procedures. Built in pairs, antineutrino detectors were filled with liquid scintillator within eight days of each other to further minimize possible differences in performance.

This paper describes all of the Daya Bay detector systems. After a brief overview of the experimental layout in Section 2, the construction, assembly and installation of the detectors are presented in Section 3. Sections 4 and 5 describe the antineutrino detector and calibration systems. The muon detector system overview is given in Section 6, followed by a description of the data acquisition system in Section 7 and offline computing in Section 8. Sev-

Figure 1: Layout of the Daya Bay experiment. The dots represent reactor cores. Each pair of cores forms a Nuclear Power Plant (NPP). The three NPPs, Daya Bay, Ling Ao, and Ling Ao-II are between the Daya Bay coast and 400-m-high mountains inland. Eight antineutrino detectors (ADs) were installed in three experimental halls (EHs). Additional halls were used for filling the detectors (LS Hall) or processing the water (Water Hall) for the experimental hall water pools. The ADs were assembled in the Surface Assembly Building (SAB) before being moved underground. Not shown are surface support buildings adjacent to the SAB containing the control room, electrical and ventilation utilities.

Table 1: Vertical overburden in meters-water-equivalent (mwe), estimated muon rate R_μ and average muon energy in the three experimental halls.

	Overburden(mwe)	$R_\mu(\text{Hz/m}^2)$	$\langle E \rangle(\text{GeV})$
EH1	250	1.27	57
EH2	265	0.95	58
EH3	860	0.056	137

67 eral years of calibration and antineutrino data measure detector performance
68 over time and allow comparisons of the ADs in Section 9.

69 2. Experimental Layout

70 As depicted in Fig. 1, the Daya Bay reactors are arranged in three pairs
71 of reactor cores spread over nearly 2 km of coastline in southern China.
72 Each pair of cores is designated a Nuclear Power Plant (NPP). Experimental
73 Hall 1 (EH1) measures the antineutrino rate and spectrum primarily from
74 the Daya Bay (DYB) NPP. Similarly, Experimental Hall 2 (EH2) measures
75 the rate and spectrum primarily from the Ling Ao and Ling Ao-II NPPs.
76 These near sites are positioned as close to the reactor cluster as possible,
77 given constraints on the desired overburden and rock condition. The location
78 of the far hall (EH3) was determined from a multi-parameter optimization
79 of desired baselines for maximum oscillation sensitivity given the mountain
80 profile and rock quality [13]. The water equivalent overburden, simulated
81 muon rate and average muon energy at each site are listed in Table 1.

82 Daya Bay deploys two ADs in each near hall and four ADs in the far hall.
83 Multiple, functionally identical, movable detectors are unique to the Daya

84 Bay experiment. Dividing the target mass in each hall into standard units,
85 built above ground, allowed the construction and assembly of the antineu-
86 trino detectors to proceed in parallel with preparation of the experimental
87 halls, considerably reducing the total project time. The large total target
88 mass at the far hall enables the unprecedented statistical precision of the
89 Daya Bay experiment. Side-by-side comparisons between detectors at the
90 near sites cross check calculations of relative detector efficiencies to better
91 than 0.2%.

92 Daya Bay ADs are constructed from three-concentric-cylindrical tanks.
93 The inner acrylic vessel (IAV) contains gadolinium doped liquid scintillator.
94 A second acrylic vessel (OAV) containing un-doped liquid scintillator sur-
95 rounds the IAV. The OAV is surrounded by a stainless steel vessel (SSV)
96 containing mineral oil. The inner cylindrical wall of the SSV supports nearly
97 two hundred photomultipliers (PMTs) pointed inward toward the OAV and
98 IAV. Reflectors above and below the OAV redirect light to the PMTs, thereby
99 reducing the total number of PMTs required.

100 The antineutrino detectors are placed in water pools. At least 2.5 meters
101 of ultra pure water surround each antineutrino detector and shield against
102 background radiation. Instrumented with PMTs, the pool is segmented into
103 two Cerenkov detector zones for active identification of cosmic rays. The
104 pool is covered by resistive plate chambers (RPCs) for additional cosmic ray
105 muon identification. Figure 2 shows three ADs in the far water pool.

106 *2.1. The reactor complex*

107 The Daya Bay nuclear power complex is located 55 km northeast of Hong
108 Kong and 45 km east of Shenzhen. The six reactor cores are pressurized

Figure 2: Far hall with 3 ADs installed. The water pools are segmented into two independent zones - an outer water shield 1m thick along the perimeter and over the floor and inner water shield containing the antineutrino detectors. Four layers of Resistive Plate Chambers over the water pool add further cosmic ray identification capability.

109 water reactors with a maximum of 2.9 GW thermal power [14]. The cores
 110 in each pair are separated by 88 m. The Ling Ao and Ling Ao-II NPPs
 111 are approximately 500 m apart, while the Daya Bay and Ling Ao NPPs are
 112 separated by ~ 1100 m.

113 The Daya Bay and Ling Ao NPPs are based on the Framatone ANP
 114 French 990 MW_e (gross electrical power) design. The Ling Ao-II NPPs have
 115 an updated design (CPR1000) with 1080 MW_e. The Daya Bay reactors
 116 utilize $\sim 4\%$ enriched ^{235}U fuel. Each reactor core contains 157 3.7-m-long fuel
 117 rod assemblies. The centroid of the power production is known to about 2 cm
 118 horizontally and 20 cm vertically. Approximately 1/3 of the fuel assemblies
 119 of the Daya Bay NPP reactors are replaced during a refueling shut-down
 120 every 18 months. In the Ling Ao NPP reactors, $\sim 1/4$ of the fuel assemblies
 121 are replaced every 12-18 months.

122 The thermal power of each core is monitored continuously by the power
123 plant with an uncertainty of 0.5% [15–17]. The fission fractions of the four
124 main isotopes contributing to power (and antineutrino) production are mod-
125 eled as a function of time to follow the burn-up of the nuclear fuel [18]. Daily
126 thermal power measurements and fission fraction estimates are provided to
127 the collaboration by the power plant.

128 *2.2. Underground Halls*

129 The detectors reside in three large experimental halls (EH) excavated
130 from granite bedrock. The halls are connected by large cross-section ($6.2 \text{ m} \times$
131 7.0 m) tunnels through which completed antineutrino detectors were moved
132 from the filling hall to the experimental halls. The rock overburden is 93 m
133 at EH1, 100 m at EH2 and 324 m at EH3. Core samples taken throughout
134 the area found the average rock density to be 2.6 g/cm^3 .

135 The two near experimental halls are 15 m wide by 45-50 m long. As
136 seen in Fig. 3 the roof of the 28 m long central section of the hall is nearly
137 15 m above the hall floor to provide space for a 120-ton crane covering the
138 water pool and detector staging areas next to the pool. An area 18 m long
139 by 3 m high provides storage space for the RPC system when rolled back
140 from the water pool. The far experimental hall housing four ADs is 21 m
141 wide by 62 m long. The experimental halls are isolated from the tunnels by
142 roll-up doors to keep dust to a minimum. Fresh, filtered, and cooled air is
143 delivered from above ground by a network of vents which run through the
144 tunnels. Cleaning crews worked constantly to keep the halls clean during
145 installation and commissioning. Auxiliary rooms for the gas system, data
146 taking, water purification and refuge are located next to each experimental

147 hall. A secondary personnel egress tunnel links these auxiliary rooms to the
148 main access tunnel.

149 Ten-meter-deep water pools house the ADs in each hall [19]. The water
150 pools, 10 m wide by 16 m long for near halls and 16 m by 16 m for the far hall,
151 were constructed in an octagonal shape, as shown Fig. 2. The water pools
152 were excavated from the rock and lined with waterproof concrete strengthened
153 ened by grounded rebar. A PermaFlexTM liner which had been extensively
154 tested for compatibility with high purity water was sprayed over the concrete
155 to seal the surface against radon penetration. A thick black rubberized sheet
156 is suspended over the pool by cables for light and gas isolation and is securely
157 fastened to the edges of the pool. Nitrogen from cryogenic liquid bottles is
158 used as an inert cover gas between the cover and water surface. The mea-
159 sured average radon concentration in the water is less than 50 Bq/m³ for the
160 three halls, about 3-8 times lower than the radon concentration in the air
161 above the pool cover.

162 Two special purpose underground halls were also built. A liquid scintil-
163 lator (LS) hall located near EH1 was used to produce and store AD liquids
164 under clean room conditions. The LS hall contains two thermoplastic lined
165 concrete storage tanks of 200 ton capacity, five 40-ton capacity acrylic tanks,
166 and related piping and pumping systems. After LS production the hall was
167 used to fill ADs. Near the intersection point of the main tunnels a water hall
168 contains the central water purification station. Highly purified city water is
169 sent to the three experimental halls with pipes of about the same length to
170 ensure equal water quality [20]. The LS and Water hall locations are shown
171 in Fig. 1.

Figure 3: Cross section of a near hall showing an antineutrino detector on a stand in the water pool beneath the RPC support structure and overhead crane.

172 2.3. Baseline and position survey

173 The distances (baselines) from each AD to each reactor core are precisely
 174 measured via two independent surveys [1, 2, 21] using Global Positioning
 175 System (GPS) receivers and modern theodolites. GPS control networks and
 176 Total Station traverse networks were installed to link measurements of the re-
 177 actor complex area to points outside each of the experimental halls. Further
 178 Total Station surveys linked each hall position to the external monuments
 179 on each AD which had been used as references for surveys of the AD interior
 180 during construction and assembly. The GPS control points combined pre-
 181 existing power plant survey points with newly installed points to obtain a
 182 comprehensive, redundant network overlapping with the Total Station tra-
 183 verse network. Shown as red lines in Fig. 4, the GPS network gives static

Figure 4: Baseline survey. Bigger black circles are reactor cores, smaller black circles are ADs. Tunnels are shown in grey. Red lines represent the GPS control network. Blue lines represent the Total Station transverse network.

184 synchronous observations based on a high-precision dual frequency GPS re-
 185 ceiver. The Total Station transverse network shown as blue lines in Fig. 4
 186 provides coordinates for all the transverse points using a Leica TDA5005
 187 Total Station.

188 To minimize propagation errors inside the tunnels several measures were
 189 adopted. Dual transverse networks were setup in each tunnel to increase
 190 reliability. An additional transverse network loop through the main Daya
 191 bay tunnel and out the construction tunnel (Exit in Fig. 4) checked for error
 192 buildup. A station was positioned on the top of the mountain to measure
 193 points at the tunnel entrance, exit, and reactors to increase conversion pre-
 194 cision between the two networks. The overlap between GPS and transverse
 195 network improved the conversion precision between these two coordinate sys-
 196 tems.

197 The measured baselines are given in Table 2. The largest baseline differ-

Table 2: Baselines from the antineutrino detector centers to the reactor core centers in the Daya Bay (D1, D2), Ling Ao (L1-2) and Ling Ao-II (L3-4) Nuclear Power Plants is given in meters. The total survey uncertainty is 18 mm.

	D1	D2	L1	L2	L3	L4
EH1-AD1	362.380	371.763	903.466	817.158	1353.618	1265.315
EH1-AD2	357.940	368.414	903.347	816.896	1354.229	1265.886
EH2-AD1	1332.479	1358.148	467.574	489.577	557.579	499.207
EH2-AD2	1337.429	1362.876	472.971	495.346	558.707	501.071
EH3-AD1	1919.632	1894.337	1533.180	1533.628	1551.384	1524.940
EH3-AD2	1917.519	1891.977	1534.919	1535.032	1554.767	1528.079
EH3-AD3	1925.255	1899.861	1538.930	1539.468	1556.344	1530.078
EH3-AD4	1923.149	1897.507	1540.667	1540.872	1559.721	1533.179

ence between the two surveys is 4 mm and the uncertainty in the baselines is determined to be less than 18 mm. The largest contribution to the overall uncertainty was the measurement error of 12.6 mm using the GPS network. Other contributions were a 9.1 mm error due to fitting uncertainties associated with the linking of the GPS and the Total Station networks, a measurement error of 6.5 mm using the Total Station transverse network, a 5.9 mm error coming from conversion from the reactor core to GPS control network, and an error of 2.8 mm coming from conversion between antineutrino detector center and the GPS control network. When combined with the uncertainties of the effective center of fission [2], these baseline uncertainties were found to make a negligible contribution to the uncertainties of the oscillation parameters.

210 3. Construction, Assembly and Installation Overview

211 Civil construction started in October, 2007 with the beginning of tunnel
212 excavation. Upon completion of the surface assembly building (SAB) in 2009,
213 test assembly of a full-size prototype antineutrino detector initiated onsite
214 assembly activities. As the tunnels and experimental halls were excavated
215 and prepared for beneficial occupancy, AD assembly and preparation of the
216 muon system PMTs and RPCs proceeded in parallel above ground in the
217 SAB. Well-coordinated underground and surface activities ensured the timely
218 start of the experiment.

219 All assembly of the antineutrino detectors except for filling and final ca-
220 bling was performed in the SAB. The north side of the SAB, used primarily
221 for testing muon system PMTs and storage, was serviced with a 10-ton over-
222 head crane. The south side of the building was serviced by a 40-ton overhead
223 crane and contained a 11 m wide by 33 m long by 11 m high clean room and
224 a smaller adjacent cleaning area. The clean room contained two four-meter
225 deep pits to allow insertion of AD components into the Stainless Steel Vessel
226 (SSV) at convenient working heights as shown in Fig. 5. Specialized plat-
227 forms provided safe access inside and around the perimeter of the AD. The
228 clean room was designed to meet ISO7 cleanliness standards (ISO 14644-1).
229 Cleaning crews were constantly at work in the clean room and cleaning area
230 to keep the floors and tables dust and dirt free. The weekly measured parti-
231 cle counts (≥ 0.5 micron) were typically $\leq 10,000$ counts/m³ which is better
232 than the more rigorous ISO6 standard. Radon levels in the clean room were
233 measured to be 147 ± 13 Bq/m³.

234 A roll-up curtain wall separated the cleaning area from the clean room.

Figure 5: Photograph of the SAB clean room showing two ADs under construction. Both construction pits contain a stainless steel vessel (with lid removed) and an outer acrylic vessel. The lower AD has two photomultiplier support ladders being test fit. The lid of the lower outer acrylic vessel has been removed to allow insertion of an inner acrylic vessel.

235 This semi-clean area was isolated from the rest of the SAB by a second
236 roll-up curtain wall which kept particulate levels low. The cleaning area
237 was equipped with a resin bed water filtration system which provided >
238 10 megaohm-cm water for cleaning items that would enter the clean room.

239 All eight ADs were constructed and assembled in the same clean room
240 over the period of 2009-2012 using the procedures developed during the pro-
241 totype assembly. The time needed for AD assembly in the SAB decreased
242 from 13 months for the first pair to 3.5 months as the production rate of sub-
243 components improved and the crew gained experience. ADs 1 and 2 were
244 moved underground, filled with liquids, installed and commissioned in EH1
245 in July - Sep., 2011. A comparison of detector performance of these ADs
246 can be found in [21]. AD3 was installed in EH2 the following October. ADs
247 4-6 were then installed in EH3. Six AD data collection started in Dec. 2011.
248 Installation of ADs 7 and 8 completed the planned deployment in Aug.-Sep.,
249 2012. All eight ADs have been collecting data since Oct. 2012.

250 The Liquid Scintillator hall was the first underground hall ready for occu-
251 pation in 2010. Two large storage tanks and five smaller acrylic tanks were
252 used to hold AD liquid raw materials and finished liquid scintillators. Multi-
253 ple tanks and pump stations were installed and kept in clean room conditions
254 for AD liquid production and testing. All liquid scintillators were produced
255 by early 2011 and monitored for nearly 1/2 year before first use in an AD.
256 ADs were brought into the LS hall one at a time for filling. A custom filling
257 system precisely measured the liquid masses pumped into each AD while
258 closely monitoring liquid levels to avoid damaging level mismatches.

259 The first near experimental hall was completed in 2011. The water pool

260 was lined with waterproof concrete, painted with three layers of PermaFlexTM,
261 and also lined with Tyvek TM sheets heat-welded together. Preassembled
262 frames were fastened to the wall and floor to support the muon PMTs. An
263 inner Tyvek layer was added to separate the inner and outer regions.

264 The muon detector assembly proceeded in parallel with AD assembly
265 and preparation of the experimental halls. PMTs were tested and placed
266 in mounting brackets and magnetic shields in the SAB. After completion of
267 the experimental halls, AD support stands were installed in the pools onto
268 precisely surveyed studs protruding through the concrete floor and anchored
269 to the underlying rock. Muon PMTs were then installed onto the support
270 structure and cabled to the electronics room. Interlaced with the PMT in-
271 stallation, RPC support structures were installed on the rails on either side
272 of the water pool. RPC modules containing four RPC layers were installed
273 onto the structural supports. The entire RPC assembly was then rolled into
274 its storage area for completion of the cabling and gas connections while the
275 ADs were installed.

276 A crucial element of the assembly plan was the custom automatic guided
277 vehicle transporter [22] (AGV) used to transport ADs and other large parts
278 within the SAB, dry ADs down the 10% grade from the SAB to LS Hall, and
279 filled ADs from the filling hall to the experimental halls. The AGV operated
280 under diesel or electric power as required and could support up to 130 tons.
281 When transporting filled ADs through the $\leq 0.5\%$ grades of the main Daya
282 Bay tunnels, the AGV adjustable height load bed was kept level used an
283 auto-leveling feature to reduce stress on AD components. Eight sets of heavy
284 duty omni-directional wheel modules provided 360 degree maneuverability for

precisely positioning heavy loads.

ADs were brought into the experimental hall by the AGV and set down on a temporary support stand by the edge of the water pool by lowering the AGV bed. The lifting fixture and video position monitoring system were mounted on top of the AD. The AD was then lifted by the overhead crane into the pool onto the support stands. Access ladders and platforms were mounted on top of the AD for removal of the lifting fixture and video system. Cables from the AD PMTs were run through dry pipes to the edges of the pool before being routed to the electronics room. Similarly, cables from the automated calibration units (ACU) and lid sensors were run through a large dry pipe to above the pool surface, and the cover gas lines were run to the gas room. Function tests of all AD systems were performed before final leak checks of all connections ensured that the detector was leak tight. A final survey of the AD position within the hall was conducted before filling the pool with water. The final installation steps were to seal the pool with a gas-tight, opaque, black cover and to roll the RPCs over the pool. More detailed descriptions of the assembly and installation processes can be found in references [19, 23].

4. Antineutrino Detectors

The antineutrino detectors are designed to detect IBDs while minimizing systematic errors. As shown in Fig. 6, an antineutrino detector is made from three concentric vessels of cylindrical shape. Cylinders are easier to construct than spheres and have a more uniform detector response than rectangular geometries. The inner target volume of gadolinium doped liquid scintillator

309 (GdLS) is contained by an inner acrylic vessel (IAV) approximately 3 m
 310 high and 3 m in diameter. Surrounding the IAV is an outer acrylic vessel
 311 (OAV) approximately 4 m high and 4 m in diameter filled with un-doped
 312 liquid scintillator (LS). The LS improves the detection efficiency of gamma
 313 rays from interactions in the GdLS target volume. The OAV is inside a
 314 SSV approximately 5m high and 5m in diameter which supports ladders of
 315 photomultipliers (PMTs) pointed at the GdLS and LS volumes and isolates
 316 the inner detectors from the water pool. The space between the SSV and
 317 OAV is filled with high purity mineral oil (MO) which suppresses radioactive
 318 backgrounds from the PMT glass and the surrounding environment.

319 Unlike similar experiments with cylindrical inner detectors [24, 25], Daya
 320 Bay uses circular reflectors above and below the OAV instead of PMTs [13].
 321 The reflectors redirect scintillation light towards the PMTs mounted on the
 322 perimeter and reduce the required number of PMTs by 50% without signif-
 323 icantly degrading the energy or position resolution. The dimensions of the
 324 three vessels were chosen to maximize the target mass without increasing
 325 the systematic uncertainty due to backgrounds. A Daya Bay AD has a ra-
 326 tio of GdLS target mass to total liquid mass of 26%, higher than similar
 327 experiments, such as Double Chooz [24] or RENO [25]. Radioactive back-
 328 ground rates are reduced sufficiently to render systematics in the accidental
 329 background subtraction negligible in the near/far oscillation analysis.

330 Mounted on the SSV lid are three ACUs which can position radioactive
 331 calibration sources or light emitting diode (LED) pulsers at different positions
 332 in the GdLS and LS volumes. The ACUs and liquid volumes are connected
 333 by calibration tubes. The central calibration tubes also connect the GdLS

Figure 6: Antineutrino detector.

and LS volumes to concentric overflow tanks on the SSV lid. All gas spaces above the overflow tanks were continuously purged with dry nitrogen from a cover gas system [26].

The eight ADs were built above ground in pairs and then moved below ground for filling. GdLS, LS, and MO were pumped simultaneously into an AD as needed while keeping the relative liquid levels within 10 cm of each other. The liquid levels were then topped up to fill the overflow tanks to about 1/3 capacity.

Although the nominal pair-wise production and assembly plan should have resulted in an orderly association of the first SSV with the first IAV and first OAV, this pattern and the orderly placement of ADs into the halls was quickly broken by practical schedule considerations. The first SSV needed extra work to obtain a suitably smooth finish in the O-ring grooves and was used in AD3. The second IAV pair hold-down tabs and calibration ports were mis-aligned, necessitating complimentary changes in the last pair of OAVs. IAV3 was damaged during an annealing accident and was replaced by IAV9.

Table 3: Naming scheme for antineutrino detector (AD) placement in the three experimental halls (EH). The major components of the ADs: stainless steel vessel (SSV), outer acrylic vessel (OAV), and inner acrylic vessel (IAV) were labeled by production order. ADs were assembled and filled in pairs. Even though the components were also typically produced in pairs, the order in which the components were used depended on special considerations described in the text.

	AD#	SSV#	OAV#	IAV#
EH1-AD1	AD1	SSV2	OAV1	IAV2
EH1-AD2	AD2	SSV3	OAV2	IAV1
EH2-AD1	AD3	SSV1	OAV3	IAV5
EH2-AD2	AD8	SSV8	OAV7	IAV4
EH3-AD1	AD4	SSV4	OAV4	IAV6
EH3-AD2	AD5	SSV5	OAV5	IAV8
EH3-AD3	AD6	SSV6	OAV6	IAV7
EH3-AD4	AD7	SSV7	OAV8	IAV9

350 To obtain the earliest physics data, it was decided to run with only 6 ADs
351 spread over the three experimental halls. The AD naming scheme and major
352 components of each AD are shown in Table 3.

353 Control of radioactive backgrounds was achieved by strict control of all
354 materials used in the AD and of all procedures used during assembly and
355 construction. Details for individual parts can be found in the related section.

Figure 7: Stainless steel vessel (SSV). Left: barrel and bottom sections. Right: inner side of the tank lid.

4.1. Stainless Steel Vessel

The stainless steel vessels (SSV) are 5000 mm high, 5000 mm diameter cylinders with 12 mm thick low radioactivity stainless steel walls. The external walls are strengthened by 12 mm thick internal ribs. Each SSV(dry) weighs about 24 ton (with an inner volume of $\approx 95 \text{ m}^3$). Figure 7 shows the structure of the barrel and lid. When loaded with the acrylic vessels, PMTs, and liquids, an AD weighs ≈ 112 tons. Deformations and mechanical strength for the SSV vessel were thoroughly studied using Finite Element Analysis (FEA) code to insure that the design had the required safety factor when a filled SSV was picked up and lowered onto its support stands or when the lid was submerged under 2.5 m of the water. SSVs were constructed in pairs by the Guangdong Shanfeng Chemical Machinery Co. LTD. A SSV barrel was welded together from three subsections: a bottom section including bottom ribs, a top section with a top flange, and a middle barrel section.

371 4.1.1. SSV O-ring and lid

372 The removable SSV lid is sealed at the circumference by two 19 mm thick
373 O-rings placed in two concentric grooves 23 mm wide by 14 mm high. With
374 the 112 lid bolts tightened to 120 N-m the O-ring is typically compressed
375 by 16-26%. Similar, but smaller, O-rings were used to seal the overflow
376 tank lids to the SSV lid. The O-rings were required to be compatible with
377 water, mineral oil and liquid scintillator, and to provide effective sealing
378 under pressure for more than five years. DuPontTM Fluorocarbon Rubber
379 Viton[®] A supplied by Shenzhen-O Technology Co., Ltd was chosen as the
380 O-Ring material and passed all requirements. O-rings were subjected to a
381 60 day accelerated aging study at elevated temperature (110-170 °C) under
382 the compression of 17% at the Aerospace Materials and Technology Research
383 Institute. Based on these studies, permanent O-ring deformation after five
384 years at nominal temperatures ($\leq 25^{\circ}\text{C}$) was projected to be $\leq 9\%$ in water.

385 The Viton-A O-rings were made using a segmented mold vulcanizing
386 process, which achieved the same performance as a molded vulcanized O-
387 ring seal without making a very big mold. The space between the O-rings
388 was connected to an external port to enable leakage tests [27] of the final
389 O-ring seal to ensure that the SSV would not leak when immersed in the
390 water pool.

391 4.1.2. Fluorocarbon painting inside SSV

392 The inside of the SSVs are coated with a specially developed black flu-
393 orocarbon paint [28] to minimize reflected photons. Although common flu-
394 orocarbon paints are compatible with mineral oil, low in radioactivity and
395 strongly adhesive, they typically require oven curing at high temperature.

Figure 8: Diffuse reflectivity of the black fluorocarbon paint developed for Daya Bay versus wavelength range.

Since available ovens were $\leq 20 \text{ m}^3$, a new kind of fluorocarbon paint that cures at normal temperatures was adopted for the $\sim 100 \text{ m}^3$ SSV. Different paint formulas were applied to 304L stainless steel samples. The chosen black-matte fluorocarbon paint, a mixture of diluting agent, solidifying agent and fluorocarbon resin, adhered well to the stainless steel, had good uniformity, and was scratch and wear resistant. The diffuse reflectivity of the black matt fluorocarbon paint was measured over wavelengths ranging from 250 to 800 nm as shown in Fig. 8. Reflectivity was roughly a constant (4.5%) above 300 nm. The radioactivity of the paint was measured by a very sensitive γ -ray detector consisting of a high-purity Ge detector and anti-Compton system to be $\leq 0.31 \text{ Bq/Kg}$ for ^{214}Pb , $\leq 0.21 \text{ Bq/Kg}$ for ^{212}Pb , $\leq 0.29 \text{ Bq/Kg}$ for ^{238}Ac , and $\leq 0.30 \text{ Bq/Kg}$ for ^{40}K .

4.1.3. Radioactivity Control

Stainless steel plates and welds typically have natural radioactive contaminants which could generate backgrounds in the inner detectors. A Monte

411 Carlo study established requirements for allowable radioactivity levels from
412 the major contaminants. Quality assurance procedures were implemented to
413 ensure that construction materials met these goals. A total of 260 tons low
414 radioactivity 304L stainless steel was specially made by the Shanxi Taigang
415 Stainless Steel Co., Ltd. in one batch. The steel sample was radioassayed
416 before the steel ingots were rolled to sheets of more than 10 different thick-
417 ness as required by the SSV design. The contamination goals and measured
418 radioactivity of materials used in SSV construction are shown in Table 4.
419 Welding materials were selected from many vendors after assaying radioac-
420 tivity in samples. The radioactivity of the chosen welding materials was
421 measured to be 0.14 ± 0.08 Bq/kg for ^{238}U , < 0.30 Bq/kg for ^{232}Th , and
422 < 0.48 Bq/kg for ^{40}K . The contribution to the overall background rate was
423 negligible due to the small mass of the welds compared to the total SSV mass.
424 A stainless steel disk sample with a welding line was also measured to verify
425 that no additional radioactivity was introduced during the welding process.
426 The radioactivity of all SSV materials was well below the requirements set
427 by Monte Carlo studies.

428 *4.2. Acrylic vessels*

429 All of the inner components of the AD which come into contact with the
430 liquid scintillator (LS or GdLS) are constructed from acrylic or teflon to en-
431 sure chemical compatibility with the LS. Short and long-term compatibility
432 tests of all materials were performed before selection for use in the AD. Ul-
433 traviolet transmitting (UVT) acrylic was chosen both for its increased trans-
434 mission of the light produced by the liquid scintillator at short wavelengths
435 and for the absence of chemical additives that protect against yellowing from

Table 4: Measured radioactivity levels in the stainless steel used in SSV construction are compared to the maximum allowable contamination levels determined from Monte Carlo Studies. The expected contribution to the AD trigger rate from each contaminant at the allowed level is also shown.

Isotope	Requirement (Bq/Kg)	Measured Value (Bq/Kg)	AD Trigger Rate (Hz > 1 MeV)
^{238}U	< 0.03	< 0.00124	8.5
^{232}Th	< 0.02	< 0.006	8.5
^{40}K	< 0.13	< 0.013	5
^{60}Co	< 0.03	< 0.002	11

UV light. Special care was taken to avoid prolonged exposure of the acrylic components to sunlight during construction and assembly [29].

Both acrylic vessels were built with conical lids to allow bubbles from the filling process to escape easily into the overflow tanks. Tanks were visually inspected during construction and assembly for any areas of cracks or crazing. Any flaws discovered were repaired by cutting out the affected area and replacing with fresh material. This visual inspection also served as the primary leak test since any imperfections were clearly visible in the optically clear materials. More detailed descriptions of the acrylic vessel design, construction and assembly are found in references [30, 31].

4.2.1. Inner Acrylic Vessels

The walls of the inner acrylic vessels (IAV) were constructed at Nakano International Limited, in Taipei, Taiwan [32] from 10 mm thick UVT sheets

Figure 9: Inner acrylic vessel (IAV).

449 manufactured by PoSiang. The bottom and top lid were made from 15-mm-
 450 thick UVT sheets. Cut sheets were first bonded into large panels and then
 451 heat formed around cylindrical or conical forms to achieve the desired shape.
 452 Ribs made from 55-mm-thick material were bonded onto the bottom and
 453 lid to stiffen the shape and to provide lifting points. All components were
 454 polished until optically clear. The lid and cylinder were bonded together as
 455 a unit. All surfaces were then cleaned inside a class 10,000 clean room before
 456 the final bond between the base of the cylinder and bottom were made. A
 457 drawing of the IAV showing the external ribs and calibration ports is shown
 458 in Fig. 9.

459 The physical size and position of the IAV within the AD was tracked to
 460 provide accurate data for simulation studies and to ensure that the calibration
 461 ports were properly aligned. The diameter of the IAVs were calculated from
 462 the measured circumference at several heights. The height was measured
 463 from the outside of the bottom to the top of the lid rim excluding ribs. Wall

464 thickness was measured by an ultrasonic thickness gauge. These data are
 465 shown in Table 5 and compared to the design dimensions. Although the
 466 specifications were mostly met, IAV5 was nearly 20 mm taller than the other
 467 ADs due to a manufacturing error. The inner volume while dry is calculated
 468 from the derived internal dimensions. Since the thin walled acrylic walls are
 469 not perfectly rigid, the GdLS filled volume is expected to differ from the dry
 470 volume. The overall IAV acrylic mass was determined from a spring scale
 471 during assembly.

Table 5: Outer IAV dimensions from the as-built surveys. Errors are the r.m.s. of the
 measured points. The acrylic mass resolution was 5 kg. The inner volume (dry) was
 estimated from the outer dimensions, wall thickness, and the design angle of the conical
 lid.

	height (mm)	diameter (mm)	wall (mm)	mass (kg)	volume (m ³)
design	3100	3120	10.0		
EH1-AD1	3101 $\pm$ 2	3123 $\pm$ 2	10.7 $\pm$ 0.8	907	23.39
EH1-AD2	3106 $\pm$ 1	3123 $\pm$ 1	10.6 $\pm$ 1.0	916	23.43
EH2-AD1	3102 $\pm$ 2	3125 $\pm$ 2	10.9 $\pm$ 0.9	915	23.42
EH2-AD2	3109 $\pm$ 2	3120 $\pm$ 1	10.7 $\pm$ 0.9	950	23.40
EH3-AD1	3101 $\pm$ 2	3122 $\pm$ 2	10.9 $\pm$ 0.8	945	23.37
EH3-AD2	3123 $\pm$ 2	3117 $\pm$ 3	no data	965	23.79
EH3-AD3	3102 $\pm$ 2	3115 $\pm$ 3	no data	920	23.57
EH3-AD4	3102 $\pm$ 2	3118 $\pm$ 2	10.8 $\pm$ 0.6	945	23.31

472 Except for the IAV prototype, IAVs were constructed in pairs using the

473 same techniques and procedures. As previously described, the IAV which
474 was used in AD7 was a replacement for an IAV damaged during an annealing
475 procedure. The replacement IAV was built one year after the construction
476 of the other IAVs. The IAV in AD8 was damaged during tests of the manual
477 calibration system. The damaged section of IAV8 was cut out of the bottom
478 floor and new acrylic was bonded in place. Neither of the affected AD's
479 performance differs significantly from other ADs in studies to date.

480 *4.2.2. Outer Acrylic Vessels*

481 The outer acrylic vessels (OAV) were manufactured by Reynolds Poly-
482 mer Technology, Inc. [33] in Grand Junction, Colorado, USA. The cylindri-
483 cal walls were formed from eight 18-mm-thick Polycast UVT acrylic sheets
484 bonded together to form half height cylinders. Reynolds cast large, thick
485 UVT sheets for the top and bottom. Two such sheets were bonded together
486 to make blanks for the 3D milling machine. The conical shape was cut into
487 the underside of the lid. The top of the lid was thinned down to 18 mm
488 thickness except for the six ribs and central hub. Holes were cut for the
489 calibration ports. Additional material was bonded to the ribs and central
490 hub to strengthen the OAV lid. The cylinder flange and a half cylinder were
491 bonded together as were the bottom and the other half cylinder. Bonding the
492 two half cylinders together completed the process. A drawing of a completed
493 OAV is shown in Fig. 10.

494 All of the acrylic surfaces were sanded smooth and polished with 1-3
495 micron grit aluminum oxide polishing powder until the surfaces were optically
496 clear. The vessels were surveyed in a manner similar to the IAVs. Leak checks
497 of the O-ring seals on the lid and calibration ports were made before the

Figure 10: Outer acrylic vessel (OAV).

498 vessels were mounted in steel shipping frames for transport. The OAVs were
 499 shipped by truck to the port of Long Beach, CA and placed on a container
 500 vessel for shipment to China. The long term testing of acrylic properties is
 501 described in [29, 34].

502 As with the IAVs, the physical size and position of the OAVs were mea-
 503 sured to provide accurate data for simulation studies and to ensure proper
 504 alignment of the ports. The outer height and diameter, wall thickness, and
 505 mass of the OAVs are shown in Table 6.

506 4.3. Overflow tanks and calibration tubes

507 All three AD liquids are connected to overflow tanks on the lid of the
 508 SSV to allow for thermal expansion or contraction of the AD liquids while
 509 keeping the internal volumes full. Figure 11 shows the concentric LS and
 510 GdLS overflow tanks connected to the central calibration tube and the two
 511 MO overflow tanks. The MO overflow tanks are part of the SSV lid structure

Table 6: OAV dimensions from the as-built surveys. Errors are the r.m.s. of the measured points. The acrylic mass resolution was 5 kg. The inner volume (dry) was estimated from the outer dimensions, wall thickness, and the design angle of the conical lid.

	height (mm)	diameter (mm)	wall (mm)	mass (kg)	volume (m ³)
design	3982.	4000.	18.0		
EH1-AD1	3981.5 ± 1.0	3997.4 ± 0.4	16.80 ± 1.06	1851	49.33
EH1-AD2	3997.4 ± 0.4	4002.3 ± 0.7	16.42 ± 0.95	1876	49.46
EH2-AD1	3986.3 ± 1.3	3998.9 ± 0.6	16.96 ± 0.70	1810	49.42
EH2-AD2	3978.8 ± 1.8	3998.0 ± 0.6	17.57 ± 0.56	1860	49.27
EH3-AD1	3980.8 ± 0.9	3999.9 ± 0.3	17.25 ± 0.70	1860	49.37
EH3-AD2	3983.0 ± 0.5	3999.8 ± 0.9	17.05 ± 0.97	1960	49.40
EH3-AD3	3983.3 ± 1.6	3998.4 ± 0.8	17.61 ± 0.76	1985	49.34
EH3-AD4	3985.0 ± 1.0	3996.3 ± 1.5	17.24 ± 0.80	1910	49.33

Figure 11: Orthogonal slices through the top of an AD showing the GdLS (green), LS (red), and MO (blue) overflow tanks and calibration tubes.

512 and are connected to the MO volume below by several 10 cm holes in the
513 SSV lid.

514 Two additional calibration tubes near the edge of the IAV and near the
515 edge of the OAV provide paths for calibration sources at different radii to
516 study the positional dependence of the detector energy measurement. The
517 OAV tube was constructed from a 10 cm convoluted teflon tube, 16.5 cm
518 long, connected to the OAV and SSV lid by acrylic transition pieces. A
519 double O-ring sliding seal at the top transition adjusts for any variation in
520 detector height and decouples the OAV from movement of the SSV lid when
521 the water pool is filled. Similar double O-ring seals on the top of the 6.4 cm
522 diameter, 65 cm long IAV calibration tube connect the tube to the lid and
523 overflow tanks.

524 As seen in Fig. 12 of the central overflow region, the IAV tube goes
525 through the OAV tube before connecting to the IAV overflow tank. Likewise,
526 the OAV tube connects to the OAV overflow tank. The tanks are constructed
527 from 10-mm-thick acrylic at the Reynolds factory to ensure chemical com-
528 patibility with the GdLS. The inner GdLS (IAV) tank has an inner diameter
529 of 1.28 m and is 134 mm tall. The outer LS (OAV) tank has an inner diam-

Figure 12: Central overflow tanks and calibration tubes are shown. The central IAV tube connects the bottom of the IAV overflow tank (brown) to the top of the IAV and is open at the top to allow the insertion of sources from the central ACU. The central OAV tube connects the bottom of the LS overflow tank (aqua) to the top of the OAV. The overflow tanks are covered by a steel lid (grey) with a mating flange for the central ACU.

530 eter of 1.82 m and is 139 mm tall. The total volume of the GdLS (LS) tank
 531 is ≈ 200 l (165 l). Both overflow tanks are covered with acrylic lids which
 532 have mounts for height, temperature, and tilt sensors. Gas volumes above
 533 and below the acrylic lids are flushed by the AD cover gas system.

534 4.4. Reflective panels

535 Daya Bay ADs were designed to use optical reflective panels at the top
 536 and bottom of the inner detector cylinder to increase the number of detected
 537 photons. In simulation studies, end reflectors increased the total light collec-
 538 tion by 40 to 50% and provided a more uniform response to light produced
 539 at different heights within the GdLS volume. Both specular and diffuse re-
 540 flectors were considered at the design stage. Monte Carlo studies show that
 541 specular and diffuse reflectors have the same total photoelectron yield if the
 542 total reflectivity is the same. Specular reflectors were chosen since position
 543 reconstruction is simpler with specular reflectors. A maximum likelihood

vertex reconstruction method [35] was developed to handle the difficulties of position reconstruction with multiple light reflections.

Highly specular reflective materials, ESR (Enhanced Specular Reflector, 3M[®]) and MIRO-Silver (silver film coated on aluminium base, MIRO[®]), were found to be the best candidates for the AD reflectors. ESR film is a highly reflective, mirror-like optical enhancement film with a typical thickness of 65 μm . It is a non-metallic film made using multi-layer polymer technology and is almost free from radioactivity. The specular reflectivity of ESR was measured to be $>98.5\%$ in the visible region using a Varian-5E spectrophotometer, as shown in Fig. 13. MIRO-Silver film can also reach 98% total reflectivity, with a diffuse reflectivity less than 6%. It was observed that the reflectivity of ESR in air is almost unchanged at large incident angles, while the reflectivity of MIRO-Silver degraded at incident angles $> 70^\circ$. Furthermore, large-size ESR sheets were commercially available, unlike MIRO-Silver. ESR was therefore selected for use in the AD reflectors.

Figure 13: Measured specular and total reflectivity of ESR.

The reflector design was first tested in a $\approx 1/3$ size prototype [36]. Good

560 agreement of the energy response between data from this prototype and
561 Monte Carlo simulations validated our understanding of the optical proper-
562 ties. In this prototype, the ESR was immersed in mineral oil for more than
563 a year, demonstrating good long-term compatibility.

564 In a Daya Bay AD, the PMT surface is 20 cm from the OAV, thus the ideal
565 reflector diameter is 4.5m, to maximize the effective PMT coverage while
566 maintaining enough clearance for installation. To avoid scratches, the ESR
567 film is sandwiched between two acrylic panels. The thickness of each acrylic
568 panel was determined to be 1cm by the FEA analysis to ensure enough me-
569 chanical strength during lifting and installation. The transmittance of each
570 acrylic panel was measured by UV-Vis spectrophotometer, and the results
571 showed very good transparency in both visible and UV region. The final
572 assembly of each reflector was in a 10,000 class clean room. Special care
573 was taken to clean the acrylic panels, the ESR and the lifting fixtures before
574 the assembly. A bulk polymerization technique using methyl methacrylate
575 (MMA) was used to seal the sandwich structure. Several small-size and full-
576 size prototypes of the AD reflector were made to resolve various technical dif-
577 ficulties in making the ESR film lie flat and attach tightly to the acrylic pan-
578 els. Excellent optical surfaces were achieved by continuous vacuum-pumping
579 of the air gap between the acrylic panels during the polymerization process
580 and the gap vacuum sealing stage.

581 A custom lifting fixture equipped with suction cups was designed to move
582 the reflectors as shown in Fig. 14. The lower reflector rests on top of the SSV
583 bottom ribs and supports the OAV. The upper reflector rests on the top of the
584 OAV lid ribs but is additionally cable-tied to the ribs to prevent movement

585 during transportation or liquid filling.

Figure 14: Installation of the bottom reflector

586 4.5. PMT system

587 Light from the liquid scintillators is measured by 192 PMTs mounted
588 on ladders positioned around the SSV perimeter. Effective photocathode
589 coverage with the top/bottom reflectors was about 12%.

590 The performance of the PMTs in an AD plays a critical role for the suc-
591 cess of the experiment since all antineutrino events are reconstructed from
592 the signals provided by the PMT arrays. Based on physics requirements,
593 specifications were developed [13] to guide our selection of the PMTs. We
594 chose the Hamamatsu [37] R5912 ten-stage 20-cm photomultiplier tubes with
595 low-radioactivity borosilicate glass and a tapered base. The R5912 can with-
596 stand pressure up to 7 atmospheres. The photocathodes of the PMTs are
597 biased with positive voltage, allowing the use of a single coaxial cable to pro-
598 vide the HV as well as the transmission of the PMT signal to the front-end
599 electronics.

600 The base of each PMT is sealed inside an acrylic shell with epoxy so that
601 it will not contaminate the mineral oil. All PMTs were thoroughly tested
602 before being installed in the ADs.

603 4.5.1. PMTs

604 After passing visual inspection, PMTs were burned-in for three days while
605 operating at a gain of about 1×10^7 to reduce infant mortality. Only 0.5%
606 of the ≈ 2500 PMTs failed the burn-in tests. A test stand capable of evalu-
607 ating 16 PMTs simultaneously was setup to ensure that each PMT met the
608 requirements listed in Table 7. An additional 1% of the PMTs failed these
609 requirements.

610 All the measured quantities were saved in the Quality Control data base.
611 PMTs passing all the requirements were randomly distributed among the
612 eight ADs to ensure statistically identical performance. Each PMT was
613 wrapped with a truncated conical magnetic shield made of FINEMET[®] to
614 reduce variation in charge collection due to the Earth’s magnetic field [38].

615 Based on five samples of glass used by Hamamatsu in fabricating the Daya
616 Bay R5912 PMTs, the average radioactivity of the PMT glass was measured
617 as 1.9 ± 0.3 Bq/kg for ^{238}U , 1.4 ± 0.4 Bq/kg for ^{232}Th , and 5.1 ± 1.4 Bq/kg
618 for ^{40}K .

619 During commissioning and early operation of the ADs about 5% of AD
620 triggers were found to be caused by “flasher” events. These events seemed to
621 be caused by electrical discharges within the PMT bases, generating visible
622 light which was then detected by other PMTs. The distribution of observed
623 PMT signals was distinctive and analysis cuts were developed to reject this
624 background source [21]. Although some PMTs were more likely than others

Table 7: PMT requirements

Quantum Efficiency	$> 25\%$ @420 nm
Gain	1×10^7
Single photo-electron (SPE) peak to valley ratio	≥ 2.5
Pulse Linearity 1×10^7	2% @40 mA with tapered base
Dark Rate	< 10 kHz
After-Pulse Ratio	$< 10\%$ for SPE main pulse
Rise and Fall Time	$t_{rise} \leq 6.5$ ns $t_{fall} \leq 10$ ns
Transit Time Spread for SPE	< 3 ns (FWHM)

625 to cause "flasher" type events, the inefficiency introduced by the analysis cut
626 ($\leq 0.02\%$) and the residual contamination of IBD events ($\leq 10^{-3}$) were low
627 enough that it was not necessary to turn off any PMTs.

628 *4.5.2. PMT support ladders*

629 The PMTs are mounted on 304-stainless steel tripod brackets which are
630 bolted to a 304-stainless steel curved-frame ladder. Each ladder holds 24
631 PMTs arranged in three columns and eight rows. To reduce the complexity
632 in event reconstruction due to reflected photons, a 3.2-mm-thick matte-black
633 acrylic radial shield (reflectivity 4-5%) is attached to the inner curved surface
634 of the ladder frame. The inactive sections of the PMTs are hidden behind
635 the radial shield. Figure 15 shows the front and back of an assembled PMT
636 ladder. A total of eight ladders holding 192 PMTs were installed in each SSV.
637 The 24 7-m-long RG305 50- Ω coaxial cables from the PMTs of each ladder
638 were routed to one of eight cable feedthroughs on the wall of the SSV. Plugs
639 and O-rings previously installed on each cable were mated to holes in the
640 feedthrough flange. These cables were later connected to 35-m-long JUDD
641 C07947 coaxial cables just outside the SSV in dry pipe elbows which are later
642 flushed with nitrogen gas [26]. Vacuum grade stainless steel bellows between
643 the elbow and the cable tray running around the water pool perimeter keep
644 each group of cables dry.

645 *4.5.3. HV system*

646 The positive high voltages of the PMTs are provided by a HV system
647 based on a CAEN SY1527 LC main frame housing up to eight 1934A 48-
648 channel high-voltage distribution modules. Each HV channel is controlled

(a)

(b)

Figure 15: Front (a) and back (b) of an assembled PMT ladder. Eight ladders hold the PMTs in a cylindrical shell centered on the IAV and OAV.

Figure 16: History of the operating PMT HV of a few typical PMTs. High voltages are set to zero during extended accesses or when recovering from power outages.

649 with a LabViewTM program. Individual voltages and currents are monitored
 650 and archived in the slow-control database. Hardware limits are set on the
 651 output voltage and current to protect the PMTs in case of any error or fail-
 652 ure. Custom-built splitter boxes holding 48 channels of decoupling circuits
 653 were installed in standard racks containing the front-end and trigger elec-
 654 tronics. The high voltage is filtered out, leaving only the analog signals of
 655 the PMTs to be sent to the front-end electronics. PMT gains were set to
 656 be 1×10^7 for normal data taking. The operating HVs for the PMTs in the
 657 three experimental halls were quite stable as shown in Fig. 16.

658 4.6. *AD liquids*

659 Liquids for the ADs were produced and stored in the dedicated under-
660 ground Liquid Scintillator hall (LS hall). The GdLS for all eight detectors
661 was produced in Jan., 2011 and stored in five 40-ton acrylic storage tanks.
662 All of the LS was produced in Mar., 2011. In total, 185 tons of gadolinium-
663 loaded (0.1% by mass) liquid scintillator and 200 tons of unloaded scintillator
664 were produced. Details of the production and characterization of the liquid
665 scintillator are given in [39].

666 4.6.1. *Gadolinium doped liquid scintillator*

667 With a high thermal neutron capture cross section of about 49000 barn,
668 gadolinium has been added to liquid scintillators in previous neutrino exper-
669 iments [6, 40] to reduce backgrounds by shortening the average capture time
670 and to provide a delayed ≈ 8 MeV energy signal from the de-excitation of the
671 Gd nucleus after the neutron is captured.

672 Dissolving inorganic salts of Gd with organic liquid scintillator is chem-
673 ically challenging. Previous studies showed that gadolinium could be either
674 extracted or dissolved into a scintillator through chelating ligands, such as
675 phosphors, diketones, or carboxylates [41, 42]. The Daya Bay liquid scintil-
676 lator uses linear alkylbenzene (LAB), a straight alkyl chain of 10-13 carbons
677 attached to a benzene ring [43], as the solvent, 3g/L of 2,5-di-phenyloxazole
678 (PPO) as the fluor, and 15 mg/L of p-bis-(-o-methylstyryl)-benzene (bis-
679 MSB) as the wavelength shifter. TMHA (3,5,5-tri-methylhexanoic acid) was
680 chosen as the ligand to complex with the gadolinium because of its high
681 solubility in LAB. The solid Gd-complex was dissolved into LAB to form a
682 Gd-LAB solution with 0.5% concentration of Gd by mass. This Gd-LAB

683 solution was then mixed with a pre-prepared concentrated liquid scintillator
684 and then diluted with more LAB to form the final GdLS.

685 Production methods were tested and developed during a prototype stage [41,
686 42]. GdLS test samples of 800 and 600 liters were separately prepared and
687 deployed at IHEP [36] and at the Aberdeen Tunnel Underground Labora-
688 tory [44] for prototype testing. Performance was monitored [42] for over a
689 year before production at Daya Bay to ensure that the chemical stability,
690 optical transparency, and light-yield of the GdLS met the required criteria.
691 The prototype production system, seen in Fig. 17, was reassembled in the LS
692 hall for Daya Bay liquid production. Production procedures can be found in
693 reference [39].

694 The quality of liquid scintillators (GdLS and LS) was controlled by purifi-
695 cation of raw materials and by an optimized LAB production procedure at
696 the manufacturer. Colored contaminants such as iron (Fe) and cobalt (Co),
697 were removed since they degrade optical transparency and affect the chemi-
698 cal stability of the scintillators. Other impurities of concern were radioactive
699 daughters from naturally occurring uranium (U) and thorium (Th) decay
700 chains. These contaminants were required to be ≤ 1 ppb.

701 A self-scavenging purification [45] using pH adjustment and fine-filtration
702 of $\text{GdCl}_3 \cdot x\text{H}_2\text{O}$ dissolved in water removed U, Th, and color impurities ef-
703 fectively. Based on an assessment of nearly 1 kg purified $\text{GdCl}_3 \cdot x\text{H}_2\text{O}$, a
704 U/Th level of ≤ 1 ppb was achieved with no trace of Fe detected. Long-lived
705 radium from the U/Th decay chain could not be removed but was unlikely to
706 complex with TMHA. A total of 1.2 tons of THMA was purified by thin-film
707 vacuum distillation operating at 20 kg per day. The UV spectra of TMHA

Figure 17: Prototype production system for the Daya Bay liquid scintillators. The upper diagram is of the prototype production system in IHEP. The lower photograph shows operation of the production system in the Daya Bay LS hall.

708 before and after purification showed a factor of two improvement in the opti-
709 cal transparency over the wavelength range in which the Hamamatsu R5912
710 PMTs are sensitive.

711 As the solvent, the LAB determines the optical properties of the GdLS.
712 The LAB manufacturer, Jinling Petro-chemical Corporation in Nanjing, China,
713 provided 388 tons of high-quality LAB using an optimized production pro-
714 cess. Heavier alkylbenzene components were removed by narrowing the al-
715 lowed range of distillation temperatures and altering the timing of catalyst
716 use in the production cycle. Transfer pipes, containers and transportation
717 were also carefully planned to avoid contamination. PPO, provided by the
718 Joint Institute for Nuclear Research (JINR), Russia, was further purified by
719 an analytical laboratory (Huashuo Technology Co., Ltd. in Wuhan, China)
720 via filtration after melting, distillation and recrystallization. A total of 1.4
721 tons of PPO was purified and passed the Quality Assurance (QA) require-
722 ment. Water for production processes was produced by a resin-bed water
723 purification system outside the LS hall. The resistivity of the purified water
724 was about 15 M Ω -cm with no detectable metallic impurities.

725 4.6.2. *Liquid Scintillator*

726 LS production used the same production system as the GdLS produc-
727 tion. All parts which had contacted GdLS were cleaned thoroughly before
728 LS production by acid extraction and rinsing with deionized water. Rinsing
729 with a strong acid such as a diluted HCl solution extracts the gadolinium
730 carboxylate into an aqueous phase. The equipment was then washed with
731 deionized water several times to remove the gadolinium. After washing, the
732 water was measured by X-ray fluorescence (XRF) to verify the absence of

733 residual gadolinium. The LS was produced by adding LAB to a concen-
734 trated PPO-bis-MSB LAB solution. Fifty batches of LS were pumped into
735 a 200-ton-capacity storage bag made of nylon and polyethylene composite
736 membrane (PA/PE). LS is compatible with nylon based on our compatibil-
737 ity studies. The bag was placed in a concrete storage pool which is coated
738 with thermoplastic Olefin as a secondary containment.

739 *4.6.3. Mineral Oil*

740 Optically transparent, low radioactivity mineral oil fills the space between
741 the SSV and the outer acrylic vessel to attenuate radiation from the PMT
742 glass, SSV materials and other sources outside the detector module. With a
743 minimum of ~ 20 cm of space between the photocathode and the OAV, the
744 detector response is relatively uniform for events occurring in the LS.

745 *4.6.4. Liquid procurement, storage and handling*

746 As shown in Fig. 18, there are two concrete storage tanks (L1 and L2) of
747 200 ton capacity, five 40-ton storage tanks (G1-G5) for GdLS, and related
748 piping and pumping systems in the LS hall. Each concrete pool is about
749 250 m^3 (10.5 m long, 4.75 m wide, 5 m high) and is lined with a rectangular
750 nylon bag. Bags are made from a $200 \text{ }\mu\text{m}$ thick co-extruded multi-layer of
751 polyamide (PA) and polyethylene (PE) composite film, with the PA surface
752 contacting the liquid. Since the PA cannot be thermo-melt bonded, the
753 outside of the bag is PE, which allows thermo-bonding of the seams. Limiting
754 the intrusion of air into the bags helps keep the liquids below the required
755 radioactivity upper limits. The bags were tested for leak-tightness, with
756 liquid leak rates less than $5 \times 10^{-2} \text{ cm}^3/\text{s}$, corresponding to an air leak rate

Figure 18: Footprint of the LS hall. G1-G5 label the five 40-ton storage tanks for the GdLS. L1 and L2 are two 200-ton storage pools. Mixing equipment and filling equipment are installed close to the door. There is a Quality Assurance (QA)/ Quality Control (QC) laboratory and an area for nitrogen and water.

757 of $\sim 1 \text{ cm}^3/\text{s}$.

758 The GdLS storage tanks are 4 m in diameter, 4 m high, 15 mm thick
 759 acrylic vessels, reinforced with 5 mm thick fiber reinforced plastic and alu-
 760 minum belts. The GdLS tanks were sealed and purged with nitrogen gas
 761 before filling and continuously purged with nitrogen during liquid storage.

762 The 388 tons of LAB were synthesized by Jinling Petro-chemical Cor-
 763 poration, Nanjing, China, in a single batch over three days. The synthesis
 764 parameters and procedures were adjusted (based on more than 2 years of
 765 collaborative research with the company) to satisfy Daya Bay requirements.
 766 The synthesized LAB was directly pumped into shipping tanks via tempo-
 767 rary PVDF piping. The ISO shipping tank was lined with a 20 ton PE/PA
 768 bag. A total of 21 tanks were shipped via sea and truck to Daya Bay. The
 769 LAB was unloaded and pumped into the two 200-ton storage pools, L1 and
 770 L2, also using PVDF piping (Fig. 18). All wetted parts of the diaphragm

771 pump were made of fluoridated plastics. The LAB produced in this manner
772 has an attenuation length of 15 meter, significantly more transparent than
773 normally available LAB.

774 The LAB stored in L2 was used to make the 185 tons of GdLS in 50
775 batches. The synthesized GdLS was pumped into the storage tanks G1-G5
776 in sequence. After GdLS production, the L2 tank was emptied and the piping
777 system was switched to send LAB from the L1 tank to the mixing equipment
778 for synthesizing LS which was then stored in L2.

779 The Hangzhou Sinopec Co. Ltd produced 305 tons of mineral oil in one
780 production run. The MO was custom made for Daya Bay to be between
781 99% and 100% of the LS density. The density (at 20°C) of the Daya Bay
782 GdLS, LS, and MO liquids are 0.860 g/cm³, 0.859 g/cm³, and 0.851 g/cm³,
783 respectively. The production MO was split into two batches. The first half
784 was shipped to Daya Bay using liquid bags similar to the LAB bags, and
785 emptied into a new PE/PA bag in the L1 storage tank. The remaining oil
786 was stored at Hangzhou and later shipped to Daya Bay after the first 4 ADs
787 were filled.

788 4.6.5. *Liquid scintillator performance and stability*

789 Many liquid properties, such as the gadolinium concentration in the
790 GdLS, optical transmission, carbon to hydrogen ratios, liquid density, light
791 yield, and radioactivity, were measured when the ADs were filled and mon-
792 itored over time. Liquid samples [GdLS (two samples, before and after fill-
793 ing), LS, and MO] were taken from each AD for these measurements. The Gd
794 concentration was measured using X-ray fluorescence spectrometers with an
795 uncertainty of about 2%. Two years of monitoring the Gd concentration in

Figure 19: Gd content of a GdLS sample taken during AD filling as a function of time

796 GdLS samples at the BNL and IHEP labs show that the Gd concentration re-
 797 mains stable with time as shown in Fig. 19. A second, indirect measurement
 798 of the GD concentration relies on the correlation between the Gd concentra-
 799 tion and the neutron capture time. Since neutron capture times in all ADs
 800 are stable [39], the Gd concentration must also be stable.

801 Figure 20 shows the absorbance spectra measured by a UV-vis spectrom-
 802 eter of GdLS samples taken from storage tanks G1 to G5 which had been
 803 filled from different production batches of GdLS. The curves are nearly iden-
 804 tical, indicating the uniformity of the produced GdLS optical-transmission
 805 properties. The optical transmission of the AD samples were also studied
 806 over time. A representative curve of GdLS absorption at 430 nm is shown in
 807 Fig. 21 and has been stable for more than 500 days.

808 The emission spectrum from the GdLS excited at 260 nm were consis-
 809 tent with the bis-MSB emission spectrum [39]. LAB-based liquid scintillator
 810 usually has a high light yield. According to our preliminary research, the
 811 addition of the Gd-TMHA complex to the LS does not significantly degrade
 812 the light yield. The light yield of both GdLS and LS are about half of that
 813 of anthracene, as measured at BNL and IHEP.

Figure 20: Comparison of absorbance spectra of GdLS samples from the five 40-ton storage tanks at Daya Bay. The spectra are nearly identical with the last curve obscuring the other curves.

Figure 21: Optical absorbance at 430 nm of GdLS as a function of time

Table 8: Percentage weights of Gd, C, and H in GdLS samples from the eight ADs. The relative Gd concentration error is less than 1%. The relative error in the computed C/H ratio is less than 0.7%.

	Gd concentr.(%)	Carbon(%)	Hydrogen(%)
EH1-AD1	0.1031 ± 0.0003	87.85 ± 0.94	12.01 ± 0.31
EH1-AD2	0.1029 ± 0.0006	87.91 ± 0.91	11.97 ± 0.38
EH2-AD1	0.1032 ± 0.0009	88.02 ± 0.92	11.95 ± 0.14
EH2-AD2	0.1034 ± 0.0005	87.90 ± 0.92	11.95 ± 0.13
EH3-AD1	0.1034 ± 0.0009	87.97 ± 0.92	12.00 ± 0.13
EH3-AD2	0.1031 ± 0.0009	88.05 ± 0.93	11.97 ± 0.13
EH3-AD3	0.1029 ± 0.0009	87.86 ± 0.95	12.04 ± 0.43
EH3-AD4	0.1020 ± 0.0010	87.81 ± 0.95	11.98 ± 0.16

814 C/H/N weight and atomic ratios were obtained directly from combustion
815 analysis. Gd concentration together with C/H weight ratios of the GdLS
816 samples from the eight ADs are summarized in Table 8. Table 9 lists the
817 percentage weights of C and H in LS samples from the eight ADs.

818 4.6.6. Filling system and target mass measurement

819 Each AD was moved underground to the filling hall and filled with pre-
820 cisely measured amounts of GdLS, LS, and MO [46]. This irreversible op-
821 eration was the single highest risk element of an AD's construction as there
822 was no feasible method to remove liquid from the AD which would not com-
823 promise the mass measurements or the liquid purity. The ultimate precision
824 of the relative measurements at the near and far detectors is directly driven
825 by the GdLS mass measurements.

Table 9: Percentage weights of C and H in LS samples from the eight ADs.

Samples	Carbon	Hydrogen
EH1-AD1	88.00 ± 0.93	11.91 ± 0.13
EH1-AD2	88.09 ± 0.92	11.98 ± 0.12
EH2-AD1	88.16 ± 0.92	11.89 ± 0.13
EH2-AD2	88.08 ± 0.88	11.93 ± 0.13
EH3-AD1	88.23 ± 0.92	11.96 ± 0.13
EH3-AD2	87.90 ± 0.92	11.97 ± 0.13
EH3-AD3	88.14 ± 0.92	12.02 ± 0.13
EH3-AD4	88.09 ± 0.93	11.95 ± 0.13

826 The filling system was made of three metering pump circuits, each with
 827 a Coriolis flowmeter, Teflon plumbing, filling lines and level probes which
 828 were inserted into the AD. The LS and MO circuits pumped directly from
 829 the storage tanks. The GdLS circuit utilized a PFA Teflon-lined ISO tank
 830 to collect equal amounts of GdLS from the five acrylic GdLS storage tanks.
 831 The ISO tank rested on precision weigh-bridge load cells. Weighing the tank
 832 before and after detector filling determined the GdLS mass pumped into the
 833 detector. A peristaltic pump was used to top-up the GdLS overflow tank
 834 before the final weight measurement.

835 All components of the filling system were tested for compatibility with
 836 the AD liquids. Piping was constructed from PVDF (polyvinylidene fluoride)
 837 or Teflon materials. Particular care was taken with the GdLS pump circuit
 838 which used a non-metallic PVDF pump head. Remaining stainless steel parts
 839 in the GdLS pumps and valving were passivated with nitric acid.

840 The filling process took several days to complete. First, MO was pumped
841 into the SSV until the MO reached the bottom of the OAV. Then MO and
842 LS were pumped at speeds adjusted to keep the liquid levels within ± 5 cm
843 of each other. When the LS reached the bottom of the IAV the GdLS pump
844 was started. Filling-level probes and special cameras inside the AD were
845 used to monitor the filling levels. Transitions between the filling stages were
846 identified by changes of slope in the liquid height vs pumped mass plots.
847 Filling was stopped periodically to add and remove four one-ton calibration
848 masses to the ISO tank to track any calibration drifts.

849 The GdLS pump was stopped when the top of the IAV was reached.
850 Likewise the LS pump was stopped when the OAV lid was reached. MO
851 pumping continued until the MO partially filled the overflow tanks on the
852 SSV lid. The added pressure of the MO on the non-rigid OAV and IAV lids
853 and walls squeezed the inner volumes. As a result both the GdLS and LS
854 liquid levels rose into their respective overflow tanks without further pump-
855 ing. However, the resulting liquid levels were not well matched. Additional
856 liquids were added so that all three overflow tanks were at the same level
857 and were filled to about 1/3 capacity. It was necessary to add several liters
858 of MO and LS after the ADs were moved into the experimental halls since
859 trapped gas bubbles at the top of the SSV and OAV were dislodged during
860 transport. It was not necessary to add GdLS.

861 The dominant uncertainty in the GdLS mass measurement was due to
862 drifts in the load cell readings. Studies over several days with empty and
863 full ISO tanks saw a maximum variation of ± 2 kg (0.01% of 20 ton). The
864 calibration mass measurements determined a 0.18% correction to the absolute

865 mass scale, probably due to different values of gravitational acceleration (g)
 866 at Daya Bay and the manufacturer's location. An additional 0.13% correction
 867 was made to correct for the weight of the nitrogen gas which filled the empty
 868 ISO tank volume. The AD target mass was defined as the mass of the GdLS
 869 in the IAV vessel. Thus the estimated mass of the GdLS in the overflow tank
 870 and calibration tubes is subtracted from the total GdLS mass determined
 871 from the before and after load cell measurements. The masses of each AD
 872 are shown in Table 10. The error in the GdLS mass, ± 3 kg (0.015%), is well
 873 below the design goal of $\pm 0.2\%$.

Table 10: AD liquid masses. The GdLS mass is calculated from the change in weight of
 the ISO storage container before and after filling with corrections for the amount of liquid
 in the overflow tank and connection tubes and for the weight of the gas in the ISO tank.
 The LS and MO mass are determined from Coriolis flow meters.

	GdLS mass (kg)	LS mass (kg)	MO mass (kg)
EH1-AD1	19941 ± 3	21623 ± 28	36444 ± 76
EH1-AD2	19967 ± 3	21570 ± 28	36472 ± 76
EH2-AD1	19891 ± 4	21637 ± 24	36240 ± 76
EH2-AD2	19944 ± 5	21500 ± 24	36348 ± 77
EH3-AD1	19917 ± 4	21616 ± 24	36292 ± 76
EH3-AD2	19989 ± 3	21459 ± 28	36248 ± 76
EH3-AD3	19892 ± 3	21702 ± 28	36414 ± 76
EH3-AD4	19931 ± 3	21524 ± 24	36520 ± 133

874 4.7. AD monitoring

875 An AD lid monitoring system [47] measures the heights of the liquids
876 in the overflow tanks, temperatures of the AD liquids and possible tilts of
877 the AD lid. Temperature sensors mounted at various depths inside the AD
878 monitor the MO temperature. In addition to these monitors which provide
879 continuous monitoring during normal data taking, there are several special-
880 ized systems which take data as needed. Cameras and lights [47] view the
881 liquid levels in the off-center calibration ports during calibration periods,
882 providing a cross-check of the LS and GdLS heights from the level sensors.
883 A mineral oil clarity system mounted on the SSV lid shines light at different
884 wavelengths through the mineral oil to monitor mineral oil clarity. Two other
885 sets of cameras and lights [48] were used to monitor the AD interior liquid
886 levels during filling.

887 4.7.1. Liquid level sensors

888 Two sets of liquid level sensors are incorporated into the central overflow
889 tanks as shown in Fig. 22. Each tank contains an ultrasonic level sensor
890 as well as a capacitance level sensor to provide redundant measurements of
891 the GdLS or LS heights. The ultrasonic sensors (Senix Corporation model
892 TSPC-30S1) are mounted parallel to the liquid surface on the overflow tank
893 lid and view the liquid level via a flat acrylic mirror built into the mount.
894 Sensors were individually calibrated in air. Reflections off the bottom of the
895 unfilled acrylic tanks were used to define the bottom of the measurable range.
896 Calibration errors were less than 1 mm over the 400 mm range. Monitor
897 readings during data taking are corrected for the difference in the speed of
898 sound between air and the dry nitrogen AD cover gas.

Figure 22: AD level monitoring sensors. (a) Ultrasonic, (b) temperature, (c) capacitance, (d) calibration port cameras

899 As a cross check of the height versus volume calculation, one of the over-
 900 flow tanks was filled with water in precisely measured increments while mon-
 901 itoring the reported height. The largest deviation from the expected volume
 902 versus height curve was 1.5 liters which is taken as the systematic error in
 903 the overflow tank liquid volume measurement.

904 The calibration port cameras shown in Fig. 22(d) are used once a month
 905 to verify the GdLS and LS liquid levels in the off-axis calibration ports. The
 906 camera views the backlit liquid level in the clear acrylic port next to a fixed
 907 scale. Long-term drifts in the level sensors are measured by comparing the
 908 pictures and sensor data. No significant drift has been observed in any of the
 909 AD ultrasonic sensors.

910 A time history of the GdLS levels in the near hall detectors is shown in

Figure 23: Time history of the GdLS levels in the near detector AD overflow tanks as measured by the ultrasonic sensor. The rapid variations in Aug.-Sep., 2012 data correspond to multiple EH1 pool draining and filling cycles during manual calibration tests. EH2 was drained and filled once in the same time frame to install EH2-AD2.

Fig. 23. The large observed fluctuations in the liquid level are generated when the SSV is compressed by the weight of the water during pool filling. Smaller fluctuations are correlated with changes in the water pool temperature.

Custom capacitance sensors were constructed by Gill [49] from Polytetrafluoroethylene (PTFE) for compatibility with the LS. Calibration was provided by the factory for the LAB base of the LS and GdLS. A standard stainless steel capacitance sensor (Gill Type R) was installed in one of the two MO overflow tanks. The capacitance sensors lack the precision of the ultrasonic sensors, and have proved to be more susceptible to long term drifts. Therefore they are used as backups to the ultrasonic sensors.

4.7.2. Temperature monitoring

AD liquid temperatures are measured by Pt100 platinum resistance thermometers in the GdLS and LS overflow tanks and at four depths in the MO volume. The sensors have a nominal accuracy of $\pm 0.2^\circ\text{C}$ at 25°C . A time history of recorded temperatures in AD1 can be seen in Fig. 24. The tem-

Figure 24: Time history of liquid temperatures in EH1-AD1. Temperatures measured in the GdLS overflow tank (red), LS overflow tank (green), and at two heights in the MO volume(blue or black) track each other very well.

peratures clearly track each other and are correlated to the small changes of liquid level observed in Fig. 23.

4.7.3. Mineral Oil clarity sensor

An automated MO clarity system was designed and installed on the AD lid to monitor the optical transmission properties of the MO. The system utilizes a high power LED to send light pulses to the AD through a 50 m optical fiber. A stepping motor driven monochromator and collimator produce beams at different wavelengths. The light is directed vertically down into the MO parallel to the PMT faces. A corner cube retro-reflector mounted above the bottom AD reflector reflects the light back to the top of the AD, thus doubling the light path in the mineral oil to around 8 m. Reflected light is received by a 2" PMT and digitized in a flash ADC. To monitor the light intensity, a small mirror in the light path generates a secondary beam of light directed to the 2" PMT without entering the MO. The ratio of reflected and reference signals at each wavelength measures the light attenuation in the MO as a function of wavelength.

942 A second system within the MO clarity enclosure uses a different method
943 to monitor the attenuation in the MO. A diffuser ball mounted beneath the
944 AD lid is driven by a broad band LED pulse. The ball illuminates the AD
945 PMTs directly below. Since the PMTs are at different MO path lengths
946 from the diffuser ball the ratio of PMT responses is sensitive to the MO
947 attenuation length. Systematic differences in the PMT gain and acceptance
948 limit the precision of an absolute attenuation length measurement. However,
949 changes with time of the ratios are a clear indication of changes in the MO
950 attenuation length.

951 One diffuser ball run and nine monochromator/collimator runs (from 390
952 nm to 430 nm, at 5 nm intervals) are performed weekly. LEDs for the diffuser
953 ball or collimator are triggered by a pulser board operating at 20 Hz for 2000
954 triggers in each run. In collimator runs, the total sampling time window of
955 the FADC is 1280 ns, of which 100 ns is occupied by two clear peaks. The
956 reference signal and the corner cube reflection are shown in Fig. 25. The
957 time separation between the two signals is 40 ns, consistent with the speed
958 of light in the MO over the ≈ 8 m path length. The area of the two signals
959 is obtained by fitting the peaks with a double Crystal Ball function [50] or
960 by numerical integration. The ratio of the first peak to the second peak is
961 monitored for the stability of the MO attenuation. Tests demonstrate that
962 this ratio is stable even if the PMT gain changes by a factor of two from the
963 nominal operating value of 1×10^7 .

964 4.8. AD gas system

965 Liquids in the overflow tanks and calibration tubes are covered by inert
966 nitrogen gas supplied by an AD gas system. The cover gas system constantly

Figure 25: : Averaged waveform of 2000 samples of a typical MO monitoring run in AD2, at 410 nm wavelength. The unit of the vertical axis is the voltage resolution and the unit of the horizontal axis is the time resolution of the FADC

967 flushes the gas volumes above the liquid scintillator with dry nitrogen to min-
 968 imize oxidation of the scintillator over the five year lifetime of the experiment.
 969 This constant flush also prevents the infiltration of radon or other contam-
 970 inants into these detecting liquids, keeping the internal backgrounds low.
 971 Since the Daya Bay antineutrino detectors are immersed in the large water
 972 pools of the muon veto system, other gas volumes are needed to protect vital
 973 detector cables or gas lines. These volumes are also purged with dry gas.
 974 Return gas is monitored for oxygen content and humidity to provide early
 975 warning of potentially damaging leaks.

976 The cover gas circuit flushes the gas volumes above the overflow tanks
 977 and calibration tubes (as shown in Fig. 11), the ACUs, Mineral Oil clarity
 978 Monitor, and the LED cabling port, all of which are located on the SSV lid.
 979 The cover gas volume of approximately 1500 l is refreshed at the rate of two
 980 volume exchanges per day. The single 1/2 inch supply line is split ten ways

981 by a manifold on the lid. Separate 3/8 inch supply and return lines go to
982 each detector volume. Another tenfold manifold combines the return flows.
983 This manifold arrangement and the low flow rate ensure that no significant
984 pressure differential exists between the gas volumes. As a further safety
985 measure the cover gas is kept at ≈ 1 cm of water equivalent pressure above
986 ambient atmosphere by an oil bubbler in the return line.

987 Since the ADs are beneath about 2.5 m of water, considerable design
988 effort went into preventing a single seal failure from causing a water leak
989 into the AD liquids. Almost all of the lid and ACU seals are of a double
990 O-ring design. The cover gas volumes are isolated from other gas volumes on
991 the lid by electrical and gas feed-through flanges. All of the cover gas lines
992 are protected by vacuum grade stainless steel bellows which are flushed with
993 an additional purge gas circuit. Similarly, electrical lines for the ACUs, lid
994 sensors, and LEDs are run through bellows flushed by a third gas circuit. A
995 fourth gas circuit flushes the eight bellows and dry box elbows which protect
996 the PMT high voltage lines from the AD to above the water level.

997 **5. AD Calibration**

998 Detailed characterization of the AD detector properties and frequent mon-
999 itoring of the detector performance are required to understand detector differ-
1000 ences at the 0.1% level and to identify 0.1% changes in a particular detector
1001 module over time. This is achieved via full and frequent detector calibration.

1002 *5.1. Automated Calibration Units*

1003 Three fully automated calibration units (ACUs) were mounted and sealed
1004 on the top of each detector. Deployment positions are along the AD central

Figure 26: An illustration of the Daya Bay detector showing the three ACUs.

axis, near the edge of the GdLS cylindrical volume ($r = 1350$ mm), and in the LS between the inner and outer acrylic vessels ($r = 1773$ mm), as shown in Fig. 26. All eight ADs (using one ACU at a time) can be calibrated simultaneously. ACUs deploy one of three possible calibration sources within the ACU into the detector at different z positions. Multiple calibration runs for each antineutrino detector were taken every week for ≈ 3 hours during 2011-2012.

Calibration sources are deployed with a precision of 7 mm along the vertical axis (z -axis) [51]. Figure 27 shows the major ACU components. The three calibration sources are a LED, a ^{68}Ge source and a ^{241}Am - ^{13}C and ^{60}Co combined source. ^{68}Ge decays into ^{68}Ga via electron capture with a half life of 271 days. The ^{68}Ga decays via positron emission with a half life of 68 minutes. The positron typically annihilates within the source enclosure, yielding two 511 keV gammas. The ^{68}Ge source is thus effectively a 10 Hz

1019 gamma source. The ^{60}Co sources emit two gammas at 1.173 and 1.332 MeV
1020 at about 100 Hz. Alphas from the 28 μCi ^{241}Am source interact with ^{13}C
1021 within the source, emitting neutrons at a rate of 0.7 Hz [52].

1022 As shown in Fig. 28, each source is enclosed in an acrylic shell and con-
1023 nected to the turntable by a stainless steel (radioactive sources) or coaxial
1024 cable (LED) in Teflon sleeves. All materials having contact with liquids were
1025 thoroughly studied for compatibility. Radioactivity of the components meets
1026 the required specification of ≤ 0.6 Bq/kg of ^{238}U , 0.4 Bq/kg of ^{233}Th and
1027 2.6 Bq/kg of ^{40}K .

1028 During the summer shutdown of 2012, the calibration sources used in each
1029 ACU were modified based on experience with the 6 AD data set. Neutrons
1030 from the off-axis ^{241}Am - ^{13}C sources could produce IBD-like signals in the
1031 AD, even with the source in the parked position above the SSV lid. Source
1032 neutrons could produce gammas in the LS or GdLS either by scattering or
1033 capture in the Fe-Cr-Mn-Ni SSV walls followed by the subsequent capture
1034 of the neutron in the GdLS. The ^{241}Am - ^{13}C sources in the EH3 off-center
1035 ACUs were removed to reduce the background level. ^{68}Ge and ^{60}Co sources
1036 were swapped in some ACUs to improve the neutron source calibration data
1037 for H/Gd ratio studies. The final sources inside all 8 AD ACUs are listed in
1038 Table 11.

1039 A special source made from ^{137}Cs deposited at the center of a spherical
1040 scintillator was used during dry detector commissioning in the SAB before
1041 the ADs were filled. A detailed description of the automated calibration
1042 system can be found in [51].

Table 11: Calibrations sources in the ACUs after the start of the 8-AD data-taking in Dec. 2012. The name of an ACU stands for the installed position (A/B/C) in an AD. See Fig 26 for details.

AD-ACU	1	2	3
EH1-AD1-1A	LED	$^{241}\text{Am}-^{13}\text{C}/^{68}\text{Ge}$	^{60}Co
EH1-AD1-1B	LED	$^{241}\text{Am}-^{13}\text{C}/^{60}\text{Co}$	^{68}Ge
EH1-AD1-1C	LED	$^{241}\text{Am}-^{13}\text{C}/^{68}\text{Ge}$	^{60}Co
EH1-AD2-2A	LED	$^{241}\text{Am}-^{13}\text{C}/^{68}\text{Ge}$	^{60}Co
EH1-AD2-2B	LED	$^{241}\text{Am}-^{13}\text{C}/^{60}\text{Co}$	^{68}Ge
EH1-AD2-3C	LED	$^{241}\text{Am}-^{13}\text{C}/^{68}\text{Ge}$	^{60}Co
EH2-AD1-3A	LED	$^{241}\text{Am}-^{13}\text{C}/^{60}\text{Co}$	^{68}Ge
EH2-AD1-3B	LED	$^{241}\text{Am}-^{13}\text{C}/^{60}\text{Co}$	^{68}Ge
EH2-AD1-2C	LED	$^{241}\text{Am}-^{13}\text{C}/^{68}\text{Ge}$	^{60}Co
EH2-AD2-8A	LED	$^{241}\text{Am}-^{13}\text{C}/^{68}\text{Ge}$	^{60}Co
EH2-AD2-4B	LED	$^{241}\text{Am}-^{13}\text{C}/^{68}\text{Ge}$	^{60}Co
EH2-AD2-4C	LED	$^{241}\text{Am}-^{13}\text{C}/^{68}\text{Ge}$	^{60}Co
EH3-AD1-4A	LED	$^{241}\text{Am}-^{13}\text{C}/^{60}\text{Co}$	^{68}Ge
EH3-AD1-8B	LED	^{60}Co	^{68}Ge
EH3-AD1-8C	^{40}K	^{60}Co	^{68}Ge
EH3-AD2-5A	LED	$^{241}\text{Am}-^{13}\text{C}/^{60}\text{Co}$	^{68}Ge
EH3-AD2-5B	LED	^{60}Co	^{68}Ge
EH3-AD2-5C	LED	^{60}Co	^{68}Ge
EH3-AD3-6A	LED	$^{241}\text{Am}-^{13}\text{C}/^{60}\text{Co}$	^{68}Ge
EH3-AD3-6B	LED	^{60}Co	^{68}Ge
EH3-AD3-6C	LED	^{60}Co	^{68}Ge
EH3-AD4-7A	LED	$^{241}\text{Am}-^{13}\text{C}/^{68}\text{Ge}$	^{60}Co
EH3-AD4-7B	LED	$^{67}_{60}\text{Co}$	^{68}Ge
EH3-AD4-7C	LED	^{60}Co	^{68}Ge

Figure 27: An overview picture of the ACU

Figure 28: Calibration sources

5.2. Manual Calibration System

A manual calibration system (MCS) was built to extend source calibrations to the full detector volume. The MCS consisted of three main components: a calibration tower, a source rod assembly and control system. Figure 29 shows a schematic of the MCS installed on an AD. The calibration tower protected all of the inner components from the water pool since manual calibrations were performed with normal water levels. The rotatable source rod could be moved up or down to control the height of the calibration source and rotated to any desired azimuthal (Φ) position. The source was moved inside the hollow source arm (an acrylic tube) by a teflon coated stainless steel wire and pulley system to position the source at the desired radius. The source rod, articulation joint, and elbow were made of stainless steel coated with Teflon for compatibility with AD liquids. The source could be deployed to almost any position within the IAV except very near the IAV wall. Calibration sources were placed in the full volume of the AD with a positional accuracy of 25 mm radially, 12 mm vertically and 0.5° in azimuth.

A combined ^{60}Co and ^{238}Pu - ^{13}C source was used in the MCS. The source strength was selected to minimize data acquisition time while avoiding saturation of the PMT or readout systems. The ^{60}Co source strength was 150 Bq. The ^{238}Pu - ^{13}C source produced 2000 Bq of neutrons and 40 Bq of 6.13 MeV gamma rays. The overall trigger rate at a trigger threshold of 0.4 MeV was about 3 kHz. With these sources, calibrations at energies from 1 MeV to 8 MeV were possible, using 1.02 MeV of gammas from positron annihilation, 2.2 MeV gamma rays from neutron capture on hydrogen, 4.43 MeV from $^{12}\text{C}(\text{n},\text{n})^{12}\text{C}^*$, 4.95 MeV from gamma ray emission of the $^{12}\text{C}(\text{n},\gamma)$ reac-

Figure 29: Schematics of the MCS deployment

tion, 6.13 MeV of the de-excitation energy release of $^{16}\text{O}^*$ from $^{13}\text{C}(\alpha, n)^{16}\text{O}^*$ reaction and 8 MeV from neutron capture on gadolinium.

A manual calibration was performed on AD1 in EH1 during the summer shutdown of 2012. The performance is detailed in section 9. A detailed description of the MCS can be found in [53].

5.3. 2-inch PMT calibration system

Three pairs of Hamamatsu 2-inch R7724 PMTs with standard Borosilicate glass were mounted in each AD to monitor the attenuation length of the liquids. The radioactivity of the PMTs was measured to be ≤ 15 Bq/PMT from ^{40}K , ≤ 1.5 Bq/PMT from ^{238}U , and ≤ 0.25 Bq/PMT from ^{232}Th , contributing negligibly to the AD singles rate. One PMT of each pair was installed on the top lid facing downwards and the other was installed on the bottom of the steel vessel facing upwards. 10-cm circular optical apertures were cut into top and bottom reflectors to allow the PMTs to see into the liquids. The glass windows of the PMTs have a nominal clearance of 28 (90) mm to the top (bottom) reflector. The radial position of the three PMT pairs was chosen to be 370.5, 1790.0, and 2150.0 mm, to look into the GdLS target, LS gamma catcher, and MO liquid zones. The attenuation length of the liquids can be monitored by the ratio of the charge collected by the top and bottom PMTs with a mono-energetic calibration source deployed in the AD.

6. Water pool and Muon Detectors

The photograph in Fig. 30 shows a partially filled water pool in the far experimental hall and the major components of the water pool and muon

systems. The highly purified water attenuates environmental background radiation while allowing the detection of cosmic muons via Cerenkov radiation. Two groups of PMTs monitor the water pool and detect muons. White TyvekTM sheets mounted on stainless steel Unistrut frames separate the pool into two regions. The central region, containing the ADs, has inward facing PMTs and is designated as the Inner Water Shield (IWS). A 1m-thick Outer Water Shield (OWS) forms an open shell around the IWS (except for the top). Rolled off to the top right side of Fig. 30 is an array of modules, each containing four layers of resistive plate chamber (RPC), which provide additional muon detection. Two modules of RPCs are mounted on the hall walls to provide a subset of muons with well measured angular parameters. Not seen is the black rubberized cover supported by cables which is pulled over the water pool after filling. Nitrogen gas flows beneath the cover to reduce the intrusion of radon which is commonly found in underground granite environments. Detailed descriptions of the muon and water purifying systems can be found in [19] and [20].

6.1. Water system

The antineutrino detectors are surrounded by more than 2.5-meter of water shielding in all directions reducing backgrounds from rock radioactivity by nearly a factor of 10^6 . The effectiveness of the water shield can be clearly seen in Fig. 31 which shows the reconstructed position of background hits within an AD when the water pool was about 1/2 full. Backgrounds below the water level are strongly suppressed.

The near (far) experimental halls contain nearly 1250 tons (2000 tons) of highly purified water with an optical attenuation length of about 40 m at 420

Figure 30: The partially filled far hall water pool containing three ADs is shown. Inner Water Shield (IWS) PMTs are visible above and below the water line. Outer Water Shield (OWS) PMTs are hidden beneath the white TyvekTM sheets separating the IWS and OWS. Previously installed RPC arrays are rolled into the RPC garage area on the top-right to provide access to the water pool. Two sets of telescope RPCs are mounted on the hall walls as shown.

Figure 31: Reconstructed positions of AD background triggers with the water pool partially filled showing the strong suppression of background events by the water.

1117 nm (based on indirect studies [19]). Producing and maintaining this water
 1118 are several water processing systems. A central treatment system in the
 1119 Water Hall (WH) produces ultra-pure water from the city supply for filling
 1120 and topping off the muon pools. Polishing systems, located in utility rooms
 1121 of each experimental hall, recirculate 5-8 m³ of water per hour. Pool water
 1122 passes through 1 and 0.1 micron filters, a de-gasser, heat exchanger, and is
 1123 exposed to UV at 185 and 254 nm to sterilize any biological contaminants.
 1124 The water is returned to the pool at 22.7°C with a resistivity of 18.2 MΩ-cm
 1125 and ≤ 10 ppb dissolved oxygen. Radon levels measured after polishing in
 1126 the polishing station are between 30 – 50 Bq/m³ [19].

1127 *6.2. Muon PMTs*

1128 The water pools EH1 and EH2 are instrumented with 8" PMTs, with 121
 1129 PMTs in the IWS and 167 PMTs in the OWS. The larger, EH3 contains
 1130 160 PMTs in the IWS and 224 PMTs in the OWS. The PMTs are 619
 1131 Hamamatsu R5912 assemblies with water-proof bases and 341 EMI 9350KA
 1132 and D642KB assemblies [54] recycled from the MACRO experiment. Positive
 1133 high voltage and PMT signals are transmitted via a 52-m-long coaxial cable.
 1134 FINEMET shields, identical to those used with AD PMTs, protect muon
 1135 PMTs against ambient magnetic fields. The EMI PMTs were used primarily
 1136 in the upper region of the water pool because they were shown to fail at
 1137 pressures corresponding to depths of about 5 m. All muon PMTs were tested
 1138 to the same standards as the AD PMTs. Characterization of each PMT
 1139 determined the HV setting for the nominal data-taking gain of 1.0×10^7 ,
 1140 which yields about 20 ADC counts per photoelectron.

1141 During normal data taking the PMT gain is constantly calibrated with

Figure 32: Mean PMT multiplicity per muon vs time. The drop of mean PMT multiplicity in EH2-IWS and EH3-IWS during the 8 AD period is due to the reduced mean muon track lengths since one more AD was put into the water pool. The instability of EH3-OWS is induced by electronics noise from unknown sources, which create triggers that have much lower PMT multiplicity. This instability has negligible impact on muon detection efficiency, as shown in Fig. 43.

1142 dark noise based on data collected via periodic triggers. The average PMT
 1143 gains are relatively stable with a slight upward trend over time. As an
 1144 indicator of the muon tagging efficiency, the mean PMT multiplicity per
 1145 muon is monitored over time as shown in Fig. 32.

1146 6.3. Resistive Plate Chamber Detectors

1147 Muon detection modules utilizing resistive plate chambers (RPCs) are
 1148 mounted on a rolling frame which is positioned over the water pool [19] for
 1149 additional muon detection. The modules [55], $2.17 \text{ m} \times 2.20 \text{ m} \times 8 \text{ cm}$,
 1150 contain four layers of RPCs constructed from non-oiled Bakelite sheets [56],
 1151 each instrumented with eight $26 \text{ cm} \times 2.10 \text{ m}$ long readout strips oriented in
 1152 either the x or y direction to provide a position resolution of about 10 cm.

1153 Modules are purged with dry nitrogen to protect the enclosed RPCs

1154 against the moist underground environment. The RPC gases are mixed,
 1155 distributed and monitored locally by a gas system located in a dedicated
 1156 utility room near each hall. Argon, freon(R134a), isobutane and SF_6 are
 1157 mixed in the ratio 65.5:30:4:0.5 [57]. The gas mixture is periodically verified
 1158 with a gas chromatograph to ensure stable performance. RPC layers of the
 1159 same module are connected to different HV channels to ensure that only one
 1160 layer is lost during the failure of a HV channel.

1161 **7. Electronics and data acquisition**

1162 The primary data for the Daya Bay experiment originates in seventeen
 1163 independent readout crates (eight ADs, three inner and three outer water
 1164 shields, and three RPC detectors) in the three experimental halls. Collect-
 1165 ing, compiling, and coordinating these data requires a configurable, synchro-
 1166 nized, and flexible data acquisition (DAQ) system, one that can record data
 1167 from a single readout crate, or combine data from any number of readout
 1168 crates. Data from readout crates within a single detector hall are time or-
 1169 dered and merged to create a data stream. The streams from each hall are
 1170 saved independently for offline analysis. Each detector system (AD, IWS,
 1171 OWS, RPC) can self-trigger as well as receive triggers from other detector
 1172 units via a master trigger board (one per hall) that coordinates hall wide
 1173 trigger or calibration requests.

1174 The primary function of the PMT electronics is to amplify, discriminate,
 1175 shape, digitize and record the PMT waveform, and to provide precise timing
 1176 information between PMTs. To simplify design and maintenance, all PMT-
 1177 based detector systems use identical electronics. The PMT readout system

Figure 33: The PMT readout system block diagram is shown. PMT signals are sent to the FEE channels via coaxial cables after passing through an AC decoupling unit.

1178 consists of one MVME Power PC controller, one Local Trigger Board (LTB),
 1179 one trigger and clock fan-out board (FAN), one flash ADC board and up
 1180 to fourteen Front End Electronics (FEE) boards. The functionality of the
 1181 readout system is shown in Fig. 33. RPC detectors have their own electronics
 1182 system to record muon hit, time, and location information.

1183 All readout electronics systems follow the VME 64xp standard. There is
 1184 one VME crate per AD, two VME crates per experimental hall to read out
 1185 the IWS and OWS, and one VME crate per experimental hall for the RPC
 1186 system. Figure 34 is a photograph taken within the electronics room of EH2.

1187 7.1. PMT-based FEE

1188 Raw signals from PMTs are transmitted via a coaxial cable to an AC
 1189 decoupler unit that filters out the HV and sends the fast PMT signals to the
 1190 front end electronic (FEE) cards. Essential FEE functions are:

- 1191 • To calculate the number of over-threshold PMT signals within a given
 1192 time window (N_{hit}) and send the information to the trigger system.

Figure 34: Photograph of the readout electronics in EH2.

- 1193 • To create the linear sum of signals (E_{sum}) within the card and send it
1194 to the trigger system.
- 1195 • To shape and digitize the height of each PMT signal, values that cor-
1196 respond to the energy deposition in the detector.
- 1197 • To digitize and record the precise timing data for each over-threshold
1198 PMT signal for position reconstruction and background rejection.

1199 Detailed requirements are documented in [13]. Design and testing infor-
1200 mation have also been documented elsewhere [58]. In summary, each channel
1201 in the FEE board accepts signals from up to sixteen PMTs. Signals are am-
1202 plified for use in three distinct circuits: a discrimination circuit, a charge
1203 summing circuit, and a pulse shaping circuit. Discriminator outputs are

Figure 35: Charge Measurement Scheme of the FEE

1204 used by the local trigger board to form a multiplicity trigger and to start
 1205 onboard time-to-digital converters (TDC). Outputs of the FEE charge sum-
 1206 ming circuits are used by the trigger board as inputs to an energy summing
 1207 circuit. The pulse-shaping circuit shapes and then samples the signal using
 1208 a flash analog-to-digital converter (ADC). A trigger signal serves as the com-
 1209 mon stop for all the TDCs throughout the VME crate, and initiates read out
 1210 of the ADC and TDC data.

1211 7.1.1. Charge Measurement

1212 Typical output signals from the front-end linear amplifiers in each FEE
 1213 channel have widths between 10 and 20 ns and heights proportional to the
 1214 number of photo electrons (P.E.) produced at the photocathode of the PMT.
 1215 Amplified PMT signals are shaped via 4-resistor-capacitor network after a
 1216 capacitor-resistor (CR - (RC)⁴) circuit before being digitized by two 12-bit
 1217 ADCs with a sampling rate of 40 MHz. One ADC is set to have a fine-grain
 1218 resolution for measuring small amplitude signals. The second ADC range is
 1219 set to sample large signals. The ADC samples are fed into an FPGA which
 1220 selects and stores the largest value (peak value) and the baseline value, as
 1221 shown in Fig. 35. Further details are available in [58].

7.1.2. Time Measurement

The arrival time of the PMT signals is measured relative to the trigger signal. Signals within 1300 ns prior to the trigger are recorded. This time window comfortably contains the range of times expected between the first and the last PMTs to see either direct or reflected light from an IBD event, and also allows for trigger latency. Precision of the timing measurements are limited by jitter in the rising edge of the amplified PMT signal (2-3 ns) and by time walk of the signal (~ 1 ns). The digitization process begins when the amplified PMT signal is fed into a high-speed discriminator circuit. The discriminator output starts a TDC. To adjust for possible differences in PMT gains the threshold for each channel can be set via a VME controller. TDC functionality is programmed into the FEE's FPGA. Major components of the TDC are two high speed gray-code counters, the first changes at the rising edge of a 320 MHz clock, while the other changes at the falling edge. A discriminator output starts the TDC, while the trigger signal is used to stop the TDC. The time difference between the start signal and stop signal is measured via these two counters with a time resolution of less than 0.8 ns. Further details can be found in [58].

7.2. Flash ADC Board

An independent eight channel 1 GHz Flash ADC board is used to cross-check the triggers based on Esum signals. Each channel employs an 8-bit ADC, although the effective number of bits is seven. Each channel samples the output of an FEE's charge summing circuit, which is an integration of all sixteen amplified PMT signals on an FEE board. The output is recorded so that offline cross-checks can be carried out. Additional information on the

Flash ADC board can be found in [59].

7.3. Trigger system

The trigger system initiates electronic recording of detector information. Triggers are formed primarily within an individual readout crate, but may also be initiated by a different crate (called a cross-trigger), or from other sources, such as the calibration system (called a calibration trigger). The primary physics requirement is that the system achieve a greater than 99% efficiency for detecting events with an energy deposition of 0.7 MeV or greater in an AD. Detailed information can be found elsewhere [60]. In brief, each PMT readout crate contains a local trigger board (LTB) that receives information as described in section 7.1. The LTB can generate internal triggers or respond to external trigger requests. Various trigger types can be enabled or disabled via VME interface. The enabled triggers are ORed to generate a final trigger signal (local trigger) that is distributed to the FEEs by front-panel cables. This design is flexible, reproducible, and redundant. The RPC sub-detector employs a self-triggering scheme. Each hall also possesses one master trigger board (MTB) that coordinates external requests and cross-triggers. For example, calibration trigger requests originating in the calibration system are initially sent to the MTB, which forwards the trigger to the appropriate LTB. The trigger decisions implemented are:

- Physics trigger: $N_{hit} \geq 45$ or $E_{sum} \geq 65$ P.E. in an AD, which corresponds to about 0.4 MeV in deposited energy.
- Periodic trigger: A 10 Hz period trigger is turned on during normal physics data-taking to monitor detector stability and check random

Figure 36: Layout of the trigger system. In each PMT-based detector system (AD, IWS, or OWS), the number of PMTs over the threshold (N_{hit}) and the analog sum of PMT signals (E_{sum}) are sent to the Local Trigger Board (LTB), which will generate a local trigger once the prescribed trigger conditions are met. DAQ software reads PMT signal information from the FEEs and trigger information from the LTB upon receiving a trigger. The LTB in one subsystem can also initiate a trigger in a different subsystem by sending a trigger request to the Master Trigger Board (MTB), which can send trigger requests (cross triggers and calibration triggers) to all of the subsystems in the same experimental hall. The RPC trigger board differs in detail, but can also send and receive triggers to the MTB.

1271 backgrounds.

1272 • Calibration trigger: Generated by the LED system in ACUs to rou-
1273 tinely monitor PMT gains and timing.

1274 • Cross trigger: Used to correlate activity in different detector modules
1275 (such as water pool versus AD). An important cross trigger is the "look-
1276 back" trigger which is issued when two physics triggers within a single
1277 AD coincide within a time interval of 200 μs . The LTB for the AD gen-
1278 erates a cross trigger (if enabled) which is sent to water pool detectors
1279 to initiate readout of all hit information.

1280 • RPC trigger: Any three of the four RPC layers see a signal above the
1281 threshold.

1282 7.4. Global timing system

1283 The timing system provides a global time reference for synchronizing
1284 triggers from individual detectors. It consists of three major parts: a GPS
1285 driver module, a central clock generator (CCG) and local clock fanout boards
1286 (CFB). A commercial GPS driver is located in the control room, providing
1287 an IRIG-B coding signal with absolute UTC time information and a pulse
1288 per second (PPS) signal to the CCG. To achieve frequency accuracy and
1289 long-term stability, the 10 MHz signal generated by the Rubidium oscillator
1290 in the CCG is synchronized by the PPS signal from the GPS module. The
1291 10 MHz frequency is quadrupled to 40 MHz via PLL (Phase-locked loops).
1292 The 40 MHz clock and the IRIG-B information are distributed separately
1293 to each detector hall via optical fibers. The CFB in each hall converts the

optical signal and fanout timing information to all FEE boards in the hall. The CFB returns the IRIG-B signal immediately to the control room via another fiber link to measure the link delay. The GPS driver is equipped with an embedded system and acts as a Tier-1 network time protocol server for all experimental computers. A backup 10 MHz oscillator was built in the CCG in case the Rubidium clock is not available.

7.5. Data acquisition

The DAQ architecture is a multi-level system using embedded linux, advanced commercial computers and distributed network technology. Figure 37 shows the hardware deployed in the DAQ system. Additional details can be found in reference [61].

The front-end of the DAQ system is a real-time embedded system based on the VME bus. Each VME crate holds a Motorola MVME 5500 system controller. The MVME 5500 is an embedded single-board computer based on a PowerPC MPC 7455 CPU with a 1GHz clock and Universe II chip for interfacing to the VME bus. The computer uses a TimeSys operating system (real-time Linux, kernel version 2.6.9). This front-end system can manage a 2.5 kHz event rate (2 kilobyte event size) and meets all experimental requirements [62]. The DAQ back-end is a blade server based computing farm [63] utilizing a gigabit ethernet network. Also included are two x3650 file servers, and nine blade servers for data gathering and data quality monitoring. Daya Bay online software was customized and migrated to the front-end Linux and PPC platforms from the ATLAS online software [64] with three high level components (packages) - Control, Databases, and Information Sharing. These packages work together to provide the functionality for the various

Figure 37: Hardware deployment diagram of the DAQ system (Fiber connections from online computer room to EH1 and EH2 are removed from the figure for simplicity).

Figure 38: The data flow software. Details can be found in Sec 7.5.

1319 configurations of the DAQ and detector instrumentation.

1320 The data flow is shown schematically in Fig. 38, and was developed using
1321 BESIII front-end readout software and ATLAS back-end data flow software [65].
1322 Identical readout systems (ROS) run on each MVME5500 controller, reading and
1323 concatenating data to build individual events, packing and sending the data to the
1324 back-end [66]. The Event Flow Distributor (EFD) runs on a blade server. It
1325 parses data, fills and publishes histograms for data quality monitoring, and
1326 sends data to the SubFarm Output (SFO). The SFO runs on file servers, merges
1327 data from EFDs, sort events by time, and saves files to a disk array.
1328

1329 The DAQ supports six run modes: physics, pedestal, electronics diagnosis,
1330 AD calibration, AD mineral oil monitoring, and water pool calibration. The
1331 DAQ exchanges information with the Distributed Information Management (DIM)
1332 system [67] and with external subsystems such as the calibration system.
1333 Online monitoring was implemented to aid shifters in spotting problems in
1334 real time with multiple graphic interfaces for monitoring and control. Web-
1335 based remote monitoring and control software was developed to enable
1336 remote shifting using CORBA Java API for DAQ access, and RESTful for

Figure 39: Snapshot of part of the overview page of the DAQ web site. Event rate history as well as the related run information for EH1 are shown.

web service [68]. Figure 39 displays an example of web-based DAQ monitoring. An online supernova trigger [69] has been running in the DAQ system since August 2013.

7.6. Monitoring and control system

The Daya Bay Detector Control System (DCS) monitors and controls the experimental hardware and environment. A commercial software package (Supervisory Control And Data Acquisition), implementing the supervisory, control, and data acquisition standard for the DCS, uses LabVIEW with Data logging and Supervisory control modules. The framework of the detector control system is divided into two parts, a Local Control Layer (LCL) and a Global Control Layer (GCL). The LCL is composed of multiple local control IPCs (Industrial Personal Computer) which are responsible for the subsystem data acquisition and control. The GCL is composed of three parts: the Global Control Station (GCS), a Database Server (DBS), and a Web Server (WS). The status and alarm information of all subsystems are collected into the GCS and stored into the DBS. Physics data recorded in the DCS can be

1353 retrieved by the Web Server.

1354 The DCS monitors a wide array of detector parameters, including AD
1355 liquid levels, temperatures, humidity and air pressure, RPC high voltages,
1356 as well as photomultiplier high voltages and currents. It measures gas flow
1357 and pressures in the gas purification system and in the gas storage system.
1358 In addition, the DCS monitors environmental parameters and auxiliary sys-
1359 tems, such as the electronics crate, radon gas concentration inside and outside
1360 the experiment halls, extra environmental temperatures during the detector
1361 commission and installation process. The state of the water pool system (wa-
1362 ter pressure, water level, etc.) is also monitored. Data from environmental
1363 parameters and detector conditions (e.g. power supply voltages, tempera-
1364 ture/humidity, gas mixtures, radiation, etc.) are collected from local IPCs
1365 and put into a data pool by DIM [67] and monitored by global control layer
1366 IPCs. Details of the DCS system design and implementation can be found
1367 in [70].

1368 Some safety systems, such as rack protection and fast interlocks, are
1369 also included in the DCS. Safety interlocks can be activated to prevent un-
1370 desired states in the state machine, based on information from the RPC
1371 HV, water and cover gas systems. DCS can control the RPC HV sta-
1372 tus (on/standby/off). VME power is shutdown for over-current or over-
1373 temperature conditions. Interlock signals can shutdown the gas system in
1374 the event of a HV trip. Likewise RPC HV can be shutdown in the event of
1375 a gas system alarm.

1376 A remote monitor and query system [71] was developed in the DCS to
1377 provide various functions including history data querying, real-time display,

1378 historical charts, threshold display and realtime alarm information.

1379 **8. Offline software overview**

1380 NuWa, the Daya Bay offline software adaptation of the Gaudi frame-
1381 work [72], provides the full functionality required for simulation, reconstruc-
1382 tion and physics analysis. NuWa employs Gaudi's event data service as
1383 the data manager. Raw data, as well as other offline data objects, can be
1384 accessed from the Transient Event Store (TES). The prompt-delayed coinci-
1385 dence analysis requires looking-back in time, which is fulfilled via a specific
1386 implemented Archive Event Store (AES). All the data objects in both TES
1387 and AES can be written into or read back from ROOT [73] files through
1388 various Gaudi converters.

1389 An alternative Lightweight Analysis Framework (LAF) was designed and
1390 implemented by Daya Bay to improve the analysis efficiency. LAF is com-
1391 patible with NuWa data objects with higher I/O performance by the simpler
1392 data conversion, the implementation of lazy loading, and the flexible cy-
1393 cling mechanism. LAF allows access to events both backwards and forwards
1394 through a data buffer, which also serves to exchange data among multiple
1395 analysis modules. The NuWa and LAF packages are available to collabora-
1396 tors using the Subversion (SVN) code management system [74].

1397 Daya Bay simulation is based on GEANT4 [75] with critical features val-
1398 idated against external data or other simulation packages. The Monte Carlo
1399 simulation is tuned to match the observed detector response. To simulate
1400 time correlations of unrelated events, different categories of simulated events
1401 are mixed in the output file. Reconstruction algorithms have been developed

Figure 40: Network topology

Figure 41: Schematic diagram of the offline computing environment at the Daya Bay site described in Sec 8.

1402 to construct the energy and vertex of the antineutrino event from the charge
 1403 pattern of the PMTs. The detector-related parameters and calibration con-
 1404 stants needed by the reconstruction are stored in an offline central database
 1405 with a number of mirror sites located at different institutes. The algorithms
 1406 can access the contents in the database via an interface software package
 1407 called DBI.

1408 An automated system ensures real-time delivery of raw data and databases
 1409 to the offline processes. Data are first transferred to the onsite farm, then

1410 to the Institute of High Energy Physics (IHEP) in Beijing and Lawrence
1411 Berkeley National Laboratory (LBNL) in California for central storage and
1412 processing, and distributed to other institutions for validation and analysis.
1413 Figure 40 shows the network topology. A dedicated small-size offline system
1414 is deployed at the Daya Bay site, consisting of a file server, a data transfer
1415 server, an offline database server, a web server, two data monitoring servers
1416 and five servers forming user farms, as shown in Fig. 41. Two large-size of-
1417 fline systems are deployed at IHEP and LBNL, separately. Each of them has
1418 about 1 PB of disk space capacity and 800 CPU cores. Additional computing
1419 resources have been planned to accommodate the increasing data volume.

1420 A Performance Quality Monitoring system (PQM) [76] runs onsite, using
1421 fast reconstruction algorithms and analysis modules in NuWa, to monitor the
1422 physics performance with a latency of around 40 minutes. A “keep-up” data
1423 processing takes place as soon as the data reach IHEP or LBNL, using the
1424 full NuWa reconstruction and the latest calibration constants. The generated
1425 detector monitoring plots are distributed for collaboration-wide review and
1426 archived through an Offline Data Monitoring system with a latency of around
1427 3 hours. The extracted data quality information are filled in a dedicated
1428 database for long-term monitoring.

1429 **9. Operation and Performance**

1430 Daya Bay has collected antineutrino data since Sep. 2011, first with 2
1431 ADs, then with six ADs starting in Dec. 2011, and most recently with 8 ADs
1432 since Oct. 2012. Over one million (150 K) antineutrinos have been detected
1433 at the near (far) site detectors by the end of 2013.

Figure 42: Sample trigger rates in the near experimental halls

1434 Typical data taking includes a ≈ 48 hr physics run, followed by ≈ 1 -
 1435 minute pedestal and electronic diagnostic runs, as shown in Fig. 42. Data
 1436 from each of the three halls is recorded in separate runs using a universal
 1437 clock to record the begin/end run times. Triggers from any of the detectors
 1438 within a hall can initiate readout of the electronics in that hall. The typical
 1439 trigger rate for all detectors at each site in a physics run is 1.3 kHz for EH1,
 1440 1.0 kHz for EH2, and 0.6 kHz for EH3. Variation between halls is driven by
 1441 the different overburden and muon rate at each hall. One example of the
 1442 trigger rates for each subsystem is listed in Table 12. Electronic diagnostic
 1443 runs give additional information on the noise and gain of each channel for
 1444 nearby physics runs. About 320 raw data files, 1 GB per file, are generated
 1445 per day and transferred to the computing farms. Radioactive source and
 1446 LED pulser calibrations are also routinely performed. Detector live time for
 1447 recording antineutrino events was greater than 92% with the majority of the
 1448 down time dedicated to calibration.

Table 12: Trigger rates recorded for run 54261(EH1), run 54262(EH2) and run 54260(EH3)

Site	Total rate	Subsystem	Trigger rate
	(Hz)		(Hz)
EH1	1301	AD1	273
		AD2	268
		IWS	220
		OWS	325
		RPC	215
EH2	966	AD1	215
		AD2	211
		IWS	192
		OWS	245
		RPC	103
EH3	635	AD1	131
		AD2	124
		AD3	120
		AD4	131
		IWS	39
		OWS	54
		RPC	36

Figure 43: WP muon detection efficiencies from Dec. 2011 to Dec. 2013 (See Sec. 9.1 for details). The gap corresponds to the summer shutdown in 2012.

9.1. Muon System Performance

The water pool systems provides essentially 100% efficiency in detecting muons. This efficiency is studied using muons going through the ADs, which occur at about 20 Hz in EH1, 15 Hz in EH2 and 1 Hz in EH3. The mean IWS efficiencies for all three halls is $99.98 \pm 0.01\%$ and is quite stable over time as shown in Fig. 43(a). The OWS efficiencies shown in Fig. 43(b) are underestimated since there is no correction for stopping muons which deposit energy in the AD or IWS but not in the OWS.

9.2. AD performance - energy response

To take full advantage of the unprecedented statistical precision of this data set, careful monitoring and control of systematic errors are required. ADs are calibrated periodically using the ACUs on the lid of each detector. Combined radioactive sources ($\text{Am-}^{13}\text{C}/^{60}\text{Co}$ or $\text{Am-}^{13}\text{C}/^{68}\text{Ge}$) are lowered to five vertical locations for all three ACUs separately, followed by a similar scan using the other radioactive sources (^{68}Ge or ^{60}Co). Data are collected for four minutes at each position, accumulating about 27×10^3 (4.8×10^3) ^{60}Co

(^{68}G) events in each AD during early 2012. Low intensity LED runs collect data which can be used in the ADC calibration for each PMT readout channel. The demonstrated stability of the AD performance allowed a reduction in the number of calibrations starting Jan. 11, 2013. The weekly calibration was reduced to calibrations with the LED and ^{60}Co source at the detector center. The full calibration previously described was performed every fourth week.

Single photoelectron (SPE) hit data from physics data runs were selected and fitted after baseline subtraction with a convolution of a Poisson distribution with a Gaussian function as shown below:

$$S(x) = \sum_n \frac{\mu^n e^{-\mu}}{n!} \frac{1}{\sigma_1 \sqrt{2n\pi}} \exp\left(-\frac{(x - nQ_1)^2}{2n\sigma_1^2}\right) \quad (5)$$

where μ is the expected value of the Poisson distribution determined by the PMT hit occupancy, σ_1 and Q_1 are the resolution and mean value of the SPE distribution. The summation over the number of photo-electrons is truncated at $n=2$. The SPE typically has a mean value of about 19 ADC counts with σ_1/Q_1 of about 33%.

Figure 44 shows an example ADC distribution for SPE hits with fitting results overlaid.

Combined with low intensity LED run data, the PMT gain and timing are monitored continuously. The stability of the ADC gains was measured every ≈ 6 hours using SPE data from physics data runs as shown Fig. 45. Slight gain drifts are seen in all ADs and have been crosschecked with low-intensity LED data. Many parameters were investigated as possible causes of the observed gain drifts. Temperature drifts were partially correlated to the observed drifts but were not the only cause. Jumps correlated among the

Figure 44: Example baseline-subtracted SPE distribution with fit using the function in Eq. 5. The x-axis is the peak ADC value minus the ADC average baseline before the pulse.

1489 ADs are suspected to be caused by power cycling of the HV mainframes. It
 1490 is likely that the PMT SPE pulse height is slowly increasing with time since
 1491 periodic calibrations of the FEE card ADCs by onboard DACs are stable to
 1492 better than a percent per year.

1493 The number of P.E. per MeV (energy scale) based on ^{60}Co calibrations
 1494 at the detector center is monitored over time as shown in Fig. 46. A slight
 1495 downward slope ($\approx 1.5\%$ per year) is seen. Similar behavior is seen in energy
 1496 scales derived from spallation neutrons. The reason for the downward slope
 1497 of the energy scale is under study. A downward slope could be consistent with
 1498 a decrease in LS and GdLS attenuation lengths, however, no degradation of
 1499 light yield or attenuation length has been observed to date in the LS and
 1500 GdLS production samples preserved and tested in the lab. Further study
 1501 is clearly needed to fully understand all long-term trends. However, as will
 1502 be shown in later plots, the AD energy scales after calibration are stable to
 1503 0.1%.

Figure 45: Average number of ADC counts for single photoelectrons in each AD versus time. There were fewer measurements during the summer shutdown of 2012 (between the two grey vertical dash lines).

Figure 46: The energy scale from ^{60}Co calibrations at the detector center is monitored over time. The 2012 shutdown period occurs between the two grey dashed lines.

1504 Many special calibration runs were performed during the shutdown in
1505 2012 with nonstandard sources. A strong 60 Hz ^{241}Am - ^{13}C neutron source
1506 (same design as the weaker ACU sources) was deployed on the top lid of AD5
1507 in EH3 to better measure the backgrounds in physics data due to the ^{241}Am -
1508 ^{13}C neutron sources while in the parked position inside the ACU.

1509 Other dedicated calibrations were performed in EH1 and EH2 in parallel.
1510 Gamma source (^{68}Ge , ^{60}Co) and neutron source scans were performed with
1511 fine steps in z. Two ACUs (9A and 8C) were instrumented with special
1512 calibration sources and mounted on AD1 and AD2 in EH1. The sources in
1513 ACU 9A were AmC, AmBe, and a combined $^{60}\text{Co}/^{137}\text{Cs}$ source. PuC, ^{60}Co
1514 and ^{40}K were mounted in ACU 8C. The ^{137}Cs source in ACU 9A was replaced
1515 with a ^{54}Mn source at a later time.

1516 Additional calibration data included full detector vertex scans with the
1517 gamma sources in the z range from 1.8 m to -1.5 m in 10 cm steps and 7-
1518 hour-long neutron scans at four z positions (1.775 m, 1.35 m, 0 m, -1.35 m).
1519 Data taking without calibration sources was also performed.

1520 These tests provided:

- 1521 • Multiple gamma source data for studying the non-linearity of the GdLS
1522 energy response along the center axes.
- 1523 • Neutron source data at various positions for studying the H/Gd ratio,
1524 the prompt and delayed spectra in AD1/AD2, and to study the spill-
1525 in/out correction uncertainties.
- 1526 • High statistics ^{40}K data for non-linearity studies at the detector center.

1527 A manual calibration system was installed in AD1 in EH1 during the 2012

Figure 47: Energy spectrum from muon induced spallation neutrons. Points from detectors in the same hall are nearly identical and may be plotted over each other.

shutdown (Sec 5.2). Calibrations were performed over the full GdLS detector volume (r and z).

Data from the calibration sources as well as data from cosmogenic events such as spallation neutrons (recorded with the physics data) are used to study and track the energy response of the detectors. Spallation neutrons are cleanly selected by searching a time window from 20 to 200 μ s after a muon passes through an AD. The energy spectra of spallation neutrons in all 8 ADs are shown in Fig. 47. The 8 MeV peak from neutron capture on Gd and the 2.2 MeV peak from neutron capture on hydrogen are clearly identified. The neutron capture spectra of all eight ADs are in excellent agreement. The calibrated energy is a convolution of the ADC per PE and PE/MeV calibrations (see Fig. 45 and Fig. 46 as examples). Although both calibrations change with time, the calibrated energy should be stable. Figure 48 shows that the average energy of spallation neutrons capturing on Gd for the eight ADs is very stable.

Figure 48: The time dependence of the reconstructed energy of spallation neutrons captured on gadolinium.

1543 The neutron capture time is directly related to the Gd concentration
 1544 in the detector and can be tracked to look for time variation in the Gd
 1545 doping. Both neutrons from IBD interactions and muon spallation events
 1546 are used to study the capture time spectra. Figure 49 shows the capture
 1547 times of neutrons from IBD samples in the four near ADs. The four far ADs
 1548 are added together in the bottommost plot to improve the statistical error.
 1549 Capture times are stable, reaffirming that the Gd concentration in the GdLS
 1550 does not change with time.

1551 Relative energy scale uncertainties between ADs are a major source of
 1552 systematic errors. Particles at different energies are selected to study the
 1553 relative energy scale in all ADs. Clean samples of alphas are selected from
 1554 decays of ^{214}Po , ^{212}Po and ^{215}Po using the $\beta - \alpha$ and $\alpha - \alpha$ time and energy
 1555 correlations. Calibration gamma and neutron sources at specific locations
 1556 inside the AD, as well as uniformly distributed IBD and spallation neutron
 1557 sources are studied in all eight ADs. Figure 50 shows that the relative energy

Figure 49: Average capture time of IBD events in the four near hall ADs are shown as a function of time. From the top the ADs are: EH1-AD1, EH1-AD2, EH2-AD1 and EH2-AD2. The bottom graph averages all four far hall ADs together for better statistical precision. The dashed line shows the mean capture time of the AD(s).

Figure 50: Comparison of the reconstructed energy between antineutrino detectors for a variety of calibration references. E_{AD} is the reconstructed energy measured in each AD, and $\langle E \rangle$ is the 8-detector average. Error bars are statistical only. Systematic variations between detectors for all calibration references were $< 0.2\%$. The energy scale of each detector was determined using the ~ 8 MeV n-Gd capture gamma peaks from Am-C sources, hence these points show zero deviation.

scale differences among the eight ADs are all within a $\pm 0.2\%$ band.

A precise check of relative detector efficiencies can be made from the ratio of detected neutrino events in the nearly identical detectors in each hall. For the detectors in the same hall, background and reactor uncertainties cancel out, leaving only small corrections for differing baselines and target masses. For IBD-like events with more than 0.7 MeV prompt energy (nearly 100% of all prompt IBD positrons), the expected ratio in EH1 (AD1/AD2) is 0.982 and the measured value is 0.981 ± 0.004 , agreeing within error. The same is true in EH2, where the ratio of detected IBD events in AD3 and AD8 is measured to be 1.019 ± 0.004 , consistent with the expected value of 1.012 within error.

9.3. EH3-AD1 LS leak

AD operating conditions were generally stable during the first three years of operation. Return cover gas humidities and oxygen levels gradually declined in all ADs with no indications of water leaks. Liquid levels in the overflow tanks track small variations in the water pool temperature as the AD liquids expand or contract. The largest change of GdLS levels from Dec. 2013 to May 2015 was in EH3-AD1 with a 11.4 mm decrease corresponding to a change in the AD target mass of 0.06%.

After the Aug. 2012 access to EH3 to install EH3-AD4 it was observed that the LS level in EH3-AD1 was dropping while the MO level was rising as shown in Fig. 51a. This feature was also accompanied by changes in the MO absorbance vs wavelength measured by the MO clarity monitor in EH3-AD1. This trend continued for many months before reaching new equilibrium values in 2014. Taking the difference in AD liquid levels removes most of the

1583 temperature variation effects seen in Fig. 51(a). Figure 51(b) shows the
1584 difference between the LS and MO overflow tanks levels for the EH3-AD1-3.
1585 Unlike the near constant levels seen in EH3-AD2 and EH3-AD3, EH3-AD1
1586 shows a clear decline starting during the August access.

1587 This behavior can be understood as the consequence of a small crack at
1588 the bottom of the OAV that allows the LS to leak into the MO volume. The
1589 crack must be at or near the bottom of the OAV since the leak continued
1590 even when the MO level was several cm higher than the LS level. Since LS
1591 is $\approx 0.9\%$ denser than the MO, the differential pressure between the LS and
1592 MO volumes changes with height. Near the top of the OAV the MO is at
1593 higher pressure since the overflow tank levels are higher than the LS. The
1594 situation was reversed at the bottom of the OAV at the start of the leak. As
1595 the MO levels rose and LS levels dropped the leak slowed as the differential
1596 pressure between the LS and MO volumes decreased. The leak stopped when
1597 the LS and MO pressures across the crack equalized. At the bottom of the
1598 OAV the pressure due to the denser LS is offset by the 4.8 cm higher MO
1599 level.

1600 LS is expected to dissolve completely in the MO and to diffuse through
1601 the entire MO volume. Since the LS absorbance at low wavelengths (≤ 410
1602 nm) is much higher than MO, the absorbance of the MO in EH3-AD1 is
1603 expected to change as a function of time. Above 410 nm, the MO and
1604 LS absorbance curves have a very similar dependence on wavelength. The
1605 absorbance measured by the MO clarity monitor at 390 and 420 nm divided
1606 by the absorbance at 430 nm is plotted in Fig. 51(c). In the wavelength
1607 region where the MO and LS have nearly identical absorbance, the ratio of

the 420/430 nm data is flat. In the region where the absorbance of the LS and MO differ significantly, the ratio of the 390/430 nm absorbances exhibits a similar time dependence as the LS level, supporting the interpretation of a LS leak, as an explanation of the changing overflow tank levels.

An estimated 50 liters of LS has leaked into the 42800 liters of MO (0.12%) in EH3-AD1. Several studies were made to check if the AD performance was changed in any measurable way by this leak. Since almost all of the energy from Gd captures is contained within the GdLS and LS volumes, no changes to any of the oscillation analyses have been observed or are expected. The average energy of neutron captures on hydrogen particularly for captures near the edge of the LS should in principle increase, however, the effect is well below the systematic errors of the detector to detector energy response variation (0.2%). An analysis of cosmogenic muon distributions in the ADs has seen increased light yield of tracks traversing part of MO volume in EH3-AD1 compared to other ADs. The light yield increases with time as expected from the predicted LS contamination.

10. Conclusion

The Daya Bay experiment is unique in using multiple antineutrino detectors at the near and far experimental hall sites. Two antineutrino detectors are deployed in each of the two near detector halls to monitor the 6 Daya Bay nuclear reactor cores. Four antineutrino detectors are deployed in the far experimental hall built to measure the antineutrino oscillation. Interchangeable antineutrino detectors were built and assembled above ground using standardized procedures to minimize efficiency differences between de-

Figure 51: (a) GdLS, LS, and MO levels in EH3-AD1 versus time. (b) Difference between the LS and MO levels versus time for the original 3 ADs in EH3. (c) Ratio of the absorbance at 395 nm (420) to that at 430 nm as measured by the EH3-AD1 MO clarity monitor. The liquid levels and MO clarity data are both consistent with a small leak of LS from the OAV volume to the MO volume which started during the access in Aug. 2012.

tectors. Construction of the underground tunnels and halls in parallel with detector assembly minimized total project time and resulted in the first ≥ 5 standard deviation measurement of $\sin^2 2\theta_{13}$. Side-by-side comparisons between detectors at the near sites cross check calculations of relative detector efficiencies to better than 0.2%. The large target mass provided by the four detectors at the far hall enables an unmatched statistical precision, yielding the world's most precise measurements of $\sin^2 2\theta_{13}$ and the effective mass splitting Δm_{ee}^2 , now, and for the foreseeable future.

11. Acknowledgements

The Daya Bay Experiment is supported in part by the Ministry of Science and Technology of China, the United States Department of Energy, the Chinese Academy of Sciences, the CAS Center for Excellence in Particle Physics, the National Natural Science Foundation of China, the Guangdong provincial government, the Shenzhen municipal government, the China General Nuclear Power Group, the Research Grants Council of the Hong Kong Special Administrative Region of China, the MOST fund support from Taiwan, the U.S. National Science Foundation, Yale University, the Ministry of Education, Youth and Sports of the Czech Republic, the Joint Institute of Nuclear Research in Dubna, Russia, the NSFC-RFBR joint research program, the National Commission for Scientific and Technological Research of Chile. We acknowledge Yellow River Engineering Consulting Co., Ltd. and China Railway 15th Bureau Group Co., Ltd. for building the underground laboratory. We are grateful for the ongoing cooperation from the China Guangdong Nuclear Power Group and China Light & Power Company.

1656 References:

- 1657 [1] F. P. An et. al., Phys. Rev. Lett. 108 (2012) 171803, [arXiv:1203.1669].
- 1658 [2] F.P. An et al., Chin. Phys. C37 (2013) 011001, [arXiv:1210.6327].
- 1659 [3] F.P. An et al., Phys. Rev. Lett. 112 (2014) 061801, [arXiv:1310.6732].
- 1660 [4] B. Pontecorvo, Sov. Phys. JETP 6 (1957) 429 and 26 (1968) 984.
- 1661 [5] Z. Maki, M. Nakagawa, and S. Sakata, Prog. Theor. Phys. 28 (1962) 870
- 1662 .
- 1663 [6] M. Apollonio, et al., Phys. Lett. B466 (1999) 415 ; Eur. Phys. J. C 27
- 1664 (2003) 331.
- 1665 [7] K. Abe, et al., Phys. Rev. Lett. 107 (2011) 041801.
- 1666 [8] P. Adamson, et al., Phys. Rev. Lett. 107 (2011) 181802.
- 1667 [9] Y. Abe, et al., Phys. Rev. Lett. 108(2012) 131801 [arXiv:1112.6353]
- 1668 (2012).
- 1669 [10] T. A. Mueller et al., Phys. Rev. C 83(2011) 054615.
- 1670 [11] P. Huber, Phys. Rev. C84(2011) 024617.
- 1671 [12] L.A. Mikaelyan, V.V. Sinev, Physics of Atomic Nuclei 63 (2000) 1002.
- 1672 [arxiv:hep-ex/9908047].
- 1673 [13] Daya Bay Collaboration, Proposal of Daya Bay Experiment, 2007 [
- 1674 arXiv:hep-ex/0701029].

- 1675 [14] <<http://www.cgnpc.com.cn>>
- 1676 [15] J. Cao, Nucl. Phys. B (Proceeding Supplements) 229-232 205-209(2012)
1677 [arXiv:1101.2266]
- 1678 [16] S. F. E. Tournu et al., EPRI Report No. EPRI 2001.1001470, Palo Alto,
1679 CA (2001).
- 1680 [17] C. Xu, X. N. Song, L. M. Chen, and K. Yang, Chin. J. Nucl. Sci. and
1681 Eng. 23, 26 (2003).
- 1682 [18] F.P. An et al., [arXiv:1508.04233].
- 1683 [19] F.P. An et al., Nucl. Instr. and Meth. A 773(2015) 8-20,
1684 [arXiv:1407.0275].
- 1685 [20] J. Wilhelmi, et al., Journal of Water Process Engineering , Vol 5, 127-
1686 135(2015) [arXiv:1408.1302]
- 1687 [21] F. P. An et al., Nucl. Instr. and Meth. A 685 (2012) 78-97,
1688 [arXiv:1202.6181].
- 1689 [22] Doerfer Companies/TDS Automation, Waverly, Iowa. U.S.A.
- 1690 [23] H.R. Band et al., JINST 8 (2013) T11006, [arXiv:1309.1557].
- 1691 [24] Y. Abe, et al., arXiv:hep-ex/0606025.
- 1692 [25] J. K. Ahn et al., [arXiv:1003.1391] (2010).
- 1693 [26] H.R. Band et al., JINST 7 (2012) P11029, [arXiv:1210.0643].
- 1694 [27] X. Chen, et al., Science China Technological Sciences 56 (2013) 148.

- 1695 [28] R. Wang, et al., Science China Technological Sciences, 55 (2012) 1572.
- 1696 [29] B. Littlejohn, et al., JINST 4 (2009) T09001 [arXiv:0907.3706].
- 1697 [30] H.R. Band et al., JINST 7 (2012) P06004, [arXiv:1202.2000].
- 1698 [31] B. Littlejohn, Ph.D. Thesis, University of Wisconsin - Madison (2012).
- 1699 [32] Nakano International Limited. No. 3, Ln. 483, Sec. 2, Zhongshan Rd.,
1700 Taishan Township, Taipei County 243, Taiwan (R.O.C).
- 1701 [33] Reynolds Polymer Technology, Inc. 607 Hollingsworth St. Grand Junc-
1702 tion, CO 81505 USA, <<http://www.reynoldspolymer.com/>>.
- 1703 [34] M. Krohn et al., JINST 7 T08001 (2012).
- 1704 [35] L.J. Wen, et al., Nucl. Instr. and Meth. A 629 (2011) 296.
- 1705 [36] Zhimin Wang, et al., Nucl. Instr. and Meth. A 602 (2009) 489.
- 1706 [37] Hamamatsu Photonics, Electron Tube Divison. 314-5
1707 Shimokanzo, Iwata City, Shizuoka Pref., 438-0193, Japan,
1708 <<http://www.hamamatsu.com/>>.
- 1709 [38] P. DeVore et al., Nucl. Instr. and Meth. A 737 (2014) 222.
- 1710 [39] Wanda Beriguete, et al., Nucl. Instr. and Meth. A 763 (2014) 82-88
- 1711 [40] F. Boehm et al., Phys. Rev. D62(2000), 072002.
- 1712 [41] M. Yeh, A Garnov, R.L. Hahn, Nucl. Instr. and Meth. A 578 (2007) 329.
- 1713 [42] Yayun Ding, et al., Nucl. Instr. and Meth. A 584 (2008) 238.

- 1714 [43] M.C. Chen, Nuclear Physics B (Proceedings Supplements) 145 (2005)
1715 65.
- 1716 [44] S.C. Blyth, et al., Nucl. Instr. and Meth. A 723 (2013) 67.
- 1717 [45] M. Yeh, et al., Nucl. Instr. and Meth. A 618 (2010) 124.
- 1718 [46] H.R. Band, et al., JINST 8 (2013) P09015, [arXiv:1307.1089].
- 1719 [47] H.R. Band, et al., JINST 8 (2013) T04001, [arXiv:1206.7082].
- 1720 [48] H.R. Band, et al., JINST 7 (2012) P08005, [arXiv:1204.3500].
- 1721 [49] Gill Sensors, Saltmarsh Park, 67 Gosport Street, Lymington, Hants,
1722 SO41 9EG, UK. <http://www.gillsensors.co.uk/>
- 1723 [50] T. Skwarnicki, Ph.D Thesis, DESY F31-86-02(1986), Appendix E; M.J.
1724 Oreglia, Ph.D Thesis, SLAC-236(1980), Appendix D; J.E. Gaiser, Ph.D
1725 Thesis, SLAC-255(1982), Appendix F.
- 1726 [51] J. Liu et al., Nucl. Instr. and Meth. A 750(2014) 19-37,[arXiv 1305.2248].
- 1727 [52] J. Liu et al., arXiv:1504.07911.
- 1728 [53] H. X. Huang et al., JINST 8 (2013) P09013, [arXiv:1305.2343].
- 1729 [54] K. Chow et al., Nucl. Instr. and Meth. A 794 (2015)25.
- 1730 [55] J. L. Xu, et al., Chin. Phys. C 35, (2011) 844.
- 1731 [56] L. Ma, et al., Nucl. Instr. and Meth. A 659 (2011)154.
- 1732 [57] L. Ma et al., Chin. Phys. C34 (2010) 1116.

- 1733 [58] Z. Wang et al., IEEE NSS/MIC conference N30-456 (2008); Q.J.Li et al.,
1734 IEEE Nuclear Science Symposium Conference Record, N25-240 (2009)
1735 1821.
- 1736 [59] H. Xiang, et al., Nuclear Electronics & Detection Technology. Vol.29
1737 No.1, Jan. 2009 (in Chinese).
- 1738 [60] Hui Gong et al., Nucl. Instr. and Meth. A 637 (2011) 138.
- 1739 [61] F. Li et al., IEEE Transactions on Nuclear Science NS-58 (4) (2011)
1740 1723 .
- 1741 [62] X. Ji et al. 2008 IEEE NUCLEAR SCIENCE SYMPOSIUM AND MED-
1742 ICAL IMAGING CONFERENCE, VOLS 1-9; Xiaolu Ji et al. Real Time
1743 Conference (RT), 2012 18th IEEE-NPSS.
- 1744 [63] F. Li et al., Design and Implementation of BESIII Online Farm, in
1745 2008 IEEE Nuclear Science Symposium Conference Record, N30-72, pp.
1746 2112-2113, Dresden, Germany, 2008.
- 1747 [64] I. Alexandrov et al., "Online software for the ATLAS test beam data
1748 acquisition system" IEEE Trans. Nucl. Sci., vol. 51, no.3, Jun.2004.
- 1749 [65] F. Li et al., Nucl. Electron. Detection Technol. vol.26 , no. 5 (2006), 605
1750 (In Chinese).
- 1751 [66] X. Ji et al., Real Time Conference (RT), 2012 18th IEEE-NPSS.
- 1752 [67] <<http://dim.web.cern.ch/dim>>.

- 1753 [68] F. Li et al., 19th IEEE NPSS Real Time Conference, Nara, Japan, May
1754 2014.
- 1755 [69] H. Wei, et al., [arXiv:1505.02501].
- 1756 [70] S. Zhang et al., Science China Techological Sciences 56 (2013) 1966; X.
1757 Chen, et al., Proceedings of the 13th National Conference on Nuclear
1758 Electronics & Nuclear Detection Technology. October, 2006. P.108-113.
- 1759 [71] X. Du, M.S. thesis (in Chinese), University of Chinese Academy (2013).
- 1760 [72] <<http://cern.ch/gaudi>>.
- 1761 [73] <<http://root.cern.ch/drupal>>.
- 1762 [74] <<http://subversion.apache.org>>.
- 1763 [75] S. Agostinelli, et al., Nucl. Instr. and Meth. A 506 (2003) 250; J. Allison,
1764 et al., IEEE Transactions on Nuclear Science NS-53 (1) (2006) 270.
- 1765 [76] Ying-Biao Liu et al., Chinese Physics C 8(2014) 086001
1766 [arXiv:1406.2104]