The MAJORANA DEMONSTRATOR Radioassay Program

N. Abgrall^a, I.J. Arnquist^b, F.T. Avignone III^{c,d}, H.O. Back^{e,f,1},
A.S. Barabash^g, F.E. Bertrand^d, M. Boswell^h, A.W. Bradley^a, V. Brudaninⁱ,
M. Busch^{j,f}, M. Buuck^m, D. Byram^k, A.S. Caldwell^l, Y-D. Chan^a,
C.D. Christofferson^l, P.-H. Chu^h, C. Cuesta^m, J.A. Detwiler^m,
J.A. Dunmore^m, Yu. Efremenkoⁿ, H. Ejiri^o, S.R. Elliott^{*,h}, P. Finnerty^{p,f,2},
A. Galindo-Uribarri^d, V.M. Gehman^{h,6}, T. Gilliss^{p,f}, G.K. Giovanetti^{p,f}, J.
Goett^h, M.P. Green^{d,e,f}, J. Gruszko^m, I.S. Guinn^m, V.E. Guiseppe^c,
R. Henning^{p,f}, E.W. Hoppe^b, S. Howard^l, M.A. Howe^{p,f}, B.R. Jasinski^k,
R.A. Johnson^{m,3}, K.J. Keeter^r, M.F. Kidd^s, O. Kochetovⁱ, S.I. Konovalov^g,
R.T. Kouzes^b, B.D. LaFerriere^b, J. Leon^m, J.C. Loach^{t,4}, J. MacMullin^{p,f},
S. MacMullin^{p,f,5}, R.D. Martin^{k,12}, R. Massarczyk^h, S. Meijer^{p,f}, S. Mertens^a,
M.L. Miller^m, J.L. Orrell^b, C. O'Shaughnessy^{p,f}, N.R. Overman^b,
A.W.P. Poon^a, K. Pushkin^{k,7}, D.C. Radford^d, J. Rager^{p,f}, K. Rielage^h,
R.G.H. Robertson^m, E. Romero-Romero^{n,d}, M.C. Ronquest^{h,8},
A.G. Schubert^{m,9}, B. Shanks^{p,f}, M. Shirchenkoⁱ, K.J. Snavelly^{p,f,11}, N. Snyder^k,
D. Steele^{h,10}, A.M. Suriano^l, D. Tedeschi^c, J.E. Trimble^{p,f}, R.L. Varner^d,
S. Vasilyevⁿ, K. Vetter^{a,13}, K. Vorren^{p,f}, B.R. White^d, J.F. Wilkerson^{p,f,d},
C. Wiseman^c, W. Xu^h, E. Yakushevⁱ, C.-H. Yu^d, V. Yumatov^g, I. Zhitnikovⁱ

^aNuclear Science Division, Lawrence Berkeley National Laboratory, Berkeley, CA, USA

^bPacific Northwest National Laboratory, Richland, WA, USA

^cDepartment of Physics and Astronomy, University of South Carolina, Columbia, SC, USA

^dOak Ridge National Laboratory, Oak Ridge, TN, USA

^eDepartment of Physics, North Carolina State University, Raleigh, NC, USA

^fTriangle Universities Nuclear Laboratory, Durham, NC, USA

^gNational Research Center "Kurchatov Institute" Institute for Theoretical and Experimental Physics, Moscow, Russia

^hLos Alamos National Laboratory, Los Alamos, NM, USA

ⁱJoint Institute for Nuclear Research, Dubna, Russia

*Corresponding author, elliotts@lanl.gov

¹Permanent Address: Pacific Northwest National Laboratory, Richland, WA, USA

²Permanent Address: Applied Research Associates, Inc., 8537 Six Forks Road, Suite 600, Raleigh, NC, USA

³Permanent Address: Microsoft Corporation, One Microsoft Way, Redmond, WA, USA

⁴Permanent Address: Shanghai Jiao Tong University, Shanghai, China

⁵Permanent Address: Picarro Inc., 3105 Patrick Henry Dr., Santa Clara, CA, USA

⁶Permanent Address: Lawrence Berkeley National Laboratory, Berkeley, CA, USA

⁷Permanent Address: Randall Laboratory of Physics, University of Michigan, Ann Arbor, Michigan, USA

⁸Permanent Address: CCRi, 1422 Sachem Pl #1, Charlottesville, VA, USA

⁹Permanent Address: Department of Physics, Stanford University, Stanford, CA, USA

¹⁰Permanent Address: Picarro Inc., 3105 Patrick Henry Dr., Santa Clara, CA, USA

¹¹Permanent Address: IBM Cloudant, Boston, 200 State St, Boston, MA, USA

¹²Permanent Address: Department of Physics, Engineering Physics and Astronomy, Queen's University, Kingston, ON, Canada

¹³Alternate Address: Department of Nuclear Engineering, University of California, Berkeley, CA, USA

Preprint submitted to Nuclear Inst. and Methods in Physics Research, A April 3, 2016

^j*Department of Physics, Duke University, Durham, NC, USA*
^k*Department of Physics, University of South Dakota, Vermillion, SD, USA*
^l*South Dakota School of Mines and Technology, Rapid City, SD, USA*
^m*Center for Experimental Nuclear Physics and Astrophysics, and Department of Physics,*
University of Washington, Seattle, WA, USA
ⁿ*Department of Physics and Astronomy, University of Tennessee, Knoxville, TN, USA*
^o*Research Center for Nuclear Physics and Department of Physics, Osaka University,*
Ibaraki, Osaka, Japan
^p*Department of Physics and Astronomy, University of North Carolina, Chapel Hill, NC,*
USA
^q*Centre for Particle Physics, University of Alberta, Edmonton, AB, Canada*
^r*Department of Physics, Black Hills State University, Spearfish, SD, USA*
^s*Tennessee Tech University, Cookeville, TN, USA*
^t*Shanghai Jiao Tong University, Shanghai, China*

Abstract

The MAJORANA collaboration is constructing the MAJORANA DEMONSTRATOR at the Sanford Underground Research Facility at the Homestake gold mine, in Lead, SD. The apparatus will use Ge detectors, enriched in isotope ^{76}Ge , to demonstrate the feasibility of a large-scale Ge detector experiment to search for neutrinoless double beta decay. The long half-life of this postulated process requires that the apparatus be extremely low in radioactive isotopes whose decays may produce backgrounds to the search. The radioassay program conducted by the collaboration to ensure that the materials comprising the apparatus are sufficiently pure is described. The resulting measurements from gamma-ray counting, neutron activation and mass spectroscopy of the radioactive-isotope contamination for the materials studied for use in the detector are reported. We interpret these numbers in the context of the expected background for the experiment.

Key words: Radiopurity; Trace analysis; Neutron activation analysis; Mass spectrometry; Mass spectroscopy; Germanium counting; Low background; Double beta decay; MAJORANA
PACS: 82.80.Jp; 14.60.Pq; 23.40.-s; 23.40.Bw

1. Introduction: Overview of the DEMONSTRATOR

The MAJORANA collaboration [1] will search for the neutrinoless double-beta decay ($\beta\beta(0\nu)$) of ^{76}Ge . The observation of this rare decay would indicate the neutrino is its own anti-particle, demonstrate that lepton number is not conserved, and provide information on the absolute mass-scale of the neutrino (see Refs. [2, 3, 4, 5, 6, 7, 8] for recent reviews of $\beta\beta(0\nu)$). Reaching the neutrino mass-scale sensitivity associated with the inverted mass ordering (15 – 50 meV) is a goal for next-generation $\beta\beta(0\nu)$ searches. This goal will require a half-life sensitivity exceeding 10^{27} y, which corresponds to a signal on the order of a few

58 counts or fewer per tonne-year in the $\beta\beta(0\nu)$ peak (2039 keV for ^{76}Ge). To ob-
 59 serve such a rare signal, one will need to construct large-scale experiments with
 60 backgrounds in the region of interest (ROI) below 1 count per tonne of isotope
 61 per year ($<1 \text{ c}/(\text{ROI t yr})$). The MAJORANA collaboration [1] is constructing
 62 the DEMONSTRATOR, an array of high-purity germanium (HPGe) detectors at
 63 the 4850 ft level of the Sanford Underground Research Facility (SURF) in Lead,
 64 South Dakota [9, 10]. The DEMONSTRATOR will consist of a mixture of HPGe
 65 detectors including, 15 kg fabricated from natural-isotopic-abundance Ge and
 66 29.7 kg fabricated from Ge enriched to $>87\%$ in ^{76}Ge . These detectors are con-
 67 tained within two low-background copper cryostats. Each cryostat will contain
 68 seven closely-packed stacks of detectors, called strings, with up to five detectors
 69 comprising each string. Figures 1 and 2 show the DEMONSTRATOR concept.

70 The DEMONSTRATOR aims to show that a background rate lower than
 71 $3 \text{ c}/(\text{ROI t yr})$ in the 4 keV ROI surrounding the 2039 keV ^{76}Ge Q-value
 72 energy is achievable. This background level will scale to $1 \text{ c}/(\text{ROI t yr})$ in
 73 a future experiment based on simulations considering improved self-shielding,
 74 thicker inner copper shield, and improved cosmogenic isotope control. Hence
 75 the DEMONSTRATOR will establish the technology required to build a large-scale
 76 germanium based $\beta\beta(0\nu)$ experiment.

Figure 1: A cross sectional view of a MAJORANA DEMONSTRATOR cryostat. The strings within each cryostat hold a mixture of natural and enriched germanium detectors.

77 The MAJORANA collaboration uses p-type point contact (PPC) HPGe de-
 78 tectors. These detectors [11, 12, 13, 14] have been demonstrated to provide
 79 both good energy resolution ($<2.0 \text{ keV FWHM}$ at 1332 keV, $<4.0 \text{ keV FWHM}$
 80 at 2039 keV) and low-energy threshold ($\sim 500 \text{ eV}$) [12, 15]. (See e.g., Fig. 1 in
 81 Ref. [15].) Each PPC detector used in the DEMONSTRATOR has a mass in the
 82 range of 0.5-1.1 kg with a mean of 0.85 kg.

Figure 2: The MAJORANA DEMONSTRATOR is shown here with both active and passive shielding in place. The outer surface of the inner Cu shield is 50.8 cm in height and 76.2 cm in length.

This report summarizes the assay program conducted by the MAJORANA collaboration to ensure that the various components of the experimental apparatus have radioactive isotope contamination sufficiently low to meet the background goal. Section 2 discusses the strategy to reach the required background levels in the experiment. Section 3 describes the methods and facilities used for the assays. Section 4 summarizes the results giving the levels of radioactive contamination found in the various materials studied and describes some special techniques or materials that were investigated. Finally, Section 5 provides some conclusions based on these studies.

Numerous other such studies have been done and, in concert with the results here, there is a wealth of information available to help select materials for future projects. Other reports can be found in Refs. [16, 17, 18, 19, 20, 21, 22, 23, 24]. Heusser [25] wrote a nice review on low background counting techniques. The results presented in this manuscript will be made available on the online database at radiopurity.org following publication.

2. The DEMONSTRATOR Low-Background Strategy and Model

The projected background in the DEMONSTRATOR is significantly lower than previous generation experiments. This reduction is achieved in part by fielding the detectors in large arrays that share a cryostat, allowing for the minimization of the amount of interstitial passive material. Further background suppression is achieved through the aggressive reduction of radioactive impurities in construction materials and by minimizing exposure to cosmic rays. MAJORANA will also make use of event signatures including pulse-shape characteristics, detector hit granularity, cosmic ray veto tags, and single-site time correlations to reduce backgrounds that do appear. In this section we describe these aspects of the MAJORANA DEMONSTRATOR design and their impact on the projected

backgrounds and physics sensitivity. In future sections we add detail about processes that are not published elsewhere and cite those that are.

The production process for enriched germanium detectors (enrichment, zone refining, and crystal growth) efficiently removes natural radioactive impurities from the bulk germanium. The cosmogenic activation isotopes, ^{60}Co and ^{68}Ge , are produced in the crystals while they are above ground, but can be sufficiently reduced by limiting the time above ground and by the use of passive shielding during transport and storage.

For the main structural material in the innermost region of the apparatus, MAJORANA chose copper for its lack of naturally occurring radioactive isotopes and its excellent physical properties. Starting with the cleanest copper stock identified and electroforming it underground reduces primordial radioactivity and cosmogenically-produced ^{60}Co . This electroformed copper will also be employed for the innermost passive, high-Z shield. Commercial oxygen-free high-conductivity (OFHC) copper stock is clean enough for use as the next layer of shielding. For all uses of copper, MAJORANA certified the cleanliness of samples via assay. Modern lead is available with sufficient purity for use as the bulk shielding material outside of the copper layers.

Several clean plastics are available for electrical and thermal insulation. For the detector supports we use a pure polytetrafluoroethylene (PTFE), DuPontTM Teflon[®] NXT-85. For the few weight-bearing plastic components requiring higher rigidity, we have sourced pure stocks of PEEK[®] (polyether ether ketone), produced by Victrex[®], and Vespel[®] (all grade SP-1), produced by DuPontTM. After machining, these parts are leached to remove surface activity. Thin layers of low-radioactivity parylene are used as a coating on copper threads to prevent galling, and for providing electrical insulation.

The high material purities required for the MAJORANA DEMONSTRATOR necessitated the development of improved assay capabilities. These capabilities are needed not only to ensure that the required purities can be achieved, but to also monitor construction processes to verify that cleanliness is maintained. We rely primarily on three assay methods: γ -ray counting, mass spectrometry, and neutron activation analysis.

Unlike activation analysis and mass spectrometry, γ -ray counting has the benefit of being non-destructive so that specific assayed parts can be used for the apparatus, hence avoiding a reliance on sampling. However, it is not sensitive enough for most parts of the DEMONSTRATOR and is used primarily for prescreening and for parts requiring only modest radiopurity. Furthermore, it is also useful to supplement mass spectrometry with γ -ray counting to cross-check for potential disequilibria in the natural decay chains. Mass spectrometry and neutron activation analysis are destructive methods that only measure the concentrations of the isotopes ^{238}U and ^{232}Th , which can elucidate the concentrations of ^{214}Bi and ^{208}Tl when secular equilibrium can be assumed. However, ^{238}U and ^{232}Th are at the top of these chains, unlike ^{214}Bi and ^{208}Tl , which produce the γ rays that actually comprise the radioactive background. By contrast, γ -ray counting measures the ^{214}Bi and ^{208}Tl activities directly.

3. Description of Assay Methods and Facilities

3.1. γ -Ray Counting

For γ -ray counting, the MAJORANA collaboration primarily used three facilities. Each had its own Monte Carlo for estimating efficiencies for each sample counted. A well-characterized sample was counted on all systems and analyzed blind as the operators did not know the activity of this test sample. All groups found similar results to within about 20%. The uncertainties in the results were dominated by counting statistics and simulation details. All the detectors were enclosed within a Cu shield surrounded by a Pb brick enclosure. All of the cavities were purged of Rn with boil-off gas from liquid nitrogen.

The three facilities have their key characteristics summarized in Table 1 and were:

- The low-background counting facility located underground at the Waste Isolation Pilot Plant (WIPP) near Carlsbad, NM operated by Los Alamos National Laboratory. (See, e.g., Ref. [26].) The Ge detector used at this facility was fabricated in 1985 and placed underground at WIPP in 1998. It is an n-type semi-coax design. This facility was decommissioned in 2014.
- The Kimballton Underground Research Facility (KURF) [27] is located at Lhoist North America's Kimballton mine in Ripplemead, VA and is operated by the Virginia Polytechnic Institute and State University. The counting facility, operated by the University of North Carolina and Triangle Universities Nuclear Laboratory, consists of two HPGe detectors specifically designed for low-background assay work. The first detector, named VT-1, is a commercial ORTEC LLB (very low-background) series coaxial detector. The second detector, named MELISSA, is a Canberra LB (low background) coaxial detector.
- The low background facility (LBF) at Oroville [28, 29], operated by Lawrence Berkeley National Laboratory (LBNL), was located in the Edward Hyatt Power Plant of the Oroville Dam in Oroville, CA. The p-type HPGe/ULB detector (manufactured by ORTEC) has been underground since ~ 1995 . This apparatus was moved to SURF in 2014. A separate low background facility in Building 72 of LBNL was used to count samples that do not require high sensitivity and to pre-screen samples for more sensitive counting at other facilities. The principal detector is an n-type germanium detector.

3.2. Mass Spectrometry Analysis

Both Glow Discharge Mass Spectrometry (GDMS) and Inductively-Coupled Plasma Mass Spectrometry (ICPMS) were used for MAJORANA DEMONSTRATOR assays. Each of these measure specific elements within the decay chains and therefore give little information about the equilibrium of the chain. Our

Table 1: A summary of the key characteristics of the 5 Ge detectors housed at 3 facilities that were used for this work. The overburden is given in meters water equivalent (mwe). All the cavity footprints are square except for VT-1, which is circular.

	LBNL Surface	LBNL Oroville	KURF MELISSA	KURF VT-1	WIPP
Overburden	10 mwe	600 mwe	1450 mwe	mwe	1700 mwe
Rel. Effic.	115%	85%	50%	35%	22%
Shield Cu		12.7 mm	25.4 mm	0.3 mm	50 mm
Cavity Height	14.5 cm	43 cm	38 cm	41 cm	15 cm
Cavity Area	317 cm ²	400 cm ²	1444 cm ²	616 cm ²	100 cm ²
≈ U Sens.	0.5 ppb	50 ppt	0.4 ppb	0.7 ppb	0.1 ppb
≈ Th Sens.	2.0 ppb	200 ppt	0.8 ppb	1.6 ppb	0.3 ppb

background estimates assume equilibrium and therefore, since our most sensitive assays come from ICPMS, there is an uncertainty associated with this assumption that is difficult to estimate.

3.2.1. Glow Discharge Mass Spectrometry

Glow Discharge Mass Spectrometry is useful for electrically conductive samples and requires very little sample preparation since the surface can be sputtered cleanly. Sensitivities approaching ppt (10^{-12} g/g) can be achieved for high mass elements such as U and Th. A homogenous sample is required unless surface-only or depth profiling is desired. In GDMS the bombarding ions come from a low pressure DC plasma discharge cell in which the sample is the cathode. GDMS is not particularly matrix-dependent and can be performed directly on samples with little or no preparation or separation chemistry. It also offers the advantage of quick turnaround compared to other mass spectrometric analysis methods that require time- and labor-intensive preparation. Conductive materials are easiest to analyze by GDMS, although conductive electrodes can often be formed when the sample material is non-conductive. GDMS is excellent for identifying trace elements in bulk samples down to tens of ppt. MAJORANA used the Thermo VG9000 GDMS instrument at the National Research Council of Canada (NRCC) [30] for Pb assay and an Astrum at Nu Instruments Limited in Wrexham, UK [31] to provide K values for the electroformed copper.

3.2.2. Inductively-Coupled Plasma Mass Spectrometry

A sample analyzed by Inductively-Coupled Plasma Mass Spectrometry is usually first put into solution using various combinations of acids or bases. In ICPMS, a flow of gas (usually argon) converts the liquid sample into a fine aerosol. A portion of the aerosol is then directed through the center of an argon plasma torch, where the aerosol particles are effectively atomized and ionized. In some cases the ions are then directed into a hexapole or octopole collision cell where polyatomics (molecular ions) can be dissociated or excluded before entering the final mass filter, most often a quadrupole. MAJORANA used four

facilities for its ICPMS analysis. Most of our assay analyses were performed at Pacific Northwest National Laboratory (PNNL) and Lawrence Berkeley National Laboratory (LBNL); those facilities are described below. MAJORANA also used the Institute of Microelectronic Technology and High Purity Materials in Chernogolovka, Russia [32] and Validation Resources Inc in Bend OR [33].

In general, ICPMS samples for introduction must be prepared in solution. This requires solid samples to be digested prior to assay. For the utmost in accuracy and precision, we use isotope tracer dilution methods to account for sample preparation effects on the analyte as well as plasma perturbations and instrument drift during analysis. While sample dissolution complicates the procedure, the advantage is that the results represent the bulk material which is homogenized into a solution. This also allows for solution-based chemistries to be performed, if desired, to further concentrate the analyte or remove unwanted matrices. However, the sample size is typically small, so for large items only an indication of the bulk contamination results. For compound materials, some components dissolve better than others. For MAJORANA we demonstrated sensitivities better than ppt for U and Th using ICPMS.

Samples were prepared in clean room conditions to avoid contamination. Elemental isolation or chemical purification techniques were often necessary to avoid the isobaric interferences caused by other isotopes or ion complexes within large quantities of dissolved solids. Based on the MAJORANA DEMONSTRATOR background calculations, the most stringent radiopurity goal is that for copper used in the inner shield and detector components. The effort has reached the required purity levels ($<0.3 \mu\text{Bq}$ for both ^{232}Th and ^{238}U per kg of copper or $0.024 \times 10^{-12} \text{ g } ^{238}\text{U/g Cu}$ and $0.075 \times 10^{-12} \text{ g } ^{232}\text{Th/g Cu}$). See Ref. [34, 35] and the results reported here in Rows 1 through 7 in Table 3.

At LBNL the ICPMS measurements were done at the Earth Sciences Division's Aqueous Geochemistry Laboratory. This facility is equipped with a PerkinElmer SCIEX Elan DRCII ICPMS instrument and an Anton Paar Multiwave 3000 Microwave Reaction System, as well as Class 100 flow hoods for sample and standards preparation. This equipment has been used for quantitative ppt-scale measurements in a variety of materials.

At PNNL an Agilent 7700 ICPMS and all sample preparation equipment is located in dedicated clean rooms for low background measurements. With a variety of advanced dissolution technologies, such as electrochemical sample preparation or microwave digestion, virtually anything can be brought into solution and analyzed by ICPMS down to the sub ppt or ppq level. All analysis performed at PNNL for the MAJORANA DEMONSTRATOR used isotopic tracers as the standard. The tracers used for the analysis of ^{232}Th and ^{238}U were ^{229}Th and ^{233}U , respectively.

3.3. Neutron Activation Analysis

For materials such as hydrocarbons with no long-lived neutron activation products, instrumental neutron activation analysis (NAA) can achieve substantially greater sensitivity than direct γ -ray counting. In this technique, samples are irradiated with neutrons from a nuclear reactor. When the neutrons are

268 captured on ^{238}U and ^{232}Th , the isotopes ^{239}Np and ^{233}Pa are generated. After
 269 irradiation, the samples are counted by high-resolution γ -ray detectors to search
 270 for characteristic lines at 106 keV and 312 keV from ^{239}Np and ^{233}Pa decays.
 271 Using known or calibrated neutron capture probabilities, neutron flux, irradiation
 272 time, γ -ray detector efficiencies, and sample mass, the concentrations of U
 273 and Th in a sample can be calculated.

274 For high-sensitivity NAA it is important to have samples free from contami-
 275 nants. Plastic samples were prepared for assay by leaching with Optima™ grade
 276 (Fisher Scientific) acids. After leaching, samples were dried with nitrogen and
 277 handling was minimized thereafter to avoid the introduction of Na or K. These
 278 elements can easily capture neutrons producing the short-lived ^{24}Na and ^{42}K
 279 isotopes. The contribution to the γ -ray spectra from these elements can obscure
 280 lines from ^{239}Np and ^{233}Pa that indicate the presence of U and Th. Other con-
 281 taminants that can decrease sensitivity and require care to avoid include Cr,
 282 Mn, Cu, Zn, Br, W, and Au.

283 MAJORANA made use of three neutron activation facilities. These facilities
 284 have their key characteristics summarized in Table 2 and were:

- 285 • The University of California Davis' McClellan Nuclear Radiation Center
 286 (MNRC) provides an array of options for sample irradiations including
 287 both in-core and out-of-core locations with varying encapsulation require-
 288 ments, neutron spectra, and total neutron fluence limits. For this work
 289 we employed MNRC's Pneumatic Transfer System, in which small sam-
 290 ples are placed inside polyethylene containers (rabbits) and transferred
 291 into and out of the reactor core via a pneumatic system. Consideration
 292 of the structural integrity of the rabbits limits irradiations for in-core
 293 position to <1 MW-hour. Neutron fluences were calibrated by simultane-
 294 ously irradiating a sample of "standard pottery" ceramic material, with
 295 well-characterized U and Th content, in the same rabbit as the sample
 296 of interest. Post-irradiation sample counting was performed at the low-
 297 background counting facilities at LBNL. (See Section 3.1.)
- 298 • North Carolina State University's research and teaching reactor, in its
 299 Nuclear Engineering Department, uses the 1 MW PULSTAR reactor fa-
 300 cility for irradiations. Our samples were irradiated in an out-of-core port
 301 in which the sample was rotated to provide a smooth irradiation profile
 302 across the sample holder. Our samples were irradiated for 12 MW-hours
 303 along with samples of standard pottery and other flux-monitoring mate-
 304 rials used by the NC State facility. Assay of the samples post-irradiation
 305 was performed on-site via HPGe γ -ray detectors.
- 306 • The High Flux Isotope Reactor (HFIR) at Oak Ridge National Laboratory
 307 (ORNL) is an 85 MW enriched fuel reactor located at ORNL. Samples are
 308 irradiated at HFIR by pneumatically inserting rabbits into the beryllium
 309 reflector at a distance of 180 mm from the edge of the fuel region. The
 310 neutron flux at the irradiation location is 2.8×10^{14} thermal n/(cm² s).
 311 The thermal-to-epithermal flux ratio is approximately 40. Sample size

Table 2: A summary of the key characteristics of the NAA facilities used for this work.

Facility	MNRC	NCSU	HFIR
Reactor	TRIGA	PULSTAR	Enriched Fuel
Power	2 MW	1 MW	85 MW
Thermal n Flux	1×10^{13} n/cm ²	1×10^{13} n/(cm ² s)	2.8×10^{14} n/(cm ² s)
Fast n Flux	5×10^{12} n/cm ²	1×10^{12} n/(cm ² s)	7×10^{12} n/(cm ² s)

is limited by the size of the rabbit and is usually at the level of a few grams or less. Irradiation time is limited to 10-20 minutes to prevent sample melting due to extensive γ -ray heating. After irradiation, the samples were left overnight in a hot cell to allow highly-radioactive short-lived isotopes to decay and then counted on-site via HPGe γ -ray detectors. Some samples were also subsequently counted underground at KURF with a low background Ge detector. (See Section 3.1.)

4. Assay Results

The results of our assay program are listed in Table 3. This Table lists the fractional content of K, U and Th contained within the numerous materials and products assayed during development and assembly of the DEMONSTRATOR. In addition, a number of parts or processes warrant special description due to the techniques developed for their assay. In this section we discuss such developmental work in a series of subsections.

Table 3: Radioactive isotope levels within various materials and their 68% CL uncertainties. Values for K were not always provided by the analysis.

#	Material	Method	K (10^{-9} g/g)	^{232}Th (10^{-12} g/g)	^{238}U (10^{-12} g/g)
<i>Metals</i>					
1	Cu Electroformed Stock Sample	ICPMS		<0.17	
2	Cu Electroformed Stock Sample	ICPMS		0.011 $\pm$ 0.005	0.017 $\pm$ 0.003
3	Cu Electroformed Stock Sample	GDMS	<2.2	<50	<70
4	Cu Electroformed Stock Sample	ICPMS		<0.029	<0.008
5	Cu Electroformed Stock Sample	ICPMS		<0.029	<0.009
6	Cu Electroformed Stock Sample	ICPMS		<0.029	<0.008
7	Cu Electroformed Stock Sample	ICPMS		<0.030	<0.009
8	Cu Electroformed, machined part, guide clip	ICPMS		0.330 $\pm$ 0.022	0.123 $\pm$ 0.005
9	Cu Electroformed, machined part, guide clip	ICPMS		0.112 $\pm$ 0.009	0.078 $\pm$ 0.002
10	Cu Electroformed, machined part, guide clip	ICPMS		0.170 $\pm$ 0.008	0.087 $\pm$ 0.002
11	Cu Electroformed, machined part, spring clip	ICPMS		0.215 $\pm$ 0.009	0.130 $\pm$ 0.010
12	Cu Electroformed, machined part, hex bolt	ICPMS		0.118 $\pm$ 0.011	0.035 $\pm$ 0.004
13	Cu Electroformed, machined part, hex bolt	ICPMS		0.119 $\pm$ 0.014	0.041 $\pm$ 0.003
14	Cu Electroformed, machined part, hex bolt	ICPMS		0.148 $\pm$ 0.021	0.051 $\pm$ 0.002
15	Cu, C10100 cake stock, (source for Rows 16,17)	ICPMS		0.46 $\pm$ 0.06	0.21 $\pm$ 0.06
16	Cu, C10100 2.5" plate stock, exterior sample	ICPMS		0.27 $\pm$ 0.05	0.10 $\pm$ 0.02
17	Cu, C10100 2.5" plate stock, interior sample	ICPMS		0.27 $\pm$ 0.05	0.12 $\pm$ 0.02
18	Cu, C10100 1" plate stock, saw cut (same stock Row 19)	ICPMS		10.2 $\pm$ 1.0	6.62 $\pm$ 0.58
19	Cu, C10100 1" plate stock, machined surfaces	ICPMS		1.88 $\pm$ 0.45	3.11 $\pm$ 0.39
20	Cu, C10100 1" x 2" bar stock, machined surfaces	ICPMS		2.12 $\pm$ 0.39	2.25 $\pm$ 0.15
21	Cu, C10100 1" plate stock	ICPMS		<0.029	0.013 $\pm$ 0.002
22	Cu, C10100 2.5" plate stock	ICPMS		<0.030	0.017 $\pm$ 0.003
23	Cu, C10100 2.5" plate stock	ICPMS		0.049 $\pm$ 0.010	0.061 $\pm$ 0.006
24	Cu, C10100 0.5" plate stock	ICPMS		<0.030	0.009 $\pm$ 0.001
25	Cu wire, California Fine Wire	ICPMS	<25000	<87	<40

Table 3: Radioactive isotope levels within various materials and their 68% CL uncertainties. Values for K were not always provided by the analysis.

#	Material	Method	K (10^{-9} g/g)	^{232}Th (10^{-12} g/g)	^{238}U (10^{-12} g/g)
26	Pb, smelted from virgin ore, Sullivan Metals	γ count	<60	<100	<30
27	Pb, UW	γ count	<190	<200	<500
28	Pb, UW	γ count	<160	<170	<400
29	Pb, smelted from virgin ore, Sullivan Metals	γ count	<160	<173	<241
30	Pb, smelted from virgin ore, Sullivan Metals	GDMS	4 ± 2	<10	<10
31	Pb, UW	GDMS	23 ± 11	<8	<10
32	Pb, UW	GDMS	<0.4	<8	<9
33	Pb, archeological ingot, UChicago	GDMS	<0.3	<9	<9
34	Pb, archeological sample prepared by Mifer Brick	GDMS	<0.2	<8	<10
35	Pb (Average from Brick samples)	ICPMS		1.3 ± 1.3	2.9 ± 2.0
36	Sn, sample of unknown origin	γ count	800 ± 450	<760	<137
37	Sn, sample of unknown origin	ICPMS	<108	940 ± 50	1190 ± 170
38	Sn, sample supplied by Canberra	ICPMS	<108	760 ± 70	1150 ± 350
39	6-way SS conflat intersection, MDC Vac. Prod., LLC	γ count	<840	3200 ± 1000	<400
40	TIG-Ce welding rods	γ count	$(1.60 \pm 0.14) \times 10^5$	$(1.68 \pm 0.31) \times 10^7$	<72000
41	TIG-Zr welding rods	γ count	5500 ± 4600	$(1.08 \pm 0.05) \times 10^5$	19300 ± 1600
42	Cr, stock used for vapor depositions	ICPMS	<7000	<20000	<5000
43	Au, sputtering target	ICPMS	<270		570 ± 130
44	Au (4N8), sputtered at LBNL	ICPMS	47000 ± 1000	1980 ± 370	2000 ± 300
45	Al, sputtered, sample film provided by ORTEC	ICPMS	$(1.42 \pm 0.51) \times 10^7$	2000 ± 250	5730 ± 300
46	Al, sputtered, sample film provided by ORTEC	ICPMS	$(1.10 \pm 0.01) \times 10^5$	2210 ± 460	4390 ± 340
47	Ge, sputtered, sample film provided by ORTEC	ICPMS	<430	207 ± 38	843 ± 62
48	Ge, sputtered, sample film provided by ORTEC	ICPMS	<215	349 ± 80	1340 ± 120
49	amorphous Ge, sputtered at LBNL	ICPMS	4800 ± 230	2370 ± 690	1680 ± 350
50	Cr, sputtered at LBNL	ICPMS	<1900	5240 ± 1290	5030 ± 700
51	Ti film, sputtered at LBNL	ICPMS		<400	<100

Table 3: Radioactive isotope levels within various materials and their 68% CL uncertainties. Values for K were not always provided by the analysis.

#	Material	Method	K (10^{-9} g/g)	^{232}Th (10^{-12} g/g)	^{238}U (10^{-12} g/g)
<i>Plastics</i>					
52	Teflon [®] TE-6742	NAA	0.15±0.02	0.025±0.002	<0.4
53	Peek [®] , Victrex [®]	NAA	180±110	<400	<5100
54	Vespel [®] , Dupont [™] rod, ThyssenKrupp Materials	NAA	350±300	<2.9	<84.4
55	Vespel [®] , Dupont [™] rod, Professional Plastics Inc.	NAA	191±31	<49	<45
56	Vespel [®] , Dupont [™] plate, Professional Plastics Inc.	NAA	<200	<42	<930
57	Parylene N dimer, Para Tech Coating Inc.	NAA		<50	<30
58	Parylene, Speciality Coating Systems [™] Inc.	NAA	5800±1300	<850	<1700
59	Parylene C, dimer, Speciality Coating Systems [™] Inc.	ICPMS	2110±15	390±30	6230±110
60	Parylene C, dimer pre cleaned, Spec. Coat. Sys. Inc.	ICPMS	<108	37±3	4230±60
61	Parylene C, infusion rod bump, Spec. Coat. Sys. Inc.	ICPMS	<320	250±60	46±20
62	Parylene C, inlet, Speciality Coating Systems [™] Inc.	ICPMS	<430	140±30	83±24
63	Parylene C, table, Speciality Coating Systems [™] Inc.	ICPMS	923±86	530±30	250±60
64	5% boron HDPE, Plasti-Shield [®] , King Plastic	γ count	<8000	<8000	<3000
65	Densetec HDPE, Polymer Industries	γ count	1400 ± 200	3000 ± 200	1600 ± 100
66	PTFE, 0.5" tubing, Cole-Palmer	ICPMS		1.5±1.8	<3.1
67	Kalrez [®] , Dupont [™] , O-ring material	NAA		4700±200	<4100
68	Kalrez [®] , Dupont [™] , O-ring material	ICPMS		1785±890	754±57
69	FEP shrink tubing ACCU-GLASS	NAA	<19	<34	<100
70	FEP shrink tubing NEWARK	NAA	20.6±2.3	<12	8.0±3.4
71	FEP tubing	NAA		<45	<150
72	PTFE 0.002-in sheet	NAA		<5.1	<7.6
<i>Cables and parts</i>					
73	FEP, Dupont [™] NP20, outer jacket Axon'	NAA	<1100	<28	<170

Table 3: Radioactive isotope levels within various materials and their 68% CL uncertainties. Values for K were not always provided by the analysis.

#	Material	Method	K (10^{-9} g/g)	^{232}Th (10^{-12} g/g)	^{238}U (10^{-12} g/g)
74	FEP, Dupont TM NP20, inner dielectric Axon'	NAA	<3700	<27	<140
75	Axon' Picocoax [®] , Model PCX46P10EP, 0.12 g/m	γ count	22400 $\pm$ 6400	14000 $\pm$ 2000	<28000
76	Axon' Picocoax [®] , same as row 75	ICPMS		< 2×10^6	< 8×10^5
77	Cu wire, bare AWG34, MWS Wire Industries	ICPMS	<320	33.2 $\pm$ 8.9	26.6 $\pm$ 10.1
78	Cu wire, bare AWG40, MWS Wire Industries	ICPMS	<270	38.9 $\pm$ 8.4	14.1 $\pm$ 12.5
79	Mini Coax cable, Cooner Wire Inc.	ICPMS		< 2×10^6	< 8×10^5
80	Mini Coax cable, Cooner Wire Inc.	γ count	11600 $\pm$ 5000	<4000	<21000
81	Handmade Parylene coated Cu wire	ICPMS		815 $\pm$ 81	330 $\pm$ 33
82	Axon' Picocoax [®] , same as row 75	γ count	21000 $\pm$ 2000	<8000	3000 $\pm$ 1000
83	Axon' Picocoax [®] , same as row 75	ICPMS		566 $\pm$ 57	3100 $\pm$ 310
84	Axon' Picocoax [®] , 3 g/m, custom HV cable	ICPMS		470 $\pm$ 110	5900 $\pm$ 300
85	Axon' Picocoax [®] , 0.4 g/m custom LV cable	ICPMS		220 $\pm$ 71	940 $\pm$ 37
86	Axon' Picocoax [®] , 3 g/m, custom HV cable	ICPMS		0.54 $\pm$ 0.05	11.7 $\pm$ 1.2
87	Axon' Picocoax [®] , 0.4 g/m custom LV cable	γ count	<400	<700	<200
88	Axon' Signal Cable (final)	γ count	<2000	<1000	<800
<i>Connectors, front-end electronics, and small parts</i>					
89	Silver epoxy part 1, TRA-DUCT 2902, Tra-Con, Inc.	ICPMS	<200	330 $\pm$ 4	64 $\pm$ 4
90	Silver epoxy part 2, TRA-DUCT 2902, Tra-Con, Inc.	ICPMS	<200	578 $\pm$ 24	349 $\pm$ 12
91	Silver epoxy part 1, TRA-DUCT 2902, Tra-Con, Inc.	ICPMS	107 $\pm$ 4	26.4 $\pm$ 4.4	20 $\pm$ 3
92	Silver epoxy part 2, TRA-DUCT 2902, Tra-Con, Inc.	ICPMS	518 $\pm$ 10	56.7 $\pm$ 7	67 $\pm$ 8
93	Silver epoxy, TRA-DUCT 2902, Tra-Con, Inc.	γ count	<30000	<70000	<10000
94	Silver epoxy packs, TRA-DUCT 2902, Tra-Con, Inc.	γ count	84000 $\pm$ 8000	<8000	<4000
95	Silver epoxy, TRA-DUCT 2902, Tra-Con, Inc.	γ count	<3000	<5000	20000 $\pm$ 3000
96	Silver epoxy hard., TRA-DUCT 2902, Tra-Con, Inc.	ICPMS	556 $\pm$ 11	56.7 $\pm$ 7	67 $\pm$ 8
97	Silver epoxy, TRA-DUCT 2902, Tra-Con, Inc.	ICPMS	<70	79 $\pm$ 4	11 $\pm$ 1

Table 3: Radioactive isotope levels within various materials and their 68% CL uncertainties. Values for K were not always provided by the analysis.

#	Material	Method	K (10^{-9} g/g)	^{232}Th (10^{-12} g/g)	^{238}U (10^{-12} g/g)
98	Silver Epoxy, Emerson-Cumings	γ count	<2000	<3000	<1000
99	SnAg Solder, NCD	GDMS	<73	<70	<200
100	Indium Corp. Solder	GDMS	<120	<100	<300
101	Abietic acid, tech grade, Sigma Aldrich	ICPMS	<108	<90	93 $\pm$ 16
102	Abietic acid, tech grade, Sigma Aldrich	NAA	<28	<28	<60
103	Soap solution, Micro-90 [®]	γ count	42000 $\pm$ 2000	<1500	<600
104	Soap solution, Liquinox [®] Alconox Inc.	γ count	1×10^7	<10000	<2000
105	Fused silica, Corning 7980 ArF	γ count	<90000	<70000	<40000
106	Fused silica, Corning 7980 KrF	ICPMS		187 $\pm$ 26	309 $\pm$ 69
107	Fused silica, Corning 7980 KrF	ICPMS		74 $\pm$ 23	99 $\pm$ 12
108	Fused silica, Corning 7980 std. polish	ICPMS		97 $\pm$ 41	54 $\pm$ 20
109	Fused silica, Corning 8650 ArF	ICPMS		118 $\pm$ 25	525 $\pm$ 134
110	Fused silica, Corning 7980 prototype wafer, uncleaned	ICPMS		11 $\pm$ 5	160 $\pm$ 20
111	Fused silica, Corning 7980 prototype wafer, cleaned	ICPMS		<10	100 $\pm$ 10
112	CFW Al-Si bonding wire	ICPMS	3280 $\pm$ 27	91000 $\pm$ 2000	8700 $\pm$ 400
113	Pins without Beryllium/Copper Contacts	ICPMS	29000 $\pm$ 530	1500 $\pm$ 400	310 $\pm$ 70
114	Pins with Beryllium/Copper Contacts	ICPMS	30000 $\pm$ 510	9900 $\pm$ 300	64000 $\pm$ 100
115	Vespel [®] , in-house machined housing	NAA	<935	310 $\pm$ 140	<190
116	In-house machined female connector, full body assay	ICPMS		96.2 $\pm$ 1.7	43.4 $\pm$ 0.7
117	In-house machined male connector, full body assay	ICPMS		7.1 $\pm$ 0.5	10.5 $\pm$ 0.9
118	Sapphire C-Plane, 0.35 mm thick, Marketech Intern. Inc.	NAA	<376	<21	<300
119	JFET dies, MX-17A, MOKTEK	ICPMS		<1900	<140
120	LMFE1, LMFE prototype without JFET	ICPMS	386 $\pm$ 12	130 $\pm$ 10	430 $\pm$ 43
121	LMFE2a, Full board, internal ID 1B0109	ICPMS	<215	2120 $\pm$ 60	850 $\pm$ 30
122	LMFE2b, Full board, internal ID 1B0110	ICPMS	<215	1610 $\pm$ 30	850 $\pm$ 30

Table 3: Radioactive isotope levels within various materials and their 68% CL uncertainties. Values for K were not always provided by the analysis.

#	Material	Method	K (10^{-9} g/g)	^{232}Th (10^{-12} g/g)	^{238}U (10^{-12} g/g)
<i>Miscellaneous</i>					
123	Precision Urethane Drive Roller (70a durometer)	γ count	<500000	<175000	<65000
124	Wipes, KIMTECH PURE [®] W4, Kimberly-Clark Prof. [®]	γ count	44800 $\pm$ 12800	(1.00 $\pm$ 0.15) $\times 10^6$	<56000
125	Charcoal, 102022, finer size grain, Blücher	ICPMS	20400 $\pm$ 170	369 $\pm$ 30	1870 $\pm$ 70
126	Charcoal, 101135 Saratoga, 0.47 mm, Blücher	ICPMS	11300 $\pm$ 46	181 $\pm$ 18	385 $\pm$ 16
127	Charcoal, 101135 Saratoga, 0.47 mm, Blücher	γ count	10000 $\pm$ 3400	5180 $\pm$ 246	5870 $\pm$ 960
128	Charcoal, UHP granules, Carbo-Act Int.	ICPMS	2880 $\pm$ 12	458 $\pm$ 48	647 $\pm$ 48
129	Charcoal, sample from MPI, Heidelberg	ICPMS	117 $\pm$ 2	135 $\pm$ 20	606 $\pm$ 9
130	Charcoal, K48, Silcarbon	ICPMS	(1.17 $\pm$ 0.01) $\times 10^7$	4070 $\pm$ 150	3990 $\pm$ 80
131	Charcoal, Calgon Carbon	γ count	(6.40 $\pm$ 0.14) $\times 10^5$	(2.610 $\pm$ 0.079) $\times 10^5$	83000 $\pm$ 2700
132	Charcoal, source from Canberra	γ count	(1.50 $\pm$ 0.08) $\times 10^5$	(1.545 $\pm$ 0.055) $\times 10^5$	18800 $\pm$ 1700
133	Hysol [®] 0151 [™] resin, McMaster-Carr	ICPMS	<86	130 $\pm$ 10	17 $\pm$ 5
134	Hysol [®] 0151 [™] hardener, McMaster-Carr	ICPMS	<215	140 $\pm$ 20	39 $\pm$ 7
135	Hysol [®] 1C [™] resin, McMaster-Carr	ICPMS	6200 $\pm$ 200	3190 $\pm$ 290	86300 $\pm$ 1100
136	Hysol [®] 1C [™] hardener, McMaster-Carr	ICPMS	26670 $\pm$ 260	79200 $\pm$ 800	1560 $\pm$ 1000
137	Torr Seal [®] Base, McMaster-Carr A	ICPMS	<7530	2350 $\pm$ 480	10300 $\pm$ 2000
138	Torr Seal [®] Hardener B, McMaster-Carr	ICPMS	35400 $\pm$ 250	45200 $\pm$ 700	148000 $\pm$ 2000
139	Silicone Rubber, P-4, Silicones Inc.	γ count	<2000	<3000	<6000
140	2-propanol, A461-500, Fischer Scientific	ICPMS	<10	<0.1	<0.1
141	Mix colored beads, 100780, Accu-Glass Prod. Inc.	ICPMS	(2.63 $\pm$ 0.01) $\times 10^7$	39500 $\pm$ 2000	453000 $\pm$ 6300
142	White beads, 100780, Accu-Glass Products Inc.	ICPMS	(2.82 $\pm$ 0.01) $\times 10^7$	40200 $\pm$ 1900	713000 $\pm$ 12000
143	Green beads, 100780, Accu-Glass Products Inc.	ICPMS	(3.05 $\pm$ 0.01) $\times 10^7$	(2.23 $\pm$ 0.08) $\times 10^5$	386000 $\pm$ 8000
144	Black bead leachate	ICPMS	1890 $\pm$ 10	57.2 $\pm$ 2.2	43 $\pm$ 2
145	Blue bead leachate	ICPMS	7850 $\pm$ 41	150 $\pm$ 3	259 $\pm$ 4
146	Brown bead leachate	ICPMS	2030 $\pm$ 9	64.7 $\pm$ 3	46.5 $\pm$ 1.7
147	Green bead leachate	ICPMS	3820 $\pm$ 20	261 $\pm$ 4	195 $\pm$ 5

Table 3: Radioactive isotope levels within various materials and their 68% CL uncertainties. Values for K were not always provided by the analysis.

#	Material	Method	K (10^{-9} g/g)	^{232}Th (10^{-12} g/g)	^{238}U (10^{-12} g/g)
148	Grey bead leachate	ICPMS	2120 ± 16	281 ± 6	199 ± 3
149	White bead leachate	ICPMS	2460 ± 12	67 ± 2	59 ± 1

4.1. Underground Electroformed Copper

The DEMONSTRATOR is constructed with a large quantity of copper used in the cryostat and inner shield. As a result, it has a significant impact on the background model and requires special consideration. Electroforming Cu is effective at removing impurities [34, 36, 37]. Electroforming and machining the Cu underground prevents the ingrowth of cosmogenically produced isotopes. While we were able to locate one lot of commercial copper that was adequately pure to obtain non-detect values using the ICPMS assay, this was premachined, heavily-etched, bulk material. Our search for such clean commercial copper underscores the wide range of U and Th contamination found in such materials and how critical any subsequent handling is for maintaining its purity.

Early on, the MAJORANA collaboration determined copper to be the most desirable material for detector string parts and cryostat construction due to its favorable thermal, electrical, and mechanical properties [38, 39, 40]. Although it can be obtained commercially with relatively high purity, it is contaminated by a variety of means. During the production and subsequent handling processes contamination with U and Th can occur. Copper also suffers from activation due to reactions with cosmic-ray produced secondary neutrons, e.g., $^{63}\text{Cu}(n,\alpha)^{60}\text{Co}$. ^{60}Co is long-lived (half-life = 5.3 yr) and attempting to limit this activation presents its own challenges. These considerations motivate underground electroforming of the MAJORANA DEMONSTRATOR copper, to not only produce Cu below the target purity level of 0.3 $\mu\text{Bq/kg}$ for ^{238}U and ^{232}Th , but also to minimize the cosmic ray exposure.

Copper for the MAJORANA DEMONSTRATOR was electroformed in two underground locations. Seven electrochemical baths operated at the Shallow Underground Laboratory (SUL) at PNNL with an overburden of ~ 40 feet. Ten baths operated in a dedicated clean room facility constructed at the 4850 ft level at SURF. Construction of the electroforming baths used parts selected for their long-term compatibility with sulfuric acid and copper sulfate solutions. Since the bath components were one potential source of radionuclide contaminants, background mitigation required careful design and construction. Excluding the copper anode material, the only materials permitted to contact the bath solution were acid-leached polymers. The bath materials include polypropylene, high-density polyethylene, and polytetrafluoroethylene. All of these materials underwent a rigorous cleaning and high-purity acid-leaching process to remove contaminants [41]. These techniques required the use of greater than 99.995% pure oxygen free high conductivity (OFHC) copper anodes, which are commercially available. We purchased ours from Titan International [42]. The bath electrolyte was produced using $>18\text{ M}\Omega\text{-cm}$ deionized water and OptimaTM grade sulfuric acid. The copper sulfate was produced in-house using pure acid and copper in order to achieve the desired purity because no vendor for such high purity material could be found. The gas volume over the bath solution was purged with liquid nitrogen boil-off gas to mitigate radon intrusion.

MAJORANA used potential limited, reverse-pulse electroplating techniques. To maximize the control over the electrical parameters such as potential and

371 waveform, the electroforming power supplies used in our process were designed
372 under the Department of Energy’s Small Business Innovation Research program
373 with Dynatronix, Inc.. This in turn influences purity and the material’s physical
374 properties. The copper purity required production rooms operated in a clean
375 room environment class 1000 or better.

376 The material can be produced centimeters thick and used as-is. Its mechan-
377 ical properties [38, 39, 40] are adequate and similar to commercially available
378 copper without requiring additional metallurgical treatment. It thus avoids the
379 introduction of contamination typical of further industrial processing.

380 In the absence of a non-destructive direct assay of electroformed copper
381 with adequate sensitivity for measuring Th and U to the required level, the
382 assay of the bath electrolyte was used to infer the material purity. The con-
383 centration of contaminant species such as U and Th in the electrolyte can be
384 used to predict their concentration in the deposited copper. Data obtained pre-
385 viously [36, 37] indicated that the concentration of U and Th as contaminants
386 deposited in electroformed copper are about 1000 pg/g when the concentration
387 in the electrolyte solution was 1000 ng/ml. Data [34], obtained later at a much
388 lower concentration very near the detection limit indicated a comparable al-
389 beit decreased contamination-rejection rate. From these data we determined
390 maximum allowable electrolyte U and Th concentrations for DEMONSTRATOR—
391 copper of 0.024 ng/ml and 0.075 ng/ml, respectively. Electrolyte contamination
392 below these concentration levels indicates that electroformed copper is being
393 produced with impurity levels below 0.3 $\mu\text{Bq/kg}$ and thereby surpassing the
394 DEMONSTRATOR radiopurity goals for U and Th in copper. The direct assay
395 results of electroformed Cu samples support this conclusion.

396 To ensure a stronger rejection rate for the DEMONSTRATOR-produced elec-
397 troformed copper than that measured in Ref. [34], much greater care in the
398 control and monitoring of the electroforming processes was implemented. The
399 electroforming baths used for the MAJORANA DEMONSTRATOR, given their large
400 bulk volume, were also expected to achieve a better rejection rate than the small
401 baths used previously. Contaminants were measured in the electrolyte before
402 electroforming began and then at intervals during deposition. For example,
403 early concentrations in the baths at PNNL were an average of 0.006 ng/ml U
404 and 0.022 ng/ml Th. After nearly two months of electroforming, the levels had
405 increased to 0.010 ng/ml U and 0.040 ng/ml Th due to the contributions of the
406 dissolving copper anode material. These increased levels were still below their
407 maximum allowable levels. The change is proportionately the same for both U
408 and Th, which supports the argument that the source of contamination is from
409 the dissolution of anode copper and not some other outside source. Since the
410 change is also proportional to the anode stock consumption, it indicates a strong
411 rejection of U and Th. The concentrations of U and Th increasing in the baths
412 were also within expected levels. This verifies that sufficiently high-purity acids
413 were used for the periodic replenishment of the bath electrolyte as required due
414 to bulk chemistry and trace constituent limitations.

415 The latest assay results for electroformed copper, presented in Table 3, vali-
416 date the approach of monitoring the electrolyte contaminant levels to determine

the Cu contamination. The new results from this work and Ref. [35] are presented in Table 3 in Rows 1 and 2 where Row 3 is an early assay of that material. Rows 4 through 7 represent bulk samples from electroformed Cu used within the apparatus. These samples underwent a heavy etch with little or no additional handling. The samples listed in Rows 8 through 14 represent parts produced for the DEMONSTRATOR but sacrificed for assay. These were typical parts from the production line prepared as all parts were. These parts were machined and then etched. The additional handling resulted in a higher level of contamination than in the raw stock material. The numbers are listed in the table as bulk contaminations, but are more consistent with surface contamination at the level of $2.8 \mu\text{Bq}/\text{m}^2$ and $1.8 \mu\text{Bq}/\text{m}^2$ for ^{238}U and ^{232}Th , respectively. These values were estimated from analyzing samples from various etch depths. Heavy etching results in the ultimate purity level, but part tolerance must be considered. This situation is an area of future research.

4.2. Commercially Available Copper

Commercially available copper is used as shielding material in the DEMONSTRATOR outside of the electroformed Cu shield. The majority of the commercial Cu is in the form of plates machined to a thickness of 2" as the outer Cu shield between the electroformed inner Cu shield and the Pb shield. Additionally, commercial Cu parts are in other locations where the detector module interfaces with the shield, such as a support stand to support the detector module within the Pb shield. It is also used as filler shielding throughout the penetrations of the Pb shield.

Several assays were performed during the selection and acquisition of the commercial Cu. With one exception, all commercial Cu assays represent Cu stock purchased through Southern Copper & Supply Company (USA), which sourced the Cu plate material from KME (Europe) where it is rolled to the desired plate thickness. The original Cu cake material is supplied to KME by either Aurubis or Mitsubishi Materials.

Samples of the starting cake material (prior to rolling) originating from Aurubis and the final 2.5" rolled plate were assayed prior to the purchase of the final material. Further, separate samples of the rolled plate were collected on the exterior rolled surface (a potential source of contamination) and the interior of the plate. The results of the pre-rolled cake sample (see Row 15 of Table 3) and the plate samples (see Rows 16 and 17 of Table 3) show a difference of a factor of 2, which we consider to be of no significance and likely due to sampling issues. In any case, it is an acceptable radiopurity for the outer Cu shield.

The detector module is supported within a removable portion of the Pb shield with a Cu support stand made from 1"x1" and 1"x2" Cu bars. The stock material was again plate stock cut to the desired width by the vendor. Once underground in our machine shop, the rough cut bar stock underwent a surface machining on all surfaces and was then cut to length. An initial assay of the 1" plate stock originating from Mitsubishi Materials (See Row 18) prior to our own surface machining showed undesirable purity levels, though it was noted that surface preparations for the ICPMS assay were not sufficient to remove

the evidence of saw cuts on the surface of the sample. A repeat assay of the same stock (Row 19 of Table 3) with the production machined surface did show improvement in the radio purity levels. A machined-surface sample of 1"x2" bar stock from Copper and Brass Sales (Row 20 of Table 3) showed similar radio purity levels.

In an attempt to identify cleaner material for the Cu support structure and noting the success of the acceptable purity found for the outer Cu shield 2.5" plate, a second round of plate stock was evaluated. Since the final 1"x2" portion of the Cu support stand has a more stringent radiopurity requirement, we evaluated a sample of 2.5" plate cut to a 1.25" width, finding much improved radiopurity (Row 22 of Table 3) without any pre-assay surface machining. In addition, a second 1" plate sample originating from Mitsubishi Materials without any pre-assay surface machining (Row 21 of Table 3) also shows improved radio purity. The two latter samples represent the Cu stock used for the Cu support stand.

Two remaining Cu sources are evaluated for use where the detector module interfaces with the shield. A sample of 2.5" plate stock cut to 2.5" width originating from Arubis (Table 3, Row 23) is used for 2"x4" Cu bricks around a cylindrical crossarm through the Pb shield. A sample of 0.5" plate stock originating from Mitsubishi (Row 24) is used as a shielding seal around the cylindrical crossarm.

4.3. Lead

The most abundant material, by mass, used in the DEMONSTRATOR is lead, which is the dominant component of the shield. Because it is so massive, it is not possible to directly count all the materials used for the shield. Therefore, we relied on sample testing and indirect assays to assess the lead's purity.

Two sources of Pb bricks (nominally $2" \times 4" \times 8"$) were used for the DEMONSTRATOR shield. One was a new production run from virgin Doe Run Mine Pb formed into Pb bricks by Sullivan Metals, Inc. (Holly Springs, MS) and the other was from a discontinued, low-background counting facility at the University of Washington. Because different sources were used, the history of the UW Pb was not entirely known, and because direct assay of all individual bricks was not feasible, we developed a process to verify the purity of the Pb bricks. Initially, direct assay by γ -ray counting and GDMS at the NRCC (Section 3.2.1) of a small number of Pb bricks from both sources provided confidence of the Pb radiopurity. (See Rows 26 to 34 in Table 3.)

To better understand the uniformity of the purity of the Pb, we took advantage of the fact that all the bricks were cleaned prior to delivery to our underground laboratory at a collaboration facility established at Black Hills State University. We directly sampled a selection of bricks during that cleaning process and assayed those samples using ICPMS at PNNL (Section 3.2.2). The result is given in Row 35 in Table 3.

The bricks were cleaned in a cleanroom by first soaking them in groups of 5 in a pure ACS-grade acetic acid bath for 3-13 minutes depending on the brick

condition. The etch bath was replaced when its effectiveness in cleaning the bricks was noticed to decline as evidenced by the average cleaning time. During this soak the bricks were scrubbed with soft plastic brushes for 1-2 minutes. Following the soak, the bricks were rinsed in successive $>18\text{-M}\Omega\text{-cm}$ deionized water baths for 20-30 seconds each. Water rinse baths were replaced when the rinse water appeared dirty. Next they were transferred to a bath of nitric acid and peroxide (1-3% Optima™ grade acid used once previously for leaching plastics, and 3% non-stabilized ACS grade peroxide) for 1-2 minutes followed by 2 successive 20-30 second deionized water rinses. The nitric-peroxide bath was replaced whenever it was observed that bubble formation on the bricks had diminished or that bricks were not emerging from the bath with the appropriate silvery finish. The bricks were then patted dry with cleanroom-grade wipes to remove excess water, followed by a rinse with isopropyl alcohol for 1-2 minutes. The bricks were then air dried before being triple-bagged in 6-mil polyethylene. Finally, the bricks were palletized for transfer to the DEMONSTRATOR clean room underground. The acetic acid etch typically removes about $50\text{ }\mu\text{m}$ of material but can remove up to $250\text{ }\mu\text{m}$. The nitric solution etch removes about $13\text{ }\mu\text{m}$. In total over 6800 bricks were cleaned at a rate of 50/d. The DEMONSTRATOR shield contains approximately 4500 bricks.

Bulk samples from 40 bricks, 20 from each source, were collected after the nitric-peroxide etch. A gouging tool was used to produce these samples with a mass of approximately 5 g each. These bricks were set aside for storage and not used in the shield. The average of the ICPMS measurements obtained at PNNL indicated contamination levels of 1.3 ppt and 2.9 ppt for Th and U respectively. These results meet our requirements for the Pb and are listed in Row 35 in Table 3.

4.4. Parylene

Parylene generally refers to a family of chemical vapor-deposited polymers used in industry as moisture and dielectric barriers. The MAJORANA DEMONSTRATOR makes use of parylene for multiple purposes. All hardware used to construct the inner regions of the cryostats is made from ultra-pure electroformed Cu. Because of the radiopurity requirements of the experiment, the use of oil-based lubricants is not possible. So parylene is used as a dry-film lubricant for copper nuts and bolts to prevent galling. Parylene, in the form of thin sheets, is also used as an insulator for electrical isolation of components of the copper shield and copper cryostat. Assays of parylene products from different manufacturers were performed.

Parts were coated with parylene using the PDS 2010 LABCOTER® 2 Parylene Deposition System which was manufactured by Specialty Coating Systems™ in Indianapolis, IN. See Figure 3. The deposition chamber was operated in a cleanroom located in a surface laboratory at SURF. Because of the high elevation (~ 1 mile) on the surface, parts were coated as quickly as possible and returned underground to minimize cosmic ray exposure. Prior to coating, all surfaces of the deposition chamber were wiped with a 2% solution of Micro-90® that acts as a release agent to ease later cleanup.

Figure 3: Parylene coating of copper nuts. Left: the parylene coater. Right: parts on the deposition stand.

551 The results of our parylene assays are listed in Table 3 in Rows 57 through
 552 Row 63. Mirco-90[®] release agent is listed in Row 103. We also assayed parylene
 553 film samples from three locations within the deposition chamber as indicated
 554 in Rows 61 through 63. The observed activities were already low enough for
 555 MAJORANA DEMONSTRATOR components so further investigation of possible
 556 sources of contamination was not pursued.

557 4.5. Signal and High-Voltage Cables

558 The signal and the high voltage cables used inside the MAJORANA DEMON-
 559 STRATOR cryostat have very stringent radiopurity requirements as they are in
 560 close proximity to the Ge detectors. For each detector, four coaxial signal cab-
 561 les are needed for connections to the Low Mass Front End (LMFE), and one
 562 coaxial cable is required for carrying high voltage to the n^+ contact. Axon'
 563 Cable SAS was contracted to manufacture custom miniature Picocoax[®] cables
 564 for these purposes. The inner dielectric and the outer jacket of the signal (Axon'
 565 part number TD11153A) and the high-voltage (Axon' part number TD11153B)
 566 cables were extruded from two different stocks of fluorinated ethylene propylene
 567 (FEP) — DuPont[™] FEP TE9494 and Daikin Neoflon[™] NP-20. California Fine
 568 Wire Company extruded the central AWG34.5 and AWG40.5 copper conduc-
 569 tors in the cables from the same copper stock (CFW heat number 31465), while
 570 Axon' Cable extruded the AWG50 copper wires (Axon' batch number MC5092)
 571 that form the helical ground shield of the cables. Assays of the initial cable pro-
 572 duction runs did not meet the radiopurity requirements. Several manufacturing
 573 steps were identified that were adding unwanted contaminants. These steps
 574 were modified and cleanliness controls were implemented throughout the man-
 575 ufacturing process that resulted in an acceptably radiopure product (Table 3,
 576 Rows 86 through 88.) The details of these processes are proprietary.

577 All the raw materials were individually assayed by ICPMS at PNNL for their
578 radiopurity prior to cable production. After each production step, the partially
579 complete product was washed in solvent in an ultrasonic bath and then baked
580 in a clean oven located in a cleanroom. Witness samples were collected from
581 each of these steps as quality control. The unfinished products were stored
582 in clean nylon bags in a cleanroom when not being processed. The finished
583 cables were wiped with clean isopropyl alcohol, and placed in sealed nylon bags
584 that were filled with dry nitrogen prior to delivery by Axon'. The cables were
585 stored in an ambient dry nitrogen environment subsequent to their receipt by
586 the MAJORANA collaboration.

587 Segments from different parts of the finished cables were assayed by ICPMS.
588 One spool of each cable, enough for the wirings in the MAJORANA DEMONSTRATOR,
589 was also individually assayed by γ -ray counting. The resulting assays are
590 listed in Table 3 in Rows 73 through 88. Parylene coated wire and Mini Coax
591 (Rows 79 through 81 in Table 3) were not used for the DEMONSTRATOR. The
592 Cu wire used for the parylene coated wire was also from California Fine Wire
593 (Row 25 in Table 3).

594 4.6. Connectors

595 In order to complete installation of the MAJORANA DEMONSTRATOR detec-
596 tors in a clean nitrogen-purged environment to avoid exposure of the germanium
597 crystals to trace radon levels in laboratory air, connections for the four coaxial
598 signal cables must be made within the cryostat. A low-mass, low-radioactivity
599 connector was needed to connect each LMFE's cables to a signal cable bundle
600 that extends outside the shielded space.

601 Commercially available connectors were considered, but all options investi-
602 gated at the time used beryllium-copper (BeCu) contact springs. Measurements
603 of BeCu available in the literature give U and Th specific activity levels on the
604 order of Bq/kg [18, 43] and do not satisfy the MAJORANA DEMONSTRATOR
605 radiopurity requirements. Connector pin receptacles intended for use in the
606 MAJORANA DEMONSTRATOR containing miniature BeCu contacts were assayed
607 via ICPMS at Validation Resources, Inc. and were found to have orders-of-
608 magnitude higher specific activity than the same pin receptacles without the
609 BeCu contacts (Table 3, Rows 113 and 114). As seen in Table 4, these parts
610 alone, if used, would exceed the entire MAJORANA DEMONSTRATOR background
611 budget.

612 A connector, shown in Fig. 4, was designed that uses custom Mill-Max[®]
613 gold-plated brass pin receptacles without BeCu contact springs. The female
614 socket is Mill-Max[®] pin receptacle 8210-0-00-15-00-00-03-0, while the male pin
615 uses the post of Mill-Max[®] pin receptacle 0461-2-00-15-00-00-03-0, whose outer
616 diameter matches the inner diameter of the female socket. The pins and sockets
617 are slightly misaligned radially, forcing the pin to flex; the restoring force of the
618 pin takes the role of a contact spring in maintaining the electrical connection.
619 The pins and sockets are held in Vespel[®] SP-1 (polyimide) housings, which are
620 precision-machined at the underground clean machine shop at SURF. The pins,
621 sockets, and Vespel[®] housings were ultrasonically cleaned in 1% Micro-90[®]

Figure 4: A low-radioactivity signal connector designed for the MAJORANA DEMONSTRATOR. Mill-Max[®] pin receptacles are encased in a Vespel[®] housing, clean-soldered, and strain-relieved with clean FEP heat shrink tubing. Left: MAJORANA DEMONSTRATOR 8-pin signal connector juxtaposed with a male BNC connector. Right: Male plug with end cap removed to show soldering joints and strain relief

622 soap solution, $>18\text{-M}\Omega\text{-cm}$ deionized water, and then Optima[™] grade ethanol.
 623 The Vespel[®] parts were subsequently leached in 6M Optima[™] grade nitric acid
 624 solution for 72 hours, and then soaked for 24 hours in deionized water to remove
 625 absorbed nitric acid. They were then pumped and baked in a vacuum chamber
 626 at 100°C to outgas any remaining liquid, particularly nitric acid which can
 627 corrode the brass components.

628 Following this treatment, the Vespel[®] parts were stored in a nitrogen-purged
 629 environment to prevent radon adsorption. The Axon[®] coaxial cables were sol-
 630 dered to the connector pins and sockets with a 96.5:3.5 eutectic SnAg alloy
 631 solder used in the construction of the Sudbury Neutrino Observatory (SNO)
 632 Neutral Current Detectors (NCDs) [44]. Purified abietic acid (Rows 101 and
 633 102 in Table 3) dissolved in isopropyl alcohol (Row 140) was used for flux.
 634 This process was performed in a cleanroom under a snorkel to capture solder-
 635 ing fumes. Fluorinated Ethylene-Propylene (FEP) heat-shrink tubing (Rows 69
 636 through 71) provided strain relief for the solder joints. To limit exposure of the
 637 LMFE components to laboratory air and solder fumes, the LMFE was attached
 638 to the cables after soldering was complete.

639 Individual components were assayed to estimate backgrounds and the results
 640 are shown in Table 4. Assays of the Vespel[®] and FEP were performed using
 641 NAA (Section 3.3). Assay of pin receptacles, with and without BeCu contact
 642 springs and with parts as-is from the manufacturer (no additional cleaning),
 643 was performed by Validation Resource, Inc in Bend, OR. (Section 3.2). The
 644 results are given in Table 3 in Rows 113 and 114. Particular attention was
 645 given to assaying the solder and related processes. The SnAg eutectic solder
 646 from the SNO experiment was re-assayed by the NRCC using GDMS (Row 99 in
 647 Table 3) along with another SnAg eutectic solder newly acquired from Indium
 648 Corporation[®] listed in Row 100. Our result for the NCD solder is very similar to
 649 the previous measurement. The solder was melted into a rod shape along with
 650 the pure abietic acid and isopropyl alcohol flux before being assayed. Separate
 651 solder samples were prepared using both a new soldering iron tip and a tip that
 652 was scraped with a brass scraping pad and dipped in DeoxIT[®] Tip Tinner &
 653 Cleaner, a mixture of tin and ammonium phosphate used to prevent oxidation

of the soldering iron tip. All samples yielded upper limits of <0.1 - 0.3 ppb for both U and Th. There was no significant difference observed between the samples, including the ones prepared with the soldering iron tip that was scraped and dipped in the tip cleaner. The SNO solder is being used to assemble the MAJORANA DEMONSTRATOR signal bundles.

Table 4: Summary of contaminations and background contributions for signal connector materials. Component masses reported are for one male-female pair. The column labeled Background gives the MAJORANA DEMONSTRATOR anticipated background level in the $\beta\beta(0\nu)$ ROI for the measured impurity level assuming the decay chains are in equilibrium. We assume one connector pair for each of 60 detectors in the DEMONSTRATOR. The row labeled Total indicates the specific activity expected for a connector pair based on the assays and masses of the individual components. Mass estimates for the solder, solder flux, and shrink tube are uncertain but conservative

Material	Assay Method	Mass [g]	Isotope	Specific Activity [$\mu\text{Bq/kg}$]	Background [c/(ROI t yr)]
Pins (w/BeCu)	ICPMS	0.112	^{238}U	795000 ± 12000	8.8 ± 0.1
			^{232}Th	41000 ± 1000	2.3 ± 0.1
Pins (no BeCu)	ICPMS	0.112	^{238}U	4600 ± 1500	0.05 ± 0.02
			^{232}Th	5800 ± 100	0.32 ± 0.01
Vespel [®] SP-1	NAA	0.95	^{238}U	<1000	<0.20
			^{232}Th	<12	<0.01
SnAg Solder	GDMS	0.04	^{238}U	5600 ± 1000	0.02 ± 0.004
			^{232}Th	<12	<0.0002
Solder flux	GDMS	0.04	^{238}U	1200 ± 200	0.005 ± 0.001
			^{232}Th	<400	<0.007
shrink tube	NAA	0.01	^{238}U	<1250	<0.012
			^{232}Th	<138	<0.007
Total		1.5	^{238}U	<1500	<0.28
			^{232}Th	<600	<0.36

659

One each of the completed male and female connectors underwent full-body ICPMS assay to evaluate the cleanliness of the assembly process. The cables were cut from the point at which they enter the end cap housing, and the connectors were transferred to validated quartz crucibles where they were spiked with ^{229}Th and ^{233}U tracers. The samples were heated to 700°C for four hours to ash away polymeric materials, and then digested in a combination of Optima[™] grade nitric acid, hydrochloric acid, and hydrogen peroxide solutions as described in Ref. [41]. A small amount of material was not dissolved (see Table 5), most likely due to the formation of an insoluble metal oxide precipitate. It is unlikely that the remaining solids contained significant quantities of the analytes, as all solids were dissolved in solution at some point during processing. Since the solids formed during the digestion process, the tracer isotopes account for any analyte adsorption or incorporation into the precipitate.

After the dissolved samples and precipitate materials were brought to dryness and reconstituted in 8 M nitric acid, the analytes were separated from the sample matrix solution using anion exchange resin. ICPMS was performed as described in Section 3.2 following the procedure outlined in Ref. [35]. The results are given in Table 5 and in Rows 115 through 117 in Table 3. The process blank for the female plug was lost during sample preparation, but given similar digestion steps and handling in the cleanroom, the process blank generated from the analysis of the male plug provided a good estimate of the levels that would have been seen in the female plug blank. The results are listed in Table 3 in Rows 116 and 117. Other results related to the connectors are listed in Rows 101, 102, 103, 104, and 113 through 117 in Table 3.

Table 5: Summary of specific activities and corresponding background contributions from full-body ICPMS of two manufactured plugs, one male and one female. The male plug has a larger mass because it uses larger brass pins. The column labeled Background gives the MAJORANA DEMONSTRATOR anticipated background level in the $\beta\beta(0\nu)$ ROI for the measured impurity level.

Material	Mass (g)	Post-Ash Mass (g)	Undigested Mass (g)	Isotope	Specific Activity ($\mu\text{Bq/kg}$)	Background (c/(ROI t yr))
Female plug	0.453	0.068	0.004	^{238}U	1160 ± 17	0.052
				^{232}Th	365 ± 6	0.083
Male plug	0.600	0.183	0.035	^{238}U	281 ± 24	0.017
				^{232}Th	27 ± 2	0.008
Connector totals	1.05			^{238}U	1160 ± 17	0.110
				^{232}Th	365 ± 6	0.174

684

4.7. The Low-Mass Front End Electronic Board

The readout electronics for the MAJORANA DEMONSTRATOR were developed with the twin goals of minimizing electronic noise and minimizing radioactivity. In pursuit of the former, the primary stage of amplification was placed close to the output of each detector. In the pursuit of the latter, everything but the primary stage was located outside of the cryostat and those components inside were made as small and as radiopure as possible. The primary amplification stage, the LMFE, is shown in Fig. 5 [45]. It comprises a fused silica substrate on which a photolithographic pattern was used to form conductive traces and a resistor. A JFET was affixed to a gate pad using silver epoxy, and to the drain and source traces. Ultrasonically-drilled holes provide strain relief for the cables, which were bonded to the traces using silver epoxy. The LMFE has a surface area of 145 mm^2 and a mass of 80 mg.

The task of controlling the radiopurity of the LMFE presented a number of challenges. The important ones arose from the small size of the LMFE and the novelty of its design. Specifically, determining the level of radioactive contamination in tiny components and controlling exposure to radioactive contamination during a complex production process were key issues. ICPMS was used to

Figure 5: A photograph of an LMFE with wires and JFET attached. The cables were not part of the LMFE assay measurements.

measure the radiopurity of all board component materials and the final boards themselves. ICPMS analysis of all individual components of the LMFE and of three full boards were performed at LBNL, PNNL and Chernogolovka as described in Section 3.2.2.

Table 6 shows the relative mass fractions of each component of the LMFE board. The substrate accounts for the overwhelming majority of the LMFE mass. It was made from fused silica, a synthetic glass available in extremely high purity due to its use in laser applications. The fused silica was procured in 2-inch diameter, 0.25-mm thick, highly polished wafers, ultrasonically drilled and then thoroughly cleaned before photolithography. Ultrasonic cleaning in $>18\text{-M}\Omega\text{-cm}$ deionized water and 10% nitric acid was demonstrated to decrease levels of U and Th in the wafers. A number of brands from a variety of manufacturers were tested and those produced by Corning Inc. were found to be the most promising. Measurements indicated that levels of U and Th decreased with an increase in material grade, cost, and with the distance from the bottom of the manufacturer's original boule. Cost considerations prevented the use of the highest grade material or controlling the position of the material in the boule but, nevertheless, a batch of Corning HPFS[®] 7980 standard grade fused silica was obtained with adequate purity. See Rows 105 through 109 in Table 3.

The electrical traces on the board and the resistor were sputtered and photolithographically patterned. The traces were 400 nm of gold on top of 20 nm of titanium acting as a bonding layer. The amorphous germanium resistor was 400 nm thick. These films were thin enough that direct ICPMS was impractical. Sufficient masses could only be acquired from longer, dedicated runs in which films $\sim 1\text{-}10$ microns thick could be deposited. The substrates for these depositions were the same fused silica wafers used for LMFE production and they were initially coated with Micro-90[®] soap solution as a release agent that allowed the deposited films to be floated from the wafer by slowly submerging them in water. The Micro-90[®] soap had been assayed to ensure adequate radiopurity

(Row 103 in Table 3). This procedure was effective for gold and titanium but could not be used for thick films of germanium because they readily lose their integrity with a tendency to flake. For germanium, the metal deposited onto wafers was etched, rather than floated, from the surface. The acids used during etching did not affect the fused silica substrates and the masses of metal removed were determined by before- and after- measurements of the wafer mass. Film-production assay results are listed in Rows 42 through 51 of Table 3.

Silver epoxy (TRACON TRA-DUCT BA2902) was counted using both ICPMS and γ -ray counting, with results from both of these methods found to be consistent. Due to its short shelf-life multiple batches of epoxy had to be purchased and counted during the construction of the experiment. The results of samples from the various batches are presented in Table 3. All batches have been found to be sufficiently pure though significant variations in purity have been observed. Epoxy results from several batches are listed in Rows 89 through 98 in Table 3. Part 2 of the epoxy is the hardener. We counted some epoxy in its packing because of its short shelf life and because we wanted to use assayed epoxy.

Bonding wire (99% aluminum, 1% silicon manufactured by California Fine Wire) was counted using ICPMS (Row 112 of Table 3). The cost to acquire enough JFET dies that are used in the LMFE (Moxtek Inc. MX-11) for ICPMS assay was not justified. A substantial batch of dysfunctional MX-17A JFET dies, which have a similar construct as the MX-11 JFET dies, was acquired and assayed using ICPMS (Table 3, Row 119). Because it was difficult to obtain sufficient mass of either of these components, the measurements presented are only relatively weak limits.

The second stage of the assay program was the assay of fabricated LMFEs using ICPMS. This was done as a way of accounting for contamination arising during the manufacture process and by contact of the boards with cutting devices or the variety of chemicals during photolithography. Three boards have been counted: one board (LMFE1, Row 120 in Table 3) consisting of all components except the silver epoxy, JFET and bonding wires; and two complete boards (LMFE2a/b, Rows 121 and 122 in Table 3). Boards were dissolved in a microwave reaction system using an ultra high-purity solution of hydrofluoric and nitric acids. The use of hydrochloric acid was avoided to eliminate the isobaric interference caused by the molecular ion $^{197}\text{Au}^{35}\text{Cl}^+$. Results are given in Table 6 and indicate additional contamination from the production process over what was expected from the assays of the raw materials.

4.8. TIG Welding

Early designs of the DEMONSTRATOR considered using tungsten inert gas welding for certain copper joints. Although in the end, this technique was not used, we include our study of its potential for radioactive contamination for completeness. A comparative study was made of the characteristics of zirconiated tungsten cathodes (W+1 wt% ZrO_2) and ceriated tungsten electrodes (W+2 wt% CeO_2). Eight electrodes with a radius of 0.12 cm, and lengths varying from 3.33 to 5.7 cm were used. The TIG(Zr) rods (26.5 g) were γ -ray counted

Table 6: Summary of purities and activities for LMFE components and manufactured LMFE boards. The row labeled Total is the predicted activity of a completed LMFE based on component activity and mass. Activities of completed boards are elevated compared to what one would expect from the individual components. The excess is assumed to arise from contamination during the assembly process and assembly was moved to a high-quality clean room after these results were obtained. In the first column the numbers in parenthesis refer to rows in Table 3.

Item	Mass fraction (%)	Concentration (ppb)		Activity (nBq)	
		^{232}Th	^{238}U	^{232}Th	^{238}U
Fused silica (111)	97.24	< 0.01	0.10 $\pm$ 0.01	< 3.2	99 $\pm$ 10
Amorph. Ge (50)	0.0029	2.4 $\pm$ 0.7	1.7 $\pm$ 0.4	0.02 $\pm$ 0.01	0.05 $\pm$ 0.01
Au traces (44)	0.114	47 $\pm$ 1	2.0 $\pm$ 0.3	17.9 $\pm$ 0.4	2.3 $\pm$ 0.3
Ti traces (51)	0.0013	< 0.4	< 0.1	< 0.002	< 0.001
JFET die (119)	1.17	< 2.0	< 0.14	< 7.8	< 1.7
Al-Si wire (112)	0.0049	91 $\pm$ 2	8.7 $\pm$ 0.4	< 1.50 $\pm$ 0.03	0.44 $\pm$ 0.02
Silver epoxy (95)	1.46	<5	20 $\pm$ 3	< 24	300 $\pm$ 45
Total	100.			< $\sim$ 55	< $\sim$ 402
LMFE1 (120)		0.13 $\pm$ 0.01	0.43 $\pm$ 0.04	42 $\pm$ 3	44 $\pm$ 4
LMFE2a (121)		2.12 $\pm$ 0.06	0.85 $\pm$ 0.03	706 $\pm$ 2	87 $\pm$ 3
LMFE2b (122)		1.61 $\pm$ 0.03	0.85 $\pm$ 0.03	536 $\pm$ 10	87 $\pm$ 3

for 57 days. The γ -ray count for the TIG(Ce) rods (25.3 g) was only 0.64 days due to the high activity. Most of the activity appears to be Th contamination, indicating that these rods contain approximately 0.1% ThO₂. The assay results are listed in Table 3 in Rows 40 and 41.

Once the assay results were completed, erosion measurements were made of the two different types of cylindrical cathodes. The cathodes were used on a dc-transferred torch operating at atmospheric pressure. New gas cups and lenses were used on the torch body. An ultra-high purity He (99%) gas was used as the shielding gas. Erosion measurements were taken with each cathode at current intensities ranging from 100-130 Amps, with an arc voltage of 18-19 V. The results of the erosion tests for the the TIG(Zr) and TIG(Ce) tips show that the mass loss was near the uncertainty in our balance. Using the mass measurements, upper limits on the average erosion rate for the TIG(Zr) and TIG(Ce) tips were determined to be 0.280 mg/cm and 0.568 mg/cm, respectively. The TIG tips were examined and there did not appear to be any spattering from the welds that could result in an increase in weight.

In order to study the effects of any weld contamination on the DEMONSTRATOR, a simulation of a small section of the Cu tubing was performed. This tubing was considered for use as calibration-source access to the DEMONSTRATOR. The simulated weld joint was 1.57 cm. Within the weld, we would expect to deposit less than 0.4 mg of the TIG(Zr) electrode, and less than 0.9 mg of the TIG(Ce) electrode. Based on the simulations, the experimental background was predicted to be less than 0.024 and 6.96 c/(ROI t yr) from the TIG(Zr) and

800 TIG(Ce) electrode, respectively. A previous experiment found that TIG(Zr)
 801 electrodes consume around 0.00224 g/ft in a weld in Ar gas [46]. Using this
 802 electrode consumption estimate, 0.0118 mg of the tip would be embedded in
 803 the weld resulting in a background level of 0.0008 c/(ROI t yr).

804 *4.9. Other Assayed Materials*

805 There are a number of materials in Table 3 that were used, or were under
 806 consideration for use, in the DEMONSTRATOR but were not discussed in the
 807 previous subsections. Here we summarize those materials that have not yet
 808 been described but reside in the table.

809 Teflon[®] (Row 52), Peek[®] (Row 53), and Vespel[®] (Rows 54 through 56)
 810 were all used within the detector string assembly (in addition to uses mentioned
 811 previously). These samples were assayed via NAA and were all found to be
 812 quite pure, although in several cases sensitivity to U was limited by the presence
 813 of ²⁴Na and ⁸²Br. The highest-sensitivity assay was performed on our Teflon
 814 material, which was identified by the EXO collaboration [18] and is used as the
 815 primary structural plastic in the innermost regions of the Demonstrator.

816 The electroformed Cu pins that provide electrical contact between the LMFE
 817 and the Ge crystal were coated with a thin layer of Sn. We assayed two Sn
 818 samples for this purpose (Rows 36 through 38). γ -ray counting assayed the entire
 819 Sn supply, whereas ICPMS assayed a small sample of the same source material.
 820 We do not understand fully why the two assay techniques gave different results
 821 in this case. The Sn is used in very small quantities and even at the highest
 822 assay value doesn't significantly affect our background model.

823 The shield includes a high density polyethylene layer (Row 64) and a 5%-
 824 boron-loaded high density polyethylene layer (Row 65). It also includes a layer
 825 of OFHC copper from Southern Copper (Rows 16 and 17).

826 We investigated several options for alternative cryostat seals. Indium was
 827 rejected due to its natural radioactivity, and existing measurements of butyl also
 828 indicated it would be too radioactive. We assayed o-rings made of Kalrez using
 829 NAA (Rows 67 and 68), these were somewhat cleaner but still too radioactive
 830 for use in the DEMONSTRATOR. We found that we could make seals of sufficient
 831 purity using Parylene film or PTFE (Row 72).

832 The electronic feedthroughs from the vacuum were situated on conflat flanges
 833 attached to a 6-way cube intersection (Row 39). Because there is a shine path
 834 down the cross arm into the shield, this cube was assayed and included in the
 835 background model. Shielding plates in the cross arm make us insensitive to the
 836 relatively high Th levels in this cube.

837 The calibration system uses a Teflon[®] tube (Row 66) as a guiding track
 838 that extends through the shield and encircles the cryostat. The position of a
 839 calibration line source is controlled by a motor controller and drive wheels that
 840 insert and remove the source from this track. Because the drive wheels contact
 841 the source and therefore abraded material might, in principle, be left within the
 842 track, the drive wheels were assayed (Row 123).

843 Sapphire was considered as a material for the LMFE, but was not used due
 844 to the progress using fused silica. We assayed sapphire samples using NAA and

the results are listed in Row 118. The U value was limited by the presence of ^{24}Na in the activated sample.

Our sources at SURF are checked for leakage on a regular schedule. We assayed the KIMTECH PURE[®] W4 critical task wipers from Kimberly-Clark Professionals[®] (Row 124) used to swipe the sources to better understand the background to those measurements.

Our nitrogen purge system uses N_2 gas purified by passing through a charcoal trap. We also used selected charcoal as a getter in the cryostats used for detector transport. We studied a number of charcoals for their radiopurity (Rows 125 through 132).

The purity of glass beads was determined (Rows 141 through 149). These beads were considered for use as a cable ID mechanism but was not implemented.

Epoxies were assayed to identify a product to fix the LMFE's onto their electroformed Cu mount. Hysol[®] 0151[™] and 1C[™] (Rows 133 through 136) were considered and the Hysol[®] 0151[™] was chosen. Torr Seal[®] (Rows 137 and 138) was considered for mounting the HV connector socket at the feedthrough flange. Hysol[®] was used instead.

Silicone Rubber (Row 139) was considered for use in potting the in-vacuum-side of HV feedthroughs. This was not implemented.

5. Background Projections and Conclusion.

The assay program described in this article has provided values for the radioactive contamination of the materials and components that can be used to estimate the background from trace quantities of U and Th in the DEMONSTRATOR apparatus. Such estimates were made via simulation of the geometry of the DEMONSTRATOR by assigning the measured impurity level to each component that comprises the apparatus. The simulation was based on the MaGe Monte Carlo framework described in Ref. [47]. We use the estimates to determine how many counts will remain in a fully analyzed spectrum arising from the measured levels. As mentioned in the Introduction, our goal for the DEMONSTRATOR is 3 c/(ROI t yr). We quote each contribution from each material in similar units in Table 7. We did identify that contamination arising from machining and handling will require further future research.

The U and Th impurities in detector components do not represent a full background model for the DEMONSTRATOR. For completeness, we include Table 8, which groups radiogenic background contributions by detector component. This table also includes background contributions from cosmogenic isotopes produced in materials while they resided on the Earth's surface, γ rays originating from external to the apparatus, α particles from Rn daughters plating out on surfaces of the apparatus, neutron interactions, direct passage of cosmic ray μ , and neutrino backgrounds. The quoted values are the results of our simulations and other calculation methods, which will be detailed in future publications.

The background budget for the MAJORANA DEMONSTRATOR indicates that a large experiment built with a tonne of isotope should be feasible with some modest improvements.

Table 7: The summary of the contributions to the background based on the assay results given in this article, grouped by detector material. The background values assume that the radioactive chains are in equilibrium.

Material	Typical Use	Decay Chain	Achieved Assay	
			$\mu\text{Bq/kg}$	$c/(\text{ROI t yr})$
Electroformed Cu	Inner Cu Shield, Cryostat, Coldplate, Thermal Shield, Detector Mounts	Th	0.06	0.15
		U	0.17	0.08
OFHC	Outer Copper Shield	Th	1.1	0.26
		U	1.25	0.03
Pb	Lead Shield	Th	5	0.26
		U	36	0.37
PTFE	Detector Supports	Th	0.1	0.01
		U	<5	<0.01
Vespel	Coldplate Supports, Connectors	Th	<12	<0.01
		U	<1050	<0.4
Parylene	Cu Coating, Cryostat Seals	Th	2150	0.27
		U	3110	0.09
Silica, Au, Epoxy	Front-end Electronics	Th	6530	0.32
		U	10570	0.28
Cu Wire, & FEP	Cables	Th	2.2	0.01
		U	145	0.08
Stainless Steel	Service Body	Th	13000	<0.04
		U	<5000	<0.03
Solder Flux	Connectors	Th	210	0.13
		U	335	0.06

Table 8: The summary of all the backgrounds contributing to the DEMONSTRATOR, with radiogenic backgrounds grouped by detector component. The background values assume that the radioactive chains are in equilibrium.

Background Contribution	Rate c/(ROI t yr)
Electroformed Cu	0.23
OFHC Cu Shielding	0.29
Pb Shielding	0.63
Cables and Internal Connectors	<0.38
Front Ends	0.6
U/Th within the Ge	<0.07
Plastics + Other	0.39
^{68}Ge , ^{60}Co within the ^{enr}Ge	0.07
^{60}Co within the Cu	0.09
External γ rays, (alpha,n) Reactions	0.1
Rn and Surface α Emission	0.05
Ge, Cu, Pb (n,n'gamma) Reactions	0.21
Ge(n,n') Reactions	0.17
Ge(n, γ)	0.13
Direct μ Passage	0.03
ν Induced Background	<0.01
Total	<3.5

889 Acknowledgments

890 This material is based upon work supported by the U.S. Department of
 891 Energy, Office of Science, Office of Nuclear Physics under Award Numbers DE-
 892 AC02-05CH11231, DE-AC52-06NA25396, DE-FG02-97ER41041,
 893 DE-FG02-97ER41033, DE-FG02-97ER41042, DE-SC0012612, DE-FG02-10ER41715,
 894 DE-SC0010254, and DE-FG02-97ER41020. We acknowledge support from the
 895 Particle Astrophysics Program and Nuclear Physics Program of the National
 896 Science Foundation through grant numbers PHY-0919270, PHY-1003940, 0855314,
 897 PHY-1202950, MRI 0923142 and 1003399. We acknowledge support from the
 898 Russian Foundation for Basic Research, grant No. 15-02-02919. We acknowl-
 899 edge the support of the U.S. Department of Energy through the LANL/LDRD
 900 Program.

901 This research used resources of the Oak Ridge Leadership Computing Facil-
 902 ity, which is a DOE Office of Science User Facility supported under Contract
 903 DE-AC05-00OR22725. This research used resources of the National Energy
 904 Research Scientific Computing Center, a DOE Office of Science User Facility
 905 supported under Contract No. DE-AC02-05CH11231. We thank our hosts and
 906 colleagues at the Sanford Underground Research Facility for their support. We
 907 acknowledge support from the facilities; the Waste Isolation Pilot Plant, the
 908 Kimballton Underground Research Facility and the Low Background Facility
 909 at Oroville, the Berkeley Low Background Facilities at SURF and at LBNL,
 910 the McClellan Nuclear Radiation Center, the North Carolina State University's
 911 PULSTAR research and teaching reactor, the Radiation Safety Office at Vir-
 912 ginia Tech, the Radiation Safety Department at the University of Tennessee,
 913 Knoxville, and Nu Instruments Limited in Wrexham, UK.

914 References

- 915 [1] Abgrall N *et al.* 2014 *Adv. High Energy Phys.* **2014** 365432 (*Preprint*
 916 **arXiv:1308.1633**)
- 917 [2] Elliott S R and Vogel P 2002 *Ann. Rev. Nucl. Part. Sci.* **52** 115
- 918 [3] Elliott S R and Engel J 2004 *J. Phys. G: Nucl. Part. Phys.* **30** R 183
- 919 [4] Avignone III F T, Elliott S R and Engel J 2008 *Rev. Mod. Phys.* **80** 481
- 920 [5] Rodejohann W 2011 *Int. J. Mod. Phys. E* **20** 1833
- 921 [6] Barabash A S 2011 *Phys. Part. Nucl.* **42** 613
- 922 [7] Gómez-Cadenas J J, Martín-Albo J, Mezzetto M, Monrabal F and Sorel M
 923 2012 *Riv. Nuovo Cim.* **35** 29–98 (*Preprint arXiv:1109.5515*)
- 924 [8] Vergados J D, Ejiri H and Šimkovic F 2012 *Rep. Prog. Phys.* **75** 106301
- 925 [9] Heise J 2014 *AIP Conf. Proc.* **1604** 331 (*Preprint arXiv:1401.0861*)

- [10] Heise J 2015 *J. Phys. Conf. Ser.* **606** 1 (*Preprint* [arXiv:1503.01112](https://arxiv.org/abs/1503.01112))
- [11] Luke P N, Goulding F S, Madden N W and Pehl R H 1989 *IEEE Trans. Nucl. Sci.* **36** 926
- [12] Barbeau P S, Collar J I and Tench O 2007 *JCAP* **09** 009
- [13] Aguayo E *et al.* 2013 *Nucl. Instrum. Meth. A* **701** 176
- [14] Xu W *et al.* 2015 *Phys. Proc.* **61** 807 (*Preprint* [arXiv:1404.7399](https://arxiv.org/abs/1404.7399))
- [15] Aalseth C E *et al.* 2008 *Phys. Rev. Lett.* **101** 251301
- [16] Rau W and Heusser G 2000 *Applied Radiation and Isotopes* **53** 371
- [17] Apresella C *et al.* 2002 *Astropart. Phys.* **18** 1
- [18] Leonard D S *et al.* 2008 *Nucl. Instr. Meth. A* **591** 490
- [19] Aprile E *et al.* 2011 *Astropart. Phys.* **35** 43
- [20] Armengaud E *et al.* 2013 *Astropart. Phys.* **47** 1
- [21] Ahmed Z *et al.* 2014 *JINST* **9** P01009
- [22] Akerib D *et al.* 2015 *Astropart. Phys.* **62** 33
- [23] ILIAS database URL <http://radiopurity.in2p3.fr/>
- [24] GOPHER: High-Purity Ge gamma counter at Soudan URL <http://www.hep.umn.edu/lbcf/>
- [25] Heusser G 1995 *Ann. Rev. Nucl. Part. Sci.* **45** 543
- [26] Elliott S, Guiseppe V, LaRoque B, Johnson R and Mashnik S 2010 *Phys. Rev. C* **82** 054610
- [27] Finnerty P *et al.* 2011 *Nucl. Instr. Meth. A* **642** 65
- [28] Thomas K J *et al.* 2013 *AIP Conf. Proc.* **1549** 20
- [29] Thomas K J *et al.* 2015 *AIP Conf. Proc.* **1672** 020004
- [30] National Research Council Canada, 1200 Montreal Rd., Ottawa, Ontario K1A 0R6 URL <http://www.nrc-cnrc.gc.ca>
- [31] Nu Instruments Ltd., Unit 74, Clywedog Rd S, Wrexham Industrial Estate, Wrexham, Clwyd LL13 9XS, United Kingdom URL <http://www.nu-ins.com>
- [32] Institute of Microelectronic Technology and High Purity Materials, 6 Academician Ossipyan str, Chernogolovka, Moscow Region, 142432, Russia URL <http://www.ipmt-hpm.ac.ru>

- 957 [33] VR Analytical, 63020 NE Lower Meadow Drive, Suite 3, Bend, Oregon
958 97701 URL <http://www.vranalytical.com>
- 959 [34] Hoppe E W *et al.* 2014 *Nucl. Instrum. Meth. A* **764** 116
- 960 [35] LaFerriere B D, Maiti T C, Arnquist I J and Hoppe E W 2015 *Nucl.*
961 *Instrum. Meth. A* **775** 93
- 962 [36] Hoppe E W *et al.* 2008 *J. Radioanal. Nucl. Chem.* **277** 103
- 963 [37] Hoppe E W *et al.* 2009 *J. Radioanal. Nucl. Chem.* **282** 315
- 964 [38] Overman N R, Overman C T, Kafentzis T A, Edwards D J and
965 Hoppe E W 2012 MAJORANA electroformed copper mechanical anal-
966 ysis Tech. Rep. PNNL-21315 Pacific Northwest National Laboratory
967 URL [http://www.pnnl.gov/main/publications/external/technical_](http://www.pnnl.gov/main/publications/external/technical_reports/PNNL-21315.pdf)
968 [reports/PNNL-21315.pdf](http://www.pnnl.gov/main/publications/external/technical_reports/PNNL-21315.pdf)
- 969 [39] Overman N R, Overman C T, Edwards D J and Hoppe E W 2015 *Applied*
970 *Phys. A, Materials Sci. Proc.* **120** 1181
- 971 [40] Overman N R, Neal D L, Jensen A M, Overman C T and Hoppe E W
972 2015 MAJORANA electroformed copper mechanical analysis 2 Tech. Rep.
973 PNNL-21315 Pacific Northwest National Laboratory
- 974 [41] Overman N R, Hoppe E W and Addleman R S 2013 *J. Radioanal. Nucl.*
975 *Chem.* **296** 1303
- 976 [42] Titan International, Inc., 301 Lilac Lane, Cinnaminson, NJ 08077
- 977 [43] Goldbach C and Nollez G Radioassay Measurements by Edelweiss,
978 <http://radiopurity.in2p3.fr/search.php?Material=CuBe>
- 979 [44] Amsbaugh J F *et al.* 2007 *Nucl. Instrum. Meth. A* **579** 1054
- 980 [45] Barton P *et al.* 2011 *IEEE Nucl. Sci. Symp. Conf. Rec.* 1976 proceedings
981 of 2011 IEEE Nucl. Sci. Symp. and Med. Imag. Conf. (NSS/MIC)
- 982 [46] Winsor L and Turk R 1957 *Welding J.(NY)* **36**
- 983 [47] Boswell M *et al.* 2011 *IEEE Trans. Nucl. Sci.* **58** 1212