Performance of the Versatile Array of Neutron Detectors at Low Energy (VANDLE)

W. A. Peters^{a,b,c,*}, S. Ilyushkin^d, M. Madurga^b, C. Matei^a,
S. V. Paulauskas^b, R. K. Grzywacz^b, D. W. Bardayan^{c,f}
C. R. Brune^g, J. Allen^{h,e}, J. M. Allen^b, Z. Bergstromⁱ,
J. Blackmon^j, N. T. Brewer^b, J. A. Cizewski^e, P. Copp^{e,k},
M. E. Howard^e, R. Ikeyama^{e,k}, R. L. Kozubⁱ, B. Manning^e,
T. N. Massey^g, M. Matos^j, E. Merino^e, P. D. O'Malley^{e,f},
F. Raiola^d, C. S. Reingold^e, F. Sarazin^d, I. Spassova^{a,e},
S. Taylor^b, D. Walter^d

^a*Oak Ridge Associated Universities, Oak Ridge, TN 37830*

^b*Dept. of Physics & Astronomy, University of Tennessee, Knoxville, TN 37996*

^c*Physics Division, Oak Ridge National Laboratory, Oak Ridge, TN 37831*

^d*Dept. of Physics & Astronomy, Colorado School of Mines, Golden, CO 80401*

^e*Dept. of Physics & Astronomy, Rutgers University, New Brunswick, NJ 08903*

^f*Dept. of Physics, University of Notre Dame, Notre Dame, IN 46556*

^g*Dept. of Physics & Astronomy, Ohio University, Athens, OH 45701*

^h*University of California-Berkeley, Berkeley, CA 94720*

ⁱ*Dept. of Physics, Tennessee Technological University, Cookeville, TN 38505*

^j*Dept. of Physics & Astronomy, Louisiana State University, Baton Rouge, LA 70803*

^k*Dept. of Physics & Astronomy, University of Wisconsin-La Crosse, LaCrosse, WI 54601*

Abstract

1 The Versatile Array of Neutron Detectors at Low Energy (VANDLE) is a new,
2 highly efficient plastic-scintillator array constructed for decay and transfer reaction
3 experimental setups that require neutron detection. The versatile and modular de-
4 sign allows for customizable experimental setups including beta-delayed neutron
5 spectroscopy and (d,n) transfer reactions in normal and inverse kinematics. The
6 neutron energy and prompt-photon discrimination is determined through the time
7 of flight technique. Fully digital data acquisition electronics and integrated trigger-
8 ing logic enables some VANDLE modules to achieve an intrinsic efficiency over 70%
9 for 300-keV neutrons, measured through two different methods. A custom GEANT4
10 simulation models aspects of the detector array and the experimental setups to
11 determine efficiency and detector response. A low detection threshold, due to the
12 trigger logic and digitizing data acquisition, allowed us to measure the light-yield re-
13 sponse curve from elastically-scattered carbon nuclei inside the scintillating plastic
14 from incident neutrons with kinetic energies below 2 MeV.

15 *Key words:* Neutron detection; Plastic scintillator array; Carbon-nuclei light
16 response;

17 *PACS:* 29.30.Hs, 29.40.Gx, 29.40.Mc

18 1 Introduction

19 Neutron detection is increasingly necessary in modern nuclear physics experi-
20 ments. Advancements in radioactive beams have made available for study more
21 neutron-rich nuclei, where delayed neutron emission will increasingly dominate
22 beta decay [1]. Furthermore, inverse-kinematic transfer reactions such as (d,n)
23 can probe structure not easily studied otherwise. The Versatile Array of Neu-
24 tron Detectors at Low Energy, VANDLE, facilitates such experiments.

25 VANDLE [2,3] has similarities to other existing detector arrays such as: TON-
26 neau pour NEutRons REtardes (TONNERRE) [4] at the Grand Accélérateur
27 National d'Ions Lourds (GANIL) in France; the Modular Neutron Array (MoNA)
28 [5,6], the beta decay neutron detector [7], and the Low Energy Neutron Detec-
29 tor Array (LEND) [8,9] at the National Superconducting Cyclotron Labora-
30 tory in Michigan; the Large Area Neutron Detector [10] at GSI in Germany;
31 the neutron wall at the Institute of Physical and Chemical Research (RIKEN)
32 in Japan [11]; and the neutron wall [12] at the University of Notre Dame.
33 Improvements are achieved by optimizing materials and design, and by using

* Corresponding author. Tel.: +1-865-576-6598; fax: +1-865-574-1118.
e-mail address: wapeters@nuclearemail.org

digital signal processing. VANDLE incorporates three sizes of detector modules so that an experimental setup can be customized and tailored to meet geometric and efficiency requirements.

Some neutron reactions within detectors such as these are considered “dark” and do not induce a detectable amount of light. These may be followed by an event that induces a measurable amount of light within the detector (or in a different detector), thereby creating false position and timing for those events. For many scintillator detectors, neutron scattering off carbon nuclei is a strong source of dark reactions. While it is nearly impossible to eliminate false triggering following dark reactions, by running with the lowest possible threshold they are minimized because fewer events are missed. Section 3 discusses how we get VANDLE’s detection threshold as low as possible. Section 6 details VANDLE’s sensitivity to these carbon-scattering events, reducing the amount of dark events in our data. We also present the light-yield response curve for recoil carbon nuclei within the bulk scintillator material. The design motivations and capabilities of VANDLE are discussed in Sections 2 through 5. Results from efficiency and resolution measurements as well as comparisons with a custom Monte Carlo simulation are reported in Sections 7 and 8. Section 9 outlines various experimental programs for which VANDLE is used.

2 Design

Arrays of neutron detectors must have good timing characteristics when used in experiments where the neutron energy is determined via time-of-flight (TOF). Typically scintillators are coupled to photo-multiplier tubes (PMT). To achieve higher intrinsic efficiency, the detector should be thick to expose each neutron to more material with which to interact. Unfortunately, thicker materials degrade the resolution of the path length and reduce the resolution of the measured energy (see equation 10 of [13]). The resolution of the flight path is mostly due to the thickness of the bar surface facing the target. This is the first value listed when quoting the dimensions of a VANDLE bar. The full resolution of the flight path also includes the longitudinal position resolution along the bar’s axis, which is dependent on the TOF resolution, and is discussed in more detail in Section 4.1.

VANDLE is comprised of three different sized scintillating detector modules with scintillator material dimensions of, $3 \times 3 \times 60 \text{ cm}^3$, $3 \times 6 \times 120 \text{ cm}^3$, and $5 \times 5 \times 200 \text{ cm}^3$; labeled “small”, “medium”, and “large”, respectively. This allows for custom experimental layouts to optimize the energy resolution for the energy range of neutrons important to each particular experiment. The medium size modules are a recent addition to the VANDLE suite of detector modules. Therefore, much of the characterization and analysis in this

Fig. 1. (Color online) VANDLE components. Typical of time-of-flight neutron detectors. The specific geometry, PMT, and wrapping used for each of the three VANDLE modules are detailed in the text.

document was performed with the small or large modules.

The small modules are designed to be as sensitive as possible to low-energy neutrons. The medium modules are designed to have the same angular coverage as the small modules, mounted further away for a longer TOF and better resolution for higher neutron energies (2-7 MeV). The large modules are designed to detect neutrons up to about 20 MeV. The VANDLE collaboration has assembled over 100 small modules, 60 large modules, and 45 medium modules.

Figure 1 illustrates the VANDLE module design. Scintillating plastic bars made of Eljen EJ-200 [14] or Bicron BC408 [15] are coupled to photo-multiplier tubes (PMT) made by Hamamatsu [16] using an optical epoxy (Eljen brand EJ-500). Because the PMTs are approximately the same size as the scintillator's cross section, there is less light loss than using a light-guide. The Hamamatsu model R580 PMTs on the smaller VANDLE modules have a diameter of 3.8 cm that nearly matches the diagonal of the square cross section for the small bars. For the large and medium VANDLE modules, the Hamamatsu PMT model R7724 with a diameter of 5.2 cm is used because it is nearly equal to the width of the bars. The medium bars have a 1.5 cm long taper at their ends bringing the width of 6 cm down to 5 cm to more closely match the diameter of the R7724 PMTs. These PMTs were also chosen because their timing resolution and a photomultiplier gain factor of at least 10^7 .

Ideally, scintillation light is internally reflected and guided to the ends, but for extended geometries such as the VANDLE bars, reflective wrapping improves the light collection and reduces attenuation. Different wrapping materials were tested by placing a standard γ -ray source at the center of a VANDLE bar and recording the light-yield spectrum transmitted to the ends while keeping the voltage on the PMT constant. Nitrocellulose (rolls of Cole-Palmer nitrocellulose membranes) transported more light to the ends, leading to a light-yield

spectrum 20% higher than Teflon tape, which was 15% better than aluminized Mylar foil wrapping. Since nitrocellulose or Teflon tape wrapping material are expensive, and considering the large number of modules, the nitrocellulose wrapping is only used for the inside layer on the small VANDLE modules that are designed to optimize sensitivity to low-energy events. To reduce costs, the larger sized VANDLE modules had their bars wrapped with SunFilm brand (2 mil gauge) aluminized Mylar. All VANDLE bars have a second layer consisting of Mylar that is aluminized on the inner surface and coated with an opaque white film on the outer surface (Easygrow brand ‘Lightite’ film).

Optical epoxy from Eljen (EJ-500) was used to couple the PMT to the VANDLE bars to provide a rigid coupling joint. The epoxy was mixed and then placed in an evacuated chamber for 5 minutes to remove bubbles that formed during the mixing process. Epoxy was applied to the ends of vertically mounted VANDLE bars that had been sanded to increase adhesion. The weight of the PMTs squeeze the epoxy to fill in the rough surface and results in a thin (less than 0.1 mm) layer of rigid epoxy when cured. The final assembly included a layer of shrink-wrap tubing around the ends to protect the epoxy joint and electrically insulate the PMT and its base.

3 Digital Data Acquisition

A common problem with detectors that rely on PMTs are random pulses generated by the PMT itself that can overwhelm the data acquisition (DAQ) when trigger thresholds are low. The most common unwanted pulses come from thermionic emission (electrons), that create dark current, arising from the semi-conductor photocathode [17]. Other background pulses caused by stray ions or photons typically have shorter time widths or less amplitude than a normal scintillation event in the PMT. Their frequency is random but decays after any strong PMT signal in a few hundred nanoseconds.

To deal with these background pulses, the VANDLE DAQ system takes full advantage of the digitizing electronics and integrated, programmable trigger logic. One way to reduce the background event rate is to require a coincidence between the two ends. Pixie-16 digitizing modules from XIA LLC [18] have been chosen for VANDLE and are programmable to include this coincidence within a user-selected range. We operate with a coincidence window set to a range that is twice the time for the scintillation light to travel to the opposite end of a bar. This setting reduces the spurious trigger rate enough to run all the channels with very low thresholds while still being able to quantify the non-coincident background for later subtraction.

Another advantage the XIA digitizing electronics have over conventional DAQ

electronics that use leading-edge or constant-fraction discriminators is the built in low-pass filter and the trapezoidal triggering scheme. The low-pass filters (Sallen-Key type), included within the input of the XIA modules, smooth signals appropriately for digitizing by the 12-bit Analog to Digital Converter (ADC) at 250 MS/s. For two signals of equal amplitude, but one is five times shorter in width (like a pulse from a thermionic emission), the shorter pulse will be attenuated more than the normal pulse induced by a true neutron event. Even though the overall shape of the pulses after the low-pass filter will be similar, their amplitudes will now differ since the random background pulses are generally shorter (higher frequency) than the normal pulses [19]. Since large signals can saturate the 12-bit ADC which has a 2 Volt full range, the modules have programmable options to control polarity, a voltage offset, and a low-gain mode to attenuate the signals by a factor of 4.

Additionally, a trapezoid triggering filter, applied by the on-board field-programmable gate array (FPGA), continuously parses the ADC traces and is used for setting the trigger logic. A threshold can be set on the area (or integral charge) of the pulse to reject the signals that result from high-frequency noise. These features are similar to a “shaped” signal used for preamp signals from semi-conductor detectors, the background events are significantly reduced with these features, thereby allowing for lower detection thresholds compared to DAQ systems that use PMTs without these features. To achieve the desired TOF resolution, we record the signal traces from VANDLE at 250 MS/s (generally 120 ns wide). Despite this data-taking requirement, the trigger logic and the digital data processing design in the digitizing modules means the system usually runs 100% live and the data recording rate of the DAQ system is only limited by PXI bridge to the DAQ computer. Further details of the triggering scheme for VANDLE can be found in Reference [20].

4 Timing Resolution

The timing of each VANDLE PMT signal is determined by fitting the digitized waveform, sampled at 250 MS/s, with a custom fitting function detailed in Reference [20]. This fit, along with the internal 125 MHz clock of the digitizers, yields the raw time for each left and right PMT signal, t_l and t_r . They are averaged to get the position independent time for each module by $t_i = (t_l + t_r)/2$. Then, the event TOF is determined by subtracting a “start” signal, t_s from t_i , such that $TOF_i = t_i - t_s$.

Cosmic-ray muons passing through two modules stacked on top of each other were used to determine the timing resolution. This method has the advantage that no external timing detector is needed whose resolution may be difficult to unfold from the data. Cosmic-ray muons travel at nearly the speed of light and

178 deposit 2 MeV per cm in the plastic, inducing about 6 MeVee of light yield in
 179 the small and medium VANDLE bars and about 10 MeVee in the large ones
 180 (when the bars are horizontal). By gain matching all the PMTs, the timing
 181 resolution for each end is nearly identical and is a normal distribution. The
 182 resolution of the time difference between the two modules, $t_{1-2} = t_1 - t_2$, is
 183 equal to the average resolution of the each of the four individual ends notated
 184 with subscripts $1l$, $1r$ and $2l$, $2r$ for the left or right ends of modules 1 and 2,
 185 respectively.

$$t_{1-2} = t_1 - t_2 = \frac{1}{2}[(t_{1l} + t_{1r}) - (t_{2l} + t_{2r})] \quad (1)$$

$$\delta t_{1-2} = \frac{1}{2}\sqrt{(\delta t_{1l})^2 + (\delta t_{2r})^2 + (\delta t_{1l})^2 + (\delta t_{2r})^2} \quad (2)$$

186 This measurement can be repeated for various crossing points between the
 187 two bars to ensure the resolution is position independent.

188 To measure VANDLE's timing resolution, the modules were laid parallel, hor-
 189 izontal, and on top of one another with a gap of 15 cm between them. This
 190 exposed the whole length to muon-induced coincidences. In this configuration
 191 an additional gate on the position of the hit in the top module relative to the
 192 bottom can be used to select muon events where the muons travel perpendic-
 193 ular to the detector's length [21]. Figure 2 shows the time difference between
 194 the left and right ends of two modules plotted against each other (t_{1l-r} ver-
 195 sus t_{2l-r}). Since the time difference between the ends is proportional to the
 196 distance along the bar, the intense streak along the diagonal corresponds to
 197 events that traveled straight down through both modules. By following equa-
 198 tions 1 and 2, the position-averaged timing resolution can be calculated by
 199 making a gate on only these events along the diagonal. By further gating on
 200 small ranges in either t_{1l-r} or t_{2l-r} , the timing resolution at any position along
 201 the modules can be determined.

202 The timing resolution from muons using this method and conventional DAQ
 203 components was $\delta t_{1-2} = 0.69(4)$ ns and $0.81(5)$ ns (FWHM) for the small and
 204 large modules, respectively and was found to be consistent along the length
 205 of the VANDLE modules. Using the DAQ with the 250 MS/s digitizers, the
 206 position-independent timing resolution was nearly the same, since the PMT
 207 timing response dominates the resolution. The measured resolution with the
 208 250 MS/s digitizers was $\delta t_{1-2} = 0.63(4)$ ns and $0.75(5)$ ns for the small and
 209 large VANDLE modules, respectively. Since the medium VANDLE modules
 210 use the same model PMTs as the large modules, they have nearly the same
 211 timing resolution. Direct comparisons are difficult to make for a couple of
 212 reasons. First, signal strength varies with PMT voltage. Second, the induced
 213 light yield can vary by bar type. The intrinsic timing resolution of the digitizers

Fig. 2. (Color online) Two dimensional plot of the time difference of muon events between two stacked large VANDLE modules. The intense diagonal streak is from muons traveling straight down through both bars. As described in the text, the negative time difference values correspond to events on the left side of the VANDLE bar. The units are 0.5 ns per division.

as a function of signal strength is discussed in Reference [20] and measured to be much smaller than the VANDLE δt . Therefore, the VANDLE timing resolution is dominated by the PMT response and faster digitizers would not improve the timing resolution.

4.1 Energy and Position Resolution

There are two methods for determining the longitudinal position within an extended detector like VANDLE: relative light yield or time difference between the two ends. The first takes advantage of the attenuation of the induced light as it passes to each end. For VANDLE this was less desirable since the attenuation lengths are all longer than the length of the bar and vary for each module. The time difference method for determining the position was found to be a more reliable approach. The position resolution can therefore be determined by the timing resolution and the speed of the scintillation light within the bar. For VANDLE this leads to a normal distribution with a FWHM of 4.3 cm for the smaller VANDLE modules and 6.0 cm for the medium and large VANDLE modules.

The speed of the scintillation light within the VANDLE bars was measured using standard gamma check-sources and muons. For muons, a separate coincident trigger detector to only record muons that pass through the small trigger

233 detector was placed above a VANDLE bar. For both methods the source or
 234 trigger detector was then positioned at different distances along the bar to
 235 record the time difference between the ends. The resulting speed of the scin-
 236 tillation light within the VANDLE bars depends on their shape: 12.7(5) cm/ns
 237 for the small bars, 13.5(4) cm/ns for the medium bars, and 15.2(5) cm/ns for
 238 the large bars.

239 The thickness of the detector module is the largest component of the energy
 240 resolution for VANDLE because our timing resolution (FWHM) is less than
 241 1 ns and the resolution of the longitudinal position along the bar has a neg-
 242 ligible effect on the calculated neutron path length. The small, medium, and
 243 large VANDLE modules are typically mounted 50, 100, and 150 cm from the
 244 target, respectively. Their respective thicknesses of 3, 3, and 5 cm means the
 245 energy resolution due to their thickness is 12%, 6%, and 7%. For the large
 246 VANDLE modules, which are designed to operate further from the target, the
 247 timing resolution also contributes to the overall energy resolution and depends
 248 on the energy of the neutron. To calculate the neutron TOF resolution for a
 249 given experiment, the additional TOF resolution due to a start detector used
 250 along with the VANDLE modules needs to be added in quadrature with the
 251 path-length resolution. See [13] for a more complete discussion of calculating
 252 common TOF neutron detector resolutions.

253 Figure 3 shows the energy response due to each module's thickness and $\delta t =$
 254 1 ns, from a GEANT4 simulation of mono-energetic neutrons into small VAN-
 255 DLE neutron detectors (3 cm thick) for a flight-path length of 54 cm. Figure
 256 4 shows the same parameter for the large VANDLE modules (5 cm thick)
 257 at a distance of 150 cm. The sawtooth-like shape of the peaks for the small
 258 modules is due to the dominance of the thickness on the energy resolution. At
 259 these energies the mean free path in the plastic is smaller than the thickness,
 260 so there are more events in the front half of the detector compared to the back
 261 half, while the energy is calculated for a reaction in the middle of the module.
 262 At energies where the timing resolution begins to dominate the overall energy
 263 resolution, the peaks begin to look like a normal distribution. Any additional
 264 source of timing resolution due to a start detector would need to be folded in
 265 with the resolutions shown in Figures 3 and 4.

266 5 Calibration of Light Yield into keVee

267 Due to light attenuation as it passes through the scintillating plastic bars,
 268 each PMT records a position-sensitive signal according to the standard decay
 269 equation:

Fig. 3. GEANT4 Monte Carlo simulations of energy spectra for mono-energetic neutrons in the small VANDLE modules at a flight path of 54 cm. The asymmetric shapes are due to more neutrons interacting near the front surface of the module compared to the back. The FWHM resolution is dominated by the thickness of the small modules (3 cm) and is 12% at 54 cm.

$$l_l = l_o(e^{-(\frac{L}{2}-x)/\lambda}) \quad (3)$$

where l_o is the light yield induced in the bar at the reaction point, λ is the attenuation length, $L/2$ is the half length of the bar, and x is the position along the length of the bar where the light was induced, with zero at the center of the bar. By taking the square-root of the product of the left and right deposited light yields, we get a position independent-measurement of the light yield induced in the plastic. The average attenuation lengths are all longer than the physical lengths of the bars: 100 cm for the small modules, 180 cm for the medium modules, and 220 cm for large modules. The variation of the attenuation lengths among each VANDLE module size has a standard deviation of 20%. VANDLE modules with different geometry and wrapping material account for the significant difference in attenuation lengths between

Fig. 4. GEANT4 Monte Carlo simulations for mono-energetic neutrons into the large VANDLE modules at a distance 150 cm. In this case the FWHM resolution for the large modules (5 cm thick) is near 8%. Both the resolution due to the thickness (6.7% resolution) and the timing resolution of $\delta t = 1$ ns contributes to the overall energy resolution for the large modules.

different sizes. Variations of the attenuation length between modules of the same size are likely due to slight differences in the PMT coupling, defects in the plastic, or wrinkles in the wrapping.

The light response curve from low-energy neutrons has been mapped down to a few keVee and exhibits a deviation from Birks formula [22], whereby the light response curve becomes unphysical near zero [23]. For consistency, we use the standard convention that 1 MeVee is the light yield induced in the EJ-200 scintillating plastic from a 1 MeV electron. Extrapolations of those light response curves lead to a small, but non-zero amount of light yield for zero-energy protons (scattered from neutrons). This effect is also seen with elastically scattered carbon recoils presented in Section 6 and is likely due to inaccuracies in calibrating the light yield spectrum below 20 keVee [24, 25].

For VANDLE modules it was found that at light yields above 31 keVee, the amplitude and area of the response signal is linear with keVee, as measured by three different methods: energy deposited by cosmic-ray muons, Compton scattering of γ rays, and full photoabsorption of low-energy γ rays. The energy deposited by cosmic-ray muons is calculated from the stopping power of muons listed in Ref. [26] and is 2 MeVee per cm of plastic. Full photo-peak absorption of γ rays in plastic scintillators can occur at energies less than about 80 keV leading directly to a peak in the light yield spectrum in keVee equal to the γ -ray energy in keV.

Fig. 5. Uncalibrated ADC spectrum of full photoabsorption peak from the 59.5 keV gamma from a ^{241}Am source.

Two common sources used for calibrating the low light yield spectrum for scintillators (listed by IAEA [27]) are ^{133}Ba and ^{241}Am since they both have relatively intense decays leading to photons with energies less than 80 keV. ^{133}Ba decay has two X-rays near 31 keV (30.63 and 30.97) that are sufficiently within the light yield resolution that they are observed as a single peak. ^{241}Am has an intense γ -ray emission at 59.5 keV. Figure 5 shows the raw ADC spectrum from this source. The full photoabsorption peak near channel 125 is clearly identified above the background and separated from the low energy threshold near channel 30. The width of the photoabsorption peak or the shape of the light yield curve at the Compton edge can be analyzed to determine the resolution of the light yield. We selected the Compton edge energy at 70% of the peak for our calibration points. Figure 6 shows the calibration from ^{60}Co , ^{137}Cs , and ^{241}Am sources. From the ^{60}Co source only the Compton edge from the 1.33 MeV γ -ray transition was used. The linear fit of this calibration is also plotted with a slope of 0.369 keVee per ADC channel, and an offset of -24.4 channels or -9 keVee. In Figure 6 the error bars are smaller than the points. The light yield peak at 59.5 keV peak has a resolution of 16 keVee (FWHM). Calibrations using cosmic muons require a different gain setting that would not be sensitive to the low energy point from ^{241}Am , but is linear with the calibration points from other Compton-scattered γ rays.

Improvements to this process of calibrating the light response may involve measurements of lower photoabsorption peaks near 15 or 20 keV, or by calibrating the deposition of a small amount of energy into the scintillator by also measuring the angle of scattered particles [23]. The light response curve for NaI, Anthracene, and Pilot B plastic was measured using low energy electrons impinging the surface of these scintillators [28]. The data from this Refer-

Fig. 6. (Color online) Light yield calibration for a small VANDLE module using Compton-scattered γ rays and full photoabsorption peaks from ^{60}Co , ^{137}Cs , and ^{241}Am sources. The linear fit is also plotted (values in text).

ence [28] confirms the non-linearity of the light curve for these materials below 30 keV electrons. In Section 6 we present our measurement of the carbon-recoil light response curve measured within the bulk material.

6 Light Response of Recoil Carbon Nuclei

Neutron reactions off the carbon nuclei in scintillator detectors are often considered “dark” events because the induced light-yield is below detection threshold. These “dark” reactions contribute to scattering and poor resolution. While protons can scatter with energies up to the energy of the neutron, there is a maximum amount of energy transferred to carbon nuclei for completely back-scattered neutrons due to their mass difference. By applying energy and momentum conservation the maximum recoil energy of the carbon nucleus can be determined as a function of neutron energy for fully back scattered neutrons ($\theta_n = 180^\circ$). However, at low detection thresholds, light yield induced by elastically scattered carbon nuclei (recoils) contributes significantly to the total light yield induced in the scintillator [29]. VANDLE is capable of running with thresholds low enough to identify and measure the light yield response curve for these carbon recoils.

As mentioned in Section 5, recent work [23] has identified a strong deviation from Birks’ light response equation [22]. Even the formula of Madey *et al.* [30] does not have a zero crossing point for the light yield from zero-energy charged particles. By recording the light yield versus TOF from neutron-scattered

Fig. 7. Setup of two VANDLE modules Bar1 and Bar2 for studying carbon recoil events in coincidence. The blocker material consisted of two layers. The first was over 8 inches of paraffin and the second was over 8 inches of a heavy metal such as tin or lead.

carbon recoils, the light response curve from low-energy carbon nuclei can be determined. Other studies have been done on the light response from carbon, but most are for energies above the first excited state of ^{12}C [31, 32]. These studies are also possibly corrupted by surface effects, whereas we measure the light yield from low-energy carbon recoils produced within the bulk of the plastic scintillator.

The setup for the carbon recoil study, illustrated in Figure 7, included a ^{252}Cf source positioned 1.5 meters from one small VANDLE module (Bar1). A second small VANDLE module (Bar2) was placed in between, but shielded from the ^{252}Cf source by thick layers of a heavy metal and paraffin. Four types of coincident events between both VANDLE modules are possible: random, cosmic-ray shower events, forward-scattered γ rays, and back-scattered neutrons. By gating on relative TOF for each module, we can select back-scattered neutron events when TOF_2 is greater than the TOF_1 , indicating the first reaction of this event occurred in Bar1. Since neutrons cannot back-scatter off protons, these events must be from neutron-carbon scattering. The light in-

Fig. 8. (Color online) TOF_2 versus TOF_1 . Coincident events from γ rays (*a*), neutron forward scattering (*b*), neutron backscattering (*c*), and nearly simultaneous cosmic-ray events (*d*) are marked. See text for full description of marked features.

Fig. 9. (Color online) Light yield in Bar1 versus TOF_1 . Un-gated events from γ rays (*a*), neutron-proton scattering (*b*), neutron-carbon scattering (*c*) are identified and also marked in Figure 8.

365 duced in Bar1 for these events is therefore the light from the scattered carbon
 366 nuclei. The signal in Bar2 is from any neutron event above threshold and is
 367 only used to gate the events to identify the reaction.

Fig. 10. Light yield deposited in Bar1 (in keVee) as a function of carbon kinetic energy in MeV. The data are taken from Table 1. The top x -axis shows the energy of the fully back-scattered neutron that induces the carbon recoil kinetic energy shown on the bottom x -axis. Due to uncertainties in calibrating below 31 keVee, there is an additional systematic uncertainty of about 1.5 keVee on all points (not shown).

Figure 8 displays TOF_2 versus TOF_1 for the pair of VANDLE modules. The events labeled **a** are from the γ flash of the ^{252}Cf decay reaching either VANDLE module in coincidence with a random event in the other module. The group of events labeled **b** mark a region where $\text{TOF}_1 > \text{TOF}_2$ that corresponds to scattered neutrons striking Bar2 (which is closer to the source) before Bar1. These events are greatly reduced by the thick heavy metal and paraffin shielding, but not eliminated. The opposite is true for the group of events marked as **c** where $\text{TOF}_1 < \text{TOF}_2$. These events are scattered-neutron events from Bar1 into Bar2. This can only happen via scattering by carbon nuclei or multiple proton-scattered events. Indeed, calculating the carbon-scattered energy for a neutron from Bar1 to Bar2 at 165° leads to a TOF kinematic loci that corresponds exactly to the edge of events marked by **c**, confirming these events are from neutrons scattering off carbon nuclei. Coincident events on the $\text{TOF}_1 = \text{TOF}_2$ line come from cosmic-ray showers striking both modules at nearly the same time, labeled by **d**.

The same features are marked in Figure 9 that displays the light yield in Bar1 versus TOF_1 after all coincident restrictions are removed. The strong low-light-yield kinematic curve at shorter TOF is identified as light yield from carbon recoils. When the coincidence gating is removed (as in Figure 9) there are

Table 1

Induced light yield in keVee from carbon recoils within the plastic scintillator. The electron equivalent calibration is described in Section 5.

Carbon energy (MeV)	Light yield (keVee)
0.37	12.4(9)
0.43	14.2(9)
0.48	15.3(8)
0.54	16.9(7)
0.60	17.8(6)
0.65	18.7(6)
0.71	19.0(6)
0.77	20.8(6)
0.83	21.7(6)
0.88	23.3(6)
0.94	24.8(6)
1.00	25.3(7)
1.05	26.8(7)

387 enough events to measure the upper edge corresponding to fully back-scattered
 388 neutrons off carbon nuclei. Table 1 lists the energy of the carbon recoil from a
 389 full backscatter event in Bar1 and the light yield from the carbon in electron
 390 equivalent (as calibrated by the methods described in Section 5). These TOF
 391 points were chosen because they correspond to carbon recoil energies that
 392 are less than the energy of the first excited state, ensuring all the light yield
 393 induced in Bar1 is solely from elastically scattered carbon nuclei. There are two
 394 sources of uncertainty in the listed values of Table 1. The listed uncertainties
 395 come from the statistical analysis associated with finding the edge of the
 396 carbon light yield curve. There is an estimated additional 1.5 keVee of common
 397 systematic uncertainty due to the lack of calibration points below 31 keVee. If
 398 the trend of the carbon response curve in Figure 10 is extended to lower kinetic
 399 energies, there would be a non-zero amount of light yield for carbon nuclei at
 400 rest, similar to results from [23]. This non-physical result might be due to non-
 401 linearities in the electron equivalent calibration below 31 keVee, [24, 28]. The
 402 results of our study can be used to anchor the light response curve of carbon
 403 nuclei measured at higher energies [31] and the low-energy end of light response
 404 equations [32]. Using a threshold low enough to trigger off recoil carbon nuclei
 405 improves VANDLE's detection efficiency by being sensitive to both proton
 406 and carbon recoil events. It also improves the resolution of the system by
 407 being able to identify and reject cross-talk events where the neutron scatters

Fig. 11. Flux of neutrons for $^{27}\text{Al}(d,n)$ at 120° measured at the Edwards Accelerator Laboratory (EAL) at Ohio University [34]. The smooth nature of the spectrum makes it ideal to use for calibrating the efficiency of VANDLE modules.

from one bar into a another, that would be considered “dark” reactions for detectors with a higher threshold.

7 Efficiency

7.1 Results from $^{27}\text{Al}(d,n)$

The Edwards Accelerator Laboratory (EAL) at Ohio University [33] provided a beam of deuterons at 7.44 MeV impinging on an ^{27}Al target. This reaction produces neutrons with a broad energy distribution. The neutron spectrum has been repeatedly calibrated at 120 degrees relative to the beam direction for use as a standard spectrum to measure the efficiency of neutron detectors [34]. This reaction and angle were chosen because a wide range of neutrons energies are produced and the intensity does not exhibit sharp resonant peaks with respect to neutron energy, as seen in Figure 11.

The target assembly was rotated to position the 30-meter neutron tunnel at 120 degrees relative to the incident beam. One small VANDLE module was placed 13 meters from the target inside the tunnel to expose the full 60-cm length of the module to the uniform neutron flux. A down-scaled beam pulse from the beam buncher (with a resolution about 1 ns) was used as the start timing reference for measuring the neutron time-of-flight (TOF). Figure 12 shows the TOF for a small VANDLE module in the tunnel, with the module

Fig. 12. (Color online) TOF of neutrons measured with a small VANDLE module in the Ohio neutron tunnel at EAL. The TOF scale is in units of 0.5 ns. Black data (top) were taken with the module centered. Red data (bottom) were taken with the module shifted 40 cm out of the central path of neutrons, shadowed by significant shielding at the tunnel entrance. This comparison shows no scattered neutron events. Only the TOF peak from scattered γ rays remains and is slightly delayed in the shifted (bottom) data because the γ rays must scatter before reaching the detector.

427 in the center of the tunnel (top line) and after being translated up 40 cm
 428 such that the module is shielded by the blocking material at the entrance of
 429 the tunnel near the target (bottom line). The two TOF spectra are plotted
 430 without normalization on a log scale so relative intensities of the peaks can
 431 be seen. The lack of a neutron TOF peak in the shifted (bottom) data in
 432 Figure 12 demonstrates the lack of neutron background inside the tunnel.
 433 The flat background in Figure 12 is mostly due to random γ rays from the
 434 environment and was easily subtracted to measure the neutron TOF spectrum
 435 due to the $^{27}\text{Al}(\text{d},\text{n})$ reaction.

436 The intrinsic efficiency of the small module was calculated from the intensity
 437 of the neutron spectrum, which is determined by recording the flux of the
 438 incoming deuteron beam. Figure 13 displays the intrinsic efficiency of a small
 439 VANDLE module with a software threshold set to 31 keVee.

440 A similar setup outside the neutron tunnel was used for measuring the effi-
 441 ciency of the large $5 \times 5 \times 200 \text{ cm}^3$ VANDLE modules. The large modules
 442 were mounted vertically at 120 degrees relative to another aluminum target at
 443 the end of a general purpose beam line in the large experimental hall at EAL.
 444 To estimate the background in this configuration, over 8 inches of brass was
 445 positioned to shield neutrons traveling directly to the detector module but al-

Fig. 13. Efficiency of a small VANDLE module obtained from the $^{27}\text{Al}(\text{d},\text{n})$ measurement at EAL. The software threshold was set to 31 keVee. These data were taken with the VANDLE modules run with the digitizers in low-gain mode. The solid line follows the efficiency as calculated by the VANDLE simulation code with the same threshold.

lowed for reflections off other equipment and features within the room (floors, walls, ceiling, etc.). Figure 14 shows the intrinsic efficiency for the large VANDLE module with no added software threshold. Because of their overall size, the large modules must have a higher hardware threshold to reduce random background. They have a relatively better efficiency at neutron energies above 4 MeV due to their increased thickness of 5 cm. Since the medium VANDLE modules are a recent addition, this test has not yet been performed with a medium module.

7.2 Results from ^{252}Cf

Recording the neutron spectrum from a calibrated detector at the same time as a new uncalibrated detector is another way to measure the efficiency of the new device, as demonstrated in References [8, 9, 13]. The setup involved mounting a ^{252}Cf source with a trigger detector over 3 meters above the floor in a high-ceiling room with the VANDLE module and the calibrated liquid detector 1 meter from a thin steel floor. This configuration reduces scattering and allows easy access to place a long tray with tin or lead and a layer of paraffin for a shadow-bar measurement to record the background from scattering. The TOF is calculated relative to the small trigger scintillator mounted above the ^{252}Cf source that provides the “start” signal for both the VANDLE module and the yield reference detector. We used an Eljen 2-inch organic liq-

Fig. 14. Efficiency of a large VANDLE module obtained from the $^{27}\text{Al}(\text{d},\text{n})$ measurement at EAL. Background subtraction was based on a separate run with a shadow bar made from a thick layer of paraffin and a mixture of heavy metals shadowing the VANDLE module directly but not the other scattering surfaces within the room. These data were taken with the digitizers in low-gain mode.

uid scintillator (EJ301) with a threshold set to the full photo-absorption peak from ^{241}Am at 59.5 keV for a yield reference detector. The neutron spectrum of the ^{252}Cf source is well established, and compiled in [35, 36], and therefore can be normalized by the known efficiency of the yield reference detector. This liquid scintillating material (NE213 or EJ301) has been well characterized for use as an efficiency standard [37]. This method, and motivation for choosing a 2-inch Eljen module as the yield reference detector, are detailed in Ref. [8].

There are three components to the background for this setup that must be accounted for in order to extract an efficiency for the low-energy neutrons. The first is a random TOF background that can be fitted with a flat horizontal line in TOF, similar to the background fitting for the EAL measurement from Section 7.1. The second component is due to correlated scattered neutrons. This component was measured by using a shadow bar which was comprised of 8 inches of tin or lead along with 8 inches of paraffin. The shadow bar was placed directly between the source and detector, 20 cm above the detector, allowing scattered neutrons from the room and equipment to reach the detectors. The third source of background comes from the decay of short-lived daughters of ^{252}Cf fission products. This background is a combination of many different decaying daughters, which is more intense after the gamma-flash peak and decays over many nanoseconds. This type of background is not present when a beam-induced neutron source is used. If this second source of background is not subtracted, the lowest energy bins (that have longer TOF ranges than the higher energy bins) will have more integrated counts above

Fig. 15. (Color online) Efficiency of a small VANDLE module measured with a ^{252}Cf source with various threshold cuts measured in keVee. Both ^{152}Eu and ^{133}Ba sources were used to calibrate the keVee for setting the software thresholds. Data were taken with the VANDLE modules run with the digitizers in high-gain mode. The drop in efficiency above 1.5 MeV is due to some events saturating the XIA digitizers.

the random background and an artificially high efficiency will be calculated. To account for this second background source, a range of TOF corresponding to neutron energies below our detection threshold (longer TOF) was used to estimate the contribution of these decaying daughters and was added to the background subtracted from the ^{252}Cf spectrum. The resulting efficiency plots have more uncertainty for lower neutron energies with longer TOF bins, as seen in Figures 15 and 16.

Figure 17 shows the intrinsic efficiency for a medium VANDLE module using the same method. The efficiency is slightly less than the small VANDLE modules due to the larger size and lack of nitrocellulose wrapping. However, the larger size of the medium modules means that they can be mounted at twice the distance of the small VANDLE modules to improve resolution and maintain the same geometric coverage.

7.3 Comparing Results

Comparisons can be made between the intrinsic efficiency plots of the small modules from the ^{252}Cf measurement and from the measurement at Ohio University (Figure 13) that have a peak efficiency using a 31 keVee threshold of 48% and 50%, respectively. The nearly identical efficiency for neutrons

Fig. 16. Small VANDLE module intrinsic efficiency measured with a ^{252}Cf source with a 31 keVee threshold determined by the full photo-absorption of ^{133}Ba decay. Data were taken with the VANDLE digitizers in low-gain mode. Notice the efficiency does not drop off as sharply above 1.5 MeV as it does in high gain mode (Figure 15).

Fig. 17. (Color online) Medium VANDLE module intrinsic efficiency measured with a ^{252}Cf source. Data were taken with the VANDLE digitizers in low-gain mode. The efficiency was determined with no threshold and with two software thresholds applied: 31 keVee and 59.5 keVee determined by the full photo-absorption of γ rays from sources of ^{133}Ba and ^{241}Am , respectively.

507 between 0.5 and 1.0 MeV for the two measurements with similar threshold
 508 values provides confidence in the intrinsic efficiency plots. However, there is
 509 one noticeable difference between the two plots. The efficiency at energies

510 above 1.5 MeV is reduced (for similar threshold values) for the data in Figure
 511 15 because it was recorded in high-gain mode, and some of the stronger signals
 512 associated with higher energy neutrons saturated the digitizing electronics.
 513 Without high-gain mode, the electronics do not saturate for this range of
 514 neutron energies, as can be seen in Figure 16. The gradual drop in efficiency
 515 for neutron energies above 1 MeV follows the neutron–proton cross section and
 516 is characteristic of plastic scintillators. The larger uncertainties in Figure 16
 517 are due to uncertainties with the live-time clock used to compare background
 518 runs recorded with a shadow bar. This contributes 7% to the uncertainty.
 519 In addition to statistical uncertainties, systematic uncertainties come from
 520 uncertainties in the Mannhart ^{252}Cf compilation [35, 36] estimated to be 2%,
 521 and uncertainties in the efficiency of the reference EJ301 liquid detector [8]
 522 estimated to be 3%.

523 Repeated tests confirm that after gain matching the PMTs of a single module,
 524 the corresponding efficiency curve is reproduced within measuring uncertain-
 525 ties for any applied software threshold. Consistency in the design and the
 526 materials also leads to consistency in the efficiency curve between different
 527 modules within our measuring uncertainties. Slight variations in the efficiency,
 528 due to threshold effects, can occur if the light yield resolution varies by mod-
 529 ule. This is minimized by gain matching the entire array before an experiment.
 530 The PMT gain remains stable after they have been energized and settled for
 531 a day.

532 8 Simulation

533 Determining neutron detection efficiency, intrinsic spatial resolution, neutron
 534 interaction time distribution, and other parameters requires accurate simu-
 535 lations. We have developed a GEANT4 simulation of VANDLE to optimize
 536 the array geometry for an experiment through consideration of detection effi-
 537 ciencies and backgrounds. GEANT4 is a toolkit for simulating the passage of
 538 particles and interactions in matter [38, 39].

539 Advantages of using GEANT4 include incorporating of a variety of physics
 540 models specific to the low- and high-energy domains and the ability to optimize
 541 parameters with different accuracy requirements. Recent improvements to low
 542 energy neutron transport allow us to implement the “Neutron High Precision
 543 Models and Cross Sections” module, which is a data-driven model based on
 544 the G4NDL evaluated neutron data library [40]. This model works well for
 545 neutrons with energies below 7 MeV. In addition, GEANT4 is able to simulate
 546 the geometries of complex setups permitting us to model neutron scattering by
 547 beam line components, detector supports, target, and room floors and walls.
 548 These common laboratory and experimental features are a significant source

Fig. 18. Small VANDLE module intrinsic efficiency from a ^{252}Cf source with a 31 keVee threshold measured by the full photo-absorption of ^{133}Ba decay. The solid line follows the efficiency as calculated by the VANDLE simulation code. Data are the same as the third set in Figure 15.

of unwanted background for neutron detection.

Two different position-dependent threshold effects for a valid left-right coincident VANDLE event are treated in the simulation. The first occurs when a large amount of light is deposited in the bar or when the PMTs are biased too high; the digitizing electronics saturate and the signal is discarded. For signals near this intensity, the attenuation of the scintillation light through the bar means more of these events are lost at either end because the left or right signal is too large for the digitizers. The second case occurs for low-light signals near the detection threshold. For these signals, the attenuation through the bar means that the light from events near one end will fall below the detection threshold at the other end. Both effects will cut more reactions near the ends for the same induced light yield. This effect is avoided by requiring a software cut to induce a position-independent threshold, thereby throwing away some events in order to properly calibrate the efficiency of the detectors. By tracking the intensity and transport of the induced scintillation light, the VANDLE simulation can accurately model position sensitivity of both the low threshold and saturation effects of the acquisition without rejecting otherwise good data, increasing our efficiency to low-energy neutrons. The resulting efficiency plots match our data in both high-gain and low-gain modes (see Figures 13 and 18). Further details will be presented in a comprehensive publication by S. Ilyushkin *et al.* [41].

Fig. 19. (Color online) Image of the VANDLE beta-delayed neutron decay setup from 2012 at HRIBF with 48 VANDLE modules and two HPGe clovers that were removed from their left-right mounts for this picture. The beam comes in from the right to implant on the movable tape in the center of the setup, where a thin beam pipe at the end of the tape drive is surrounded by two thin beta-particle scintillators, used as the TOF start for VANDLE.

570 9 Experimental Setups

571 VANDLE easily couples to a variety of other auxiliary detectors enabling a
 572 range of experiments involving neutron detection. Current efforts include beta-
 573 delayed neutron spectroscopy, and neutron-ejecting reaction experiments like
 574 (α, n) and proton transfer (d, n) experiments. Future plans involve additional
 575 reaction $((d, n)$ and $(\alpha, n))$ and beta-decay spectroscopy experiments at various
 576 laboratories as well as studying neutron correlations with fission fragments.

577 9.1 Beta-delayed neutron spectroscopy

578 For nuclei approaching the neutron dripline, the energy released in beta decay
 579 increases as the neutron separation energy decreases leading to the emission
 580 of a neutron. While the neutron branching ratio can sometimes be determined
 581 directly using ^3He neutron counters [42] or indirectly by identifying γ rays
 582 from the decay chain of these $(A - 1)$ daughters [43], more information about
 583 the Gamow-Teller strength distribution can be extracted by also measuring

the neutron energy spectrum. This extra information can be used to compare with and guide models for predicting the half-lives and neutron emission branching ratios of neutron-rich isotopes beyond the reach of current production methods. These half lives play an important role in determining the final abundance pattern from the r process and can even help to pin down the stellar environment that induces it [44, 45].

Figure 19 show an image of a setup using 48 modules in an arch around an ion-stopping station with room for two high-purity germanium (HPGe) clover detectors. This setup was used in a commissioning experiment to measure the beta-delayed neutron spectra from short-lived fission fragments. Fission fragments from the Holifield Radioactive Ion Beam Facility at Oak Ridge National Lab were collected on a moving tape system and the γ -ray and neutron spectra were recorded following beta decay for over 30 isotopes from near ^{78}Ni to around ^{132}Sn . A beta detector surrounded the implantation point and provided the “start” time for the neutron TOF. For many of these nuclei it was the first time their beta-delayed neutron energy spectrum had been studied. Details of this experiment and the analysis of $^{83,84}\text{Ga}$ spectra are included in a forthcoming publication by M. Madurga *et al.* [46]. VANDLE has since been used in similar beta-delayed neutron spectroscopy experiments at RIKEN and CARIBU at Argonne National Lab, with small and medium VANDLE modules.

9.2 Reaction experiments (α, n) and (d, n)

Proton transfer in inverse kinematics on radioactive nuclei can probe the single-particle structure of nuclei. Spectroscopic information can be extracted such as level energies, orbital angular momentum transfer, and spectroscopic factors. Knowing these properties for nuclei near closed shells can be important bench marks, analogous to the goals of (d, p) neutron transfer experiments [47]. Such proton transfer experiments can also inform the nucleosynthesis proton-capture reaction rates for select nuclei where the single-proton spectroscopic factors can be used to determine the resonance widths. VANDLE was used for (d, n) experiments, whose analyses are ongoing, at the National Superconducting Cyclotron Facility at Michigan State University, and at the Nuclear Science Lab at the University of Notre Dame.

VANDLE is also suited for measuring (α, n) cross sections on select nuclei that, in addition to being important for astrophysical reactions, are of interest to the nuclear safeguard community (α, n). For example, uranium-hexafluoride (UF_6) is used in many parts of the uranium fuel cycle and one technique used by nonproliferation agencies to monitor and account for ^{235}U -enriched material consists of measuring gross neutron rates induced by uranium-decay alpha

Fig. 20. (Color online) Drawing of VANDLE reaction setup used for (α,n) and (d,n) experiments. The tripod-style stands support curved mounting brackets that each hold 14 small VANDLE modules. The tripod mounts can be adjusted to hold the modules at heights of up to two meters.

particles reacting with the fluorine and emitting a neutrons. The $^{19}\text{F}(\alpha,n)^{22}\text{Na}$ cross section is vital for this non-destructive assay method. Some (α,n) reactions are also a source of background for highly sensitive underground experiments [48]. Neutrons released by (α,n) reactions with equipment or the surrounding material can become the dominant source of unknown background rates.

A challenge for these neutron-ejecting reactions is that an external timing signal related to the incoming beam is needed for VANDLE to determine the energy of the detected neutrons. Timing signals from pulsed accelerators with beam bunches smaller than 1 ns will preserve the timing resolution. Another possibility is to use a timing detector that can operate with the full beam intensity, which should be located in the beam line away from the target location to reduce background. Alternatively, after separation of a sufficiently fast recoil through a recoil separator, the timing of the reaction residue can be used. The background reduction attained by utilizing a recoil separator and tagging on heavy reaction products is necessary for most cases with radioactive beams, where the random room background can be significant compared to the desired reaction. Figure 20 shows a drawing of the curved mounts for the small VANDLE modules used for a couple $^{19}\text{F}(\alpha,n)^{22}\text{Na}$ experiments [49] and a couple (d,n) experiments still under analysis.

643 10 Conclusion

644 The Versatile Array of Neutron Detectors at Low Energy (VANDLE) is a new
645 highly efficient plastic-scintillator array built for nuclear decay and reaction
646 experiments that require neutron detection. It is comprised of three different
647 sized EJ-200 scintillating-plastic bars ($3 \times 3 \times 60 \text{ cm}^3$, $3 \times 6 \times 120 \text{ cm}^3$, and
648 $5 \times 5 \times 200 \text{ cm}^3$). The neutron energy and prompt photon discrimination is
649 determined through the time of flight technique. The versatile and modular
650 design allows for experimental setups that can be customized for the particular
651 application, including beta-delayed neutron spectroscopy and (d,n) transfer
652 reactions in inverse kinematics. Depending on the setup the energy resolution
653 of VANDLE is between 12% to 7%. By incorporating digitizing electronics
654 for the data acquisition and triggering logic, the smaller VANDLE modules
655 can achieve an intrinsic efficiency of over 70% for 300 keV neutrons. This is
656 consistent with achievements of other low-threshold neutron detectors that
657 record left-right coincident events [50]. The hardware threshold for the light
658 yield in the bar is near 8 keVee. This is low enough to measure and characterize
659 the light response curve from elastically-scattered carbon nuclei inside the
660 plastic from incident neutrons with kinetic energies below 2 MeV. The low-
661 energy end of the carbon light response curve and equations are constrained
662 by these new data. This also reduces the amount of “dark” reactions during
663 our experiments, improving efficiency and energy resolution. The VANDLE
664 system has been successfully deployed to measure neutrons from beta decay,
665 as well as (α ,n) and (d,n) reaction experiments.

666 11 Acknowledgments

667 This research was sponsored in part by the National Nuclear Security Admin-
668 istration (NNSA) under the Stewardship Science Academic Alliances program
669 through U.S. Department of Energy (DOE) Cooperative Agreements No. DE-
670 FG52-08NA28552, DE-FG52-09NA29455, and DE-NA0002132, the DOE of-
671 fice of Science No. DE-FG02-96ER40955, National Science Foundation, and
672 the NNSA Office of Defense Nuclear Nonproliferation Research and Develop-
673 ment (NA-22) project IN13-V-F(a,n)-PD2La.

674 References

- 675 [1] A. Gade, C. K. Gelbke, and T. Glasmacher. *Nuclear Physics News*, 24:28, 2014.
676 [2] C. Matei, D. W. Bardayan, J. C. Blackmon, J. A. Cizewski, R. K. Grzywacz,

- 677 S. Liddick, S. W. Padgett, W. A. Peters, and F. Sarazin. *AIP Conf. Proc.*,
678 1099:790, 2009.
- 679 [3] C. Matei, D. W. Bardayan, J. C. Blackmon, J. A. Cizewski, R. K. Grzywacz,
680 S. N. Liddick, W. A. Peters, and F. Sarazin. *Proceedings of the Ninth*
681 *International Symposium on Nuclei in the Cosmos, Proc. of Science*, 138:1,
682 2008.
- 683 [4] A. Buta, T. Martin, C. Timis, N. Achouri, J.C. Angelique, C. Borcea, I. Cruceru,
684 A. Genoux-Lubain, S. Grevy, M. Lewitowicz, E. Lienard, F.M. Marques,
685 F. Negoita, F. de Oliveira, N.A. Orr, J. Peter, and M. Sandu. *Nucl. Instr.*
686 *and Meth. A*, 455:412, 2000.
- 687 [5] B. Luther, T. Baumann, M. Thoennessen, J. Brown, P. A. DeYoung, J. Finck,
688 J. Hinnefeld, R. Howes, K. Kemper, P. Pancella, G. Peaslee, W. Rogers, and
689 S. Tabor. *Nucl. Instr. and Meth. A*, 505:33, 2003.
- 690 [6] T. Baumann, J. Boike, J. Brown, M. Bullinger, J. P. Bychowski, S. Clark,
691 K. Daum, P. A. DeYoung, J. V. Evans, J. Finck, N. Frank, A. Grant,
692 J. Hinnefeld, G. W. Hitt, R. H. Howes, B. Isselhardt, K. W. Kemper,
693 J. Longacre, Y. Lu, B. Luther, S. T. Marley, D. McCollum, E. McDonald,
694 U. Onwumene, P. V. Pancella, G. F. Peaslee, W. A. Peters, M. Rajabali,
695 J. Robertson, W. F. Rogers, S. L. Tabor, M. Thoennessen, E. Tryggstad, R. E.
696 Turner, P. J. VanWylen, and N. Walker. *Nucl. Instr. and Meth. A*, 543:517,
697 2005.
- 698 [7] R. Harkewicz, D. J. Morrissey, B. A. Brown, J. A. Nolen Jr., N. A. Orr, B. M.
699 Sherrill, J. S. Winfield, and J. A. Winger. *Phys. Rev. C*, 44:2365, 1991.
- 700 [8] G. Perdikakis, M. Sasano, S. M. Austin, D. Bazin, C. Caesar, S. Cannon, J. M.
701 Deaven, H. J. Doster, C. J. Guess, G. W. Hitt, J. Marks, R. Meharchand, D. T.
702 Nguyen, D. Peterman, A. Prinke, M. Scott, Y. Shimbara, K. Thorne, L. Valdez,
703 and R. G. T. Zegers. *Nucl. Instr. and Meth. A*, 686:117, 2012.
- 704 [9] G. Perdikakis, S. M. Austin, D. Bazin, C. Caesar, J. M. Deaven, C. J. Guess,
705 G. W. Hitt, R. T. Meharchand, D. T. Nguyen, Y. Shimbara, K. Thorne, and
706 R. G. T. Zegers. *IEEE, Trans. on Nucl. Sci.*, 56:1174, 2009.
- 707 [10] Th. Blaich, Th. W. Elze, H. Emling, H. Freiesleben, K. Grimm, W. Henning,
708 R. Holzmann, G. Ickert, J. G. Keller, H. Klingler, W. Kneissl, R. König,
709 R. Kulesa, J. V. Kratz, D. Lambrecht, J. S. Lange, Y. Leifels, E. Lubkiewicz,
710 M. Proft, W. Prokopowicz, C. Schütter, R. Schmidt, H. Spies, K. Stelzer,
711 J. Stroth, W. Walus, E. Wajda, H. J. Wollersheim, M. Zinser, and E. Zude.
712 *Nucl. Instr. and Meth. A*, 314:136, 1992.
- 713 [11] A. Ozawa, G. Raimann, R.N. Boyd, F.R. Chloupek, M. Fujimaki, K. Kimura,
714 T. Kobayashi, H. Kitagawa, T. Kobayashi, J.J. Kolata, S. Kubono, I. Tanihata,
715 Y. Watanabe, and K. Yoshida. *Nucl. Phys. A*, 592:244, 1995.
- 716 [12] J. J. Kolata, Hanan Amro, M. Cloughesy, P. A. DeYoung, J. Rieth, J. P.
717 Bychowski, and G. Peaslee. *Nucl. Instr. and Meth. A*, 557:594, 2006.

- 718 [13] N.V. Kornilov, I. Fabry, S. Oberstedt, and F.-J. Hambsch. *Nucl. Instr. and*
719 *Meth. A*, 599:226, 2009.
- 720 [14] Eljen Technology, 1300 W. Broadway, Sweetwater, TX 79556, USA,
721 www.eljentechnology.com .
- 722 [15] Saint-Gobain Crystals & Detectors, Scintillation Products, 12345 Kinsman Rd.,
723 Newbury, OH 44065, USA, www.crystals.saint-gobain.com .
- 724 [16] Hamamatsu Photonics K.K. Headquarters, 325-6, Sunayama-cho, Naka-ku,
725 Hamamatsu City, Shizuoka Pref., 430-8587, Japan www.hamamatsu.com .
- 726 [17] G. F. Knoll. *Radiation Detection and Measurement -4th ed.* John Wiley and
727 Sons, Inc., 2010.
- 728 [18] XIA LLC, 31057 Genstar Rd., Hayward, CA 94544, USA, www.xia.com .
- 729 [19] Xiaodong Zhang, J. P. Hayward, and M. A. Laubach. *Nucl. Instr. and Meth.*
730 *A*, 755:32, 2014.
- 731 [20] S. V. Paulauskas, M. Madurga, R. Grzywacz, D. Miller, S. Padgett, and H. Tan.
732 *Nucl. Instr. and Meth. A*, 737:22, 2014.
- 733 [21] W. A. Peters. *Study of Neutron Unbound States Using the Modular Neutron*
734 *Array (MoNA)*. PhD thesis, Michigan State University, 2007.
- 735 [22] J. B. Birks. *Phys. Rev.*, 86:569, 1952.
- 736 [23] B. H. Daub, V. Henzl, J. L. Matthews, M. A. Kovash, Z. W. Miller,
737 K. Shoniyozov, and H. Yang. *Nucl. Instr. and Meth. A*, 701:171, 2013.
- 738 [24] B. D. Rooney and J. D. Valentine. *IEEE, Trans. on Nucl. Sci.*, 43:1271, 1996.
- 739 [25] J. F. Williamson, J. F. Dempsey, A. S. Kirov, J. I. Monroe, W. R. Binns, and
740 H. Hedtjarn. *Phys. Med. Bio.*, 44:857, 1999.
- 741 [26] D. E. Groom, N. V. Mokhov, and S. I. Striganov. *At. Data and Nucl. Data*
742 *Tables*, 76:1, 2001.
- 743 [27] International Atomic Energy Agency. *Update of X Ray and Gamma Ray Decay*
744 *Data Standards for Detector Calibration and Other Applications*, volume 1.
745 IAEA, Vienna, Austria, 2007.
- 746 [28] F. T. Porter, M. S. Freedman, F. Wagner, and I. S. Sherman. *Nucl. Instr. and*
747 *Meth.*, 39(1):35, 1966.
- 748 [29] L. Reichhart, D.Yu. Akimov, H. M. Araujo, E. J. Barnes, V. A. Belov, A. A.
749 Burenkov, V. Chepel, A. Currie, L. DeViveiros, B. Edwards, V. Francis,
750 C. Ghag, A. Hollingsworth, M. Horn, G. E. Kalmus, A. S. Kobayakin, A. G.
751 Kovalenko, V. N. Lebedenko, A. Lindote, M. I. Lopes, R. Luscher, P. Majewski,
752 A.St J. Murphy, F. Neves, S. M. Paling, J. Pinto da Cunha, R. Preece, J. J.
753 Quenby, P. R. Scovell, C. Silva, V. N. Solovov, N. J. T. Smith, P. F. Smith,
754 V. N. Stekhanov, T. J. Sumner, C. Thorne, and R. J. Walker. *Phys. Rev. C*,
755 85:065801, 2012.

- [30] R. Madey, F. M. Waterman, A. R. Baldwin, J. Knudson, J. D. Carlson, and J. Rapaport. *Nucl. Instr. and Meth.*, 151:445, 1978.
- [31] F. D. Becchetti, C. E. Thorn, and M. J. Levine. *Nucl. Instr. and Meth.*, 138:93, 1976.
- [32] A. Mechaca-Rocha. *Nucl. Instr. and Meth. A*, 602:421, 2009.
- [33] W. B. Howard, S. M. Grimes, T. N. Massey, S. I. Al-Quraishi, D. K. Jacobs, C. E. Brient, and J. C. Yanch. *Nucl. Sci. Eng.*, 138:145, 2001.
- [34] T. N. Massey, S. Al-Quraishi, C. E. Brient, J. F. Guillemette, S. M. Grimes, D. Jacobs, J. E. O'Donnell, J. Oldendick, and R. Wheeler. *Nucl. Sci. Eng.*, 129:175, 1998.
- [35] W. Mannhart. *Proceedings of the Consultants Meeting on Physics of Neutron Emission in Fission, Mito City, Japan Report INDC(NDS)*, 220:305, 1988.
- [36] W. Mannhart. Technical report, IAEA-TECDOC-410, VIENNA, 1987.
- [37] J. Cub, E. Finckh, K. Gebhardt, K. Geissdörfer, R. Lin, J. Strate, and H. Klien. *Nucl. Instr. and Meth. A*, 274:217, 1989.
- [38] GEANT. CERN library long writeup. Technical Report W5013, CERN, 1994.
- [39] S. Agostinelli et al. *Nucl. Instr. and Meth. A*, 506:250, 2003.
- [40] J. Apostolakis et al. *Journal of Physics: Conference Series*, 160, 2009.
- [41] S. Ilyushkin et al. in preparation.
- [42] H. Franz, W. Rudolph, H. Ohm, K. L. Kratz, G. Herrmann, F. M. Nuh, D. R. Slaughter, and S. G. Prussin. *NIM*, 144:253, 1977.
- [43] J. A. Winger, S. V. Ilyushkin, K. P. Rykaczewski, C. J. Gross, J. C. Batchelder, C. Goodin, R. Grzywacz, J. H. Hamilton, A. Korgul, et al. *Phys. Rev. Lett.*, 102:142502, 2009.
- [44] R. Surman, M. Mumpower, and A. Aprahamian. The sensitivity of r-process nucleosynthesis to individual beta-delayed neutron emission probabilities. *JPS Conference Proceedings*, page 8, 2014. Advances in Radioactive Isotope Science (ARIS) Proceedings.
- [45] P. Möller, B. Pfeiffer, and K-L Kratz. *Phys. Rev. C*, 67:055802, 2003.
- [46] M. Madurga, S. V. Paulauskas, R. Grzywacz, D. Miller, D. W. Bardayan, J. C. Batchelder, N. T. Brewer, J. A. Cizewski, A. Fijalkowska, C. J. Gross, M. E. Howard, S. V. Ilyushkin, B. Manning, M. Matos, A. J. Mendez II, K. Miernik, S. W. Pagett, W. A. Peters, B. C. Rasco, A. Ratkiewicz, K. P. Rykaczewski, D. W. Stracener, E. H. Wang, M. Wolinska-Cichocka, and E. F. Zganjar. 2016. submitted to *Phys. Rev. Lett.*

- 791 [47] K. L. Jones, A. S. Adekola, D. W. Bardayan, J. C. Blackmon, K. Y. Chae,
792 K. A. Chipps, J. A. Cizewski, L. Erikson, C. Harlin, R. Hatarik, R. Kapler,
793 R. L. Kozub, J. F. Liang, R. Livesay, Z. Ma, B. H. Moazen, C. D. Nesaraja,
794 F. M. Nunes, S. D. Pain, N. P. Petterson, D. Shapira, J. F. Shriner, M. S. Smith
795 T. P. Swan, and J. S. Thomas. *Nature*, 465:454, 2010.
- 796 [48] S. Harissopulos, H. W. Becker, J. W. Hammer, A. Lagoyannis, C. Rolfs, and
797 F. Strieder. *Phys. Rev. C*, 72:062801(R), 2005.
- 798 [49] W. A. Peters et al. A kinematically complete, interdisciplinary, and co-
799 institutional measurement of the $^{19}\text{F}(\alpha, n)$ cross section for nuclear safeguards
800 science. Report: Inl/ext-16-38791, osti # 1263500, NNSA, Office of Defense
801 Nuclear Nonproliferation R & D, 2016.
- 802 [50] R. Beyer, E. Grosse, K. Heidel, J. Hutsch, A. R. Junghans, J. Klug, D. Légrády,
803 R. Nolte, S. Röttger, M. Sobiella, and A. Wagner. *Nucl. Instr. and Meth. A*,
804 575(3):449, 2007.