Title Page

Transmission positron images using imaging plates

Masao Doyama^a, A. Kogure^a, M. Inoue^a, T. Kurihara^b, X. Cao^b, M. Matsuya^c,

T. Yoshiie^d, Y. Hayashi^d, Q. Xu^d and M. Fujinami^e

^a*Teikyo University of Science and Technology, 2525 Yatsusawa, Uenohara, Yamanashi, 409-0193, JAPAN*

^b*High Energy Accelerator Research Organization (KEK), 1-1 Ono, Tsukuba-shi, Ibaraki 305-0801, JAPAN*

^c*Japan Electron Optics Laboratory (JEOL) Ltd, 3-1-2 Musashino, Akishima, 196-0021, JAPAN*

^d*Reactor Research Institute, Kyoto University, Asashiro, Kumatori, Osaka 590-0494 JAPAN*

^e*Faculty of Engineering, Chiba University, 1-33 Yayoicho, Inage, Chiba, 263-8522 JAPAN*

Corresponding Author:

Masao Doyama

Teikyo University of Science and Technology

2525 Yatsusawa, Uenohara, Yamanashi 409-0193 Japan

Tel: +81554634411

Fax: +81554634431

E-mail: doyama@ntu.ac.jp

Transmission positron images using imaging plates

Masao Doyama^a, A. Kogure^a, M. Inoue^a, T. Kurihara^b, X. Cao^b, M. Matsuya^c, T. Yoshiie^d, Y. Hayashi^d, Q. Xu^d and M. Fujinami^e

^aTeikyo University of Science and Technology, 2525 Yatsusawa, Uenohara, Yamanashi, 409-0193, JAPAN

^bHigh Energy Accelerator Research Organization (KEK), 1-1 Ono, Tsukuba-shi, Ibaraki 305-0801, JAPAN

^cJapan Electron Optics Laboratory (JEOL) Ltd, 3-1-2 Musashino, Akishima, 196-0021, JAPAN

^dReactor Research Institute, Kyoto University, Asashiro, Kumatori, Osaka 590-0494 JAPAN

^eFaculty of Engineering, Chiba University, 1-33 Yayoicho, Inage, Chiba, 263-8522 JAPAN

Abstract

As an image recording medium for transmission positron microscopes, imaging plates are quite useful and powerful. Imaging plates are also quite sensitive and the photon stimulated luminescence (PSL) is linearly proportional to the positron intensity in six digits (10^6). No bulky or expensive equipment is necessary to accommodate in vacuum. Imaging plates can be set under bright lights, this is different from the photographic films. Darkness is only required during exposure and transfer to a reader. Slow Positron Facility at KEK, Japan was used to study the effect of “mono-

chromatic" positron beam. Specimens were set just in front of an imaging plate. After a certain time of exposure, the imaging plates were processed by a reader. Used imaging plates can be used repeatedly after erased by ultra-violet lights. Images through samples can be obtained. Similar experiments using non-monochromatic (white) positrons and electrons have been performed at Teikyo University of Science and Technology (TUST) and Research Reactor Institute, Kyoto Univ. (RRI). Sealed ^{22}Na positron source can be conveniently used for non-destructive tests.

Keywords:

07.05.Pj Image processing, 07.78+s Electron, positron, and ion microscopes; electron diffractometers, 25.30.Hm Positron Scattering,

Introduction

We have been trying to construct a microscope using positrons instead of using electrons for electron microscopes [1]. Different from a transmission electron microscope, a transmission positron microscope has been constructed reforming an old scrapped transmission electron microscope. A positron is the anti-particle of an electron, so by inverting the sign of the acceleration voltage of a transmission electron microscope and by supplying a positron beam instead of an electron source, a transmission positron microscope should be able to be built. In this way, the techniques of electron microscopes highly developed by industries for many years can be utilized. It is not easy to obtain strong intensity positron beams instead of

electron beams; accelerators or atomic reactors are required to obtain strong beams of positrons. We have remolded the transmission electron source chamber of an old transmission microscope Type JEM100SX manufactured by JEOL and transported to KEK (High Energy Accelerator Organization) in Tsukuba, Japan. When we supply a positive voltage to the positron emitter, the electron microscope can be converted to a positron microscope. We had been waited for a positron beam to come to the microscope.

At KEK, a strong positron pulsed beam of $10^8 \text{ e}^+/\text{s}$ has been supplied to a Time Of Flight (TOF) Facility which has been open to Cooperative Research for several years. We have been performed positron experiments with given energies below 35 keV. We needed a separate branch of a positron beam for a positron microscope by a changeover switches. The construction of the positron beam branch requires a lot of funds. It took us a long time. A new proposal headed by Professor Fujinami "Transmission Positron Microscopes" has been accepted by the Japan Science and Technology Promotion Agency (JST) from 2005 fiscal year[2]. This project partially supported a branch out the positron beam from TOF to a new transmission positron microscope. A brand-new transmission positron microscope based on our long experience has been set at KEK removing our old positron microscope. The positron beam is expected to reach the new microscope in a few months.

While we have been waiting the new positron beam branch, we have investigated the method of recording positron images using imaging plates.

Methods obtaining positron images

There are many ways of obtaining positrons images. Photographic films are one of common methods. However, the sensitivity (optical density / fluence) is not constant with the fluence although the positional resolution is excellent. A multi-channel plate requires a high voltage supply and an electron detector. CCD (Charge Coupled Device Image Sensor) also has to be inserted in vacuum.

Imaging plates are a card coated with barium fluorohalides phosphors which shows photostimulated luminescence (PSL). Darkness is not required before exposure but darkness is only required during exposure and the time of transfer the card to an analyzer. We use Fuji Imaging Analyzer BA2500. When the exposed plate is transferred to an analyzer, darkness is required so that the plate was put in a black paper envelope through which light does not penetrate. The envelope was taken to an analyzer, and the plate was taken out in darkness and was set in the analyzer in dim light. If an analyzer is far, we can carry the envelope. Photostimulated luminescence is linearly proportional to fluence in six digits. We have succeeded to obtain positron images using imaging plates, the first in the world [3].

Fuji Imaging Plates BAS-TR2025 [4] for bio imaging analyzer BAS2500 were used. The imaging plates were cut into 30mm x 30 mm, and these imaging plates do not have protective layers to respond slow positron or electron beams. Fuji Imaging Analyzer BAS2500's at the photon Factory, KEK, at Teikyo University of Science and Technology and at Kyoto University Reactor Research Institute were used to analyze positron and electron images recorded on imaging plates.

A positron excites an electron of barium fluorohalides phosphors to a metastable state and the metastable electron decays very slowly. In an analyzer, by irradiating with scanning red laser (633 nm), blue light (400 nm, short life-time of 0.8 micro-

seconds) is emitted. This blue light is detected by photo-electron conversion element, after digital imaging process, an image is obtained.

Photo-stimulated-luminescence usually has very wide dynamic ranges. Logarithm of the detected number of photo-stimulated-luminescence is taken to shrink imaging data to narrower ranges and then quantized, which means the figure was rounded to an integer. This value is called QL (Quantum Level) by Fuji Film Co., Japan [4]. The “IPReader” (commercial name of an imaging plate analyzer by Fuji Film Co.) output is kept as QL. The grade of the number of positrons entered to one pixel can be recorded as 8 bits (Grad 256), 10 bits (Grad 1024), or 16 bits (Grad 65536). 16 bits was selected in this experiment. Photo-stimulated luminescence value PSL is proportional to the photo-stimulated-luminescence. The photo-stimulated-luminescence per area, I , can be represented as PSL/mm^2 . The relation between PSL and QL is defined as

$$PSL = (\text{Pixel Size}/100)^2 \times (4000/\text{Sensitivity}) \times 10^{(Latitude - (QL/\text{Grad} - 0.5))},$$

according to Fuji Film Co.[4] $PSL = 0$, when $QL = 0$ where $\text{Grad} = 2^{\text{bit}}$.

For the same kind of particles, IP, reading densities are the same, therefore the comparison between different exposures is possible. We have used sensitivity = 4000, latitude = 5, and Grad = 1024.

The PSL is proportional in the range 1 to one million of the fluence of positrons [3]. This is much more sensitive than films and more quantitative. It is handy to use imaging plates for positron microscopy. When an imaging plate is mounted like a photographic film, no dark place is required, only during exposure and before mounting in a reader (analyzer) after the exposure, darkness is required. No complicated instrument is required inside the exposing equipment. No coincidence

circuit is necessary. The images on imaging plates can be erased by irradiating ultra-violet light and the imaging plates can be used repeatedly.

As an example, we will describe when Fuji Imaging Plate's (BAS-TR2025) for bio imaging analyzer BAS2500 are used. The size of an imaging plate can be cut to any size we desire. 30mm x 30 mm size plates are used for the Positronium TOF facility. 100mm x 81 mm is the same as the size of the films for electron microscopy. This is convenient because we can use film loader of an electron microscope. We analyze the exposed plate by BAS2500. We have chosen the pixel size 50 μm x 50 μm . This is the smallest size of a pixel for BAS-TR2025. The positional resolution of imaging plates is as good as photographic film, but the analyzer has a limit of pixel size, for a normal use. If we use Microarray Scanner FLA-8000 with imaging plates, the smallest pixel size is 10 μm X 10 μm . 8 x 12 cm imaging plates can be used. Type FDL UR-V imaging plates are also used for ^{22}Na positron source in this experiment. This is for electron microscopy and widely used in electron microscopy.

Mono-chromatic positron beam

Using TOF facility at KEK, a beam of mono-energetic positrons up to 35 kV is available since 2000 as a Cooperative Research. We have succeeded to obtain transmission images of objects.

Figure 1 shows the positron beam intensity distribution at TOF. Before our measurement, it was not easy to obtain an intensity distribution within a positron beam. The size of the IP is 30mm x 30mm. The accelerator to produce positrons we use is a pulse type. The intensity of positrons per pulse is $2 \times 10^5 \text{ e}^+$, repetition rate is 50 Hz [5] so that usual equipment is difficult to measure the exact intensity.

Imaging plates are accumulative and can measure the total intensity of positrons without counting loss.

Non-monochromatic positron beam

We bought in 2006, a legally sealed source of ^{22}Na of 100 micro Curies which can be used legally in an ordinal research laboratory, if we are careful for gamma radiation.

We have found that transmission images by positrons can be obtained even in air. A specimen was put on an imaging plate, and the set was irradiated by positrons from the ^{22}Na sealed source as shown in Fig. 2. Figure 3 (a) shows a transmission positron image a leaf in air. Veins can be clearly seen. Figure 3(b) shows a transmission positron image of a slice of bacon. White parts are tendons.

In Vacuum

Under reduced pressure, pumped by a mechanical pump only, clearer images have been obtained. A specimen on a sheet of IP was put in a desicator with a stop cock as shown in Fig. 4. Figure 5 shows a transmission positron images of a leaf in in vacuum which many streaks can be seen. Comparing with Fig. 3 (a), the image is clearer. Figure 6 (a) transmission positron images of shrimps under vacuum. Even internal organs can be seen. The scale shown is in centi-meters. (b) is in air. The image under vacuum is clearer than that in air. In figures 5 and 6, black and white are reversed. In dark place, more positrons are absorbed. This new method has been applied for a patent [6]. The particles making images are not known at present.

Acknowledgement

We are grateful for the support by the Grant-in-aid, Scientific Research (B) from 2002-2007, (“Trial production and application of transmission positron-electron microscopes”) and also supported by Japan Science and Technology Agency (2005-2009), and Development and Cooperative Experiments by Photon Factory, KEK and RRI (Kyoto University, Research Reactor Institute).

References

- [1] Masao Doyama, in: Positron Annihilation, Eds. Jain, Singru and Gopinathan (World Science, Singapore, 1985) p. 437.
- [2] M. Fujinami, “Transmission Positron Microscopes”, Japan Science and Technology Agency.
- [3] M. Doyama, J. Takano, M. Inoue, T. Yoshiie, Y. Hayashi, M. Kiritani, and T. Oikawa, Nuclear. Instruments and Methods. Phys. Res. A394(1997), 146.
- [4] Fuji Film Co. IP manual.
- [5] T. kurihara et al. Materials Science Forum 445-446 (2004) 486.
- [6] Patent received, No. 2005-346688 on November 30, 2005 by the Japan Patent Office.

Figure Captions

Fig. 1. Intensity distribution of TOF at KEK. The positron speed is 30 keV. (a) positron intensity image, (b) positron intensity curve along a line passing through the center of the beam.

Fig. 2. A specimen was put on an imaging plate, and the set was irradiated by positrons from a ^{22}Na sealed source.

Fig. 3. (a) Transmission positron image of a leaf, (b) transmission positron image of a slice of bacon.

Fig. 4. schematic draw of an apparatus in which specimens are irradiated by positrons under vacuum.

Fig. 5. transmission positron images of shrimps. Even many streaks can be seen.

Fig. 6. transmission positron images of shrimps. (a) in vacuum, (b) in air.

Fig. 7. transmission positron image of a fish head in vacuum.

Fig. 1(a)

(a)

(b)

Fig. 2.

Fig. 3.

Fig. 3

(a)

(b)

Fig. 4.

Fig. 5.

Fig. 6.

Fig. 7. Relation between muon momentum and range. Upper curve is for copper and lower curve is for tungsten.